DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Onslow Instrument Location Onslow County

Instrument Serial No. 008819 Sheriff's Office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Randy E Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ONSLOW COUNTY ONSLOW COUNTY SD 660

Serial Number: *008819*
Test Date: *07/18/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG104004
Exp Date: 02/09/2013

Test	g/210L	Time
DIAG	Pass	11:16am
AIR BLK	.00	11:16am
ACCY CHK	.08	11:17am
AIR BLK	.00	11:18am
SUB TEST	**.00**	**11:18am**
AIR BLK	.00	11:19am
SUB TEST	**.00**	**11:21am**
AIR BLK	.00	11:22am

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY ONSLOW COUNTY SD 660

Serial Number: *008819* Test Record Number: *412*
Test Date: *07/18/2012* Test Time: *11:22am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:22am
FLO	Pass	11:22am
FC	Pass	11:23am

Temperature Tests

Test	Status	Time
FC1	Pass	11:23am
SRC	Pass	11:23am
DET	Pass	11:23am
BAR	Pass	11:23am
BT	Pass	11:23am

Blank Tests

Test	Status	Time
AIR	Pass	11:23am

Printer Tests

Test	Status	Time
PRNT	Pass	11:23am

CRC Tests

Test	Status	Time
COMP	Pass	11:23am
CAL	Pass	11:23am

Preventive Maintenance
Status: Pass

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Haywood Instrument Location Haywood Co. Jail

Instrument Serial No. 008714 Waynesville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

David R. Cuthbertson
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: *008714*
Test Date: *07/24/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *CUTLER, DANIEL R*
Permit Number: *8457E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203103
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	10:45am
AIR BLK	.00	10:46am
ACCY CHK	.08	10:46am
AIR BLK	.00	10:47am
SUB TEST	**.00**	**10:49am**
AIR BLK	.00	10:50am
SUB TEST	**.00**	**10:51am**
AIR BLK	.00	10:52am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: *008714* Test Record Number: *751*
Test Date: *07/24/2012* Test Time: *10:54am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:54am
FLO	Pass	10:54am
FC	Pass	10:54am

Temperature Tests

Test	Status	Time
FC1	Pass	10:54am
SRC	Pass	10:54am
DET	Pass	10:54am
BAR	Pass	10:54am
BT	Pass	10:54am

Blank Tests

Test	Status	Time
AIR	Pass	10:55am

Printer Tests

Test	Status	Time
PRNT	Pass	10:55am

CRC Tests

Test	Status	Time
COMP	Pass	10:55am
CAL	Pass	10:55am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Haywood Instrument Location Haywood Co. Jail

Instrument Serial No. 008712 Waynesville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: *008712*
Test Date: *07/24/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *CUTLER, DANIEL R*
Permit Number: *8457E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG109703
Exp Date: 04/07/2013

Test	g/210L	Time
DIAG	Pass	10:44am
AIR BLK	.00	10:45am
ACCY CHK	.08	10:45am
AIR BLK	.00	10:46am
SUB TEST	**.00**	**10:47am**
AIR BLK	.00	10:48am
SUB TEST	**.00**	**10:50am**
AIR BLK	.00	10:51am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: *008712* Test Record Number: *1223*
Test Date: *07/24/2012* Test Time: *10:52am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:52am
FLO	Pass	10:52am
FC	Pass	10:52am

Temperature Tests

Test	Status	Time
FC1	Pass	10:52am
SRC	Pass	10:52am
DET	Pass	10:52am
BAR	Pass	10:52am
BT	Pass	10:52am

Blank Tests

Test	Status	Time
AIR	Pass	10:53am

Printer Tests

Test	Status	Time
PRNT	Pass	10:53am

CRC Tests

Test	Status	Time
COMP	Pass	10:53am
CAL	Pass	10:53am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Swain Instrument Location Swain Co. Jail

Instrument Serial No. 008727 Bryson City, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: *008727*
Test Date: *07/23/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *CUTLER, DANIEL R*
Permit Number: *8457E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG104004
Exp Date: 02/09/2013

Test	g/210L	Time
DIAG	Pass	10:47am
AIR BLK	.00	10:48am
ACCY CHK	.07	10:48am
AIR BLK	.00	10:49am
SUB TEST	**.00**	**10:50am**
AIR BLK	.00	10:51am
SUB TEST	**.00**	**10:52am**
AIR BLK	.00	10:53am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: *008727* Test Record Number: *669*
Test Date: *07/23/2012* Test Time: *10:56am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:56am
FLO	Pass	10:56am
FC	Pass	10:57am

Temperature Tests

Test	Status	Time
FC1	Pass	10:57am
SRC	Pass	10:57am
DET	Pass	10:57am
BAR	Pass	10:57am
BT	Pass	10:57am

Blank Tests

Test	Status	Time
AIR	Pass	10:57am

Printer Tests

Test	Status	Time
PRNT	Pass	10:57am

CRC Tests

Test	Status	Time
COMP	Pass	10:57am
CAL	Pass	10:57am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Surgin Instrument Location Surgin Co. Jail

Instrument Serial No. 008723 Bryson City, NE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: *008723*
Test Date: *07/23/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *CUTLER, DANIEL R*
Permit Number: *8457E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG104101
Exp Date: 02/10/2013

Test	g/210L	Time
DIAG	Pass	10:46am
AIR BLK	.00	10:47am
ACCY CHK	.07	10:47am
AIR BLK	.00	10:48am
SUB TEST	**.00**	**10:49am**
AIR BLK	.00	10:49am
SUB TEST	**.00**	**10:51am**
AIR BLK	.00	10:52am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: *008723* Test Record Number: *423*
Test Date: *07/23/2012* Test Time: *10:53am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:53am
FLO	Pass	10:53am
FC	Pass	10:53am

Temperature Tests

Test	Status	Time
FC1	Pass	10:53am
SRC	Pass	10:53am
DET	Pass	10:53am
BAR	Pass	10:53am
BT	Pass	10:53am

Blank Tests

Test	Status	Time
AIR	Pass	10:54am

Printer Tests

Test	Status	Time
PRNT	Pass	10:54am

CRC Tests

Test	Status	Time
COMP	Pass	10:54am
CAL	Pass	10:54am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Macon Instrument Location Macon Co. Magistrate

Instrument Serial No. 008795 Highlands, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MACON COUNTY MACON CO MAGISTRATE 550

Serial Number: *008795*
Test Date: *07/17/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *CUTLER, DANIEL R*
Permit Number: *8457E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG104101
Exp Date: 02/10/2013

Test	g/210L	Time
DIAG	Pass	1:13pm
AIR BLK	.00	1:14pm
ACCY CHK	.07	1:15pm
AIR BLK	.00	1:15pm
SUB TEST	**.00**	**1:16pm**
AIR BLK	.00	1:17pm
SUB TEST	**.00**	**1:18pm**
AIR BLK	.00	1:19pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MACON COUNTY MACON CO MAGISTRATE 550

Serial Number: *008795* Test Record Number: *237*
Test Date: *07/17/2012* Test Time: *1:25pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:25pm
FLO	Pass	1:25pm
FC	Pass	1:25pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:25pm
SRC	Pass	1:25pm
DET	Pass	1:25pm
BAR	Pass	1:25pm
BT	Pass	1:25pm

Blank Tests

Test	Status	Time
AIR	Pass	1:26pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:26pm

CRC Tests

Test	Status	Time
COMP	Pass	1:26pm
CAL	Pass	1:26pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cherokee Instrument Location Cherokee Co. Jail

Instrument Serial No. 008711 Murphy, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CHEROKEE COUNTY CHEROKEE COUNTY JAIL
190

Serial Number: *008711*
Test Date: *07/12/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *CUTLER, DANIEL R*
Permit Number: *8457E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG104004
Exp Date: 02/09/2013

Test	g/210L	Time
DIAG	Pass	11:47am
AIR BLK	.00	11:48am
ACCY CHK	.07	11:48am
AIR BLK	.00	11:49am
SUB TEST	**.00**	**11:50am**
AIR BLK	.00	11:51am
SUB TEST	**.00**	**11:52am**
AIR BLK	.00	11:53am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CHEROKEE COUNTY CHEROKEE COUNTY JAIL 190

Serial Number: *008711* Test Record Number: *526*
Test Date: *07/12/2012* Test Time: *11:55am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:55am
FLO	Pass	11:55am
FC	Pass	11:55am

Temperature Tests

Test	Status	Time
FC1	Pass	11:55am
SRC	Pass	11:55am
DET	Pass	11:55am
BAR	Pass	11:55am
BT	Pass	11:55am

Blank Tests

Test	Status	Time
AIR	Pass	11:56am

Printer Tests

Test	Status	Time
PRNT	Pass	11:56am

CRC Tests

Test	Status	Time
COMP	Pass	11:56am
CAL	Pass	11:56am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cherokee Instrument Location Cherokee Co. Jail

Instrument Serial No. 008622 Murphy, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CHEROKEE COUNTY CHEROKEE COUNTY JAIL
190

Serial Number: *008622*
Test Date: *07/12/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *CUTLER, DANIEL R*
Permit Number: *8457E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG104004
Exp Date: 02/09/2013

Test	g/210L	Time
DIAG	Pass	11:46am
AIR BLK	.00	11:46am
ACCY CHK	.08	11:47am
AIR BLK	.00	11:48am
SUB TEST	**.00**	**11:49am**
AIR BLK	.00	11:50am
SUB TEST	**.00**	**11:51am**
AIR BLK	.00	11:52am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CHEROKEE COUNTY CHEROKEE COUNTY JAIL 190

Serial Number: *008622* Test Record Number: *763*
Test Date: *07/12/2012* Test Time: *11:55am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:55am
FLO	Pass	11:55am
FC	Pass	11:55am

Temperature Tests

Test	Status	Time
FC1	Pass	11:55am
SRC	Pass	11:55am
DET	Pass	11:55am
BAR	Pass	11:55am
BT	Pass	11:55am

Blank Tests

Test	Status	Time
AIR	Pass	11:56am

Printer Tests

Test	Status	Time
PRNT	Pass	11:56am

CRC Tests

Test	Status	Time
COMP	Pass	11:56am
CAL	Pass	11:56am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wilkes Instrument Location Wilkes Co. Courthouse

Instrument Serial No. 008843 Wilkesboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WILKES COUNTY WILKES CO COURTHOUSE 960

Serial Number: *008843*
Test Date: *07/25/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2011-06/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	1:27pm
AIR BLK	.00	1:28pm
ACCY CHK	.07	1:28pm
AIR BLK	.00	1:29pm
SUB TEST	**.00**	**1:30pm**
AIR BLK	.00	1:31pm
SUB TEST	**.00**	**1:32pm**
AIR BLK	.00	1:33pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WILKES COUNTY WILKES CO COURTHOUSE 960

Serial Number: *008843* Test Record Number: *1126*
Test Date: *07/25/2012* Test Time: *1:33pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:34pm
FLO	Pass	1:34pm
FC	Pass	1:34pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:34pm
SRC	Pass	1:34pm
DET	Pass	1:34pm
BAR	Pass	1:34pm
BT	Pass	1:34pm

Blank Tests

Test	Status	Time
AIR	Pass	1:35pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:35pm

CRC Tests

Test	Status	Time
COMP	Pass	1:35pm
CAL	Pass	1:35pm

Preventive Maintenance
Status: Pass

L. Kuindon

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Yadkin Instrument Location Yadkin Co. Jail

Instrument Serial No. 008944 Yadkinville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: *008944*
Test Date: *07/25/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2011-06/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	11:50am
AIR BLK	.00	11:51am
ACCY CHK	.08	11:51am
AIR BLK	.00	11:52am
SUB TEST	**.00**	**11:53am**
AIR BLK	.00	11:54am
SUB TEST	**.00**	**11:55am**
AIR BLK	.00	11:56am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: *008944* Test Record Number: *799*
Test Date: *07/25/2012* Test Time: *11:57am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:57am
FLO	Pass	11:57am
FC	Pass	11:57am

Temperature Tests

Test	Status	Time
FC1	Pass	11:57am
SRC	Pass	11:57am
DET	Pass	11:57am
BAR	Pass	11:57am
BT	Pass	11:57am

Blank Tests

Test	Status	Time
AIR	Pass	11:58am

Printer Tests

Test	Status	Time
PRNT	Pass	11:58am

CRC Tests

Test	Status	Time
COMP	Pass	11:58am
CAL	Pass	11:58am

Preventive Maintenance
Status: Pass

L. Kevin Dixon
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Yadkin Instrument Location Yadkinville Police Department

Instrument Serial No. 008925

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

L. Kevin Dion
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

YADKIN COUNTY YADKINVILLE PD 980

Serial Number: *008925*
Test Date: *07/25/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2011-06/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	11:04am
AIR BLK	.00	11:05am
ACCY CHK	.08	11:05am
AIR BLK	.00	11:07am
SUB TEST	**.00**	**11:07am**
AIR BLK	.00	11:08am
SUB TEST	**.00**	**11:09am**
AIR BLK	.00	11:11am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

YADKIN COUNTY YADKINVILLE PD 980

Serial Number: *008925* Test Record Number: *296*
Test Date: *07/25/2012* Test Time: *11:12am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:12am
FLO	Pass	11:12am
FC	Pass	11:12am

Temperature Tests

Test	Status	Time
FC1	Pass	11:12am
SRC	Pass	11:12am
DET	Pass	11:12am
BAR	Pass	11:12am
BT	Pass	11:12am

Blank Tests

Test	Status	Time
AIR	Pass	11:13am

Printer Tests

Test	Status	Time
PRNT	Pass	11:13am

CRC Tests

Test	Status	Time
COMP	Pass	11:13am
CAL	Pass	11:13am

Preventive Maintenance
Status: Pass

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Stokes Instrument Location Stokes Co. Jail

Instrument Serial No. 008596 Danbury, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

THE GREAT SEAL OF THE STATE OF NORTH CAROLINA MAY 20, 1775 APRIL 12, 1776 ESSE QUAM VIDERI

L. Kevin Dunn 642

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

STOKES COUNTY STOKES COUNTY JAIL 840

Serial Number: *008596*
Test Date: *07/26/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2011-06/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	2:08pm
AIR BLK	.00	2:09pm
ACCY CHK	.08	2:09pm
AIR BLK	.00	2:10pm
SUB TEST	**.00**	**2:11pm**
AIR BLK	.00	2:12pm
SUB TEST	**.00**	**2:13pm**
AIR BLK	.00	2:14pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

STOKES COUNTY STOKES COUNTY JAIL 840

Serial Number: *008596* Test Record Number: *554*
Test Date: *07/26/2012* Test Time: *2:15pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:16pm
FLO	Pass	2:16pm
FC	Pass	2:16pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:16pm
SRC	Pass	2:16pm
DET	Pass	2:16pm
BAR	Pass	2:16pm
BT	Pass	2:16pm

Blank Tests

Test	Status	Time
AIR	Pass	2:17pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:17pm

CRC Tests

Test	Status	Time
COMP	Pass	2:17pm
CAL	Pass	2:17pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Stokes Instrument Location King Police

Instrument Serial No. 008610 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

STOKES COUNTY KING PD 840

Serial Number: *008610*
Test Date: *07/26/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2011-06/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	3:56pm
AIR BLK	.00	3:56pm
ACCY CHK	.08	3:57pm
AIR BLK	.00	3:58pm
SUB TEST	**.00**	**3:58pm**
AIR BLK	.00	3:59pm
SUB TEST	**.00**	**4:01pm**
AIR BLK	.00	4:02pm

Reported AC: .00 g/210L

L. Kevin Dean
Signature of Chemical Analyst

Court CVR

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

STOKES COUNTY KING PD 840

Serial Number: *008610* Test Record Number: *1141*
Test Date: *07/26/2012* Test Time: *4:02pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:03pm
FLO	Pass	4:03pm
FC	Pass	4:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:03pm
SRC	Pass	4:03pm
DET	Pass	4:03pm
BAR	Pass	4:03pm
BT	Pass	4:03pm

Blank Tests

Test	Status	Time
AIR	Pass	4:04pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:04pm

CRC Tests

Test	Status	Time
COMP	Pass	4:04pm
CAL	Pass	4:04pm

Preventive Maintenance
Status: Pass

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ALAMANCE Instrument Location ALAMANCE CO. JAIL

Instrument Serial No. 008853 109 S. MAPLE ST. GRAHAM, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: *008853*
Test Date: *07/10/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SMITH, BRIAN D*
Permit Number: *08937E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203102
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	1:57pm
AIR BLK	.00	1:58pm
ACCY CHK	.08	1:59pm
AIR BLK	.00	2:00pm
SUB TEST	**.00**	**2:01pm**
AIR BLK	.00	2:02pm
SUB TEST	**.00**	**2:03pm**
AIR BLK	.00	2:04pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: *008853* Test Record Number: *927*
Test Date: *07/10/2012* Test Time: *2:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:09pm
FLO	Pass	2:09pm
FC	Pass	2:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:09pm
SRC	Pass	2:09pm
DET	Pass	2:09pm
BAR	Pass	2:09pm
BT	Pass	2:09pm

Blank Tests

Test	Status	Time
AIR	Pass	2:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:09pm

CRC Tests

Test	Status	Time
COMP	Pass	2:10pm
CAL	Pass	2:10pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ALAMANCE Instrument Location ALAMANCE Co. JAIL

Instrument Serial No. 008913 109 S. MAPLE ST. GRAHAM, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Brian D Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: *008913*
Test Date: *07/10/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SMITH, BRIAN D*
Permit Number: *08937E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG124202
Exp Date: 08/30/2013

Test	g/210L	Time
DIAG	Pass	1:58pm
AIR BLK	.00	1:59pm
ACCY CHK	.07	2:00pm
AIR BLK	.00	2:01pm
SUB TEST	**.00**	**2:01pm**
AIR BLK	.00	2:02pm
SUB TEST	**.00**	**2:04pm**
AIR BLK	.00	2:05pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: *008913* Test Record Number: *1423*
Test Date: *07/10/2012* Test Time: *2:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:09pm
FLO	Pass	2:09pm
FC	Pass	2:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:09pm
SRC	Pass	2:09pm
DET	Pass	2:09pm
BAR	Pass	2:09pm
BT	Pass	2:09pm

Blank Tests

Test	Status	Time
AIR	Pass	2:10pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:10pm

CRC Tests

Test	Status	Time
COMP	Pass	2:10pm
CAL	Pass	2:10pm

Preventive Maintenance
Status: Pass

Brian D Smith
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ALAMANCE Instrument Location BURLINGTON PD

Instrument Serial No. 008812 267 W. FRONT ST. BURLINGTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Bud D Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: *008812*
Test Date: *07/10/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SMITH, BRIAN D*
Permit Number: *08937E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG106703
Exp Date: 03/08/2013

Test	g/210L	Time
DIAG	Pass	2:47pm
AIR BLK	.00	2:48pm
ACCY CHK	.07	2:49pm
AIR BLK	.00	2:49pm
SUB TEST	**.00**	**2:50pm**
AIR BLK	.00	2:51pm
SUB TEST	**.00**	**2:52pm**
AIR BLK	.00	2:53pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: *008812* Test Record Number: *1314*
Test Date: *07/10/2012* Test Time: *2:59pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:59pm
FLO	Pass	2:59pm
FC	Pass	2:59pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:59pm
SRC	Pass	2:59pm
DET	Pass	2:59pm
BAR	Pass	2:59pm
BT	Pass	2:59pm

Blank Tests

Test	Status	Time
AIR	Pass	3:00pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:00pm

CRC Tests

Test	Status	Time
COMP	Pass	3:00pm
CAL	Pass	3:00pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ALAMANCE Instrument Location BURLINGTON PD

Instrument Serial No. 008907 267 W. FRONT ST. BURLINGTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Brad D Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: *008907*
Test Date: *07/10/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SMITH, BRIAN D*
Permit Number: *08937E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG124903
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	2:46pm
AIR BLK	.00	2:47pm
ACCY CHK	.08	2:48pm
AIR BLK	.00	2:49pm
SUB TEST	**.00**	**2:50pm**
AIR BLK	.00	2:50pm
SUB TEST	**.00**	**2:52pm**
AIR BLK	.00	2:53pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: *008907* Test Record Number: *482*
Test Date: *07/10/2012* Test Time: *2:57pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:57pm
FLO	Pass	2:57pm
FC	Pass	2:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:57pm
SRC	Pass	2:57pm
DET	Pass	2:57pm
BAR	Pass	2:57pm
BT	Pass	2:57pm

Blank Tests

Test	Status	Time
AIR	Pass	2:58pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:58pm

CRC Tests

Test	Status	Time
COMP	Pass	2:58pm
CAL	Pass	2:58pm

Preventive Maintenance
Status: Pass

Brian D Smith
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County GRANVILLE Instrument Location OXFORD P.D.

Instrument Serial No. 008923 204 E. McCLANAHAN ST. OXFORD, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GRANVILLE COUNTY OXFORD PD 380

Serial Number: *008923*
Test Date: *07/23/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG202602
Exp Date: 01/26/2014

Test	g/210L	Time
DIAG	Pass	12:23pm
AIR BLK	.00	12:24pm
ACCY CHK	.08	12:25pm
AIR BLK	.00	12:26pm
SUB TEST	**.00**	**12:26pm**
AIR BLK	.00	12:27pm
SUB TEST	**.00**	**12:28pm**
AIR BLK	.00	12:29pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GRANVILLE COUNTY OXFORD PD 380

Serial Number: *008923* Test Record Number: *769*
Test Date: *07/23/2012* Test Time: *12:30pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:31pm
FLO	Pass	12:31pm
FC	Pass	12:31pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:31pm
SRC	Pass	12:31pm
DET	Pass	12:31pm
BAR	Pass	12:31pm
BT	Pass	12:31pm

Blank Tests

Test	Status	Time
AIR	Pass	12:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:31pm

CRC Tests

Test	Status	Time
COMP	Pass	12:32pm
CAL	Pass	12:32pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County GUILFORD Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008616 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: *008616*
Test Date: *07/27/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG023601
Exp Date: 08/24/2012

Test	g/210L	Time
DIAG	Pass	11:22pm
AIR BLK	.00	11:23pm
ACCY CHK	.07	11:24pm
AIR BLK	.00	11:24pm
SUB TEST	**.00**	**11:25pm**
AIR BLK	.00	11:26pm
SUB TEST	**.00**	**11:27pm**
AIR BLK	.00	11:28pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: *008616* Test Record Number: *1442*
Test Date: *07/27/2012* Test Time: *11:29pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:29pm
FLO	Pass	11:29pm
FC	Pass	11:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:30pm
SRC	Pass	11:30pm
DET	Pass	11:30pm
BAR	Pass	11:30pm
BT	Pass	11:30pm

Blank Tests

Test	Status	Time
AIR	Pass	11:30pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:30pm

CRC Tests

Test	Status	Time
COMP	Pass	11:30pm
CAL	Pass	11:30pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County GUILFORD Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Allen Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: *008707*
Test Date: *07/27/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	11:16pm
AIR BLK	.00	11:17pm
ACCY CHK	.08	11:17pm
AIR BLK	.00	11:18pm
SUB TEST	**.00**	**11:19pm**
AIR BLK	.00	11:20pm
SUB TEST	**.00**	**11:21pm**
AIR BLK	.00	11:22pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: *008707* Test Record Number: *1500*
Test Date: *07/27/2012* Test Time: *11:27pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:27pm
FLO	Pass	11:27pm
FC	Pass	11:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:28pm
SRC	Pass	11:28pm
DET	Pass	11:28pm
BAR	Pass	11:28pm
BT	Pass	11:28pm

Blank Tests

Test	Status	Time
AIR	Pass	11:28pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:28pm

CRC Tests

Test	Status	Time
COMP	Pass	11:28pm
CAL	Pass	11:28pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County GUILFORD Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: *008647*
Test Date: *07/27/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	11:27pm
AIR BLK	.00	11:28pm
ACCY CHK	.07	11:29pm
AIR BLK	.00	11:30pm
SUB TEST	**.00**	**11:30pm**
AIR BLK	.00	11:31pm
SUB TEST	**.00**	**11:33pm**
AIR BLK	.00	11:34pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: *008647* Test Record Number: *1434*
Test Date: *07/27/2012* Test Time: *11:34pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:34pm
FLO	Pass	11:34pm
FC	Pass	11:34pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:35pm
SRC	Pass	11:35pm
DET	Pass	11:35pm
BAR	Pass	11:35pm
BT	Pass	11:35pm

Blank Tests

Test	Status	Time
AIR	Pass	11:35pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:35pm

CRC Tests

Test	Status	Time
COMP	Pass	11:35pm
CAL	Pass	11:35pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Franklin Co - Youngsville Instrument Location Bat Mobile Unit 2

Instrument Serial No. 008929

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Donya B Skinner
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FRANKLIN COUNTY BATMOBILE UNIT 2 340

Serial Number: *008929*
Test Date: *07/27/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Female*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SKINNER, TONYA B*
Permit Number: *13651E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	11:05pm
AIR BLK	.00	11:06pm
ACCY CHK	.08	11:07pm
AIR BLK	.00	11:08pm
SUB TEST	**.00**	**11:08pm**
AIR BLK	.00	11:09pm
SUB TEST	**.00**	**11:11pm**
AIR BLK	.00	11:12pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Tonya B Skinner

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FRANKLIN COUNTY BATMOBILE UNIT 2 340

Serial Number: *008929* Test Record Number: *530*
Test Date: *07/27/2012* Test Time: *11:13pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:13pm
FLO	Pass	11:13pm
FC	Pass	11:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:14pm
SRC	Pass	11:14pm
DET	Pass	11:14pm
BAR	Pass	11:14pm
BT	Pass	11:14pm

Blank Tests

Test	Status	Time
AIR	Pass	11:14pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:14pm

CRC Tests

Test	Status	Time
COMP	Pass	11:15pm
CAL	Pass	11:15pm

Preventive Maintenance
Status: Pass

Donya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Franklin Co - Youngsville Instrument Location Bat Mobile Unit 2

Instrument Serial No. 008601

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Donya B Skinner
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

FRANKLIN COUNTY BAT MOBILE UNIT 2 340

Serial Number: *008601*
Test Date: *07/27/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Female*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SKINNER, TONYA B*
Permit Number: *13651E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG023701
Exp Date: 08/25/2012

Test	g/210L	Time
DIAG	Pass	11:08pm
AIR BLK	.00	11:09pm
ACCY CHK	.07	11:10pm
AIR BLK	.00	11:11pm
SUB TEST	**.00**	**11:12pm**
AIR BLK	.00	11:12pm
SUB TEST	**.00**	**11:14pm**
AIR BLK	.00	11:15pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Tonya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FRANKLIN COUNTY BAT MOBILE UNIT 2 340

Serial Number: *008601* Test Record Number: *684*
Test Date: *07/27/2012* Test Time: *11:19pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:20pm
FLO	Pass	11:20pm
FC	Pass	11:20pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:20pm
SRC	Pass	11:20pm
DET	Pass	11:20pm
BAR	Pass	11:20pm
BT	Pass	11:20pm

Blank Tests

Test	Status	Time
AIR	Pass	11:21pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:21pm

CRC Tests

Test	Status	Time
COMP	Pass	11:21pm
CAL	Pass	11:21pm

Preventive Maintenance
Status: Pass

Tonya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Anson Instrument Location Bat Mobile Unit 2

Instrument Serial No. 008601

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Jonya B Skinner 444

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ANSON COUNTY BAT MOBILE UNIT 2 030

Serial Number: *008601*
Test Date: *07/20/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Female*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SKINNER, TONYA B*
Permit Number: *13651E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG023701
Exp Date: 08/25/2012

Test	g/210L	Time
DIAG	Pass	8:34pm
AIR BLK	.00	8:35pm
ACCY CHK	.07	8:35pm
AIR BLK	.00	8:36pm
SUB TEST	**.00**	**8:37pm**
AIR BLK	.00	8:38pm
SUB TEST	**.00**	**8:40pm**
AIR BLK	.00	8:41pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Tonya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ANSON COUNTY BAT MOBILE UNIT 2 030

Serial Number: *008601* Test Record Number: *681*
Test Date: *07/20/2012* Test Time: *8:43pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:44pm
FLO	Pass	8:44pm
FC	Pass	8:44pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:44pm
SRC	Pass	8:44pm
DET	Pass	8:44pm
BAR	Pass	8:44pm
BT	Pass	8:44pm

Blank Tests

Test	Status	Time
AIR	Pass	8:45pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:45pm

CRC Tests

Test	Status	Time
COMP	Pass	8:45pm
CAL	Pass	8:45pm

Preventive Maintenance
Status: Pass

Donya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CRAVEN Instrument Location BAT Mobile Unit 6

Instrument Serial No. 008869 S. O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY BAT MOBILE UNIT 6 240

Serial Number: *008869* Test Record Number: *808*
Test Date: *07/31/2012* Test Time: *6:33pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:33pm
FLO	Pass	6:33pm
FC	Pass	6:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:33pm
SRC	Pass	6:33pm
DET	Pass	6:33pm
BAR	Pass	6:33pm
BT	Pass	6:33pm

Blank Tests

Test	Status	Time
AIR	Pass	6:34pm

Printer Tests

Test	Status	Time
PRNT	Pass	6:34pm

CRC Tests

Test	Status	Time
COMP	Pass	6:34pm
CAL	Pass	6:34pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

CRAVEN COUNTY BAT MOBILE UNIT 6 240

Serial Number: *008869*
Test Date: *07/31/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	6:25pm
AIR BLK	.00	6:26pm
ACCY CHK	.07	6:27pm
AIR BLK	.00	6:28pm
SUB TEST	**.00**	**6:29pm**
AIR BLK	.00	6:29pm
SUB TEST	**.00**	**6:31pm**
AIR BLK	.00	6:32pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Carteret Instrument Location BAT Mobile unit 6

Instrument Serial No. 008898 Emerald Isle

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] — Signature of Certifying Official

601 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY BAT MOBILE UNIT 6 150

Serial Number: *008898* Test Record Number: *735*
Test Date: *07/03/2012* Test Time: *8:30pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:31pm
FLO	Pass	8:31pm
FC	Pass	8:31pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:31pm
SRC	Pass	8:31pm
DET	Pass	8:31pm
BAR	Pass	8:31pm
BT	Pass	8:31pm

Blank Tests

Test	Status	Time
AIR	Pass	8:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:32pm

CRC Tests

Test	Status	Time
COMP	Pass	8:32pm
CAL	Pass	8:32pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

CARTERET COUNTY BAT MOBILE UNIT 6 150

Serial Number: *008898*
Test Date: *07/03/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG201802
Exp Date: 01/18/2014

Test	g/210L	Time
DIAG	Pass	8:22pm
AIR BLK	.00	8:23pm
ACCY CHK	.07	8:23pm
AIR BLK	.00	8:24pm
SUB TEST	**.00**	**8:25pm**
AIR BLK	.00	8:26pm
SUB TEST	**.00**	**8:28pm**
AIR BLK	.00	8:29pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Carteret Instrument Location BAT Mobile unit 6

Instrument Serial No. 008869 Emerald Isle

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY BAT MOBILE UNIT 6 150

Serial Number: *008869* Test Record Number: *760*
Test Date: *07/03/2012* Test Time: *8:26pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:26pm
FLO	Pass	8:26pm
FC	Pass	8:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:27pm
SRC	Pass	8:27pm
DET	Pass	8:27pm
BAR	Pass	8:27pm
BT	Pass	8:27pm

Blank Tests

Test	Status	Time
AIR	Pass	8:27pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:27pm

CRC Tests

Test	Status	Time
COMP	Pass	8:28pm
CAL	Pass	8:28pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

CARTERET COUNTY BAT MOBILE UNIT 6 150

Serial Number: *008869*
Test Date: *07/03/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	8:18pm
AIR BLK	.00	8:19pm
ACCY CHK	.07	8:20pm
AIR BLK	.00	8:21pm
SUB TEST	**.00**	**8:21pm**
AIR BLK	.00	8:22pm
SUB TEST	**.00**	**8:24pm**
AIR BLK	.00	8:25pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location BAT Mobile unit 6

Instrument Serial No. 008898 Wilmington

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

F. C. [illegible] 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 6 640

Serial Number: *008898* Test Record Number: *739*
Test Date: *07/06/2012* Test Time: *7:49pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:50pm
FLO	Pass	7:50pm
FC	Pass	7:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:50pm
SRC	Pass	7:50pm
DET	Pass	7:50pm
BAR	Pass	7:50pm
BT	Pass	7:50pm

Blank Tests

Test	Status	Time
AIR	Pass	7:51pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:51pm

CRC Tests

Test	Status	Time
COMP	Pass	7:51pm
CAL	Pass	7:51pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY BAT MOBILE UNIT 6
640

Serial Number: *008898*
Test Date: *07/06/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG201802
Exp Date: 01/18/2014

Test	g/210L	Time
DIAG	Pass	7:42pm
AIR BLK	.00	7:43pm
ACCY CHK	.07	7:44pm
AIR BLK	.00	7:44pm
SUB TEST	**.00**	**7:45pm**
AIR BLK	.00	7:46pm
SUB TEST	**.00**	**7:48pm**
AIR BLK	.00	7:48pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location BATMobile unit 4

Instrument Serial No. 008939 Wilmington

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. Rhodes
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 6 640

Serial Number: *008939* Test Record Number: *775*
Test Date: *07/06/2012* Test Time: *7:55pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:56pm
FLO	Pass	7:56pm
FC	Pass	7:56pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:56pm
SRC	Pass	7:56pm
DET	Pass	7:56pm
BAR	Pass	7:56pm
BT	Pass	7:56pm

Blank Tests

Test	Status	Time
AIR	Pass	7:57pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:57pm

CRC Tests

Test	Status	Time
COMP	Pass	7:57pm
CAL	Pass	7:57pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY BAT MOBILE UNIT 6
640

Serial Number: *008939*
Test Date: *07/06/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	7:45pm
AIR BLK	.00	7:46pm
ACCY CHK	.07	7:46pm
AIR BLK	.00	7:47pm
SUB TEST	**.00**	**7:48pm**
AIR BLK	.00	7:48pm
SUB TEST	**.00**	**7:50pm**
AIR BLK	.00	7:51pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location BAT Mobile Unit 6

Instrument Serial No. 008869 Wilmington

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 6 640

Serial Number: *008869* Test Record Number: *764*
Test Date: *07/06/2012* Test Time: *7:50pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:50pm
FLO	Pass	7:50pm
FC	Pass	7:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:51pm
SRC	Pass	7:51pm
DET	Pass	7:51pm
BAR	Pass	7:51pm
BT	Pass	7:51pm

Blank Tests

Test	Status	Time
AIR	Pass	7:51pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:51pm

CRC Tests

Test	Status	Time
COMP	Pass	7:51pm
CAL	Pass	7:51pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY BAT MOBILE UNIT 6
640

Serial Number: *008869*
Test Date: *07/06/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	7:43pm
AIR BLK	.00	7:44pm
ACCY CHK	.07	7:45pm
AIR BLK	.00	7:46pm
SUB TEST	**.00**	**7:46pm**
AIR BLK	.00	7:47pm
SUB TEST	**.00**	**7:48pm**
AIR BLK	.00	7:49pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Craven Instrument Location BAT Mobile Un. 46

Instrument Serial No. 008869 Havelock

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY BAT MOBILE UNIT 6 240

Serial Number: *008869* Test Record Number: *794*
Test Date: *07/27/2012* Test Time: *8:54pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:54pm
FLO	Pass	8:54pm
FC	Pass	8:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:55pm
SRC	Pass	8:55pm
DET	Pass	8:55pm
BAR	Pass	8:55pm
BT	Pass	8:55pm

Blank Tests

Test	Status	Time
AIR	Pass	8:55pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:55pm

CRC Tests

Test	Status	Time
COMP	Pass	8:55pm
CAL	Pass	8:55pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

CRAVEN COUNTY BAT MOBILE UNIT 6 240

Serial Number: *008869*
Test Date: *07/27/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	8:46pm
AIR BLK	.00	8:47pm
ACCY CHK	.08	8:48pm
AIR BLK	.00	8:49pm
SUB TEST	**.00**	**8:50pm**
AIR BLK	.00	8:51pm
SUB TEST	**.00**	**8:52pm**
AIR BLK	.00	8:53pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Columbus Instrument Location BAT Mobile unit 6

Instrument Serial No. 008869 S.O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

COLUMBUS COUNTY BAT MOBILE UNIT 6 230

Serial Number: *008869* Test Record Number: *798*
Test Date: *07/28/2012* Test Time: *10:29pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:30pm
FLO	Pass	10:30pm
FC	Pass	10:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:30pm
SRC	Pass	10:30pm
DET	Pass	10:30pm
BAR	Pass	10:30pm
BT	Pass	10:30pm

Blank Tests

Test	Status	Time
AIR	Pass	10:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:31pm

CRC Tests

Test	Status	Time
COMP	Pass	10:31pm
CAL	Pass	10:31pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

COLUMBUS COUNTY BAT MOBILE UNIT 6 230

Serial Number: *008869*
Test Date: *07/28/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	10:21pm
AIR BLK	.00	10:22pm
ACCY CHK	.07	10:23pm
AIR BLK	.00	10:24pm
SUB TEST	**.00**	**10:24pm**
AIR BLK	.00	10:25pm
SUB TEST	**.00**	**10:27pm**
AIR BLK	.00	10:28pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Columbus Instrument Location BAT Mobile Unit 6

Instrument Serial No. 008898 S. d.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

COLUMBUS COUNTY BAT MOBILE UNIT 6 230

Serial Number: *008898* Test Record Number: *771*
Test Date: *07/28/2012* Test Time: *10:27pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:27pm
FLO	Pass	10:27pm
FC	Pass	10:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:28pm
SRC	Pass	10:28pm
DET	Pass	10:28pm
BAR	Pass	10:28pm
BT	Pass	10:28pm

Blank Tests

Test	Status	Time
AIR	Pass	10:28pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:28pm

CRC Tests

Test	Status	Time
COMP	Pass	10:28pm
CAL	Pass	10:28pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

COLUMBUS COUNTY BAT MOBILE UNIT 6 230

Serial Number: *008898*
Test Date: *07/28/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG201802
Exp Date: 01/18/2014

Test	g/210L	Time
DIAG	Pass	10:18pm
AIR BLK	.00	10:19pm
ACCY CHK	.08	10:20pm
AIR BLK	.00	10:21pm
SUB TEST	**.00**	**10:22pm**
AIR BLK	.00	10:23pm
SUB TEST	**.00**	**10:24pm**
AIR BLK	.00	10:25pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Instrument Location BAT Mobile Unit 6

Instrument Serial No. 068898 Sunset Beach

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: *008898* Test Record Number: *775*
Test Date: *07/29/2012* Test Time: *11:04pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:05pm
FLO	Pass	11:05pm
FC	Pass	11:05pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:05pm
SRC	Pass	11:05pm
DET	Pass	11:05pm
BAR	Pass	11:05pm
BT	Pass	11:05pm

Blank Tests

Test	Status	Time
AIR	Pass	11:05pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:05pm

CRC Tests

Test	Status	Time
COMP	Pass	11:06pm
CAL	Pass	11:06pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: *008898*
Test Date: *07/29/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG201802
Exp Date: 01/18/2014

Test	g/210L	Time
DIAG	Pass	10:55pm
AIR BLK	.00	10:56pm
ACCY CHK	.08	10:57pm
AIR BLK	.00	10:58pm
SUB TEST	**.00**	**11:00pm**
AIR BLK	.00	11:01pm
SUB TEST	**.00**	**11:02pm**
AIR BLK	.00	11:03pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Instrument Location Bat Mobile unit 6

Instrument Serial No. 008869 Sunset Beach

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: *008869* Test Record Number: *803*
Test Date: *07/29/2012* Test Time: *9:55pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:55pm
FLO	Pass	9:55pm
FC	Pass	9:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:56pm
SRC	Pass	9:56pm
DET	Pass	9:56pm
BAR	Pass	9:56pm
BT	Pass	9:56pm

Blank Tests

Test	Status	Time
AIR	Pass	9:56pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:56pm

CRC Tests

Test	Status	Time
COMP	Pass	9:56pm
CAL	Pass	9:56pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: *008869*
Test Date: *07/29/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	9:43pm
AIR BLK	.00	9:44pm
ACCY CHK	.08	9:44pm
AIR BLK	.00	9:45pm
SUB TEST	**.00**	**9:46pm**
AIR BLK	.00	9:47pm
SUB TEST	**.00**	**9:48pm**
AIR BLK	.00	9:49pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location BAT Mobile Unit 6

Instrument Serial No. 008939 S. O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

THE GREAT SEAL OF THE STATE OF NORTH CAROLINA MAY 20, 1775 APRIL 12, 1776 ESSE QUAM VIDERI

K. C. [signature] 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 6 640

Serial Number: *008939* Test Record Number: *797*
Test Date: *07/21/2012* Test Time: *7:25pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:25pm
FLO	Pass	7:25pm
FC	Pass	7:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:26pm
SRC	Pass	7:26pm
DET	Pass	7:26pm
BAR	Pass	7:26pm
BT	Pass	7:26pm

Blank Tests

Test	Status	Time
AIR	Pass	7:26pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:26pm

CRC Tests

Test	Status	Time
COMP	Pass	7:27pm
CAL	Pass	7:27pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY BAT MOBILE UNIT 6
640

Serial Number: *008939*
Test Date: *07/21/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	7:17pm
AIR BLK	.00	7:18pm
ACCY CHK	.08	7:19pm
AIR BLK	.00	7:20pm
SUB TEST	**.00**	**7:20pm**
AIR BLK	.00	7:21pm
SUB TEST	**.00**	**7:23pm**
AIR BLK	.00	7:23pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cumberland Instrument Location Cumberland Co. Detention Facility

Instrument Serial No. 008614 Fayetteville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008614* Test Record Number: *1978*
Test Date: *07/30/2012* Test Time: *11:15am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:16am
FLO	Pass	11:16am
FC	Pass	11:16am

Temperature Tests

Test	Status	Time
FC1	Pass	11:16am
SRC	Pass	11:16am
DET	Pass	11:16am
BAR	Pass	11:16am
BT	Pass	11:16am

Blank Tests

Test	Status	Time
AIR	Pass	11:16am

Printer Tests

Test	Status	Time
PRNT	Pass	11:17am

CRC Tests

Test	Status	Time
COMP	Pass	11:17am
CAL	Pass	11:17am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008614*
Test Date: *07/30/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203102
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	11:06am
AIR BLK	.00	11:07am
ACCY CHK	.07	11:08am
AIR BLK	.00	11:09am
SUB TEST	**.00**	**11:09am**
AIR BLK	.00	11:10am
SUB TEST	**.00**	**11:12am**
AIR BLK	.00	11:13am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cumberland Instrument Location Cumberland Co. Detention Facility

Instrument Serial No. 008633 Fayetteville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008633*
Test Date: *07/30/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123502
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	11:04am
AIR BLK	.00	11:05am
ACCY CHK	.07	11:05am
AIR BLK	.00	11:07am
SUB TEST	**.00**	**11:07am**
AIR BLK	.00	11:08am
SUB TEST	**.00**	**11:10am**
AIR BLK	.00	11:10am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008633* Test Record Number: *2145*
Test Date: *07/30/2012* Test Time: *11:12am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:12am
FLO	Pass	11:12am
FC	Pass	11:12am

Temperature Tests

Test	Status	Time
FC1	Pass	11:12am
SRC	Pass	11:12am
DET	Pass	11:12am
BAR	Pass	11:12am
BT	Pass	11:12am

Blank Tests

Test	Status	Time
AIR	Pass	11:13am

Printer Tests

Test	Status	Time
PRNT	Pass	11:13am

CRC Tests

Test	Status	Time
COMP	Pass	11:13am
CAL	Pass	11:13am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Harnett Instrument Location Harnett Co. Detention Ctr

Instrument Serial No. 008730 Lillington, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

6571
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

HARNETT COUNTY DETENTION CENTER 420

Serial Number: *008730*
Test Date: *07/30/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108202
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	1:08pm
AIR BLK	.00	1:09pm
ACCY CHK	.07	1:09pm
AIR BLK	.00	1:10pm
SUB TEST	**.00**	**1:11pm**
AIR BLK	.00	1:11pm
SUB TEST	**.00**	**1:13pm**
AIR BLK	.00	1:14pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HARNETT COUNTY DETENTION CENTER 420

Serial Number: *008730* Test Record Number: *1297*
Test Date: *07/30/2012* Test Time: *1:16pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:16pm
FLO	Pass	1:16pm
FC	Pass	1:17pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:17pm
SRC	Pass	1:17pm
DET	Pass	1:17pm
BAR	Pass	1:17pm
BT	Pass	1:17pm

Blank Tests

Test	Status	Time
AIR	Pass	1:17pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:17pm

CRC Tests

Test	Status	Time
COMP	Pass	1:17pm
CAL	Pass	1:17pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Hoke Instrument Location Hoke County Detention CTR

Instrument Serial No. 008852 Raeford, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31st day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

HOKE COUNTY DETENTION CENTER 460

Serial Number: *008852*
Test Date: *07/31/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG124904
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	1:43pm
AIR BLK	.00	1:43pm
ACCY CHK	.08	1:44pm
AIR BLK	.00	1:45pm
SUB TEST	**.00**	**1:45pm**
AIR BLK	.00	1:46pm
SUB TEST	**.00**	**1:48pm**
AIR BLK	.00	1:49pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HOKE COUNTY DETENTION CENTER 460

Serial Number: *008852* Test Record Number: *466*
Test Date: *07/31/2012* Test Time: *1:50pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:50pm
FLO	Pass	1:50pm
FC	Pass	1:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:50pm
SRC	Pass	1:50pm
DET	Pass	1:50pm
BAR	Pass	1:50pm
BT	Pass	1:50pm

Blank Tests

Test	Status	Time
AIR	Pass	1:51pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:51pm

CRC Tests

Test	Status	Time
COMP	Pass	1:51pm
CAL	Pass	1:51pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County HOKE Instrument Location HOKE COUNTY DETENTION CTR.

Instrument Serial No. 008855 RAEFORD NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

HOKE COUNTY DETENTION CENTER 460

Serial Number: *008855*
Test Date: *07/31/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG106703
Exp Date: 03/08/2013

Test	g/210L	Time
DIAG	Pass	1:42pm
AIR BLK	.00	1:43pm
ACCY CHK	.08	1:44pm
AIR BLK	.00	1:45pm
SUB TEST	**.00**	**1:45pm**
AIR BLK	.00	1:46pm
SUB TEST	**.00**	**1:48pm**
AIR BLK	.00	1:49pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HOKE COUNTY DETENTION CENTER 460

Serial Number: *008855* Test Record Number: *735*
Test Date: *07/31/2012* Test Time: *1:50pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:50pm
FLO	Pass	1:50pm
FC	Pass	1:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:50pm
SRC	Pass	1:50pm
DET	Pass	1:50pm
BAR	Pass	1:50pm
BT	Pass	1:50pm

Blank Tests

Test	Status	Time
AIR	Pass	1:51pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:51pm

CRC Tests

Test	Status	Time
COMP	Pass	1:51pm
CAL	Pass	1:51pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CUMBERLAND Instrument Location Cumberland Co. Detention Facility

Instrument Serial No. 008672 Fayetteville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008672*
Test Date: *07/30/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG124903
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	11:03am
AIR BLK	.00	11:04am
ACCY CHK	.07	11:04am
AIR BLK	.00	11:05am
SUB TEST	**.00**	**11:06am**
AIR BLK	.00	11:07am
SUB TEST	**.00**	**11:08am**
AIR BLK	.00	11:09am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008672* Test Record Number: *3072*
Test Date: *07/30/2012* Test Time: *11:25am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:26am
FLO	Pass	11:26am
FC	Pass	11:26am

Temperature Tests

Test	Status	Time
FC1	Pass	11:26am
SRC	Pass	11:26am
DET	Pass	11:26am
BAR	Pass	11:26am
BT	Pass	11:26am

Blank Tests

Test	Status	Time
AIR	Pass	11:27am

Printer Tests

Test	Status	Time
PRNT	Pass	11:27am

CRC Tests

Test	Status	Time
COMP	Pass	11:27am
CAL	Pass	11:27am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CUMBERLAND Instrument Location CUMBERLAND Co. DETENTION FACILITY

Instrument Serial No. 008632 Fayetteville NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008632*
Test Date: *07/30/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG124903
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	10:59am
AIR BLK	.00	11:00am
ACCY CHK	.07	11:00am
AIR BLK	.00	11:01am
SUB TEST	**.00**	**11:02am**
AIR BLK	.00	11:03am
SUB TEST	**.00**	**11:04am**
AIR BLK	.00	11:05am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008632* Test Record Number: *2100*
Test Date: *07/30/2012* Test Time: *11:06am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:06am
FLO	Pass	11:06am
FC	Pass	11:06am

Temperature Tests

Test	Status	Time
FC1	Pass	11:06am
SRC	Pass	11:06am
DET	Pass	11:06am
BAR	Pass	11:06am
BT	Pass	11:06am

Blank Tests

Test	Status	Time
AIR	Pass	11:07am

Printer Tests

Test	Status	Time
PRNT	Pass	11:07am

CRC Tests

Test	Status	Time
COMP	Pass	11:07am
CAL	Pass	11:07am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MOORE Instrument Location MOORE Co. JAIL

Instrument Serial No. 008735 CARTHAGE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MOORE COUNTY MOORE COUNTY JAIL 620

Serial Number: *008735*
Test Date: *07/24/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG106703
Exp Date: 03/08/2013

Test	g/210L	Time
DIAG	Pass	2:00pm
AIR BLK	.00	2:00pm
ACCY CHK	.07	2:01pm
AIR BLK	.00	2:02pm
SUB TEST	**.00**	**2:02pm**
AIR BLK	.00	2:03pm
SUB TEST	**.00**	**2:05pm**
AIR BLK	.00	2:06pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MOORE COUNTY MOORE COUNTY JAIL 620

Serial Number: *008735* Test Record Number: *1098*
Test Date: *07/24/2012* Test Time: *2:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:08pm
FLO	Pass	2:08pm
FC	Pass	2:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:08pm
SRC	Pass	2:08pm
DET	Pass	2:08pm
BAR	Pass	2:08pm
BT	Pass	2:08pm

Blank Tests

Test	Status	Time
AIR	Pass	2:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:09pm

CRC Tests

Test	Status	Time
COMP	Pass	2:09pm
CAL	Pass	2:09pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County HARNETT Instrument Location HARNETT CO. DETENTION CTR.

Instrument Serial No. 008729 LILLINGTON NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

HARNETT COUNTY DETENTION CENTER 420

Serial Number: *008729*
Test Date: *07/30/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123502
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	1:10pm
AIR BLK	.00	1:11pm
ACCY CHK	.07	1:12pm
AIR BLK	.00	1:12pm
SUB TEST	**.00**	**1:13pm**
AIR BLK	.00	1:14pm
SUB TEST	**.00**	**1:16pm**
AIR BLK	.00	1:16pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HARNETT COUNTY DETENTION CENTER 420

Serial Number: *008729* Test Record Number: *1523*
Test Date: *07/30/2012* Test Time: *1:18pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:18pm
FLO	Pass	1:18pm
FC	Pass	1:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:18pm
SRC	Pass	1:18pm
DET	Pass	1:18pm
BAR	Pass	1:18pm
BT	Pass	1:18pm

Blank Tests

Test	Status	Time
AIR	Pass	1:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:19pm

CRC Tests

Test	Status	Time
COMP	Pass	1:19pm
CAL	Pass	1:19pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County RANDOLPH Instrument Location RANDOLPH Co. JAIL

Instrument Serial No. 00 8850 Asheboro NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY RANDOLPH COUNTY JAIL
750

Serial Number: *008850*
Test Date: *07/24/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	11:48am
AIR BLK	.00	11:49am
ACCY CHK	.08	11:50am
AIR BLK	.00	11:51am
SUB TEST	**.00**	**11:51am**
AIR BLK	.00	11:52am
SUB TEST	**.00**	**11:54am**
AIR BLK	.00	11:55am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY RANDOLPH COUNTY JAIL 750

Serial Number: *008850* Test Record Number: *459*
Test Date: *07/24/2012* Test Time: *12:05pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:05pm
FLO	Pass	12:05pm
FC	Pass	12:05pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:06pm
SRC	Pass	12:06pm
DET	Pass	12:06pm
BAR	Pass	12:06pm
BT	Pass	12:06pm

Blank Tests

Test	Status	Time
AIR	Pass	12:06pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:06pm

CRC Tests

Test	Status	Time
COMP	Pass	12:06pm
CAL	Pass	12:06pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County RANDOLPH Instrument Location RANDOLPH Co. JAIL

Instrument Serial No. 008860 Asheboro NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY RANDOLPH COUNTY JAIL
750

Serial Number: *008860*
Test Date: *07/24/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG125602
Exp Date: 09/13/2013

Test	g/210L	Time
DIAG	Pass	11:47am
AIR BLK	.00	11:48am
ACCY CHK	.07	11:49am
AIR BLK	.00	11:49am
SUB TEST	**.00**	**11:50am**
AIR BLK	.00	11:51am
SUB TEST	**.00**	**11:52am**
AIR BLK	.00	11:53am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY RANDOLPH COUNTY JAIL 750

Serial Number: *008860* Test Record Number: *1624*
Test Date: *07/24/2012* Test Time: *11:54am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:55am
FLO	Pass	11:55am
FC	Pass	11:55am

Temperature Tests

Test	Status	Time
FC1	Pass	11:55am
SRC	Pass	11:55am
DET	Pass	11:55am
BAR	Pass	11:55am
BT	Pass	11:55am

Blank Tests

Test	Status	Time
AIR	Pass	11:56am

Printer Tests

Test	Status	Time
PRNT	Pass	11:56am

CRC Tests

Test	Status	Time
COMP	Pass	11:56am
CAL	Pass	11:56am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pitt Instrument Location Pitt Co. Detention Center

Instrument Serial No. 008646 124 Detention Dr., Greenville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PITT COUNTY PITT CO DETENTION 730

Serial Number: *008646*
Test Date: *07/24/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG120101
Exp Date: 07/20/2013

Test	g/210L	Time
DIAG	Pass	1:29pm
AIR BLK	.00	1:30pm
ACCY CHK	.07	1:30pm
AIR BLK	.00	1:31pm
SUB TEST	**.00**	**1:32pm**
AIR BLK	.00	1:33pm
SUB TEST	**.00**	**1:34pm**
AIR BLK	.00	1:35pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY PITT CO DETENTION 730

Serial Number: *008646* Test Record Number: *1660*
Test Date: *07/24/2012* Test Time: *1:36pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:36pm
FLO	Pass	1:36pm
FC	Pass	1:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:36pm
SRC	Pass	1:36pm
DET	Pass	1:36pm
BAR	Pass	1:36pm
BT	Pass	1:36pm

Blank Tests

Test	Status	Time
AIR	Pass	1:37pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:37pm

CRC Tests

Test	Status	Time
COMP	Pass	1:37pm
CAL	Pass	1:37pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Currituck Instrument Location Currituck Co.S.O. - Corolla

Instrument Serial No. 008949 1123 Ocean Trail, Corolla, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CURRITUCK COUNTY SO-COROLLA 260

Serial Number: *008949*
Test Date: *07/11/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204602
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	2:00pm
AIR BLK	.00	2:01pm
ACCY CHK	.08	2:02pm
AIR BLK	.00	2:02pm
SUB TEST	**.00**	**2:03pm**
AIR BLK	.00	2:04pm
SUB TEST	**.00**	**2:05pm**
AIR BLK	.00	2:06pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CURRITUCK COUNTY SO-COROLLA 260

Serial Number: *008949* Test Record Number: *236*
Test Date: *07/11/2012* Test Time: *2:07pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:07pm
FLO	Pass	2:07pm
FC	Pass	2:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:08pm
SRC	Pass	2:08pm
DET	Pass	2:08pm
BAR	Pass	2:08pm
BT	Pass	2:08pm

Blank Tests

Test	Status	Time
AIR	Pass	2:08pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:08pm

CRC Tests

Test	Status	Time
COMP	Pass	2:08pm
CAL	Pass	2:08pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Dare Instrument Location Dare Co. Detention Center

Instrument Serial No. 008588 1044 Driftwood Dr., Manteo, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DARE COUNTY DARE CO DETENTION CE 270

Serial Number: *008588*
Test Date: *07/23/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG120101
Exp Date: 07/20/2013

Test	g/210L	Time
DIAG	Pass	12:17pm
AIR BLK	.00	12:18pm
ACCY CHK	.07	12:18pm
AIR BLK	.00	12:19pm
SUB TEST	**.00**	**12:20pm**
AIR BLK	.00	12:21pm
SUB TEST	**.00**	**12:22pm**
AIR BLK	.00	12:23pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DARE COUNTY DARE CO DETENTION CE 270

Serial Number: *008588* Test Record Number: *641*
Test Date: *07/23/2012* Test Time: *12:24pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:24pm
FLO	Pass	12:24pm
FC	Pass	12:24pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:25pm
SRC	Pass	12:25pm
DET	Pass	12:25pm
BAR	Pass	12:25pm
BT	Pass	12:25pm

Blank Tests

Test	Status	Time
AIR	Pass	12:25pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:25pm

CRC Tests

Test	Status	Time
COMP	Pass	12:25pm
CAL	Pass	12:25pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Dare Instrument Location Dare Co. Detention Center

Instrument Serial No. 008783 1044 Driftwood Dr, Manteo, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DARE COUNTY DARE CO DETENTION CE 270

Serial Number: *008783*
Test Date: *07/23/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204602
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	12:11pm
AIR BLK	.00	12:11pm
ACCY CHK	.08	12:12pm
AIR BLK	.00	12:13pm
SUB TEST	**.00**	**12:14pm**
AIR BLK	.00	12:14pm
SUB TEST	**.00**	**12:16pm**
AIR BLK	.00	12:17pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DARE COUNTY DARE CO DETENTION CE 270

Serial Number: *008783* Test Record Number: *340*
Test Date: *07/23/2012* Test Time: *12:18pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:18pm
FLO	Pass	12:18pm
FC	Pass	12:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:18pm
SRC	Pass	12:18pm
DET	Pass	12:18pm
BAR	Pass	12:18pm
BT	Pass	12:18pm

Blank Tests

Test	Status	Time
AIR	Pass	12:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:19pm

CRC Tests

Test	Status	Time
COMP	Pass	12:19pm
CAL	Pass	12:19pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Currituck Instrument Location Currituck Co. S.O.

Instrument Serial No. 008947 407-A Maple Rd., Maple, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CURRITUCK COUNTY CURRITUCK SO-MAPLE 260

Serial Number: *008947*
Test Date: *07/25/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204602
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	11:22am
AIR BLK	.00	11:23am
ACCY CHK	.08	11:23am
AIR BLK	.00	11:24am
SUB TEST	**.00**	**11:25am**
AIR BLK	.00	11:26am
SUB TEST	**.00**	**11:27am**
AIR BLK	.00	11:28am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CURRITUCK COUNTY CURRITUCK SO-MAPLE 260

Serial Number: *008947* Test Record Number: *1047*
Test Date: *07/25/2012* Test Time: *11:29am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:30am
FLO	Pass	11:30am
FC	Pass	11:30am

Temperature Tests

Test	Status	Time
FC1	Pass	11:30am
SRC	Pass	11:30am
DET	Pass	11:30am
BAR	Pass	11:30am
BT	Pass	11:30am

Blank Tests

Test	Status	Time
AIR	Pass	11:31am

Printer Tests

Test	Status	Time
PRNT	Pass	11:31am

CRC Tests

Test	Status	Time
COMP	Pass	11:31am
CAL	Pass	11:31am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Chowan Instrument Location Chowan Co. Public Safety Center

Instrument Serial No. 008895 305 W. Freemason St., Edenton, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26th day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Kelly O — Signature of Certifying Official

643 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CHOWAN COUNTY PUBLIC SAFETY CENTER 200

Serial Number: *008895*
Test Date: *07/26/2012*
Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG104101
Exp Date: 02/10/2013

Test	g/210L	Time
DIAG	Pass	10:14am
AIR BLK	.00	10:15am
ACCY CHK	.07	10:15am
AIR BLK	.00	10:16am
SUB TEST	**.00**	**10:17am**
AIR BLK	.00	10:18am
SUB TEST	**.00**	**10:19am**
AIR BLK	.00	10:20am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CHOWAN COUNTY PUBLIC SAFETY CENTER 200

Serial Number: *008895* Test Record Number: *474*
Test Date: *07/26/2012* Test Time: *10:22am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:22am
FLO	Pass	10:22am
FC	Pass	10:22am

Temperature Tests

Test	Status	Time
FC1	Pass	10:22am
SRC	Pass	10:22am
DET	Pass	10:22am
BAR	Pass	10:22am
BT	Pass	10:22am

Blank Tests

Test	Status	Time
AIR	Pass	10:23am

Printer Tests

Test	Status	Time
PRNT	Pass	10:23am

CRC Tests

Test	Status	Time
COMP	Pass	10:23am
CAL	Pass	10:23am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Johnston Instrument Location Selma Police Dept.

Instrument Serial No. 008595 Selma, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY SELMA PD 500

Serial Number: *008595*
Test Date: *07/23/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG124904
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	12:10pm
AIR BLK	.00	12:11pm
ACCY CHK	.07	12:11pm
AIR BLK	.00	12:13pm
SUB TEST	**.00**	**12:13pm**
AIR BLK	.00	12:14pm
SUB TEST	**.00**	**12:15pm**
AIR BLK	.00	12:16pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY SELMA PD 500

Serial Number: *008595* Test Record Number: *603*
Test Date: *07/23/2012* Test Time: *12:18pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:18pm
FLO	Pass	12:18pm
FC	Pass	12:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:19pm
SRC	Pass	12:19pm
DET	Pass	12:19pm
BAR	Pass	12:19pm
BT	Pass	12:19pm

Blank Tests

Test	Status	Time
AIR	Pass	12:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:19pm

CRC Tests

Test	Status	Time
COMP	Pass	12:20pm
CAL	Pass	12:20pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: *008707*
Test Date: *07/20/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	10:08pm
AIR BLK	.00	10:09pm
ACCY CHK	.08	10:10pm
AIR BLK	.00	10:11pm
SUB TEST	**.00**	**10:12pm**
AIR BLK	.00	10:12pm
SUB TEST	**.00**	**10:14pm**
AIR BLK	.00	10:15pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: *008707* Test Record Number: *1492*
Test Date: *07/20/2012* Test Time: *10:16pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:17pm
FLO	Pass	10:17pm
FC	Pass	10:17pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:17pm
SRC	Pass	10:17pm
DET	Pass	10:17pm
BAR	Pass	10:17pm
BT	Pass	10:17pm

Blank Tests

Test	Status	Time
AIR	Pass	10:17pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:18pm

CRC Tests

Test	Status	Time
COMP	Pass	10:18pm
CAL	Pass	10:18pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Allen Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: *008647*
Test Date: *07/20/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	10:13pm
AIR BLK	.00	10:14pm
ACCY CHK	.07	10:15pm
AIR BLK	.00	10:16pm
SUB TEST	**.00**	**10:16pm**
AIR BLK	.00	10:17pm
SUB TEST	**.00**	**10:19pm**
AIR BLK	.00	10:20pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: *008647* Test Record Number: *1430*
Test Date: *07/20/2012* Test Time: *10:23pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:23pm
FLO	Pass	10:23pm
FC	Pass	10:23pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:23pm
SRC	Pass	10:23pm
DET	Pass	10:23pm
BAR	Pass	10:23pm
BT	Pass	10:23pm

Blank Tests

Test	Status	Time
AIR	Pass	10:24pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:24pm

CRC Tests

Test	Status	Time
COMP	Pass	10:24pm
CAL	Pass	10:24pm

Preventive Maintenance
Status: Pass

Allen Ray Burns
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008616 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: *008616*
Test Date: *07/20/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG023601
Exp Date: 08/24/2012

Test	g/210L	Time
DIAG	Pass	10:10pm
AIR BLK	.00	10:12pm
ACCY CHK	.07	10:12pm
AIR BLK	.00	10:13pm
SUB TEST	**.00**	**10:13pm**
AIR BLK	.00	10:14pm
SUB TEST	**.00**	**10:16pm**
AIR BLK	.00	10:17pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: *008616* Test Record Number: *1432*
Test Date: *07/20/2012* Test Time: *10:17pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:18pm
FLO	Pass	10:18pm
FC	Pass	10:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:18pm
SRC	Pass	10:18pm
DET	Pass	10:18pm
BAR	Pass	10:18pm
BT	Pass	10:18pm

Blank Tests

Test	Status	Time
AIR	Pass	10:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:19pm

CRC Tests

Test	Status	Time
COMP	Pass	10:19pm
CAL	Pass	10:19pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008600 Cary

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official 636 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008600* Test Record Number: *1043*
Test Date: *07/20/2012* Test Time: *11:30pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:30pm
FLO	Pass	11:30pm
FC	Pass	11:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:30pm
SRC	Pass	11:30pm
DET	Pass	11:30pm
BAR	Pass	11:30pm
BT	Pass	11:30pm

Blank Tests

Test	Status	Time
AIR	Pass	11:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:31pm

CRC Tests

Test	Status	Time
COMP	Pass	11:31pm
CAL	Pass	11:31pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008600*
Test Date: *07/20/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	10:59pm
AIR BLK	.00	11:00pm
ACCY CHK	.07	11:00pm
AIR BLK	.00	11:01pm
SUB TEST	**.00**	**11:03pm**
AIR BLK	.00	11:03pm
SUB TEST	**.00**	**11:05pm**
AIR BLK	.00	11:06pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008698 Cary

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008698* Test Record Number: *837*
Test Date: *07/20/2012* Test Time: *11:31pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:31pm
FLO	Pass	11:31pm
FC	Pass	11:32pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:32pm
SRC	Pass	11:32pm
DET	Pass	11:32pm
BAR	Pass	11:32pm
BT	Pass	11:32pm

Blank Tests

Test	Status	Time
AIR	Pass	11:32pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:32pm

CRC Tests

Test	Status	Time
COMP	Pass	11:32pm
CAL	Pass	11:32pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008698*
Test Date: *07/20/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	11:02pm
AIR BLK	.00	11:03pm
ACCY CHK	.07	11:04pm
AIR BLK	.00	11:04pm
SUB TEST	**.00**	**11:05pm**
AIR BLK	.00	11:06pm
SUB TEST	**.00**	**11:08pm**
AIR BLK	.00	11:08pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT MOBILE UNIT

Instrument Serial No. 008788 Cary

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008788* Test Record Number: *671*
Test Date: *07/20/2012* Test Time: *11:30pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:31pm
FLO	Pass	11:31pm
FC	Pass	11:31pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:31pm
SRC	Pass	11:31pm
DET	Pass	11:31pm
BAR	Pass	11:31pm
BT	Pass	11:31pm

Blank Tests

Test	Status	Time
AIR	Pass	11:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:32pm

CRC Tests

Test	Status	Time
COMP	Pass	11:32pm
CAL	Pass	11:32pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008788*
Test Date: *07/20/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123502
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	11:01pm
AIR BLK	.00	11:01pm
ACCY CHK	.07	11:02pm
AIR BLK	.00	11:03pm
SUB TEST	**.00**	**11:04pm**
AIR BLK	.00	11:05pm
SUB TEST	**.00**	**11:06pm**
AIR BLK	.00	11:07pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008600 WENDELL

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008600* Test Record Number: *1048*
Test Date: *07/21/2012* Test Time: *11:51pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:52pm
FLO	Pass	11:52pm
FC	Pass	11:52pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:52pm
SRC	Pass	11:52pm
DET	Pass	11:52pm
BAR	Pass	11:52pm
BT	Pass	11:52pm

Blank Tests

Test	Status	Time
AIR	Pass	11:52pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:52pm

CRC Tests

Test	Status	Time
COMP	Pass	11:52pm
CAL	Pass	11:52pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008600*
Test Date: *07/21/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	11:42pm
AIR BLK	.00	11:43pm
ACCY CHK	.07	11:43pm
AIR BLK	.00	11:45pm
SUB TEST	**.00**	**11:46pm**
AIR BLK	.00	11:47pm
SUB TEST	**.00**	**11:48pm**
AIR BLK	.00	11:49pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008698 WENDELL

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008698* Test Record Number: *840*
Test Date: *07/22/2012* Test Time: *12:18am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:18am
FLO	Pass	12:18am
FC	Pass	12:18am

Temperature Tests

Test	Status	Time
FC1	Pass	12:18am
SRC	Pass	12:18am
DET	Pass	12:18am
BAR	Pass	12:18am
BT	Pass	12:18am

Blank Tests

Test	Status	Time
AIR	Pass	12:19am

Printer Tests

Test	Status	Time
PRNT	Pass	12:19am

CRC Tests

Test	Status	Time
COMP	Pass	12:19am
CAL	Pass	12:19am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008698*
Test Date: *07/22/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	12:10am
AIR BLK	.00	12:11am
ACCY CHK	.08	12:11am
AIR BLK	.00	12:12am
SUB TEST	**.00**	**12:13am**
AIR BLK	.00	12:14am
SUB TEST	**.00**	**12:15am**
AIR BLK	.00	12:16am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008788 WENDELL

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008788* Test Record Number: *675*
Test Date: *07/21/2012* Test Time: *11:54pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:54pm
FLO	Pass	11:54pm
FC	Pass	11:54pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:54pm
SRC	Pass	11:54pm
DET	Pass	11:54pm
BAR	Pass	11:54pm
BT	Pass	11:54pm

Blank Tests

Test	Status	Time
AIR	Pass	11:55pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:55pm

CRC Tests

Test	Status	Time
COMP	Pass	11:55pm
CAL	Pass	11:55pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008788*
Test Date: *07/21/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123502
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	11:44pm
AIR BLK	.00	11:45pm
ACCY CHK	.07	11:46pm
AIR BLK	.00	11:47pm
SUB TEST	**.00**	**11:48pm**
AIR BLK	.00	11:49pm
SUB TEST	**.00**	**11:51pm**
AIR BLK	.00	11:52pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MOORE Instrument Location ROBBINS POLICE DEPT.

Instrument Serial No. 008728 ROBBINS NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MOORE COUNTY ROBBINS PD 620

Serial Number: *008728*
Test Date: *07/18/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	10:29am
AIR BLK	.00	10:30am
ACCY CHK	.08	10:31am
AIR BLK	.00	10:31am
SUB TEST	**.00**	**10:32am**
AIR BLK	.00	10:33am
SUB TEST	**.00**	**10:35am**
AIR BLK	.00	10:36am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MOORE COUNTY ROBBINS PD 620

Serial Number: *008728* Test Record Number: *211*
Test Date: *07/18/2012* Test Time: *10:37am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:37am
FLO	Pass	10:37am
FC	Pass	10:37am

Temperature Tests

Test	Status	Time
FC1	Pass	10:37am
SRC	Pass	10:37am
DET	Pass	10:37am
BAR	Pass	10:37am
BT	Pass	10:37am

Blank Tests

Test	Status	Time
AIR	Pass	10:38am

Printer Tests

Test	Status	Time
PRNT	Pass	10:38am

CRC Tests

Test	Status	Time
COMP	Pass	10:38am
CAL	Pass	10:38am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County LEE Instrument Location LEE Co. JAIL

Instrument Serial No. 008645 SANFORD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

LEE COUNTY LEE CO. LEC. 520

Serial Number: *008645* Test Record Number: *1118*
Test Date: *07/23/2012* Test Time: *11:10am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:11am
FLO	Pass	11:11am
FC	Pass	11:11am

Temperature Tests

Test	Status	Time
FC1	Pass	11:11am
SRC	Pass	11:11am
DET	Pass	11:11am
BAR	Pass	11:11am
BT	Pass	11:11am

Blank Tests

Test	Status	Time
AIR	Pass	11:12am

Printer Tests

Test	Status	Time
PRNT	Pass	11:12am

CRC Tests

Test	Status	Time
COMP	Pass	11:12am
CAL	Pass	11:12am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

LEE COUNTY LEE CO. LEC. 520

Serial Number: *008645*
Test Date: *07/23/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG104004
Exp Date: 02/09/2013

Test	g/210L	Time
DIAG	Pass	11:03am
AIR BLK	.00	11:03am
ACCY CHK	.08	11:04am
AIR BLK	.00	11:05am
SUB TEST	**.00**	**11:05am**
AIR BLK	.00	11:06am
SUB TEST	**.00**	**11:08am**
AIR BLK	.00	11:09am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Instrument Location BAT Mobile Unit 6

Instrument Serial No. 008869 5HP

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus 2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: *008869* Test Record Number: *771*
Test Date: *07/13/2012* Test Time: *9:22pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:22pm
FLO	Pass	9:22pm
FC	Pass	9:23pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:23pm
SRC	Pass	9:23pm
DET	Pass	9:23pm
BAR	Pass	9:23pm
BT	Pass	9:23pm

Blank Tests

Test	Status	Time
AIR	Pass	9:23pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:23pm

CRC Tests

Test	Status	Time
COMP	Pass	9:24pm
CAL	Pass	9:24pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: *008869*
Test Date: *07/13/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	9:14pm
AIR BLK	.00	9:15pm
ACCY CHK	.08	9:16pm
AIR BLK	.00	9:17pm
SUB TEST	**.00**	**9:17pm**
AIR BLK	.00	9:18pm
SUB TEST	**.00**	**9:20pm**
AIR BLK	.00	9:21pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Instrument Location BAT Mobile Unit 6

Instrument Serial No. 008898 S. O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: *008898* Test Record Number: *745*
Test Date: *07/07/2012* Test Time: *4:42pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:42pm
FLO	Pass	4:42pm
FC	Pass	4:42pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:42pm
SRC	Pass	4:42pm
DET	Pass	4:42pm
BAR	Pass	4:42pm
BT	Pass	4:42pm

Blank Tests

Test	Status	Time
AIR	Pass	4:43pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:43pm

CRC Tests

Test	Status	Time
COMP	Pass	4:43pm
CAL	Pass	4:43pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: *008898*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG201802
Exp Date: 01/18/2014

Test	g/210L	Time
DIAG	Pass	4:34pm
AIR BLK	.00	4:35pm
ACCY CHK	.07	4:36pm
AIR BLK	.00	4:36pm
SUB TEST	**.00**	**4:37pm**
AIR BLK	.00	4:38pm
SUB TEST	**.00**	**4:40pm**
AIR BLK	.00	4:41pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Instrument Location BAT Mobile unit 6

Instrument Serial No. 008869 S.O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: *008869* Test Record Number: *768*
Test Date: *07/07/2012* Test Time: *4:42pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:43pm
FLO	Pass	4:43pm
FC	Pass	4:43pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:43pm
SRC	Pass	4:43pm
DET	Pass	4:43pm
BAR	Pass	4:43pm
BT	Pass	4:43pm

Blank Tests

Test	Status	Time
AIR	Pass	4:44pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:44pm

CRC Tests

Test	Status	Time
COMP	Pass	4:44pm
CAL	Pass	4:44pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: *008869*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	4:34pm
AIR BLK	.00	4:35pm
ACCY CHK	.07	4:36pm
AIR BLK	.00	4:37pm
SUB TEST	**.00**	**4:38pm**
AIR BLK	.00	4:39pm
SUB TEST	**.00**	**4:41pm**
AIR BLK	.00	4:42pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Instrument Location BAT Mobile unit

Instrument Serial No. 008939 S.O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 6 640

Serial Number: *008939* Test Record Number: *779*
Test Date: *07/07/2012* Test Time: *4:43pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:44pm
FLO	Pass	4:44pm
FC	Pass	4:44pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:44pm
SRC	Pass	4:44pm
DET	Pass	4:44pm
BAR	Pass	4:44pm
BT	Pass	4:44pm

Blank Tests

Test	Status	Time
AIR	Pass	4:45pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:45pm

CRC Tests

Test	Status	Time
COMP	Pass	4:45pm
CAL	Pass	4:45pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 6 640

Serial Number: *008939*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	4:35pm
AIR BLK	.00	4:36pm
ACCY CHK	.07	4:37pm
AIR BLK	.00	4:38pm
SUB TEST	**.00**	**4:39pm**
AIR BLK	.00	4:39pm
SUB TEST	**.00**	**4:41pm**
AIR BLK	.00	4:42pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Instrument Location BAT Mobile Unit 6

Instrument Serial No. 008898 SHT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: *008898* Test Record Number: *748*
Test Date: *07/13/2012* Test Time: *9:22pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:22pm
FLO	Pass	9:22pm
FC	Pass	9:22pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:22pm
SRC	Pass	9:22pm
DET	Pass	9:22pm
BAR	Pass	9:22pm
BT	Pass	9:22pm

Blank Tests

Test	Status	Time
AIR	Pass	9:23pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:23pm

CRC Tests

Test	Status	Time
COMP	Pass	9:23pm
CAL	Pass	9:23pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: *008898*
Test Date: *07/13/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG201802
Exp Date: 01/18/2014

Test	g/210L	Time
DIAG	Pass	9:14pm
AIR BLK	.00	9:15pm
ACCY CHK	.08	9:15pm
AIR BLK	.00	9:16pm
SUB TEST	**.00**	**9:17pm**
AIR BLK	.00	9:18pm
SUB TEST	**.00**	**9:19pm**
AIR BLK	.00	9:20pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Instrument Location BAT Mobile unit 6

Instrument Serial No. 008939 SHP

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 6 640

Serial Number: *008939* Test Record Number: *783*
Test Date: *07/13/2012* Test Time: *9:23pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:24pm
FLO	Pass	9:24pm
FC	Pass	9:24pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:24pm
SRC	Pass	9:24pm
DET	Pass	9:24pm
BAR	Pass	9:24pm
BT	Pass	9:24pm

Blank Tests

Test	Status	Time
AIR	Pass	9:25pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:25pm

CRC Tests

Test	Status	Time
COMP	Pass	9:25pm
CAL	Pass	9:25pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

BRUNSWICK COUNTY BAT MOBILE UNIT 6 640

Serial Number: *008939*
Test Date: *07/13/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	9:15pm
AIR BLK	.00	9:16pm
ACCY CHK	.08	9:17pm
AIR BLK	.00	9:18pm
SUB TEST	**.00**	**9:18pm**
AIR BLK	.00	9:19pm
SUB TEST	**.00**	**9:21pm**
AIR BLK	.00	9:22pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location BAT Mobile unit 6

Instrument Serial No. 008898 SHP

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. [illegible]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 6 640

Serial Number: *008898* Test Record Number: *754*
Test Date: *07/14/2012* Test Time: *7:32pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:33pm
FLO	Pass	7:33pm
FC	Pass	7:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:33pm
SRC	Pass	7:33pm
DET	Pass	7:33pm
BAR	Pass	7:33pm
BT	Pass	7:33pm

Blank Tests

Test	Status	Time
AIR	Pass	7:33pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:34pm

CRC Tests

Test	Status	Time
COMP	Pass	7:34pm
CAL	Pass	7:34pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY BAT MOBILE UNIT 6
640

Serial Number: *008898*
Test Date: *07/14/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG201802
Exp Date: 01/18/2014

Test	g/210L	Time
DIAG	Pass	7:23pm
AIR BLK	.00	7:24pm
ACCY CHK	.07	7:25pm
AIR BLK	.00	7:26pm
SUB TEST	**.00**	**7:26pm**
AIR BLK	.00	7:27pm
SUB TEST	**.00**	**7:29pm**
AIR BLK	.00	7:30pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location BAT Mobile unit 6

Instrument Serial No. 008939 SHP

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 6 640

Serial Number: *008939* Test Record Number: *788*
Test Date: *07/14/2012* Test Time: *7:37pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:38pm
FLO	Pass	7:38pm
FC	Pass	7:38pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:38pm
SRC	Pass	7:38pm
DET	Pass	7:38pm
BAR	Pass	7:38pm
BT	Pass	7:38pm

Blank Tests

Test	Status	Time
AIR	Pass	7:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:39pm

CRC Tests

Test	Status	Time
COMP	Pass	7:39pm
CAL	Pass	7:39pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY BAT MOBILE UNIT 6
640

Serial Number: *008939*
Test Date: *07/14/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	7:25pm
AIR BLK	.00	7:26pm
ACCY CHK	.07	7:27pm
AIR BLK	.00	7:28pm
SUB TEST	**.00**	**7:28pm**
AIR BLK	.00	7:29pm
SUB TEST	**.00**	**7:31pm**
AIR BLK	.00	7:31pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location BATMobile unit 6

Instrument Serial No. 008869 SHP

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 6 640

Serial Number: *008869* Test Record Number: *776*
Test Date: *07/14/2012* Test Time: *7:37pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:37pm
FLO	Pass	7:37pm
FC	Pass	7:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:38pm
SRC	Pass	7:38pm
DET	Pass	7:38pm
BAR	Pass	7:38pm
BT	Pass	7:38pm

Blank Tests

Test	Status	Time
AIR	Pass	7:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:38pm

CRC Tests

Test	Status	Time
COMP	Pass	7:38pm
CAL	Pass	7:38pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY BAT MOBILE UNIT 6
640

Serial Number: *008869*
Test Date: *07/14/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	7:24pm
AIR BLK	.00	7:25pm
ACCY CHK	.07	7:26pm
AIR BLK	.00	7:27pm
SUB TEST	**.00**	**7:27pm**
AIR BLK	.00	7:28pm
SUB TEST	**.00**	**7:30pm**
AIR BLK	.00	7:31pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Robeson Instrument Location BAT Mobile unit 6

Instrument Serial No. 008869 Red Springs

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] — Signature of Certifying Official

601 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY BAT MOBILE UNIT 6 770

Serial Number: *008869* Test Record Number: *787*
Test Date: *07/20/2012* Test Time: *5:36pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:36pm
FLO	Pass	5:36pm
FC	Pass	5:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:36pm
SRC	Pass	5:36pm
DET	Pass	5:36pm
BAR	Pass	5:36pm
BT	Pass	5:36pm

Blank Tests

Test	Status	Time
AIR	Pass	5:37pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:37pm

CRC Tests

Test	Status	Time
COMP	Pass	5:37pm
CAL	Pass	5:37pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

ROBESON COUNTY BAT MOBILE UNIT 6 770

Serial Number: *008869*
Test Date: *07/20/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	5:27pm
AIR BLK	.00	5:28pm
ACCY CHK	.07	5:28pm
AIR BLK	.00	5:29pm
SUB TEST	**.00**	**5:30pm**
AIR BLK	.00	5:31pm
SUB TEST	**.00**	**5:32pm**
AIR BLK	.00	5:33pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location BAT Mobile Unit 6

Instrument Serial No. 008898 S.O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K.C. [illegible]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 6 640

Serial Number: *008898* Test Record Number: *766*
Test Date: *07/21/2012* Test Time: *7:24pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:25pm
FLO	Pass	7:25pm
FC	Pass	7:25pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:25pm
SRC	Pass	7:25pm
DET	Pass	7:25pm
BAR	Pass	7:25pm
BT	Pass	7:25pm

Blank Tests

Test	Status	Time
AIR	Pass	7:26pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:26pm

CRC Tests

Test	Status	Time
COMP	Pass	7:26pm
CAL	Pass	7:26pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY BAT MOBILE UNIT 6
640

Serial Number: *008898*
Test Date: *07/21/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG201802
Exp Date: 01/18/2014

Test	g/210L	Time
DIAG	Pass	7:16pm
AIR BLK	.00	7:17pm
ACCY CHK	.08	7:17pm
AIR BLK	.00	7:18pm
SUB TEST	**.00**	**7:19pm**
AIR BLK	.00	7:20pm
SUB TEST	**.00**	**7:21pm**
AIR BLK	.00	7:22pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location BAT Mobile Unit 6

Instrument Serial No. 008869 S.O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 6 640

Serial Number: *008869* Test Record Number: *790*
Test Date: *07/21/2012* Test Time: *7:24pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:24pm
FLO	Pass	7:24pm
FC	Pass	7:25pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:25pm
SRC	Pass	7:25pm
DET	Pass	7:25pm
BAR	Pass	7:25pm
BT	Pass	7:25pm

Blank Tests

Test	Status	Time
AIR	Pass	7:25pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:25pm

CRC Tests

Test	Status	Time
COMP	Pass	7:26pm
CAL	Pass	7:26pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY BAT MOBILE UNIT 6
640

Serial Number: *008869*
Test Date: *07/21/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	7:16pm
AIR BLK	.00	7:17pm
ACCY CHK	.07	7:18pm
AIR BLK	.00	7:19pm
SUB TEST	**.00**	**7:19pm**
AIR BLK	.00	7:20pm
SUB TEST	**.00**	**7:22pm**
AIR BLK	.00	7:23pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County NASH Instrument Location NASH Co. JAIL

Instrument Serial No. 008630 NASHVILLE, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NASH COUNTY NASH COUNTY JAIL 630

Serial Number: *008630*
Test Date: *07/19/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG125603
Exp Date: 09/13/2013

Test	g/210L	Time
DIAG	Pass	1:40pm
AIR BLK	.00	1:41pm
ACCY CHK	.08	1:42pm
AIR BLK	.00	1:43pm
SUB TEST	**.00**	**1:43pm**
AIR BLK	.00	1:44pm
SUB TEST	**.00**	**1:46pm**
AIR BLK	.00	1:46pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NASH COUNTY NASH COUNTY JAIL 630

Serial Number: *008630* Test Record Number: *2450*
Test Date: *07/19/2012* Test Time: *1:47pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:47pm
FLO	Pass	1:47pm
FC	Pass	1:48pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:48pm
SRC	Pass	1:48pm
DET	Pass	1:48pm
BAR	Pass	1:48pm
BT	Pass	1:48pm

Blank Tests

Test	Status	Time
AIR	Pass	1:48pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:48pm

CRC Tests

Test	Status	Time
COMP	Pass	1:48pm
CAL	Pass	1:48pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County NASH Instrument Location Rocky Mount P.D.

Instrument Serial No. 008740 #1 Goverment Piaza Rocky Mount, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: *008740*
Test Date: *07/19/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG104101
Exp Date: 02/10/2013

Test	g/210L	Time
DIAG	Pass	11:30am
AIR BLK	.00	11:31am
ACCY CHK	.07	11:32am
AIR BLK	.00	11:33am
SUB TEST	**.00**	**11:34am**
AIR BLK	.00	11:35am
SUB TEST	**.00**	**11:36am**
AIR BLK	.00	11:37am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: *008740* Test Record Number: *371*
Test Date: *07/19/2012* Test Time: *11:38am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:38am
FLO	Pass	11:38am
FC	Pass	11:39am

Temperature Tests

Test	Status	Time
FC1	Pass	11:39am
SRC	Pass	11:39am
DET	Pass	11:39am
BAR	Pass	11:39am
BT	Pass	11:39am

Blank Tests

Test	Status	Time
AIR	Pass	11:39am

Printer Tests

Test	Status	Time
PRNT	Pass	11:39am

CRC Tests

Test	Status	Time
COMP	Pass	11:40am
CAL	Pass	11:40am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County NASH Instrument Location ROCKY MOUNT P.D

Instrument Serial No. 008741 #1 GOVERMENT PLAZA ROCKY MOUNT, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: *008741*
Test Date: *07/19/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG125603
Exp Date: 09/13/2013

Test	g/210L	Time
DIAG	Pass	11:27am
AIR BLK	.00	11:27am
ACCY CHK	.07	11:28am
AIR BLK	.00	11:29am
SUB TEST	**.00**	**11:29am**
AIR BLK	.00	11:30am
SUB TEST	**.00**	**11:32am**
AIR BLK	.00	11:33am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: *008741* Test Record Number: *1058*
Test Date: *07/19/2012* Test Time: *11:33am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:34am
FLO	Pass	11:34am
FC	Pass	11:34am

Temperature Tests

Test	Status	Time
FC1	Pass	11:34am
SRC	Pass	11:34am
DET	Pass	11:34am
BAR	Pass	11:34am
BT	Pass	11:34am

Blank Tests

Test	Status	Time
AIR	Pass	11:35am

Printer Tests

Test	Status	Time
PRNT	Pass	11:35am

CRC Tests

Test	Status	Time
COMP	Pass	11:35am
CAL	Pass	11:35am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location APEX P.S.S. #4

Instrument Serial No. 008621 1615 E. WILLIAMS ST. APEX

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WAKE COUNTY APEX PD
Serial Number: *008621*
Test Date: *07/17/2012*
Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123502
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	3:57pm
AIR BLK	.00	3:58pm
ACCY CHK	.07	3:58pm
AIR BLK	.00	3:59pm
SUB TEST	**.00**	**4:00pm**
AIR BLK	.00	4:01pm
SUB TEST	**.00**	**4:02pm**
AIR BLK	.00	4:03pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY APEX PD
Serial Number: *008621* Test Record Number: *1056*
Test Date: *07/17/2012* Test Time: *4:06pm EDT*
System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:07pm
FLO	Pass	4:07pm
FC	Pass	4:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:07pm
SRC	Pass	4:07pm
DET	Pass	4:07pm
BAR	Pass	4:07pm
BT	Pass	4:07pm

Blank Tests

Test	Status	Time
AIR	Pass	4:08pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:08pm

CRC Tests

Test	Status	Time
COMP	Pass	4:08pm
CAL	Pass	4:08pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WARREN Instrument Location WARREN CO. JAIL

Instrument Serial No. 008793 HWY 58 WARRENTON, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WARREN COUNTY WARREN COUNTY JAIL 920

Serial Number: *008793*
Test Date: *07/16/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	12:38pm
AIR BLK	.00	12:38pm
ACCY CHK	.08	12:39pm
AIR BLK	.00	12:40pm
SUB TEST	**.00**	**12:41pm**
AIR BLK	.00	12:41pm
SUB TEST	**.00**	**12:43pm**
AIR BLK	.00	12:44pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WARREN COUNTY WARREN COUNTY JAIL 920

Serial Number: *008793* Test Record Number: *562*
Test Date: *07/16/2012* Test Time: *12:45pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:45pm
FLO	Pass	12:45pm
FC	Pass	12:45pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:45pm
SRC	Pass	12:45pm
DET	Pass	12:45pm
BAR	Pass	12:45pm
BT	Pass	12:45pm

Blank Tests

Test	Status	Time
AIR	Pass	12:46pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:46pm

CRC Tests

Test	Status	Time
COMP	Pass	12:46pm
CAL	Pass	12:46pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WARREN Instrument Location NORLINA P.D

Instrument Serial No. 008945 101 MAIN ST. NORLINA, N.C

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WARREN COUNTY NORLINA POLICE DEPT 920

Serial Number: *008945*
Test Date: *07/16/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG104101
Exp Date: 02/10/2013

Test	g/210L	Time
DIAG	Pass	11:30am
AIR BLK	.00	11:31am
ACCY CHK	.07	11:32am
AIR BLK	.00	11:33am
SUB TEST	**.00**	**11:33am**
AIR BLK	.00	11:34am
SUB TEST	**.00**	**11:36am**
AIR BLK	.00	11:37am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WARREN COUNTY NORLINA POLICE DEPT 920

Serial Number: *008945* Test Record Number: *229*
Test Date: *07/16/2012* Test Time: *11:38am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:38am
FLO	Pass	11:38am
FC	Pass	11:39am

Temperature Tests

Test	Status	Time
FC1	Pass	11:39am
SRC	Pass	11:39am
DET	Pass	11:39am
BAR	Pass	11:39am
BT	Pass	11:39am

Blank Tests

Test	Status	Time
AIR	Pass	11:39am

Printer Tests

Test	Status	Time
PRNT	Pass	11:39am

CRC Tests

Test	Status	Time
COMP	Pass	11:39am
CAL	Pass	11:39am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location WAKE FOREST P.D.

Instrument Serial No. 008738 401 OWEN ST. WAKE FOREST, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WAKE COUNTY WAKE FOREST PD 910

Serial Number: *008738*
Test Date: *07/12/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG106703
Exp Date: 03/08/2013

Test	g/210L	Time
DIAG	Pass	3:20pm
AIR BLK	.00	3:21pm
ACCY CHK	.07	3:21pm
AIR BLK	.00	3:22pm
SUB TEST	**.00**	**3:23pm**
AIR BLK	.00	3:24pm
SUB TEST	**.00**	**3:25pm**
AIR BLK	.00	3:26pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY WAKE FOREST PD 910

Serial Number: *008738* Test Record Number: *222*
Test Date: *07/12/2012* Test Time: *3:27pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:27pm
FLO	Pass	3:27pm
FC	Pass	3:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:28pm
SRC	Pass	3:28pm
DET	Pass	3:28pm
BAR	Pass	3:28pm
BT	Pass	3:28pm

Blank Tests

Test	Status	Time
AIR	Pass	3:28pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:28pm

CRC Tests

Test	Status	Time
COMP	Pass	3:28pm
CAL	Pass	3:28pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location DETENTION CENTER

Instrument Serial No. 008686 3301 HAMMOND RD. RALEIGH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WAKE COUNTY DETENTION CENTER 910

Serial Number: *008686*
Test Date: *07/11/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	10:43am
AIR BLK	.00	10:44am
ACCY CHK	.08	10:45am
AIR BLK	.00	10:46am
SUB TEST	**.00**	**10:47am**
AIR BLK	.00	10:48am
SUB TEST	**.00**	**10:49am**
AIR BLK	.00	10:50am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY DETENTION CENTER 910

Serial Number: *008686* Test Record Number: *3476*
Test Date: *07/11/2012* Test Time: *10:51am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:51am
FLO	Pass	10:51am
FC	Pass	10:51am

Temperature Tests

Test	Status	Time
FC1	Pass	10:51am
SRC	Pass	10:51am
DET	Pass	10:51am
BAR	Pass	10:51am
BT	Pass	10:51am

Blank Tests

Test	Status	Time
AIR	Pass	10:52am

Printer Tests

Test	Status	Time
PRNT	Pass	10:52am

CRC Tests

Test	Status	Time
COMP	Pass	10:52am
CAL	Pass	10:52am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Bat Mobile Unit 2

Instrument Serial No. D08736

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Donya B Skinner
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 2 910

Serial Number: *008736*
Test Date: *07/13/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SKINNER, TONYA B*
Permit Number: *13651E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	10:31pm
AIR BLK	.00	10:32pm
ACCY CHK	.07	10:33pm
AIR BLK	.00	10:33pm
SUB TEST	**.00**	**10:34pm**
AIR BLK	.00	10:35pm
SUB TEST	**.00**	**10:36pm**
AIR BLK	.00	10:37pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Tonya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 2 910

Serial Number: *008736* Test Record Number: *479*
Test Date: *07/13/2012* Test Time: *10:38pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:39pm
FLO	Pass	10:39pm
FC	Pass	10:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:39pm
SRC	Pass	10:39pm
DET	Pass	10:39pm
BAR	Pass	10:39pm
BT	Pass	10:39pm

Blank Tests

Test	Status	Time
AIR	Pass	10:40pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:40pm

CRC Tests

Test	Status	Time
COMP	Pass	10:40pm
CAL	Pass	10:40pm

Preventive Maintenance
Status: Pass

Donya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Bat Mobile Unit 2

Instrument Serial No. 008929

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Donya B Skinner
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BATMOBILE UNIT 2 910

Serial Number: *008929* Test Record Number: *525*
Test Date: *07/13/2012* Test Time: *10:39pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:39pm
FLO	Pass	10:39pm
FC	Pass	10:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:40pm
SRC	Pass	10:40pm
DET	Pass	10:40pm
BAR	Pass	10:40pm
BT	Pass	10:40pm

Blank Tests

Test	Status	Time
AIR	Pass	10:40pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:40pm

CRC Tests

Test	Status	Time
COMP	Pass	10:41pm
CAL	Pass	10:41pm

Preventive Maintenance
Status: Pass

Tonya B Skinner

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BATMOBILE UNIT 2 910

Serial Number: *008929*
Test Date: *07/13/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SKINNER, TONYA B*
Permit Number: *13651E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	10:29pm
AIR BLK	.00	10:30pm
ACCY CHK	.08	10:31pm
AIR BLK	.00	10:32pm
SUB TEST	**.00**	**10:33pm**
AIR BLK	.00	10:34pm
SUB TEST	**.00**	**10:35pm**
AIR BLK	.00	10:36pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Tonya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County DURHAM Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008698 DURHAM

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7TH day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY BAT MOBILE UNIT 5 310

Serial Number: *008698* Test Record Number: *826*
Test Date: *07/07/2012* Test Time: *10:33pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:34pm
FLO	Pass	10:34pm
FC	Pass	10:34pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:34pm
SRC	Pass	10:34pm
DET	Pass	10:34pm
BAR	Pass	10:34pm
BT	Pass	10:34pm

Blank Tests

Test	Status	Time
AIR	Pass	10:35pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:35pm

CRC Tests

Test	Status	Time
COMP	Pass	10:35pm
CAL	Pass	10:35pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

DURHAM COUNTY BAT MOBILE UNIT 5 310

Serial Number: *008698*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	9:47pm
AIR BLK	.00	9:48pm
ACCY CHK	.07	9:48pm
AIR BLK	.00	9:49pm
SUB TEST	**.00**	**9:50pm**
AIR BLK	.00	9:51pm
SUB TEST	**.00**	**9:52pm**
AIR BLK	.00	9:53pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BAT MOBILE UNIT #3

Instrument Serial No. 00 8600 RALEIGH

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6TH day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008600* Test Record Number: *1008*
Test Date: *07/06/2012* Test Time: *12:12am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:12am
FLO	Pass	12:12am
FC	Pass	12:12am

Temperature Tests

Test	Status	Time
FC1	Pass	12:12am
SRC	Pass	12:12am
DET	Pass	12:12am
BAR	Pass	12:12am
BT	Pass	12:12am

Blank Tests

Test	Status	Time
AIR	Pass	12:13am

Printer Tests

Test	Status	Time
PRNT	Pass	12:13am

CRC Tests

Test	Status	Time
COMP	Pass	12:13am
CAL	Pass	12:13am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008600*
Test Date: *07/06/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	12:02am
AIR BLK	.00	12:03am
ACCY CHK	.08	12:03am
AIR BLK	.00	12:04am
SUB TEST	**.00**	**12:06am**
AIR BLK	.00	12:07am
SUB TEST	**.00**	**12:09am**
AIR BLK	.00	12:10am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008698 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008698* Test Record Number: *817*
Test Date: *07/06/2012* Test Time: *12:13am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:13am
FLO	Pass	12:13am
FC	Pass	12:13am

Temperature Tests

Test	Status	Time
FC1	Pass	12:14am
SRC	Pass	12:14am
DET	Pass	12:14am
BAR	Pass	12:14am
BT	Pass	12:14am

Blank Tests

Test	Status	Time
AIR	Pass	12:14am

Printer Tests

Test	Status	Time
PRNT	Pass	12:14am

CRC Tests

Test	Status	Time
COMP	Pass	12:14am
CAL	Pass	12:14am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008698*
Test Date: *07/06/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	12:04am
AIR BLK	.00	12:05am
ACCY CHK	.08	12:06am
AIR BLK	.00	12:07am
SUB TEST	**.00**	**12:08am**
AIR BLK	.00	12:08am
SUB TEST	**.00**	**12:10am**
AIR BLK	.00	12:11am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008600 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008600* Test Record Number: *1015*
Test Date: *07/07/2012* Test Time: *12:23am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:23am
FLO	Pass	12:23am
FC	Pass	12:23am

Temperature Tests

Test	Status	Time
FC1	Pass	12:23am
SRC	Pass	12:23am
DET	Pass	12:23am
BAR	Pass	12:23am
BT	Pass	12:23am

Blank Tests

Test	Status	Time
AIR	Pass	12:24am

Printer Tests

Test	Status	Time
PRNT	Pass	12:24am

CRC Tests

Test	Status	Time
COMP	Pass	12:24am
CAL	Pass	12:24am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008600*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	12:13am
AIR BLK	.00	12:14am
ACCY CHK	.07	12:15am
AIR BLK	.00	12:16am
SUB TEST	**.00**	**12:17am**
AIR BLK	.00	12:17am
SUB TEST	**.00**	**12:19am**
AIR BLK	.00	12:20am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT Mobile Unit #5

Instrument Serial No. 008788 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008788* Test Record Number: *654*
Test Date: *07/07/2012* Test Time: *12:25am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:25am
FLO	Pass	12:25am
FC	Pass	12:25am

Temperature Tests

Test	Status	Time
FC1	Pass	12:25am
SRC	Pass	12:25am
DET	Pass	12:25am
BAR	Pass	12:25am
BT	Pass	12:25am

Blank Tests

Test	Status	Time
AIR	Pass	12:26am

Printer Tests

Test	Status	Time
PRNT	Pass	12:26am

CRC Tests

Test	Status	Time
COMP	Pass	12:26am
CAL	Pass	12:26am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008788*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123502
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	12:16am
AIR BLK	.00	12:17am
ACCY CHK	.07	12:18am
AIR BLK	.00	12:19am
SUB TEST	**.00**	**12:19am**
AIR BLK	.00	12:20am
SUB TEST	**.00**	**12:23am**
AIR BLK	.00	12:24am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008698 RALEIGH

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7TH day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008698* Test Record Number: *821*
Test Date: *07/07/2012* Test Time: *12:40am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:40am
FLO	Pass	12:40am
FC	Pass	12:40am

Temperature Tests

Test	Status	Time
FC1	Pass	12:40am
SRC	Pass	12:40am
DET	Pass	12:40am
BAR	Pass	12:40am
BT	Pass	12:40am

Blank Tests

Test	Status	Time
AIR	Pass	12:41am

Printer Tests

Test	Status	Time
PRNT	Pass	12:41am

CRC Tests

Test	Status	Time
COMP	Pass	12:41am
CAL	Pass	12:41am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008698*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	12:32am
AIR BLK	.00	12:33am
ACCY CHK	.07	12:33am
AIR BLK	.00	12:34am
SUB TEST	**.00**	**12:35am**
AIR BLK	.00	12:36am
SUB TEST	**.00**	**12:37am**
AIR BLK	.00	12:38am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Batmobile Unit #5

Instrument Serial No. 008600 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official 636 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008600* Test Record Number: *1039*
Test Date: *07/15/2012* Test Time: *12:14am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:15am
FLO	Pass	12:15am
FC	Pass	12:15am

Temperature Tests

Test	Status	Time
FC1	Pass	12:15am
SRC	Pass	12:15am
DET	Pass	12:15am
BAR	Pass	12:15am
BT	Pass	12:15am

Blank Tests

Test	Status	Time
AIR	Pass	12:15am

Printer Tests

Test	Status	Time
PRNT	Pass	12:16am

CRC Tests

Test	Status	Time
COMP	Pass	12:16am
CAL	Pass	12:16am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008600*
Test Date: *07/15/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	12:05am
AIR BLK	.00	12:06am
ACCY CHK	.07	12:07am
AIR BLK	.00	12:08am
SUB TEST	**.00**	**12:09am**
AIR BLK	.00	12:10am
SUB TEST	**.00**	**12:12am**
AIR BLK	.00	12:13am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT MOBile Unit #5

Instrument Serial No. 008698 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008698* Test Record Number: *835*
Test Date: *07/15/2012* Test Time: *12:17am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:17am
FLO	Pass	12:17am
FC	Pass	12:18am

Temperature Tests

Test	Status	Time
FC1	Pass	12:18am
SRC	Pass	12:18am
DET	Pass	12:18am
BAR	Pass	12:18am
BT	Pass	12:18am

Blank Tests

Test	Status	Time
AIR	Pass	12:18am

Printer Tests

Test	Status	Time
PRNT	Pass	12:18am

CRC Tests

Test	Status	Time
COMP	Pass	12:19am
CAL	Pass	12:19am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008698*
Test Date: *07/15/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	12:09am
AIR BLK	.00	12:10am
ACCY CHK	.07	12:11am
AIR BLK	.00	12:12am
SUB TEST	**.00**	**12:12am**
AIR BLK	.00	12:13am
SUB TEST	**.00**	**12:15am**
AIR BLK	.00	12:16am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008788 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008788* Test Record Number: *668*
Test Date: *07/15/2012* Test Time: *12:16am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:16am
FLO	Pass	12:16am
FC	Pass	12:16am

Temperature Tests

Test	Status	Time
FC1	Pass	12:16am
SRC	Pass	12:16am
DET	Pass	12:16am
BAR	Pass	12:16am
BT	Pass	12:16am

Blank Tests

Test	Status	Time
AIR	Pass	12:17am

Printer Tests

Test	Status	Time
PRNT	Pass	12:17am

CRC Tests

Test	Status	Time
COMP	Pass	12:17am
CAL	Pass	12:17am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008788*
Test Date: *07/15/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123502
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	12:07am
AIR BLK	.00	12:08am
ACCY CHK	.07	12:09am
AIR BLK	.00	12:10am
SUB TEST	**.00**	**12:10am**
AIR BLK	.00	12:11am
SUB TEST	**.00**	**12:13am**
AIR BLK	.00	12:14am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008600 RALEIGH

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13TH day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008600* Test Record Number: *1030*
Test Date: *07/13/2012* Test Time: *11:29pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:29pm
FLO	Pass	11:29pm
FC	Pass	11:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:29pm
SRC	Pass	11:29pm
DET	Pass	11:29pm
BAR	Pass	11:29pm
BT	Pass	11:29pm

Blank Tests

Test	Status	Time
AIR	Pass	11:30pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:30pm

CRC Tests

Test	Status	Time
COMP	Pass	11:30pm
CAL	Pass	11:30pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008600*
Test Date: *07/13/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	11:19pm
AIR BLK	.00	11:20pm
ACCY CHK	.08	11:21pm
AIR BLK	.00	11:22pm
SUB TEST	**.00**	**11:23pm**
AIR BLK	.00	11:24pm
SUB TEST	**.00**	**11:26pm**
AIR BLK	.00	11:27pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008698 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

036
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008698* Test Record Number: *828*
Test Date: *07/13/2012* Test Time: *11:31pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:32pm
FLO	Pass	11:32pm
FC	Pass	11:32pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:32pm
SRC	Pass	11:32pm
DET	Pass	11:32pm
BAR	Pass	11:32pm
BT	Pass	11:32pm

Blank Tests

Test	Status	Time
AIR	Pass	11:32pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:32pm

CRC Tests

Test	Status	Time
COMP	Pass	11:33pm
CAL	Pass	11:33pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008698*
Test Date: *07/13/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	11:23pm
AIR BLK	.00	11:24pm
ACCY CHK	.08	11:24pm
AIR BLK	.00	11:25pm
SUB TEST	**.00**	**11:26pm**
AIR BLK	.00	11:27pm
SUB TEST	**.00**	**11:28pm**
AIR BLK	.00	11:29pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008788 RALEIGH

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13TH day of July, 2017 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official 636 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008788* Test Record Number: *662*
Test Date: *07/13/2012* Test Time: *11:30pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:30pm
FLO	Pass	11:30pm
FC	Pass	11:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:30pm
SRC	Pass	11:30pm
DET	Pass	11:30pm
BAR	Pass	11:30pm
BT	Pass	11:30pm

Blank Tests

Test	Status	Time
AIR	Pass	11:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:31pm

CRC Tests

Test	Status	Time
COMP	Pass	11:31pm
CAL	Pass	11:31pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008788*
Test Date: *07/13/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123502
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	11:21pm
AIR BLK	.00	11:22pm
ACCY CHK	.07	11:23pm
AIR BLK	.00	11:24pm
SUB TEST	**.00**	**11:25pm**
AIR BLK	.00	11:26pm
SUB TEST	**.00**	**11:28pm**
AIR BLK	.00	11:29pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008616 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: *008616*
Test Date: *07/12/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG023601
Exp Date: 08/24/2012

Test	g/210L	Time
DIAG	Pass	11:06pm
AIR BLK	.00	11:07pm
ACCY CHK	.08	11:08pm
AIR BLK	.00	11:08pm
SUB TEST	**.00**	**11:09pm**
AIR BLK	.00	11:10pm
SUB TEST	**.00**	**11:12pm**
AIR BLK	.00	11:13pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: *008616* Test Record Number: *1421*
Test Date: *07/12/2012* Test Time: *11:13pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:14pm
FLO	Pass	11:14pm
FC	Pass	11:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:14pm
SRC	Pass	11:14pm
DET	Pass	11:14pm
BAR	Pass	11:14pm
BT	Pass	11:14pm

Blank Tests

Test	Status	Time
AIR	Pass	11:14pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:15pm

CRC Tests

Test	Status	Time
COMP	Pass	11:15pm
CAL	Pass	11:15pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: *008707*
Test Date: *07/12/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	11:02pm
AIR BLK	.00	11:03pm
ACCY CHK	.08	11:04pm
AIR BLK	.00	11:05pm
SUB TEST	**.00**	**11:05pm**
AIR BLK	.00	11:06pm
SUB TEST	**.00**	**11:07pm**
AIR BLK	.00	11:08pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: *008707* Test Record Number: *1484*
Test Date: *07/12/2012* Test Time: *11:09pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:09pm
FLO	Pass	11:09pm
FC	Pass	11:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:10pm
SRC	Pass	11:10pm
DET	Pass	11:10pm
BAR	Pass	11:10pm
BT	Pass	11:10pm

Blank Tests

Test	Status	Time
AIR	Pass	11:10pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:10pm

CRC Tests

Test	Status	Time
COMP	Pass	11:10pm
CAL	Pass	11:10pm

Preventive Maintenance
Status: Pass

Allen Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

048
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: *008647*
Test Date: *07/12/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	11:39pm
AIR BLK	.00	11:40pm
ACCY CHK	.07	11:40pm
AIR BLK	.00	11:41pm
SUB TEST	**.00**	**11:42pm**
AIR BLK	.00	11:43pm
SUB TEST	**.00**	**11:45pm**
AIR BLK	.00	11:45pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: *008647* Test Record Number: *1426*
Test Date: *07/12/2012* Test Time: *11:46pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:46pm
FLO	Pass	11:46pm
FC	Pass	11:46pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:46pm
SRC	Pass	11:46pm
DET	Pass	11:46pm
BAR	Pass	11:46pm
BT	Pass	11:46pm

Blank Tests

Test	Status	Time
AIR	Pass	11:47pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:47pm

CRC Tests

Test	Status	Time
COMP	Pass	11:47pm
CAL	Pass	11:47pm

Preventive Maintenance
Status: Pass

Alan Ray Bass
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location Greensboro Jail

Instrument Serial No. 008794

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: *008794*
Test Date: *07/11/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2011-06/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	3:51pm
AIR BLK	.00	3:51pm
ACCY CHK	.08	3:52pm
AIR BLK	.00	3:53pm
SUB TEST	**.00**	**3:53pm**
AIR BLK	.00	3:54pm
SUB TEST	**.00**	**3:56pm**
AIR BLK	.00	3:56pm

Reported AC: .00 g/210L

L. Kevin Dean
Signature of Chemical Analyst

Court CVR

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: *008794* Test Record Number: *2337*
Test Date: *07/11/2012* Test Time: *3:58pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:59pm
FLO	Pass	3:59pm
FC	Pass	3:59pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:59pm
SRC	Pass	3:59pm
DET	Pass	3:59pm
BAR	Pass	3:59pm
BT	Pass	3:59pm

Blank Tests

Test	Status	Time
AIR	Pass	4:00pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:00pm

CRC Tests

Test	Status	Time
COMP	Pass	4:00pm
CAL	Pass	4:00pm

Preventive Maintenance
Status: Pass

L. Kevin Dean

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location Greensboro Jail

Instrument Serial No. 008790

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: *008790*
Test Date: *07/11/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2011-06/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	3:25pm
AIR BLK	.00	3:25pm
ACCY CHK	.07	3:26pm
AIR BLK	.00	3:27pm
SUB TEST	**.00**	**3:27pm**
AIR BLK	.00	3:28pm
SUB TEST	**.00**	**3:30pm**
AIR BLK	.00	3:31pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: *008790* Test Record Number: *3119*
Test Date: *07/11/2012* Test Time: *3:31pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:31pm
FLO	Pass	3:31pm
FC	Pass	3:32pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:32pm
SRC	Pass	3:32pm
DET	Pass	3:32pm
BAR	Pass	3:32pm
BT	Pass	3:32pm

Blank Tests

Test	Status	Time
AIR	Pass	3:32pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:32pm

CRC Tests

Test	Status	Time
COMP	Pass	3:32pm
CAL	Pass	3:32pm

Preventive Maintenance
Status: Pass

J. Kevin Dean

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Davidson Instrument Location Lexington Police Department

Instrument Serial No. 008883

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY LEXINGTON PD 280

Serial Number: *008883*
Test Date: *07/10/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2011-06/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG125602
Exp Date: 09/13/2013

Test	g/210L	Time
DIAG	Pass	1:35pm
AIR BLK	.00	1:36pm
ACCY CHK	.08	1:37pm
AIR BLK	.00	1:38pm
SUB TEST	**.00**	**1:38pm**
AIR BLK	.00	1:39pm
SUB TEST	**.00**	**1:41pm**
AIR BLK	.00	1:42pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY LEXINGTON PD 280

Serial Number: *008883* Test Record Number: *1008*
Test Date: *07/10/2012* Test Time: *1:43pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:43pm
FLO	Pass	1:43pm
FC	Pass	1:43pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:43pm
SRC	Pass	1:43pm
DET	Pass	1:43pm
BAR	Pass	1:43pm
BT	Pass	1:43pm

Blank Tests

Test	Status	Time
AIR	Pass	1:44pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:44pm

CRC Tests

Test	Status	Time
COMP	Pass	1:44pm
CAL	Pass	1:44pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Davidson Instrument Location Davidson Co Jail

Instrument Serial No. 008845 Lexington, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

THE GREAT SEAL OF THE STATE OF NORTH CAROLINA MAY 20, 1775 APRIL 12, 1776 ESSE QUAM VIDERI

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY DAVIDSON CO JAIL 280

Serial Number: *008845*
Test Date: *07/10/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2011-06/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG125603
Exp Date: 09/13/2013

Test	g/210L	Time
DIAG	Pass	2:44pm
AIR BLK	.00	2:45pm
ACCY CHK	.08	2:45pm
AIR BLK	.00	2:47pm
SUB TEST	**.00**	**2:47pm**
AIR BLK	.00	2:48pm
SUB TEST	**.00**	**2:50pm**
AIR BLK	.00	2:51pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY DAVIDSON CO JAIL 280

Serial Number: *008845* Test Record Number: *1214*
Test Date: *07/10/2012* Test Time: *2:51pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:52pm
FLO	Pass	2:52pm
FC	Pass	2:52pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:52pm
SRC	Pass	2:52pm
DET	Pass	2:52pm
BAR	Pass	2:52pm
BT	Pass	2:52pm

Blank Tests

Test	Status	Time
AIR	Pass	2:52pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:52pm

CRC Tests

Test	Status	Time
COMP	Pass	2:53pm
CAL	Pass	2:53pm

Preventive Maintenance
Status: Pass

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Davidson Instrument Location Thomasville Police

Instrument Serial No. 008872 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

S. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY THOMASVILLE PD 280

Serial Number: *008872*
Test Date: *07/10/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2011-06/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG125603
Exp Date: 09/13/2013

Test	g/210L	Time
DIAG	Pass	3:48pm
AIR BLK	.00	3:49pm
ACCY CHK	.07	3:49pm
AIR BLK	.00	3:50pm
SUB TEST	**.00**	**3:51pm**
AIR BLK	.00	3:52pm
SUB TEST	**.00**	**3:54pm**
AIR BLK	.00	3:55pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY THOMASVILLE PD 280

Serial Number: *008872* Test Record Number: *913*
Test Date: *07/10/2012* Test Time: *3:56pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:56pm
FLO	Pass	3:56pm
FC	Pass	3:56pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:56pm
SRC	Pass	3:56pm
DET	Pass	3:56pm
BAR	Pass	3:56pm
BT	Pass	3:56pm

Blank Tests

Test	Status	Time
AIR	Pass	3:57pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:57pm

CRC Tests

Test	Status	Time
COMP	Pass	3:57pm
CAL	Pass	3:57pm

Preventive Maintenance
Status: Pass

S. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wilson Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008717

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

K82E
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WILSON COUNTY BAT MOBILE UNIT 4 970

Serial Number: *008717*
Test Date: *07/06/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG920302
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	11:08pm
AIR BLK	.00	11:09pm
ACCY CHK	.07	11:09pm
AIR BLK	.00	11:10pm
SUB TEST	**.00**	**11:11pm**
AIR BLK	.00	11:12pm
SUB TEST	**.00**	**11:14pm**
AIR BLK	.00	11:14pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WILSON COUNTY BAT MOBILE UNIT 4 970

Serial Number: *008717* Test Record Number: *285*
Test Date: *07/06/2012* Test Time: *11:17pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:18pm
FLO	Pass	11:18pm
FC	Pass	11:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:18pm
SRC	Pass	11:18pm
DET	Pass	11:18pm
BAR	Pass	11:18pm
BT	Pass	11:18pm

Blank Tests

Test	Status	Time
AIR	Pass	11:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:19pm

CRC Tests

Test	Status	Time
COMP	Pass	11:19pm
CAL	Pass	11:19pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wilson Co Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008434

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

482E
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WILSON COUNTY BAT MOBILE UNIT 4 970

Serial Number: *008734*
Test Date: *07/06/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG920302
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	11:05pm
AIR BLK	.00	11:06pm
ACCY CHK	.07	11:06pm
AIR BLK	.00	11:07pm
SUB TEST	**.00**	**11:08pm**
AIR BLK	.00	11:09pm
SUB TEST	**.00**	**11:10pm**
AIR BLK	.00	11:11pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WILSON COUNTY BAT MOBILE UNIT 4 970

Serial Number: *008734* Test Record Number: *552*
Test Date: *07/06/2012* Test Time: *11:13pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:13pm
FLO	Pass	11:13pm
FC	Pass	11:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:14pm
SRC	Pass	11:14pm
DET	Pass	11:14pm
BAR	Pass	11:14pm
BT	Pass	11:14pm

Blank Tests

Test	Status	Time
AIR	Pass	11:14pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:14pm

CRC Tests

Test	Status	Time
COMP	Pass	11:14pm
CAL	Pass	11:14pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Beaufort Co Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008717

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

682E
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BEAUFORT COUNTY BAT MOBILE UNIT 4 060

Serial Number: *008717*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG920302
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	9:25pm
AIR BLK	.00	9:26pm
ACCY CHK	.07	9:27pm
AIR BLK	.00	9:28pm
SUB TEST	**.00**	**9:28pm**
AIR BLK	.00	9:29pm
SUB TEST	**.00**	**9:30pm**
AIR BLK	.00	9:31pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BEAUFORT COUNTY BAT MOBILE UNIT 4 060

Serial Number: *008717* Test Record Number: *287*
Test Date: *07/07/2012* Test Time: *9:33pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:33pm
FLO	Pass	9:33pm
FC	Pass	9:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:33pm
SRC	Pass	9:33pm
DET	Pass	9:33pm
BAR	Pass	9:33pm
BT	Pass	9:33pm

Blank Tests

Test	Status	Time
AIR	Pass	9:34pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:34pm

CRC Tests

Test	Status	Time
COMP	Pass	9:34pm
CAL	Pass	9:34pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Beaufort Co Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008434

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

7682E
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BEAUFORT COUNTY BAT MOBILE UNIT 4 060

Serial Number: *008734*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG920302
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	9:20pm
AIR BLK	.00	9:21pm
ACCY CHK	.07	9:22pm
AIR BLK	.00	9:22pm
SUB TEST	**.00**	**9:24pm**
AIR BLK	.00	9:25pm
SUB TEST	**.00**	**9:26pm**
AIR BLK	.00	9:27pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BEAUFORT COUNTY BAT MOBILE UNIT 4 060

Serial Number: *008734* Test Record Number: *554*
Test Date: *07/07/2012* Test Time: *9:29pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:29pm
FLO	Pass	9:29pm
FC	Pass	9:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:29pm
SRC	Pass	9:29pm
DET	Pass	9:29pm
BAR	Pass	9:29pm
BT	Pass	9:29pm

Blank Tests

Test	Status	Time
AIR	Pass	9:30pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:30pm

CRC Tests

Test	Status	Time
COMP	Pass	9:30pm
CAL	Pass	9:30pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Beaufort Co Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008841

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

2682E
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BEAUFORT COUNTY BAT MOBILE UNIT 4 060

Serial Number: *008871*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	9:22pm
AIR BLK	.00	9:23pm
ACCY CHK	.07	9:23pm
AIR BLK	.00	9:24pm
SUB TEST	**.00**	**9:25pm**
AIR BLK	.00	9:26pm
SUB TEST	**.00**	**9:28pm**
AIR BLK	.00	9:29pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BEAUFORT COUNTY BAT MOBILE UNIT 4 060

Serial Number: *008871* Test Record Number: *510*
Test Date: *07/07/2012* Test Time: *9:15pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:16pm
FLO	Pass	9:16pm
FC	Pass	9:16pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:16pm
SRC	Pass	9:16pm
DET	Pass	9:16pm
BAR	Pass	9:16pm
BT	Pass	9:16pm

Blank Tests

Test	Status	Time
AIR	Pass	9:17pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:17pm

CRC Tests

Test	Status	Time
COMP	Pass	9:17pm
CAL	Pass	9:17pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Robeson Instrument Location St. Pauls Police Dept.

Instrument Serial No. 008814 St. Pauls, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 03 day of July, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROBESON COUNTY ST. PAULS PD. 770

Serial Number: *008814*
Test Date: *07/03/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG124904
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	12:30pm
AIR BLK	.00	12:30pm
ACCY CHK	.07	12:31pm
AIR BLK	.00	12:32pm
SUB TEST	**.00**	**12:33pm**
AIR BLK	.00	12:33pm
SUB TEST	**.00**	**12:35pm**
AIR BLK	.00	12:36pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY ST. PAULS PD. 770

Serial Number: *008814* Test Record Number: *354*
Test Date: *07/03/2012* Test Time: *12:38pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:38pm
FLO	Pass	12:38pm
FC	Pass	12:38pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:38pm
SRC	Pass	12:38pm
DET	Pass	12:38pm
BAR	Pass	12:38pm
BT	Pass	12:38pm

Blank Tests

Test	Status	Time
AIR	Pass	12:39pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:39pm

CRC Tests

Test	Status	Time
COMP	Pass	12:39pm
CAL	Pass	12:39pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ROBESON Instrument Location PEMBROKE POLICE DEPT.

Instrument Serial No. 008837 PEMBROKE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 03 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROBESON COUNTY PEMBROKE POLICE DEPT
770

Serial Number: *008837*
Test Date: *07/03/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG106701
Exp Date: 03/08/2013

Test	g/210L	Time
DIAG	Pass	10:47am
AIR BLK	.00	10:48am
ACCY CHK	.07	10:49am
AIR BLK	.00	10:50am
SUB TEST	**.00**	**10:50am**
AIR BLK	.00	10:51am
SUB TEST	**.00**	**10:53am**
AIR BLK	.00	10:54am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY PEMBROKE POLICE DEPT 770

Serial Number: *008837* Test Record Number: *333*
Test Date: *07/03/2012* Test Time: *10:55am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:55am
FLO	Pass	10:55am
FC	Pass	10:55am

Temperature Tests

Test	Status	Time
FC1	Pass	10:56am
SRC	Pass	10:56am
DET	Pass	10:56am
BAR	Pass	10:56am
BT	Pass	10:56am

Blank Tests

Test	Status	Time
AIR	Pass	10:56am

Printer Tests

Test	Status	Time
PRNT	Pass	10:56am

CRC Tests

Test	Status	Time
COMP	Pass	10:56am
CAL	Pass	10:56am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ROBESON Instrument Location RED SPRINGS POLICE DEPT.

Instrument Serial No. 008857 RED SPRINGS NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 03 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROBESON COUNTY RED SPRINGS PD 770

Serial Number: *008857*
Test Date: *07/03/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG106701
Exp Date: 03/08/2013

Test	g/210L	Time
DIAG	Pass	9:53am
AIR BLK	.00	9:54am
ACCY CHK	.07	9:55am
AIR BLK	.00	9:56am
SUB TEST	**.00**	**9:56am**
AIR BLK	.00	9:57am
SUB TEST	**.00**	**9:59am**
AIR BLK	.00	10:00am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY RED SPRINGS PD 770

Serial Number: *008857* Test Record Number: *280*
Test Date: *07/03/2012* Test Time: *10:00am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:01am
FLO	Pass	10:01am
FC	Pass	10:01am

Temperature Tests

Test	Status	Time
FC1	Pass	10:01am
SRC	Pass	10:01am
DET	Pass	10:01am
BAR	Pass	10:01am
BT	Pass	10:01am

Blank Tests

Test	Status	Time
AIR	Pass	10:02am

Printer Tests

Test	Status	Time
PRNT	Pass	10:02am

CRC Tests

Test	Status	Time
COMP	Pass	10:02am
CAL	Pass	10:02am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Moore Instrument Location Pinehurst Police Dept.

Instrument Serial No. 008710 Pinehurst, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

THE GREAT SEAL OF THE STATE OF NORTH CAROLINA MAY 20, 1775 APRIL 12, 1776 ESSE QUAM VIDERI

Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MOORE COUNTY PINEHURST PD. 620

Serial Number: *008710*
Test Date: *07/02/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG023602
Exp Date: 08/24/2012

Test	g/210L	Time
DIAG	Pass	11:24am
AIR BLK	.00	11:24am
ACCY CHK	.07	11:25am
AIR BLK	.00	11:26am
SUB TEST	**.00**	**11:26am**
AIR BLK	.00	11:27am
SUB TEST	**.00**	**11:29am**
AIR BLK	.00	11:30am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MOORE COUNTY PINEHURST PD. 620

Serial Number: *008710* Test Record Number: *843*
Test Date: *07/02/2012* Test Time: *11:32am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:32am
FLO	Pass	11:32am
FC	Pass	11:32am

Temperature Tests

Test	Status	Time
FC1	Pass	11:32am
SRC	Pass	11:32am
DET	Pass	11:32am
BAR	Pass	11:32am
BT	Pass	11:32am

Blank Tests

Test	Status	Time
AIR	Pass	11:33am

Printer Tests

Test	Status	Time
PRNT	Pass	11:33am

CRC Tests

Test	Status	Time
COMP	Pass	11:33am
CAL	Pass	11:33am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Moore Instrument Location Southern Pines Police Dept.

Instrument Serial No. 008720 Southern Pines NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MOORE COUNTY SOUTHERN PINES PD. 620

Serial Number: *008720*
Test Date: *07/02/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG102602
Exp Date: 01/26/2013

Test	g/210L	Time
DIAG	Pass	12:38pm
AIR BLK	.00	12:38pm
ACCY CHK	.07	12:39pm
AIR BLK	.00	12:40pm
SUB TEST	**.00**	**12:40pm**
AIR BLK	.00	12:42pm
SUB TEST	**.00**	**12:43pm**
AIR BLK	.00	12:44pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MOORE COUNTY SOUTHERN PINES PD. 620

Serial Number: *008720* Test Record Number: *598*
Test Date: *07/02/2012* Test Time: *12:45pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:45pm
FLO	Pass	12:45pm
FC	Pass	12:45pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:45pm
SRC	Pass	12:45pm
DET	Pass	12:45pm
BAR	Pass	12:45pm
BT	Pass	12:45pm

Blank Tests

Test	Status	Time
AIR	Pass	12:46pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:46pm

CRC Tests

Test	Status	Time
COMP	Pass	12:46pm
CAL	Pass	12:46pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008616 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: *008616*
Test Date: *07/03/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG023601
Exp Date: 08/24/2012

Test	g/210L	Time
DIAG	Pass	4:20pm
AIR BLK	.00	4:21pm
ACCY CHK	.07	4:21pm
AIR BLK	.00	4:22pm
SUB TEST	**.00**	**4:23pm**
AIR BLK	.00	4:24pm
SUB TEST	**.00**	**4:26pm**
AIR BLK	.00	4:27pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: *008616* Test Record Number: *1410*
Test Date: *07/03/2012* Test Time: *4:30pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:30pm
FLO	Pass	4:30pm
FC	Pass	4:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:30pm
SRC	Pass	4:30pm
DET	Pass	4:30pm
BAR	Pass	4:30pm
BT	Pass	4:30pm

Blank Tests

Test	Status	Time
AIR	Pass	4:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:31pm

CRC Tests

Test	Status	Time
COMP	Pass	4:31pm
CAL	Pass	4:31pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: *008707*
Test Date: *07/03/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	4:06pm
AIR BLK	.00	4:07pm
ACCY CHK	.08	4:08pm
AIR BLK	.00	4:09pm
SUB TEST	**.00**	**4:10pm**
AIR BLK	.00	4:11pm
SUB TEST	**.00**	**4:13pm**
AIR BLK	.00	4:14pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: *008707* Test Record Number: *1468*
Test Date: *07/03/2012* Test Time: *4:15pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:15pm
FLO	Pass	4:15pm
FC	Pass	4:15pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:15pm
SRC	Pass	4:15pm
DET	Pass	4:15pm
BAR	Pass	4:15pm
BT	Pass	4:15pm

Blank Tests

Test	Status	Time
AIR	Pass	4:16pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:16pm

CRC Tests

Test	Status	Time
COMP	Pass	4:16pm
CAL	Pass	4:16pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County IREDELL Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 MOORESVILLE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

IREDELL COUNTY BAT MOBILE UNIT 3 480

Serial Number: *008647*
Test Date: *07/04/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	3:13pm
AIR BLK	.00	3:14pm
ACCY CHK	.07	3:15pm
AIR BLK	.00	3:16pm
SUB TEST	**.00**	**3:16pm**
AIR BLK	.00	3:17pm
SUB TEST	**.00**	**3:19pm**
AIR BLK	.00	3:20pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY BAT MOBILE UNIT 3 480

Serial Number: *008647* Test Record Number: *1413*
Test Date: *07/04/2012* Test Time: *3:20pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:20pm
FLO	Pass	3:20pm
FC	Pass	3:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:21pm
SRC	Pass	3:21pm
DET	Pass	3:21pm
BAR	Pass	3:21pm
BT	Pass	3:21pm

Blank Tests

Test	Status	Time
AIR	Pass	3:21pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:21pm

CRC Tests

Test	Status	Time
COMP	Pass	3:22pm
CAL	Pass	3:22pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County DAVIDSON Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 SOUTHMONT, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY BAT MOBILE UNIT 3 280

Serial Number: *008647*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	7:55pm
AIR BLK	.00	7:56pm
ACCY CHK	.07	7:57pm
AIR BLK	.00	7:58pm
SUB TEST	**.00**	**7:58pm**
AIR BLK	.00	7:59pm
SUB TEST	**.00**	**8:01pm**
AIR BLK	.00	8:02pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY BAT MOBILE UNIT 3 280

Serial Number: *008647* Test Record Number: *1419*
Test Date: *07/07/2012* Test Time: *8:02pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:03pm
FLO	Pass	8:03pm
FC	Pass	8:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:03pm
SRC	Pass	8:03pm
DET	Pass	8:03pm
BAR	Pass	8:03pm
BT	Pass	8:03pm

Blank Tests

Test	Status	Time
AIR	Pass	8:03pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:04pm

CRC Tests

Test	Status	Time
COMP	Pass	8:04pm
CAL	Pass	8:04pm

Preventive Maintenance
Status: Pass

Allen Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County DAVIDSON Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 SOUTHMONT, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alvin Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY BAT MOBILE UNIT 3 280

Serial Number: *008707*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	7:58pm
AIR BLK	.00	7:59pm
ACCY CHK	.08	8:00pm
AIR BLK	.00	8:00pm
SUB TEST	**.00**	**8:01pm**
AIR BLK	.00	8:02pm
SUB TEST	**.00**	**8:04pm**
AIR BLK	.00	8:05pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY BAT MOBILE UNIT 3 280

Serial Number: *008707* Test Record Number: *1477*
Test Date: *07/07/2012* Test Time: *8:05pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:05pm
FLO	Pass	8:05pm
FC	Pass	8:05pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:05pm
SRC	Pass	8:05pm
DET	Pass	8:05pm
BAR	Pass	8:05pm
BT	Pass	8:05pm

Blank Tests

Test	Status	Time
AIR	Pass	8:06pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:06pm

CRC Tests

Test	Status	Time
COMP	Pass	8:06pm
CAL	Pass	8:06pm

Preventive Maintenance
Status: Pass

Alan Ray Bennett
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County DAVIDSON Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008616 SOUTHMONT, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of JULY, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

THE GREAT SEAL OF THE STATE OF NORTH CAROLINA
MAY 20, 1775
APRIL 12, 1776
ESSE QUAM VIDERI

Alvin Ray Burns
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY BAT MOBILE UNIT 3 280

Serial Number: *008616*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG023601
Exp Date: 08/24/2012

Test	g/210L	Time
DIAG	Pass	8:26pm
AIR BLK	.00	8:27pm
ACCY CHK	.07	8:28pm
AIR BLK	.00	8:28pm
SUB TEST	**.00**	**8:29pm**
AIR BLK	.00	8:30pm
SUB TEST	**.00**	**8:32pm**
AIR BLK	.00	8:33pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY BAT MOBILE UNIT 3 280

Serial Number: *008616* Test Record Number: *1417*
Test Date: *07/07/2012* Test Time: *8:33pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:34pm
FLO	Pass	8:34pm
FC	Pass	8:34pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:34pm
SRC	Pass	8:34pm
DET	Pass	8:34pm
BAR	Pass	8:34pm
BT	Pass	8:34pm

Blank Tests

Test	Status	Time
AIR	Pass	8:35pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:35pm

CRC Tests

Test	Status	Time
COMP	Pass	8:35pm
CAL	Pass	8:35pm

Preventive Maintenance
Status: Pass

Allen Ray Barnes

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ALEXANDER Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 HICKORY, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of JULY, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ALEXANDER COUNTY BAT MOBILE UNIT 3 010

Serial Number: *008707*
Test Date: *07/08/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	3:18pm
AIR BLK	.00	3:19pm
ACCY CHK	.08	3:20pm
AIR BLK	.00	3:21pm
SUB TEST	**.00**	**3:21pm**
AIR BLK	.00	3:22pm
SUB TEST	**.00**	**3:24pm**
AIR BLK	.00	3:25pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALEXANDER COUNTY BAT MOBILE UNIT 3 010

Serial Number: *008707* Test Record Number: *1481*
Test Date: *07/08/2012* Test Time: *3:25pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:26pm
FLO	Pass	3:26pm
FC	Pass	3:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:26pm
SRC	Pass	3:26pm
DET	Pass	3:26pm
BAR	Pass	3:26pm
BT	Pass	3:26pm

Blank Tests

Test	Status	Time
AIR	Pass	3:27pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:27pm

CRC Tests

Test	Status	Time
COMP	Pass	3:27pm
CAL	Pass	3:27pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008788 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008788* Test Record Number: *650*
Test Date: *07/06/2012* Test Time: *12:17am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:17am
FLO	Pass	12:17am
FC	Pass	12:17am

Temperature Tests

Test	Status	Time
FC1	Pass	12:17am
SRC	Pass	12:17am
DET	Pass	12:17am
BAR	Pass	12:17am
BT	Pass	12:17am

Blank Tests

Test	Status	Time
AIR	Pass	12:18am

Printer Tests

Test	Status	Time
PRNT	Pass	12:18am

CRC Tests

Test	Status	Time
COMP	Pass	12:18am
CAL	Pass	12:18am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: *008788*
Test Date: *07/06/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123502
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	12:09am
AIR BLK	.00	12:10am
ACCY CHK	.07	12:11am
AIR BLK	.00	12:11am
SUB TEST	**.00**	**12:12am**
AIR BLK	.00	12:13am
SUB TEST	**.00**	**12:14am**
AIR BLK	.00	12:15am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County DURHAM Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008600 DURHAM

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY BAT MOBILE UNIT 5 310

Serial Number: *008600* Test Record Number: *1027*
Test Date: *07/07/2012* Test Time: *10:32pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:33pm
FLO	Pass	10:33pm
FC	Pass	10:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:33pm
SRC	Pass	10:33pm
DET	Pass	10:33pm
BAR	Pass	10:33pm
BT	Pass	10:33pm

Blank Tests

Test	Status	Time
AIR	Pass	10:33pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:34pm

CRC Tests

Test	Status	Time
COMP	Pass	10:34pm
CAL	Pass	10:34pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

DURHAM COUNTY BAT MOBILE UNIT 5 310

Serial Number: *008600*
Test Date: *07/07/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	9:44pm
AIR BLK	.00	9:45pm
ACCY CHK	.07	9:46pm
AIR BLK	.00	9:47pm
SUB TEST	**.00**	**9:48pm**
AIR BLK	.00	9:49pm
SUB TEST	**.00**	**9:50pm**
AIR BLK	.00	9:51pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Union Instrument Location Union County S.O.

Instrument Serial No. 008866 3344 Presson Rd., Monroe
704-283-3770

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

UNION COUNTY UNION COUNTY SD 890

Serial Number: *008866*
Test Date: *07/24/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG125603
Exp Date: 09/13/2013

Test	g/210L	Time
DIAG	Pass	10:40am
AIR BLK	.00	10:41am
ACCY CHK	.08	10:42am
AIR BLK	.00	10:43am
SUB TEST	**.00**	**10:43am**
AIR BLK	.00	10:44am
SUB TEST	**.00**	**10:46am**
AIR BLK	.00	10:47am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

UNION COUNTY UNION COUNTY SD 890

Serial Number: *008866* Test Record Number: *1286*
Test Date: *07/24/2012* Test Time: *10:51am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:51am
FLO	Pass	10:51am
FC	Pass	10:51am

Temperature Tests

Test	Status	Time
FC1	Pass	10:51am
SRC	Pass	10:51am
DET	Pass	10:51am
BAR	Pass	10:51am
BT	Pass	10:51am

Blank Tests

Test	Status	Time
AIR	Pass	10:52am

Printer Tests

Test	Status	Time
PRNT	Pass	10:52am

CRC Tests

Test	Status	Time
COMP	Pass	10:52am
CAL	Pass	10:52am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Gaston Instrument Location Gaston Co. S.O.

Instrument Serial No. 008684 425 N. Marietta St. Gastonia
704-869-6800

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Mark D. Hay
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GASTON COUNTY GASTON COUNTY SD 350

Serial Number: *008684*
Test Date: *07/18/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108203
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	4:29pm
AIR BLK	.00	4:30pm
ACCY CHK	.07	4:31pm
AIR BLK	.00	4:32pm
SUB TEST	**.00**	**4:32pm**
AIR BLK	.00	4:33pm
SUB TEST	**.00**	**4:35pm**
AIR BLK	.00	4:36pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GASTON COUNTY GASTON COUNTY SD 350

Serial Number: *008684* Test Record Number: *2110*
Test Date: *07/18/2012* Test Time: *4:41pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:41pm
FLO	Pass	4:41pm
FC	Pass	4:41pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:41pm
SRC	Pass	4:41pm
DET	Pass	4:41pm
BAR	Pass	4:41pm
BT	Pass	4:41pm

Blank Tests

Test	Status	Time
AIR	Pass	4:42pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:42pm

CRC Tests

Test	Status	Time
COMP	Pass	4:42pm
CAL	Pass	4:42pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Stanly Instrument Location Stanly County SO

Instrument Serial No. 008824 126 S. 3rd St., Albemarle
704-986-3734

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

STANLY COUNTY STANLY COUNTY SD 830

Serial Number: *008824*
Test Date: *07/16/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG109703
Exp Date: 04/07/2013

Test	g/210L	Time
DIAG	Pass	10:38am
AIR BLK	.00	10:39am
ACCY CHK	.08	10:39am
AIR BLK	.00	10:41am
SUB TEST	**.00**	**10:41am**
AIR BLK	.00	10:42am
SUB TEST	**.00**	**10:44am**
AIR BLK	.00	10:45am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

STANLY COUNTY STANLY COUNTY SD 830

Serial Number: *008824* Test Record Number: *726*
Test Date: *07/16/2012* Test Time: *10:49am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:49am
FLO	Pass	10:49am
FC	Pass	10:49am

Temperature Tests

Test	Status	Time
FC1	Pass	10:49am
SRC	Pass	10:49am
DET	Pass	10:49am
BAR	Pass	10:49am
BT	Pass	10:49am

Blank Tests

Test	Status	Time
AIR	Pass	10:50am

Printer Tests

Test	Status	Time
PRNT	Pass	10:50am

CRC Tests

Test	Status	Time
COMP	Pass	10:50am
CAL	Pass	10:50am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CARTERET Instrument Location CARTERET County

Instrument Serial No. 008605 SHERIFF'S OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Randy E Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CARTERET COUNTY CARTERET COUNTY SD 150

Serial Number: *008605*
Test Date: *07/02/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	2:14pm
AIR BLK	.00	2:14pm
ACCY CHK	.08	2:15pm
AIR BLK	.00	2:16pm
SUB TEST	**.00**	**2:16pm**
AIR BLK	.00	2:17pm
SUB TEST	**.00**	**2:19pm**
AIR BLK	.00	2:20pm

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY CARTERET COUNTY SD 150

Serial Number: *008605* Test Record Number: *2587*
Test Date: *07/02/2012* Test Time: *2:20pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:21pm
FLO	Pass	2:21pm
FC	Pass	2:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:21pm
SRC	Pass	2:21pm
DET	Pass	2:21pm
BAR	Pass	2:21pm
BT	Pass	2:21pm

Blank Tests

Test	Status	Time
AIR	Pass	2:22pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:22pm

CRC Tests

Test	Status	Time
COMP	Pass	2:22pm
CAL	Pass	2:22pm

Preventive Maintenance
Status: Pass

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cabarrus Instrument Location Cabarrus County SD

Instrument Serial No. 008792 30 Corban Avenue, Concord
704-920-3000

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

6570
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CABARRUS COUNTY CABARRUS COUNTY SD 120

Serial Number: *008792*
Test Date: *07/25/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG201801
Exp Date: 01/18/2014

Test	g/210L	Time
DIAG	Pass	3:56pm
AIR BLK	.00	3:57pm
ACCY CHK	.08	3:58pm
AIR BLK	.00	3:59pm
SUB TEST	**.00**	**4:00pm**
AIR BLK	.00	4:01pm
SUB TEST	**.00**	**4:03pm**
AIR BLK	.00	4:04pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CABARRUS COUNTY CABARRUS COUNTY SD 120

Serial Number: *008792* Test Record Number: *740*
Test Date: *07/25/2012* Test Time: *4:06pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:07pm
FLO	Pass	4:07pm
FC	Pass	4:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:07pm
SRC	Pass	4:07pm
DET	Pass	4:07pm
BAR	Pass	4:07pm
BT	Pass	4:07pm

Blank Tests

Test	Status	Time
AIR	Pass	4:08pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:08pm

CRC Tests

Test	Status	Time
COMP	Pass	4:08pm
CAL	Pass	4:08pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cabarrus Instrument Location Cabarrus County SD

Instrument Serial No. 008625 30 Corban Ave, Concord

704-920-3000

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Joseph E. Hatch
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CABARRUS COUNTY CABARRUS COUNTY SD 120

Serial Number: *008625*
Test Date: *07/25/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG124903
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	3:51pm
AIR BLK	.00	3:52pm
ACCY CHK	.07	3:53pm
AIR BLK	.00	3:53pm
SUB TEST	**.00**	**3:55pm**
AIR BLK	.00	3:56pm
SUB TEST	**.00**	**3:58pm**
AIR BLK	.00	3:59pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CABARRUS COUNTY CABARRUS COUNTY SD 120

Serial Number: *008625* Test Record Number: *2981*
Test Date: *07/25/2012* Test Time: *4:01pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:02pm
FLO	Pass	4:02pm
FC	Pass	4:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:02pm
SRC	Pass	4:02pm
DET	Pass	4:02pm
BAR	Pass	4:02pm
BT	Pass	4:02pm

Blank Tests

Test	Status	Time
AIR	Pass	4:03pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:03pm

CRC Tests

Test	Status	Time
COMP	Pass	4:03pm
CAL	Pass	4:03pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Matthews PD

Instrument Serial No. 008699 1201 Crews Rd, Matthews
(704) 847-4069

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Joseph E. Hatch
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY MATTHEWS PD 590

Serial Number: *008699*
Test Date: *07/18/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG201801
Exp Date: 01/18/2014

Test	g/210L	Time
DIAG	Pass	3:07pm
AIR BLK	.00	3:08pm
ACCY CHK	.08	3:09pm
AIR BLK	.00	3:10pm
SUB TEST	**.00**	**3:10pm**
AIR BLK	.00	3:11pm
SUB TEST	**.00**	**3:13pm**
AIR BLK	.00	3:14pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY MATTHEWS PD 590

Serial Number: *008699* Test Record Number: *1801*
Test Date: *07/18/2012* Test Time: *3:03pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:04pm
FLO	Pass	3:04pm
FC	Pass	3:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:04pm
SRC	Pass	3:04pm
DET	Pass	3:04pm
BAR	Pass	3:04pm
BT	Pass	3:04pm

Blank Tests

Test	Status	Time
AIR	Pass	3:05pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:05pm

CRC Tests

Test	Status	Time
COMP	Pass	3:05pm
CAL	Pass	3:05pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cleveland Instrument Location Cleveland County SD - Annex

Instrument Serial No. 008887 407 McBrayer St, Shelby
704-484-4888

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3rd day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Joseph E. Hatch
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CLEVELAND COUNTY CLEVELAND SD-ANNEX 220

Serial Number: *008887*
Test Date: *07/03/2012*

Citation Number: *M0000000-0*
Subject's Name: *PREVENTIVE, MAINTENANCE*
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective: *10/01/2011-10/01/2013*

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG109703
Exp Date: 04/07/2013

Test	g/210L	Time
DIAG	Pass	2:27pm
AIR BLK	.00	2:28pm
ACCY CHK	.07	2:29pm
AIR BLK	.00	2:29pm
SUB TEST	**.00**	**2:30pm**
AIR BLK	.00	2:31pm
SUB TEST	**.00**	**2:32pm**
AIR BLK	.00	2:33pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CLEVELAND COUNTY CLEVELAND SD-ANNEX 220

Serial Number: *008887* Test Record Number: *1144*
Test Date: *07/03/2012* Test Time: *2:34pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:35pm
FLO	Pass	2:35pm
FC	Pass	2:35pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:35pm
SRC	Pass	2:35pm
DET	Pass	2:35pm
BAR	Pass	2:35pm
BT	Pass	2:35pm

Blank Tests

Test	Status	Time
AIR	Pass	2:35pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:35pm

CRC Tests

Test	Status	Time
COMP	Pass	2:36pm
CAL	Pass	2:36pm

Preventive Maintenance
Status: Pass

J. E. W.

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cleveland Instrument Location Cleveland County SD

Instrument Serial No. 008893 100 Justice Place, Shelby
704-484-4888

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3rd day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Joseph E. Hutt
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CLEVELAND COUNTY CLEVELAND COUNTY SD 220

Serial Number: *008893*
Test Date: *07/03/2012*

Citation Number: *M0000000-0*
Subject's Name: *PREVENTIVE, MAINTENANCE*
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective: *10/01/2011-10/01/2013*

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	2:03pm
AIR BLK	.00	2:04pm
ACCY CHK	.08	2:04pm
AIR BLK	.00	2:05pm
SUB TEST	**.00**	**2:06pm**
AIR BLK	.00	2:07pm
SUB TEST	**.00**	**2:08pm**
AIR BLK	.00	2:09pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CLEVELAND COUNTY CLEVELAND COUNTY SD 220

Serial Number: *008893* Test Record Number: *1012*
Test Date: *07/03/2012* Test Time: *2:10pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:10pm
FLO	Pass	2:10pm
FC	Pass	2:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:11pm
SRC	Pass	2:11pm
DET	Pass	2:11pm
BAR	Pass	2:11pm
BT	Pass	2:11pm

Blank Tests

Test	Status	Time
AIR	Pass	2:11pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:11pm

CRC Tests

Test	Status	Time
COMP	Pass	2:12pm
CAL	Pass	2:12pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Union Instrument Location Waxhaw PD

Instrument Serial No. 008598 703 W. South Main Street, Waxhaw
704-843-0353

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of July, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Joseph E. Hutch
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

UNION COUNTY WAXHAW PD 890

Serial Number: *008598*
Test Date: *07/02/2012*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG108202
Exp Date: 03/23/2013

Test	g/210L	Time
DIAG	Pass	01:02pm
AIR BLK	.00	01:03pm
ACCY CHK	.07	01:03pm
AIR BLK	.00	01:04pm
SUB TEST	**.00**	**01:05pm**
AIR BLK	.00	01:06pm
SUB TEST	**.00**	**01:08pm**
AIR BLK	.00	01:08pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

UNION COUNTY WAXHAW PD 890

Serial Number: *008598* Test Record Number: *374*
Test Date: *07/02/2012* Test Time: *01:09pm*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	01:10pm
FLO	Pass	01:10pm
FC	Pass	01:10pm

Temperature Tests

Test	Status	Time
FC1	Pass	01:10pm
SRC	Pass	01:10pm
DET	Pass	01:10pm
BAR	Pass	01:10pm
BT	Pass	01:10pm

Blank Tests

Test	Status	Time
AIR	Pass	01:11pm

Printer Tests

Test	Status	Time
PRNT	Pass	01:11pm

CRC Tests

Test	Status	Time
COMP	Pass	01:11pm
CAL	Pass	01:11pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007