DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CRAVEN Instrument Location Havelock P.D.

Instrument Serial No. 008800

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. Rhodes
Signature of Certifying Official

604
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CRAVEN COUNTY HAVELOCK PD 240

Serial Number: *008800*
Test Date: *06/26/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG308702
Exp Date: 03/28/2015

Test	g/210L	Time
DIAG	Pass	1:44pm
AIR BLK	.00	1:44pm
ACCY CHK	.07	1:45pm
AIR BLK	.00	1:46pm
SUB TEST	**.00**	**1:47pm**
AIR BLK	.00	1:48pm
SUB TEST	**.00**	**1:49pm**
AIR BLK	.00	1:50pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY HAVELOCK PD 240

Serial Number: *008800* Test Record Number: *717*
Test Date: *06/26/2013* Test Time: *1:52pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:52pm
FLO	Pass	1:52pm
FC	Pass	1:52pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:52pm
SRC	Pass	1:52pm
DET	Pass	1:52pm
BAR	Pass	1:52pm
BT	Pass	1:52pm

Blank Tests

Test	Status	Time
AIR	Pass	1:53pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:53pm

CRC Tests

Test	Status	Time
COMP	Pass	1:53pm
CAL	Pass	1:53pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Matthews PD

Instrument Serial No. 008699 1201 Crews Rd., Matthews
704-847-4069

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY MATTHEWS PD 590

Serial Number: *008699*
Test Date: *06/27/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG201801
Exp Date: 01/18/2014

Test	g/210L	Time
DIAG	Pass	9:56am
AIR BLK	.00	9:57am
ACCY CHK	.07	9:57am
AIR BLK	.00	9:58am
SUB TEST	**.00**	**9:59am**
AIR BLK	.00	10:00am
SUB TEST	**.00**	**10:02am**
AIR BLK	.00	10:02am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY MATTHEWS PD 590

Serial Number: *008699* Test Record Number: *2001*
Test Date: *06/27/2013* Test Time: *9:49am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:49am
FLO	Pass	9:49am
FC	Pass	9:49am

Temperature Tests

Test	Status	Time
FC1	Pass	9:49am
SRC	Pass	9:49am
DET	Pass	9:49am
BAR	Pass	9:49am
BT	Pass	9:49am

Blank Tests

Test	Status	Time
AIR	Pass	9:50am

Printer Tests

Test	Status	Time
PRNT	Pass	9:50am

CRC Tests

Test	Status	Time
COMP	Pass	9:50am
CAL	Pass	9:50am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Union Instrument Location Union County SD

Instrument Serial No. 008866 3344 Presson Rd., Monroe
704-283-3770

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

UNION COUNTY UNION COUNTY SD 890

Serial Number: *008866*
Test Date: *06/27/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG125603
Exp Date: 09/13/2013

Test	g/210L	Time
DIAG	Pass	11:12am
AIR BLK	.00	11:12am
ACCY CHK	.08	11:13am
AIR BLK	.00	11:14am
SUB TEST	**.00**	**11:14am**
AIR BLK	.00	11:15am
SUB TEST	**.00**	**11:17am**
AIR BLK	.00	11:17am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

UNION COUNTY UNION COUNTY SD 890

Serial Number: *008866* Test Record Number: *1467*
Test Date: *06/27/2013* Test Time: *11:06am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:07am
FLO	Pass	11:07am
FC	Pass	11:07am

Temperature Tests

Test	Status	Time
FC1	Pass	11:07am
SRC	Pass	11:07am
DET	Pass	11:07am
BAR	Pass	11:07am
BT	Pass	11:07am

Blank Tests

Test	Status	Time
AIR	Pass	11:07am

Printer Tests

Test	Status	Time
PRNT	Pass	11:07am

CRC Tests

Test	Status	Time
COMP	Pass	11:08am
CAL	Pass	11:08am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Stanly Instrument Location Stanly County SD

Instrument Serial No. 008824 126 S. 3rd St. Albemarle
704-986-3734

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27th day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

STANLY COUNTY STANLY COUNTY SD 830

Serial Number: *008824* Test Record Number: *860*
Test Date: *06/27/2013* Test Time: *12:43pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:44pm
FLO	Pass	12:44pm
FC	Pass	12:44pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:44pm
SRC	Pass	12:44pm
DET	Pass	12:44pm
BAR	Pass	12:44pm
BT	Pass	12:44pm

Blank Tests

Test	Status	Time
AIR	Pass	12:45pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:45pm

CRC Tests

Test	Status	Time
COMP	Pass	12:45pm
CAL	Pass	12:45pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

STANLY COUNTY STANLY COUNTY SD 830

Serial Number: *008824*
Test Date: *06/27/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG308002
Exp Date: 03/21/2015

Test	g/210L	Time
DIAG	Pass	12:52pm
AIR BLK	.00	12:53pm
ACCY CHK	.08	12:53pm
AIR BLK	.00	12:55pm
SUB TEST	**.00**	**12:55pm**
AIR BLK	.00	12:56pm
SUB TEST	**.00**	**12:58pm**
AIR BLK	.00	12:58pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Avery Instrument Location Banner Elk PD

Instrument Serial No. 008724 Banner Elk, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

AVERY COUNTY BANNER ELK PD 050

Serial Number: *008724*
Test Date: *06/20/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BURNETTE, ANTHONY J*
Permit Number: *11304E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203103
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	4:26pm
AIR BLK	.00	4:26pm
ACCY CHK	.07	4:27pm
AIR BLK	.00	4:28pm
SUB TEST	**.00**	**4:29pm**
AIR BLK	.00	4:29pm
SUB TEST	**.00**	**4:31pm**
AIR BLK	.00	4:32pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

AVERY COUNTY BANNER ELK PD 050

Serial Number: *008724* Test Record Number: *324*
Test Date: *06/20/2013* Test Time: *4:33pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:33pm
FLO	Pass	4:33pm
FC	Pass	4:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:33pm
SRC	Pass	4:33pm
DET	Pass	4:33pm
BAR	Pass	4:33pm
BT	Pass	4:33pm

Blank Tests

Test	Status	Time
AIR	Pass	4:34pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:34pm

CRC Tests

Test	Status	Time
COMP	Pass	4:34pm
CAL	Pass	4:34pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Watauga Instrument Location Boone PD

Instrument Serial No. 008716 Boone NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WATAUGA COUNTY BOONE P D 940

Serial Number: *008716*
Test Date: *06/26/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BURNETTE, ANTHONY J*
Permit Number: *11304E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309105
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	3:19pm
AIR BLK	.00	3:20pm
ACCY CHK	.08	3:21pm
AIR BLK	.00	3:22pm
SUB TEST	**.00**	**3:22pm**
AIR BLK	.00	3:23pm
SUB TEST	**.00**	**3:24pm**
AIR BLK	.00	3:25pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WATAUGA COUNTY BOONE P D 940

Serial Number: *008716* Test Record Number: *1415*
Test Date: *06/26/2013* Test Time: *3:26pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:26pm
FLO	Pass	3:26pm
FC	Pass	3:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:26pm
SRC	Pass	3:26pm
DET	Pass	3:26pm
BAR	Pass	3:26pm
BT	Pass	3:26pm

Blank Tests

Test	Status	Time
AIR	Pass	3:27pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:27pm

CRC Tests

Test	Status	Time
COMP	Pass	3:27pm
CAL	Pass	3:27pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Graham Instrument Location Graham Co. S.O.

Instrument Serial No. 008915 Robbinsville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GRAHAM COUNTY GRAHAM COUNTY SD 370

Serial Number: *008915*
Test Date: *06/10/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *CUTLER, DANIEL R*
Permit Number: *8457E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203103
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	12:10pm
AIR BLK	.00	12:11pm
ACCY CHK	.07	12:12pm
AIR BLK	.00	12:13pm
SUB TEST	**.00**	**12:14pm**
AIR BLK	.00	12:15pm
SUB TEST	**.00**	**12:17pm**
AIR BLK	.00	12:18pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GRAHAM COUNTY GRAHAM COUNTY SD 370

Serial Number: *008915* Test Record Number: *552*
Test Date: *06/10/2013* Test Time: *12:20pm*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:21pm
FLO	Pass	12:21pm
FC	Pass	12:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:21pm
SRC	Pass	12:21pm
DET	Pass	12:21pm
BAR	Pass	12:21pm
BT	Pass	12:21pm

Blank Tests

Test	Status	Time
AIR	Pass	12:21pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:21pm

CRC Tests

Test	Status	Time
COMP	Pass	12:22pm
CAL	Pass	12:22pm

Preventive Maintenance
Status: Pass

David R. Cathon

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Swain Instrument Location Swain Co. Jail

Instrument Serial No. 008727 Bryson City, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of June, 20 12 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: *008727*
Test Date: *06/05/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *CUTLER, DANIEL R*
Permit Number: *8457E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG300202
Exp Date: 01/02/2015

Test	g/210L	Time
DIAG	Pass	10:05am
AIR BLK	.00	10:06am
ACCY CHK	.08	10:06am
AIR BLK	.00	10:07am
SUB TEST	**.00**	**10:07am**
AIR BLK	.00	10:08am
SUB TEST	**.00**	**10:10am**
AIR BLK	.00	10:11am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: *008727* Test Record Number: *750*
Test Date: *06/05/2013* Test Time: *10:11am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:12am
FLO	Pass	10:12am
FC	Pass	10:12am

Temperature Tests

Test	Status	Time
FC1	Pass	10:12am
SRC	Pass	10:12am
DET	Pass	10:12am
BAR	Pass	10:12am
BT	Pass	10:12am

Blank Tests

Test	Status	Time
AIR	Pass	10:12am

Printer Tests

Test	Status	Time
PRNT	Pass	10:13am

CRC Tests

Test	Status	Time
COMP	Pass	10:13am
CAL	Pass	10:13am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Swain Instrument Location Swain Co. Jail

Instrument Serial No. 008723 Bryson City, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: *008723*
Test Date: *06/05/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *CUTLER, DANIEL R*
Permit Number: *8457E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206602
Exp Date: 03/06/2014

Test	g/210L	Time
DIAG	Pass	10:04am
AIR BLK	.00	10:05am
ACCY CHK	.08	10:05am
AIR BLK	.00	10:06am
SUB TEST	**.00**	**10:07am**
AIR BLK	.00	10:07am
SUB TEST	**.00**	**10:09am**
AIR BLK	.00	10:10am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: *008723* Test Record Number: *480*
Test Date: *06/05/2013* Test Time: *10:11am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:11am
FLO	Pass	10:11am
FC	Pass	10:11am

Temperature Tests

Test	Status	Time
FC1	Pass	10:11am
SRC	Pass	10:11am
DET	Pass	10:11am
BAR	Pass	10:11am
BT	Pass	10:11am

Blank Tests

Test	Status	Time
AIR	Pass	10:12am

Printer Tests

Test	Status	Time
PRNT	Pass	10:12am

CRC Tests

Test	Status	Time
COMP	Pass	10:12am
CAL	Pass	10:12am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Madison Instrument Location Mars Hill PD

Instrument Serial No. 008582 Mars Hill, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official 649 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MADISON COUNTY MARS HILL PD 560

Serial Number: *008582*
Test Date: *06/07/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BURNETTE, ANTHONY J*
Permit Number: *11304E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG303502
Exp Date: 02/04/2015

Test	g/210L	Time
DIAG	Pass	1:22pm
AIR BLK	.00	1:23pm
ACCY CHK	.08	1:23pm
AIR BLK	.00	1:24pm
SUB TEST	**.00**	**1:25pm**
AIR BLK	.00	1:26pm
SUB TEST	**.00**	**1:27pm**
AIR BLK	.00	1:28pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MADISON COUNTY MARS HILL PD 560

Serial Number: *008582* Test Record Number: *939*
Test Date: *06/07/2013* Test Time: *1:29pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:30pm
FLO	Pass	1:30pm
FC	Pass	1:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:30pm
SRC	Pass	1:30pm
DET	Pass	1:30pm
BAR	Pass	1:30pm
BT	Pass	1:30pm

Blank Tests

Test	Status	Time
AIR	Pass	1:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:31pm

CRC Tests

Test	Status	Time
COMP	Pass	1:31pm
CAL	Pass	1:31pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Henderson Instrument Location Henderson Co. Detention

Instrument Serial No. 008806 Hendersonville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

HENDERSON COUNTY DENTENTION 440

Serial Number: *008806*
Test Date: *06/19/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BURNETTE, ANTHONY J*
Permit Number: *11304E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206102
Exp Date: 03/01/2014

Test	g/210L	Time
DIAG	Pass	4:21pm
AIR BLK	.00	4:22pm
ACCY CHK	.07	4:23pm
AIR BLK	.00	4:24pm
SUB TEST	**.00**	**4:25pm**
AIR BLK	.00	4:26pm
SUB TEST	**.00**	**4:27pm**
AIR BLK	.00	4:28pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HENDERSON COUNTY DENTENTION 440

Serial Number: *008806* Test Record Number: *1351*
Test Date: *06/19/2013* Test Time: *4:30pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:30pm
FLO	Pass	4:30pm
FC	Pass	4:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:30pm
SRC	Pass	4:30pm
DET	Pass	4:30pm
BAR	Pass	4:30pm
BT	Pass	4:30pm

Blank Tests

Test	Status	Time
AIR	Pass	4:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:31pm

CRC Tests

Test	Status	Time
COMP	Pass	4:31pm
CAL	Pass	4:31pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Henderson Instrument Location Henderson Co. Detention

Instrument Serial No. 008822 Hendersonville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HENDERSON COUNTY DETENTION 440

Serial Number: *008822*
Test Date: *06/19/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BURNETTE, ANTHONY J*
Permit Number: *11304E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG200401
Exp Date: 01/04/2014

Test	g/210L	Time
DIAG	Pass	4:23pm
AIR BLK	.00	4:24pm
ACCY CHK	.07	4:24pm
AIR BLK	.00	4:25pm
SUB TEST	**.00**	**4:26pm**
AIR BLK	.00	4:26pm
SUB TEST	**.00**	**4:28pm**
AIR BLK	.00	4:29pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HENDERSON COUNTY DETENTION 440

Serial Number: *008822* Test Record Number: *1439*
Test Date: *06/19/2013* Test Time: *4:30pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:31pm
FLO	Pass	4:31pm
FC	Pass	4:31pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:31pm
SRC	Pass	4:31pm
DET	Pass	4:31pm
BAR	Pass	4:31pm
BT	Pass	4:31pm

Blank Tests

Test	Status	Time
AIR	Pass	4:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:32pm

CRC Tests

Test	Status	Time
COMP	Pass	4:32pm
CAL	Pass	4:32pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mitchell Instrument Location Spruce Pine PD

Instrument Serial No. 008726 Spruce Pine NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: *008726*
Test Date: *06/17/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BURNETTE, ANTHONY J*
Permit Number: *11304E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309105
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	5:02pm
AIR BLK	.00	5:03pm
ACCY CHK	.08	5:03pm
AIR BLK	.00	5:04pm
SUB TEST	**.00**	**5:05pm**
AIR BLK	.00	5:05pm
SUB TEST	**.00**	**5:07pm**
AIR BLK	.00	5:08pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: *008726* Test Record Number: *608*
Test Date: *06/17/2013* Test Time: *6:05pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:05pm
FLO	Pass	6:05pm
FC	Pass	6:05pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:06pm
SRC	Pass	6:06pm
DET	Pass	6:06pm
BAR	Pass	6:06pm
BT	Pass	6:06pm

Blank Tests

Test	Status	Time
AIR	Pass	6:06pm

Printer Tests

Test	Status	Time
PRNT	Pass	6:06pm

CRC Tests

Test	Status	Time
COMP	Pass	6:06pm
CAL	Pass	6:06pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Buncombe Instrument Location Buncombe Co. Jail

Instrument Serial No. 008697 Asheville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of June, 2015 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BUNCOMBE COUNTY BUNCOMBE COUNTY JAIL
100

Serial Number: *008697*
Test Date: *06/13/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BURNETTE, ANTHONY J*
Permit Number: *11304E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309101
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	4:58pm
AIR BLK	.00	4:59pm
ACCY CHK	.07	5:00pm
AIR BLK	.00	5:01pm
SUB TEST	**.00**	**5:01pm**
AIR BLK	.00	5:02pm
SUB TEST	**.00**	**5:04pm**
AIR BLK	.00	5:05pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BUNCOMBE COUNTY BUNCOMBE COUNTY JAIL 100

Serial Number: *008697* Test Record Number: *2063*
Test Date: *06/13/2013* Test Time: *5:07pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:07pm
FLO	Pass	5:07pm
FC	Pass	5:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:07pm
SRC	Pass	5:07pm
DET	Pass	5:07pm
BAR	Pass	5:07pm
BT	Pass	5:07pm

Blank Tests

Test	Status	Time
AIR	Pass	5:08pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:08pm

CRC Tests

Test	Status	Time
COMP	Pass	5:08pm
CAL	Pass	5:08pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Buncombe Instrument Location Buncombe Co. Jail

Instrument Serial No. 008798 Asheville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BUNCOMBE COUNTY BUNCOMBE COUNTY JAIL
100

Serial Number: *008798*
Test Date: *06/13/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BURNETTE, ANTHONY J*
Permit Number: *11304E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG205401
Exp Date: 02/23/2014

Test	g/210L	Time
DIAG	Pass	4:47pm
AIR BLK	.00	4:48pm
ACCY CHK	.08	4:48pm
AIR BLK	.00	4:49pm
SUB TEST	**.00**	**4:50pm**
AIR BLK	.00	4:51pm
SUB TEST	**.00**	**4:52pm**
AIR BLK	.00	4:53pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BUNCOMBE COUNTY BUNCOMBE COUNTY JAIL 100

Serial Number: *008798* Test Record Number: *2692*
Test Date: *06/13/2013* Test Time: *4:54pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:55pm
FLO	Pass	4:55pm
FC	Pass	4:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:55pm
SRC	Pass	4:55pm
DET	Pass	4:55pm
BAR	Pass	4:55pm
BT	Pass	4:55pm

Blank Tests

Test	Status	Time
AIR	Pass	4:55pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:56pm

CRC Tests

Test	Status	Time
COMP	Pass	4:56pm
CAL	Pass	4:56pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Bat Mobile Cary

Instrument Serial No. 008778 Cary

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008778* Test Record Number: *1022*
Test Date: *06/28/2013* Test Time: *11:34pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:35pm
FLO	Pass	11:35pm
FC	Pass	11:35pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:35pm
SRC	Pass	11:35pm
DET	Pass	11:35pm
BAR	Pass	11:35pm
BT	Pass	11:35pm

Blank Tests

Test	Status	Time
AIR	Pass	11:36pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:36pm

CRC Tests

Test	Status	Time
COMP	Pass	11:36pm
CAL	Pass	11:36pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008778*
Test Date: *06/28/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG315701
Exp Date: 06/06/2015

Test	g/210L	Time
DIAG	Pass	11:16pm
AIR BLK	.00	11:17pm
ACCY CHK	.08	11:17pm
AIR BLK	.00	11:18pm
SUB TEST	**.00**	**11:19pm**
AIR BLK	.00	11:20pm
SUB TEST	**.00**	**11:21pm**
AIR BLK	.00	11:22pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Sampson Instrument Location BAT MOBILE UNIT #7

Instrument Serial No. 008778

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

SAMPSON COUNTY BAT MOBILE UNIT 7 810

Serial Number: *008778* Test Record Number: *1028*
Test Date: *06/29/2013* Test Time: *10:31pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:31pm
FLO	Pass	10:31pm
FC	Pass	10:31pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:31pm
SRC	Pass	10:31pm
DET	Pass	10:31pm
BAR	Pass	10:31pm
BT	Pass	10:31pm

Blank Tests

Test	Status	Time
AIR	Pass	10:32pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:32pm

CRC Tests

Test	Status	Time
COMP	Pass	10:32pm
CAL	Pass	10:32pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

SAMPSON COUNTY BAT MOBILE UNIT 7 810

Serial Number: *008778*
Test Date: *06/29/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG315701
Exp Date: 06/06/2015

Test	g/210L	Time
DIAG	Pass	10:20pm
AIR BLK	.00	10:21pm
ACCY CHK	.08	10:22pm
AIR BLK	.00	10:23pm
SUB TEST	**.00**	**10:24pm**
AIR BLK	.00	10:25pm
SUB TEST	**.00**	**10:26pm**
AIR BLK	.00	10:27pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Sampson Instrument Location BAT MOBILE UNIT #7

Instrument Serial No. 008612

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

SAMPSON COUNTY BAT MOBILE UNIT 7 810

Serial Number: *008612* Test Record Number: *1355*
Test Date: *06/29/2013* Test Time: *9:44pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:45pm
FLO	Pass	9:45pm
FC	Pass	9:45pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:45pm
SRC	Pass	9:45pm
DET	Pass	9:45pm
BAR	Pass	9:45pm
BT	Pass	9:45pm

Blank Tests

Test	Status	Time
AIR	Pass	9:46pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:46pm

CRC Tests

Test	Status	Time
COMP	Pass	9:46pm
CAL	Pass	9:46pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

SAMPSON COUNTY BAT MOBILE UNIT 7 810

Serial Number: *008612*
Test Date: *06/29/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG303502
Exp Date: 02/04/2015

Test	g/210L	Time
DIAG	Pass	9:32pm
AIR BLK	.00	9:33pm
ACCY CHK	.07	9:34pm
AIR BLK	.00	9:35pm
SUB TEST	**.00**	**9:35pm**
AIR BLK	.00	9:36pm
SUB TEST	**.00**	**9:38pm**
AIR BLK	.00	9:39pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Sampson Instrument Location BAT MOBILE UNIT #7

Instrument Serial No. 008623

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

6356
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

SAMPSON COUNTY BAT MOBILE UNIT 7 810
Serial Number: *008623* Test Record Number: *2727*
Test Date: *06/29/2013* Test Time: *10:18pm EDT*
System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:19pm
FLO	Pass	10:19pm
FC	Pass	10:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:19pm
SRC	Pass	10:19pm
DET	Pass	10:19pm
BAR	Pass	10:19pm
BT	Pass	10:19pm

Blank Tests

Test	Status	Time
AIR	Pass	10:20pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:20pm

CRC Tests

Test	Status	Time
COMP	Pass	10:20pm
CAL	Pass	10:20pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

SAMPSON COUNTY BAT MOBILE UNIT 7 810
Serial Number: *008623*
Test Date: *06/29/2013*
Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	10:08pm
AIR BLK	.00	10:09pm
ACCY CHK	.07	10:09pm
AIR BLK	.00	10:10pm
SUB TEST	**.00**	**10:11pm**
AIR BLK	.00	10:12pm
SUB TEST	**.00**	**10:13pm**
AIR BLK	.00	10:14pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Sampson Instrument Location BAT Mobile Unit #7

Instrument Serial No. CC 8760

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

SAMPSON COUNTY BAT MOBILE UNIT 7 810

Serial Number: *008760* Test Record Number: *455*
Test Date: *06/29/2013* Test Time: *10:09pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:10pm
FLO	Pass	10:10pm
FC	Pass	10:10pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:10pm
SRC	Pass	10:10pm
DET	Pass	10:10pm
BAR	Pass	10:10pm
BT	Pass	10:10pm

Blank Tests

Test	Status	Time
AIR	Pass	10:11pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:11pm

CRC Tests

Test	Status	Time
COMP	Pass	10:11pm
CAL	Pass	10:11pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

SAMPSON COUNTY BAT MOBILE UNIT 7 810

Serial Number: *008760*
Test Date: *06/29/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	9:55pm
AIR BLK	.00	9:56pm
ACCY CHK	.07	9:57pm
AIR BLK	.00	9:58pm
SUB TEST	**.00**	**9:58pm**
AIR BLK	.00	9:59pm
SUB TEST	**.00**	**10:02pm**
AIR BLK	.00	10:03pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Sampson Instrument Location BAT MOBILE UNIT #7

Instrument Serial No. 008577

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

SAMPSON COUNTY BAT MOBILE UNIT 7 810

Serial Number: *008577* Test Record Number: *876*
Test Date: *06/29/2013* Test Time: *9:57pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:58pm
FLO	Pass	9:58pm
FC	Pass	9:58pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:58pm
SRC	Pass	9:58pm
DET	Pass	9:58pm
BAR	Pass	9:58pm
BT	Pass	9:58pm

Blank Tests

Test	Status	Time
AIR	Pass	9:59pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:59pm

CRC Tests

Test	Status	Time
COMP	Pass	9:59pm
CAL	Pass	9:59pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

SAMPSON COUNTY BAT MOBILE UNIT 7 810

Serial Number: *008577*
Test Date: *06/29/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	9:40pm
AIR BLK	.00	9:41pm
ACCY CHK	.07	9:41pm
AIR BLK	.00	9:42pm
SUB TEST	**.00**	**9:43pm**
AIR BLK	.00	9:44pm
SUB TEST	**.00**	**9:45pm**
AIR BLK	.00	9:46pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County PERSON Instrument Location PERSON CO. LEC

Instrument Serial No. 008693 120 COURT ST. ROXBORO, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PERSON COUNTY PERSON CO. LEC 720

Serial Number: *008693*
Test Date: *06/12/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG303502
Exp Date: 02/04/2015

Test	g/210L	Time
DIAG	Pass	11:33am
AIR BLK	.00	11:34am
ACCY CHK	.07	11:34am
AIR BLK	.00	11:35am
SUB TEST	**.00**	**11:36am**
AIR BLK	.00	11:36am
SUB TEST	**.00**	**11:38am**
AIR BLK	.00	11:38am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PERSON COUNTY PERSON CO. LEC 720

Serial Number: *008693* Test Record Number: *936*
Test Date: *06/12/2013* Test Time: *11:40am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:40am
FLO	Pass	11:40am
FC	Pass	11:40am

Temperature Tests

Test	Status	Time
FC1	Pass	11:40am
SRC	Pass	11:40am
DET	Pass	11:40am
BAR	Pass	11:40am
BT	Pass	11:40am

Blank Tests

Test	Status	Time
AIR	Pass	11:41am

Printer Tests

Test	Status	Time
PRNT	Pass	11:41am

CRC Tests

Test	Status	Time
COMP	Pass	11:41am
CAL	Pass	11:41am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County PERSON Instrument Location PERSON Co. LEC

Instrument Serial No. 008880 120 COURT ST ROXBORO, N.C

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PERSON COUNTY PERSON CO. LEC 720

Serial Number: *008880*
Test Date: *06/12/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206103
Exp Date: 03/01/2014

Test	g/210L	Time
DIAG	Pass	11:30am
AIR BLK	.00	11:30am
ACCY CHK	.08	11:31am
AIR BLK	.00	11:32am
SUB TEST	**.00**	**11:32am**
AIR BLK	.00	11:33am
SUB TEST	**.00**	**11:35am**
AIR BLK	.00	11:35am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PERSON COUNTY PERSON CO. LEC 720

Serial Number: *008880* Test Record Number: *552*
Test Date: *06/12/2013* Test Time: *11:36am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:37am
FLO	Pass	11:37am
FC	Pass	11:37am

Temperature Tests

Test	Status	Time
FC1	Pass	11:37am
SRC	Pass	11:37am
DET	Pass	11:37am
BAR	Pass	11:37am
BT	Pass	11:37am

Blank Tests

Test	Status	Time
AIR	Pass	11:38am

Printer Tests

Test	Status	Time
PRNT	Pass	11:38am

CRC Tests

Test	Status	Time
COMP	Pass	11:38am
CAL	Pass	11:38am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County GRANVILLE Instrument Location OXFORD P.D.

Instrument Serial No. 008923 204 E. McCLANAHAN ST. OXFORD, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GRANVILLE COUNTY OXFORD PD 380

Serial Number: *008923*
Test Date: *06/20/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG202602
Exp Date: 01/26/2014

Test	g/210L	Time
DIAG	Pass	11:48am
AIR BLK	.00	11:50am
ACCY CHK	.07	11:50am
AIR BLK	.00	11:51am
SUB TEST	**.00**	**11:52am**
AIR BLK	.00	11:52am
SUB TEST	**.00**	**11:55am**
AIR BLK	.00	11:55am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GRANVILLE COUNTY OXFORD PD 380

Serial Number: *008923* Test Record Number: *945*
Test Date: *06/20/2013* Test Time: *11:57am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:57am
FLO	Pass	11:57am
FC	Pass	11:57am

Temperature Tests

Test	Status	Time
FC1	Pass	11:57am
SRC	Pass	11:57am
DET	Pass	11:57am
BAR	Pass	11:57am
BT	Pass	11:57am

Blank Tests

Test	Status	Time
AIR	Pass	11:58am

Printer Tests

Test	Status	Time
PRNT	Pass	11:58am

CRC Tests

Test	Status	Time
COMP	Pass	11:58am
CAL	Pass	11:58am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ROBESON Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008616 RED SPRINGS, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROBESON COUNTY BAT MOBILE UNIT 3 770

Serial Number: *008616*
Test Date: *06/01/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	10:59pm
AIR BLK	.00	11:00pm
ACCY CHK	.08	11:01pm
AIR BLK	.00	11:01pm
SUB TEST	**.00**	**11:02pm**
AIR BLK	.00	11:03pm
SUB TEST	**.00**	**11:05pm**
AIR BLK	.00	11:06pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY BAT MOBILE UNIT 3 770

Serial Number: *008616* Test Record Number: *1655*
Test Date: *06/01/2013* Test Time: *11:07pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:07pm
FLO	Pass	11:07pm
FC	Pass	11:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:07pm
SRC	Pass	11:07pm
DET	Pass	11:07pm
BAR	Pass	11:07pm
BT	Pass	11:07pm

Blank Tests

Test	Status	Time
AIR	Pass	11:08pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:08pm

CRC Tests

Test	Status	Time
COMP	Pass	11:08pm
CAL	Pass	11:08pm

Preventive Maintenance
Status: Pass

Alan Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ROBESON Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 RED SPRINGS, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROBESON COUNTY BAT MOBILE UNIT 3 770

Serial Number: *008647*
Test Date: *06/01/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	10:01pm
AIR BLK	.00	10:02pm
ACCY CHK	.07	10:03pm
AIR BLK	.00	10:04pm
SUB TEST	**.00**	**10:04pm**
AIR BLK	.00	10:05pm
SUB TEST	**.00**	**10:07pm**
AIR BLK	.00	10:07pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY BAT MOBILE UNIT 3 770

Serial Number: *008647* Test Record Number: *1673*
Test Date: *06/01/2013* Test Time: *10:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:08pm
FLO	Pass	10:08pm
FC	Pass	10:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:09pm
SRC	Pass	10:09pm
DET	Pass	10:09pm
BAR	Pass	10:09pm
BT	Pass	10:09pm

Blank Tests

Test	Status	Time
AIR	Pass	10:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:09pm

CRC Tests

Test	Status	Time
COMP	Pass	10:09pm
CAL	Pass	10:09pm

Preventive Maintenance
Status: Pass

Alan Ray Barnes

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ROBESON Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 RED SPRINGS, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Burns
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROBESON COUNTY BAT MOBILE UNIT 3 770

Serial Number: *008707*
Test Date: *06/01/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	9:49pm
AIR BLK	.00	9:50pm
ACCY CHK	.07	9:50pm
AIR BLK	.00	9:51pm
SUB TEST	**.00**	**9:52pm**
AIR BLK	.00	9:53pm
SUB TEST	**.00**	**9:55pm**
AIR BLK	.00	9:56pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY BAT MOBILE UNIT 3 770

Serial Number: *008707* Test Record Number: *1736*
Test Date: *06/01/2013* Test Time: *9:57pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:57pm
FLO	Pass	9:57pm
FC	Pass	9:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:57pm
SRC	Pass	9:57pm
DET	Pass	9:57pm
BAR	Pass	9:57pm
BT	Pass	9:57pm

Blank Tests

Test	Status	Time
AIR	Pass	9:58pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:58pm

CRC Tests

Test	Status	Time
COMP	Pass	9:58pm
CAL	Pass	9:58pm

Preventive Maintenance
Status: Pass

Allen Ray Bauer

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County GUILFORD Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: *008647*
Test Date: *06/07/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	9:01pm
AIR BLK	.00	9:02pm
ACCY CHK	.07	9:03pm
AIR BLK	.00	9:04pm
SUB TEST	**.00**	**9:04pm**
AIR BLK	.00	9:05pm
SUB TEST	**.00**	**9:07pm**
AIR BLK	.00	9:08pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: *008647* Test Record Number: *1677*
Test Date: *06/07/2013* Test Time: *9:09pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:09pm
FLO	Pass	9:09pm
FC	Pass	9:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:09pm
SRC	Pass	9:09pm
DET	Pass	9:09pm
BAR	Pass	9:09pm
BT	Pass	9:09pm

Blank Tests

Test	Status	Time
AIR	Pass	9:10pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:10pm

CRC Tests

Test	Status	Time
COMP	Pass	9:10pm
CAL	Pass	9:10pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County GUILFORD Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008616 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: *008616*
Test Date: *06/07/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	8:58pm
AIR BLK	.00	8:59pm
ACCY CHK	.08	8:59pm
AIR BLK	.00	9:00pm
SUB TEST	**.00**	**9:01pm**
AIR BLK	.00	9:02pm
SUB TEST	**.00**	**9:03pm**
AIR BLK	.00	9:04pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: *008616* Test Record Number: *1659*
Test Date: *06/07/2013* Test Time: *9:04pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:05pm
FLO	Pass	9:05pm
FC	Pass	9:05pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:05pm
SRC	Pass	9:05pm
DET	Pass	9:05pm
BAR	Pass	9:05pm
BT	Pass	9:05pm

Blank Tests

Test	Status	Time
AIR	Pass	9:06pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:06pm

CRC Tests

Test	Status	Time
COMP	Pass	9:06pm
CAL	Pass	9:06pm

Preventive Maintenance
Status: Pass

Alan Ray Bennett
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CABARRUS Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 KANNAPOLIS, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CABARRUS COUNTY BAT MOBILE UNIT 3 120

Serial Number: *008707*
Test Date: *06/14/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	9:29pm
AIR BLK	.00	9:30pm
ACCY CHK	.08	9:31pm
AIR BLK	.00	9:32pm
SUB TEST	**.00**	**9:32pm**
AIR BLK	.00	9:33pm
SUB TEST	**.00**	**9:35pm**
AIR BLK	.00	9:36pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CABARRUS COUNTY BAT MOBILE UNIT 3 120

Serial Number: *008707* Test Record Number: *1741*
Test Date: *06/14/2013* Test Time: *9:36pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:37pm
FLO	Pass	9:37pm
FC	Pass	9:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:37pm
SRC	Pass	9:37pm
DET	Pass	9:37pm
BAR	Pass	9:37pm
BT	Pass	9:37pm

Blank Tests

Test	Status	Time
AIR	Pass	9:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:38pm

CRC Tests

Test	Status	Time
COMP	Pass	9:38pm
CAL	Pass	9:38pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CABARRUS Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008616 KANNAPOLIS, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CABARRUS COUNTY BAT MOBILE UNIT 3 120

Serial Number: *008616*
Test Date: *06/14/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	9:30pm
AIR BLK	.00	9:31pm
ACCY CHK	.08	9:32pm
AIR BLK	.00	9:33pm
SUB TEST	**.00**	**9:33pm**
AIR BLK	.00	9:34pm
SUB TEST	**.00**	**9:36pm**
AIR BLK	.00	9:37pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CABARRUS COUNTY BAT MOBILE UNIT 3 120

Serial Number: *008616* Test Record Number: *1664*
Test Date: *06/14/2013* Test Time: *9:39pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:40pm
FLO	Pass	9:40pm
FC	Pass	9:40pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:40pm
SRC	Pass	9:40pm
DET	Pass	9:40pm
BAR	Pass	9:40pm
BT	Pass	9:40pm

Blank Tests

Test	Status	Time
AIR	Pass	9:40pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:40pm

CRC Tests

Test	Status	Time
COMP	Pass	9:41pm
CAL	Pass	9:41pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CABARRUS Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 CONCORD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes — Signature of Certifying Official

648 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CABARRUS COUNTY BAT MOBILE UNIT 3 120

Serial Number: *008647*
Test Date: *06/21/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	10:03pm
AIR BLK	.00	10:04pm
ACCY CHK	.07	10:05pm
AIR BLK	.00	10:06pm
SUB TEST	**.00**	**10:06pm**
AIR BLK	.00	10:07pm
SUB TEST	**.00**	**10:09pm**
AIR BLK	.00	10:10pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CABARRUS COUNTY BAT MOBILE UNIT 3 120

Serial Number: *008647* Test Record Number: *1685*
Test Date: *06/21/2013* Test Time: *10:12pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:12pm
FLO	Pass	10:12pm
FC	Pass	10:13pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:13pm
SRC	Pass	10:13pm
DET	Pass	10:13pm
BAR	Pass	10:13pm
BT	Pass	10:13pm

Blank Tests

Test	Status	Time
AIR	Pass	10:13pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:13pm

CRC Tests

Test	Status	Time
COMP	Pass	10:14pm
CAL	Pass	10:14pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CABARRUS Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008616 CONCORD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CABARRUS COUNTY BAT MOBILE UNIT 3 120

Serial Number: *008616*
Test Date: *06/21/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	10:00pm
AIR BLK	.00	10:01pm
ACCY CHK	.08	10:01pm
AIR BLK	.00	10:02pm
SUB TEST	**.00**	**10:03pm**
AIR BLK	.00	10:04pm
SUB TEST	**.00**	**10:05pm**
AIR BLK	.00	10:06pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CABARRUS COUNTY BAT MOBILE UNIT 3 120

Serial Number: *008616* Test Record Number: *1669*
Test Date: *06/21/2013* Test Time: *10:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:08pm
FLO	Pass	10:08pm
FC	Pass	10:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:08pm
SRC	Pass	10:08pm
DET	Pass	10:08pm
BAR	Pass	10:08pm
BT	Pass	10:08pm

Blank Tests

Test	Status	Time
AIR	Pass	10:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:09pm

CRC Tests

Test	Status	Time
COMP	Pass	10:09pm
CAL	Pass	10:09pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CABARRUS Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 CONCORD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CABARRUS COUNTY BAT MOBILE UNIT 3 120

Serial Number: *008707*
Test Date: *06/21/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	9:57pm
AIR BLK	.00	9:58pm
ACCY CHK	.08	9:59pm
AIR BLK	.00	10:00pm
SUB TEST	**.00**	**10:00pm**
AIR BLK	.00	10:01pm
SUB TEST	**.00**	**10:03pm**
AIR BLK	.00	10:04pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CABARRUS COUNTY BAT MOBILE UNIT 3 120

Serial Number: *008707* Test Record Number: *1748*
Test Date: *06/21/2013* Test Time: *10:04pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:04pm
FLO	Pass	10:04pm
FC	Pass	10:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:04pm
SRC	Pass	10:04pm
DET	Pass	10:04pm
BAR	Pass	10:04pm
BT	Pass	10:04pm

Blank Tests

Test	Status	Time
AIR	Pass	10:05pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:05pm

CRC Tests

Test	Status	Time
COMP	Pass	10:05pm
CAL	Pass	10:05pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County DAVIDSON Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 SOUTHMONT, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY BAT MOBILE UNIT 3 280

Serial Number: *008647*
Test Date: *06/22/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	3:20pm
AIR BLK	.00	3:22pm
ACCY CHK	.07	3:22pm
AIR BLK	.00	3:23pm
SUB TEST	**.00**	**3:23pm**
AIR BLK	.00	3:24pm
SUB TEST	**.00**	**3:26pm**
AIR BLK	.00	3:27pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY BAT MOBILE UNIT 3 280

Serial Number: *008647* Test Record Number: *1688*
Test Date: *06/22/2013* Test Time: *3:28pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:29pm
FLO	Pass	3:29pm
FC	Pass	3:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:29pm
SRC	Pass	3:29pm
DET	Pass	3:29pm
BAR	Pass	3:29pm
BT	Pass	3:29pm

Blank Tests

Test	Status	Time
AIR	Pass	3:30pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:30pm

CRC Tests

Test	Status	Time
COMP	Pass	3:30pm
CAL	Pass	3:30pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008616 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: *008616*
Test Date: *06/27/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	11:25pm
AIR BLK	.00	11:26pm
ACCY CHK	.08	11:26pm
AIR BLK	.00	11:27pm
SUB TEST	**.00**	**11:28pm**
AIR BLK	.00	11:29pm
SUB TEST	**.00**	**11:30pm**
AIR BLK	.00	11:31pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: *008616* Test Record Number: *1674*
Test Date: *06/27/2013* Test Time: *11:32pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:32pm
FLO	Pass	11:32pm
FC	Pass	11:32pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:32pm
SRC	Pass	11:32pm
DET	Pass	11:32pm
BAR	Pass	11:32pm
BT	Pass	11:32pm

Blank Tests

Test	Status	Time
AIR	Pass	11:33pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:33pm

CRC Tests

Test	Status	Time
COMP	Pass	11:33pm
CAL	Pass	11:33pm

Preventive Maintenance
Status: Pass

Alan Ray Beans

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: *008707*
Test Date: *06/27/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	11:23pm
AIR BLK	.00	11:24pm
ACCY CHK	.08	11:25pm
AIR BLK	.00	11:26pm
SUB TEST	**.00**	**11:27pm**
AIR BLK	.00	11:28pm
SUB TEST	**.00**	**11:29pm**
AIR BLK	.00	11:30pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: *008707* Test Record Number: *1753*
Test Date: *06/27/2013* Test Time: *11:31pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:31pm
FLO	Pass	11:31pm
FC	Pass	11:31pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:31pm
SRC	Pass	11:31pm
DET	Pass	11:31pm
BAR	Pass	11:31pm
BT	Pass	11:31pm

Blank Tests

Test	Status	Time
AIR	Pass	11:32pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:32pm

CRC Tests

Test	Status	Time
COMP	Pass	11:32pm
CAL	Pass	11:32pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: *008647*
Test Date: *06/27/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	11:26pm
AIR BLK	.00	11:27pm
ACCY CHK	.07	11:28pm
AIR BLK	.00	11:29pm
SUB TEST	**.00**	**11:29pm**
AIR BLK	.00	11:30pm
SUB TEST	**.00**	**11:32pm**
AIR BLK	.00	11:33pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: *008647* Test Record Number: *1692*
Test Date: *06/27/2013* Test Time: *11:33pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:34pm
FLO	Pass	11:34pm
FC	Pass	11:34pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:34pm
SRC	Pass	11:34pm
DET	Pass	11:34pm
BAR	Pass	11:34pm
BT	Pass	11:34pm

Blank Tests

Test	Status	Time
AIR	Pass	11:34pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:34pm

CRC Tests

Test	Status	Time
COMP	Pass	11:35pm
CAL	Pass	11:35pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County GRANVILLE Instrument Location CREEDMOOR P.D.

Instrument Serial No. 008641 111 MASONIC ST. CREEDMOOR, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GRANVILLE COUNTY CREEDMOOR PD 380

Serial Number: *008641*
Test Date: *06/20/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG124903
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	12:51pm
AIR BLK	.00	12:52pm
ACCY CHK	.07	12:53pm
AIR BLK	.00	12:54pm
SUB TEST	**.00**	**12:54pm**
AIR BLK	.00	12:55pm
SUB TEST	**.00**	**12:57pm**
AIR BLK	.00	12:57pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location WAKE FOREST P.D.

Instrument Serial No. 008700 401 OWEN ST. WAKE FOREST, N.C

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WAKE COUNTY WAKE FOREST PD 910

Serial Number: *008700*
Test Date: *06/20/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG300202
Exp Date: 01/02/2015

Test	g/210L	Time
DIAG	Pass	2:18pm
AIR BLK	.00	2:18pm
ACCY CHK	.08	2:19pm
AIR BLK	.00	2:20pm
SUB TEST	**.00**	**2:20pm**
AIR BLK	.00	2:21pm
SUB TEST	**.00**	**2:23pm**
AIR BLK	.00	2:23pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY WAKE FOREST PD 910

Serial Number: *008700* Test Record Number: *569*
Test Date: *06/20/2013* Test Time: *2:24pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:24pm
FLO	Pass	2:24pm
FC	Pass	2:24pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:24pm
SRC	Pass	2:24pm
DET	Pass	2:24pm
BAR	Pass	2:24pm
BT	Pass	2:24pm

Blank Tests

Test	Status	Time
AIR	Pass	2:25pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:25pm

CRC Tests

Test	Status	Time
COMP	Pass	2:25pm
CAL	Pass	2:25pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ORANGE Instrument Location HILLSBOROUGH P.D.

Instrument Serial No. 008799 127 N. CHURCH ST. HILLSBOROUGH, N.C

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: *008799*
Test Date: *06/26/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206602
Exp Date: 03/06/2014

Test	g/210L	Time
DIAG	Pass	12:12pm
AIR BLK	.00	12:12pm
ACCY CHK	.07	12:13pm
AIR BLK	.00	12:14pm
SUB TEST	**.00**	**12:15pm**
AIR BLK	.00	12:15pm
SUB TEST	**.00**	**12:17pm**
AIR BLK	.00	12:18pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: *008799* Test Record Number: *1410*
Test Date: *06/26/2013* Test Time: *12:19pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:19pm
FLO	Pass	12:19pm
FC	Pass	12:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:19pm
SRC	Pass	12:19pm
DET	Pass	12:19pm
BAR	Pass	12:19pm
BT	Pass	12:19pm

Blank Tests

Test	Status	Time
AIR	Pass	12:20pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:20pm

CRC Tests

Test	Status	Time
COMP	Pass	12:20pm
CAL	Pass	12:20pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County DURHAM Instrument Location DURHAM CO. JAIL

Instrument Serial No. 008891 217 S. MANGUM ST. DURHAM, N.C

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: *008891*
Test Date: *06/26/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206102
Exp Date: 03/01/2014

Test	g/210L	Time
DIAG	Pass	11:16am
AIR BLK	.00	11:16am
ACCY CHK	.08	11:17am
AIR BLK	.00	11:18am
SUB TEST	**.00**	**11:18am**
AIR BLK	.00	11:19am
SUB TEST	**.00**	**11:21am**
AIR BLK	.00	11:22am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: *008891* Test Record Number: *2386*
Test Date: *06/26/2013* Test Time: *11:23am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:23am
FLO	Pass	11:23am
FC	Pass	11:23am

Temperature Tests

Test	Status	Time
FC1	Pass	11:23am
SRC	Pass	11:23am
DET	Pass	11:23am
BAR	Pass	11:23am
BT	Pass	11:23am

Blank Tests

Test	Status	Time
AIR	Pass	11:24am

Printer Tests

Test	Status	Time
PRNT	Pass	11:24am

CRC Tests

Test	Status	Time
COMP	Pass	11:24am
CAL	Pass	11:24am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ORANGE Instrument Location CHAPEL HILL P.D.

Instrument Serial No. 008839 828 MARTIN LUTHER KING JR. BLVD
CHAPEL HILL, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: *008839*
Test Date: *06/26/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	1:00pm
AIR BLK	.00	1:01pm
ACCY CHK	.08	1:02pm
AIR BLK	.00	1:03pm
SUB TEST	**.00**	**1:04pm**
AIR BLK	.00	1:05pm
SUB TEST	**.00**	**1:06pm**
AIR BLK	.00	1:07pm

Reported AC: .00 g/~~210L~~

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: *008839* Test Record Number: *1072*
Test Date: *06/26/2013* Test Time: *1:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:09pm
FLO	Pass	1:09pm
FC	Pass	1:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:09pm
SRC	Pass	1:09pm
DET	Pass	1:09pm
BAR	Pass	1:09pm
BT	Pass	1:09pm

Blank Tests

Test	Status	Time
AIR	Pass	1:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:10pm

CRC Tests

Test	Status	Time
COMP	Pass	1:10pm
CAL	Pass	1:10pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ORANGE Instrument Location CHAPEL HILL P.D

Instrument Serial No. 008856 822 MARTIN LUTHER KING JR. BLVD
CHAPEL HILL, N.C

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: *008856*
Test Date: *06/26/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206102
Exp Date: 03/01/2014

Test	g/210L	Time
DIAG	Pass	12:56pm
AIR BLK	.00	12:57pm
ACCY CHK	.08	12:58pm
AIR BLK	.00	12:59pm
SUB TEST	**.00**	**1:00pm**
AIR BLK	.00	1:00pm
SUB TEST	**.00**	**1:02pm**
AIR BLK	.00	1:03pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: *008856* Test Record Number: *1199*
Test Date: *06/26/2013* Test Time: *1:12pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:13pm
FLO	Pass	1:13pm
FC	Pass	1:13pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:13pm
SRC	Pass	1:13pm
DET	Pass	1:13pm
BAR	Pass	1:13pm
BT	Pass	1:13pm

Blank Tests

Test	Status	Time
AIR	Pass	1:14pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:14pm

CRC Tests

Test	Status	Time
COMP	Pass	1:14pm
CAL	Pass	1:14pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location NCSHP D2

Instrument Serial No. 008865 2239 Bishop Rd. Greensboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



R. Kevin L Dean — Signature of Certifying Official

642 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY NCSHP D2 400

Serial Number: *008865*
Test Date: *06/26/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG315701
Exp Date: 06/06/2015

Test	g/210L	Time
DIAG	Pass	2:04pm
AIR BLK	.00	2:04pm
ACCY CHK	.07	2:05pm
AIR BLK	.00	2:06pm
SUB TEST	**.00**	**2:07pm**
AIR BLK	.00	2:08pm
SUB TEST	**.00**	**2:09pm**
AIR BLK	.00	2:10pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY NCSHP D2 400

Serial Number: *008865* Test Record Number: *272*
Test Date: *06/26/2013* Test Time: *1:59pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:00pm
FLO	Pass	2:00pm
FC	Pass	2:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:00pm
SRC	Pass	2:00pm
DET	Pass	2:00pm
BAR	Pass	2:00pm
BT	Pass	2:00pm

Blank Tests

Test	Status	Time
AIR	Pass	2:01pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:01pm

CRC Tests

Test	Status	Time
COMP	Pass	2:01pm
CAL	Pass	2:01pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Co. Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008434

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

7682E
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 4 090

Serial Number: *008734*
Test Date: *06/30/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG920302
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	1:39pm
AIR BLK	.00	1:40pm
ACCY CHK	.07	1:41pm
AIR BLK	.00	1:42pm
SUB TEST	**.00**	**1:43pm**
AIR BLK	.00	1:44pm
SUB TEST	**.00**	**1:45pm**
AIR BLK	.00	1:46pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 4 090

Serial Number: *008734* Test Record Number: *694*
Test Date: *06/30/2013* Test Time: *1:47pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:47pm
FLO	Pass	1:47pm
FC	Pass	1:47pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:48pm
SRC	Pass	1:48pm
DET	Pass	1:48pm
BAR	Pass	1:48pm
BT	Pass	1:48pm

Blank Tests

Test	Status	Time
AIR	Pass	1:48pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:48pm

CRC Tests

Test	Status	Time
COMP	Pass	1:48pm
CAL	Pass	1:48pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Carteret Co. Instrument Location BAT Mobile Unit #4

Instrument Serial No. 008871

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

7682E
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CARTERET COUNTY BAT MOBILE UNIT 4 150

Serial Number: *008871*
Test Date: *06/28/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	3:50pm
AIR BLK	.00	3:51pm
ACCY CHK	.07	3:52pm
AIR BLK	.00	3:53pm
SUB TEST	**.00**	**3:54pm**
AIR BLK	.00	3:55pm
SUB TEST	**.00**	**3:57pm**
AIR BLK	.00	3:57pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY BAT MOBILE UNIT 4 150

Serial Number: *008871* Test Record Number: *674*
Test Date: *06/28/2013* Test Time: *3:58pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:59pm
FLO	Pass	3:59pm
FC	Pass	3:59pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:59pm
SRC	Pass	3:59pm
DET	Pass	3:59pm
BAR	Pass	3:59pm
BT	Pass	3:59pm

Blank Tests

Test	Status	Time
AIR	Pass	4:00pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:00pm

CRC Tests

Test	Status	Time
COMP	Pass	4:00pm
CAL	Pass	4:00pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Carteret Co Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008434

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

7682E
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CARTERET COUNTY BAT MOBILE UNIT 4 150

Serial Number: *008734*
Test Date: *06/28/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG920302
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	3:52pm
AIR BLK	.00	3:53pm
ACCY CHK	.07	3:54pm
AIR BLK	.00	3:55pm
SUB TEST	**.00**	**3:55pm**
AIR BLK	.00	3:56pm
SUB TEST	**.00**	**3:58pm**
AIR BLK	.00	3:58pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY BAT MOBILE UNIT 4 150

Serial Number: *008734* Test Record Number: *690*
Test Date: *06/28/2013* Test Time: *3:59pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:00pm
FLO	Pass	4:00pm
FC	Pass	4:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:00pm
SRC	Pass	4:00pm
DET	Pass	4:00pm
BAR	Pass	4:00pm
BT	Pass	4:00pm

Blank Tests

Test	Status	Time
AIR	Pass	4:01pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:01pm

CRC Tests

Test	Status	Time
COMP	Pass	4:01pm
CAL	Pass	4:01pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Rockingham Instrument Location Madison Police Department

Instrument Serial No. 008802

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROCKINGHAM COUNTY MADISON PD 780

Serial Number: *008802*
Test Date: *06/11/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206603
Exp Date: 03/06/2014

Test	g/210L	Time
DIAG	Pass	11:45am
AIR BLK	.00	11:46am
ACCY CHK	.08	11:47am
AIR BLK	.00	11:48am
SUB TEST	**.00**	**11:48am**
AIR BLK	.00	11:49am
SUB TEST	**.00**	**11:51am**
AIR BLK	.00	11:52am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROCKINGHAM COUNTY MADISON PD 780

Serial Number: *008802* Test Record Number: *522*
Test Date: *06/11/2013* Test Time: *11:53am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:53am
FLO	Pass	11:53am
FC	Pass	11:53am

Temperature Tests

Test	Status	Time
FC1	Pass	11:53am
SRC	Pass	11:53am
DET	Pass	11:53am
BAR	Pass	11:53am
BT	Pass	11:53am

Blank Tests

Test	Status	Time
AIR	Pass	11:54am

Printer Tests

Test	Status	Time
PRNT	Pass	11:54am

CRC Tests

Test	Status	Time
COMP	Pass	11:54am
CAL	Pass	11:54am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Alleghany Instrument Location Alleghany County Jail

Instrument Serial No. 008890 Sparta, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALLEGHANY COUNTY ALLEGHANY CO JAIL 020

Serial Number: *008890*
Test Date: *06/19/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG200905
Exp Date: 01/09/2014

Test	g/210L	Time
DIAG	Pass	1:20pm
AIR BLK	.00	1:20pm
ACCY CHK	.07	1:21pm
AIR BLK	.00	1:22pm
SUB TEST	**.00**	**1:22pm**
AIR BLK	.00	1:23pm
SUB TEST	**.00**	**1:25pm**
AIR BLK	.00	1:26pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALLEGHANY COUNTY ALLEGHANY CO JAIL 020

Serial Number: *008890* Test Record Number: *413*
Test Date: *06/19/2013* Test Time: *1:27pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:27pm
FLO	Pass	1:27pm
FC	Pass	1:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:27pm
SRC	Pass	1:27pm
DET	Pass	1:27pm
BAR	Pass	1:27pm
BT	Pass	1:27pm

Blank Tests

Test	Status	Time
AIR	Pass	1:28pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:28pm

CRC Tests

Test	Status	Time
COMP	Pass	1:28pm
CAL	Pass	1:28pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Ashe Instrument Location Ashe County Jail

Instrument Serial No. 008849 Jefferson, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ASHE COUNTY ASHE COUNTY JAIL 040

Serial Number: *008849*
Test Date: *06/19/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG200905
Exp Date: 01/09/2014

Test	g/210L	Time
DIAG	Pass	11:35am
AIR BLK	.00	11:36am
ACCY CHK	.07	11:36am
AIR BLK	.00	11:37am
SUB TEST	**.00**	**11:38am**
AIR BLK	.00	11:39am
SUB TEST	**.00**	**11:40am**
AIR BLK	.00	11:41am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ASHE COUNTY ASHE COUNTY JAIL 040

Serial Number: *008849* Test Record Number: *731*
Test Date: *06/19/2013* Test Time: *11:42am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:42am
FLO	Pass	11:42am
FC	Pass	11:42am

Temperature Tests

Test	Status	Time
FC1	Pass	11:43am
SRC	Pass	11:43am
DET	Pass	11:43am
BAR	Pass	11:43am
BT	Pass	11:43am

Blank Tests

Test	Status	Time
AIR	Pass	11:43am

Printer Tests

Test	Status	Time
PRNT	Pass	11:43am

CRC Tests

Test	Status	Time
COMP	Pass	11:43am
CAL	Pass	11:43am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Surry Instrument Location Surry County Jail

Instrument Serial No. 008934 Dobson, N. C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

SURRY COUNTY SURRY CO JAIL 850

Serial Number: *008934*
Test Date: *06/18/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG205401
Exp Date: 02/23/2014

Test	g/210L	Time
DIAG	Pass	11:46am
AIR BLK	.00	11:46am
ACCY CHK	.07	11:47am
AIR BLK	.00	11:48am
SUB TEST	**.00**	**11:48am**
AIR BLK	.00	11:49am
SUB TEST	**.00**	**11:51am**
AIR BLK	.00	11:52am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

SURRY COUNTY SURRY CO JAIL 850

Serial Number: *008934* Test Record Number: *1103*
Test Date: *06/18/2013* Test Time: *11:53am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:53am
FLO	Pass	11:53am
FC	Pass	11:53am

Temperature Tests

Test	Status	Time
FC1	Pass	11:54am
SRC	Pass	11:54am
DET	Pass	11:54am
BAR	Pass	11:54am
BT	Pass	11:54am

Blank Tests

Test	Status	Time
AIR	Pass	11:54am

Printer Tests

Test	Status	Time
PRNT	Pass	11:54am

CRC Tests

Test	Status	Time
COMP	Pass	11:54am
CAL	Pass	11:54am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Surry Instrument Location Mount Airy

Instrument Serial No. 008943 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

SURRY COUNTY MOUNT AIRY PD 850

Serial Number: *008943*
Test Date: *06/18/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG125603
Exp Date: 09/13/2013

Test	g/210L	Time
DIAG	Pass	10:39am
AIR BLK	.00	10:39am
ACCY CHK	.08	10:40am
AIR BLK	.00	10:41am
SUB TEST	**.00**	**10:41am**
AIR BLK	.00	10:42am
SUB TEST	**.00**	**10:44am**
AIR BLK	.00	10:45am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

SURRY COUNTY MOUNT AIRY PD 850

Serial Number: *008943* Test Record Number: *1461*
Test Date: *06/18/2013* Test Time: *10:46am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:46am
FLO	Pass	10:46am
FC	Pass	10:46am

Temperature Tests

Test	Status	Time
FC1	Pass	10:46am
SRC	Pass	10:46am
DET	Pass	10:46am
BAR	Pass	10:46am
BT	Pass	10:46am

Blank Tests

Test	Status	Time
AIR	Pass	10:47am

Printer Tests

Test	Status	Time
PRNT	Pass	10:47am

CRC Tests

Test	Status	Time
COMP	Pass	10:47am
CAL	Pass	10:47am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Surry Instrument Location Pilot Mountain Police Department

Instrument Serial No. 008938

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

SURRY COUNTY PILOT MOUNTAIN PD 850

Serial Number: *008938*
Test Date: *06/17/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	11:43am
AIR BLK	.00	11:43am
ACCY CHK	.07	11:44am
AIR BLK	.00	11:45am
SUB TEST	**.00**	**11:45am**
AIR BLK	.00	11:46am
SUB TEST	**.00**	**11:48am**
AIR BLK	.00	11:49am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

SURRY COUNTY PILOT MOUNTAIN PD 850

Serial Number: *008938* Test Record Number: *416*
Test Date: *06/17/2013* Test Time: *11:50am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:50am
FLO	Pass	11:50am
FC	Pass	11:50am

Temperature Tests

Test	Status	Time
FC1	Pass	11:51am
SRC	Pass	11:51am
DET	Pass	11:51am
BAR	Pass	11:51am
BT	Pass	11:51am

Blank Tests

Test	Status	Time
AIR	Pass	11:51am

Printer Tests

Test	Status	Time
PRNT	Pass	11:51am

CRC Tests

Test	Status	Time
COMP	Pass	11:51am
CAL	Pass	11:51am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Forsyth Instrument Location Forsyth County Detention

Instrument Serial No. 008854 Winston-Salem, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: *008854*
Test Date: *06/20/2013*

Citation Number: *M0000000-0*
Subject's Name: *PREVENTIVE, MAINTENANCE*
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective: *12/01/2012-12/01/2014*

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309105
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	11:19am
AIR BLK	.00	11:19am
ACCY CHK	.07	11:20am
AIR BLK	.00	11:21am
SUB TEST	**.00**	**11:21am**
AIR BLK	.00	11:22am
SUB TEST	**.00**	**11:23am**
AIR BLK	.00	11:24am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: *008854* Test Record Number: *92*
Test Date: *06/20/2013* Test Time: *11:14am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:15am
FLO	Pass	11:15am
FC	Pass	11:15am

Temperature Tests

Test	Status	Time
FC1	Pass	11:15am
SRC	Pass	11:15am
DET	Pass	11:15am
BAR	Pass	11:15am
BT	Pass	11:15am

Blank Tests

Test	Status	Time
AIR	Pass	11:16am

Printer Tests

Test	Status	Time
PRNT	Pass	11:16am

CRC Tests

Test	Status	Time
COMP	Pass	11:16am
CAL	Pass	11:16am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Forsyth Instrument Location Forsyth County Detention

Instrument Serial No. 008944 Winston-Salem, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: *008944*
Test Date: *06/20/2013*

Citation Number: *M0000000-0*
Subject's Name: *PREVENTIVE, MAINTENANCE*
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective: *12/01/2012-12/01/2014*

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	11:11am
AIR BLK	.00	11:12am
ACCY CHK	.08	11:12am
AIR BLK	.00	11:13am
SUB TEST	**.00**	**11:14am**
AIR BLK	.00	11:15am
SUB TEST	**.00**	**11:16am**
AIR BLK	.00	11:17am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: *008944* Test Record Number: *904*
Test Date: *06/20/2013* Test Time: *11:20am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:20am
FLO	Pass	11:20am
FC	Pass	11:21am

Temperature Tests

Test	Status	Time
FC1	Pass	11:21am
SRC	Pass	11:21am
DET	Pass	11:21am
BAR	Pass	11:21am
BT	Pass	11:21am

Blank Tests

Test	Status	Time
AIR	Pass	11:21am

Printer Tests

Test	Status	Time
PRNT	Pass	11:21am

CRC Tests

Test	Status	Time
COMP	Pass	11:22am
CAL	Pass	11:22am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location NC SHP D2

Instrument Serial No. 008660 2239 Bishop Rd
Greensboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY NCSHP D2 400

Serial Number: *008660*
Test Date: *06/26/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309105
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	2:52pm
AIR BLK	.00	2:53pm
ACCY CHK	.07	2:54pm
AIR BLK	.00	2:54pm
SUB TEST	**.00**	**2:55pm**
AIR BLK	.00	2:56pm
SUB TEST	**.00**	**2:57pm**
AIR BLK	.00	2:58pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY NCSHP D2 400

Serial Number: *008660* Test Record Number: *3234*
Test Date: *06/26/2013* Test Time: *3:01pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:01pm
FLO	Pass	3:01pm
FC	Pass	3:01pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:01pm
SRC	Pass	3:01pm
DET	Pass	3:01pm
BAR	Pass	3:01pm
BT	Pass	3:01pm

Blank Tests

Test	Status	Time
AIR	Pass	3:02pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:02pm

CRC Tests

Test	Status	Time
COMP	Pass	3:02pm
CAL	Pass	3:02pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location NC SHP D2

Instrument Serial No. 008718 2239 Bishop Rd.
Greensboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY NCSHP D2 400

Serial Number: *008718*
Test Date: *06/26/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG315701
Exp Date: 06/06/2015

Test	g/210L	Time
DIAG	Pass	2:06pm
AIR BLK	.00	2:07pm
ACCY CHK	.07	2:07pm
AIR BLK	.00	2:08pm
SUB TEST	**.00**	**2:09pm**
AIR BLK	.00	2:10pm
SUB TEST	**.00**	**2:11pm**
AIR BLK	.00	2:12pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY NCSHP D2 400

Serial Number: *008718* Test Record Number: *829*
Test Date: *06/26/2013* Test Time: *2:01pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:02pm
FLO	Pass	2:02pm
FC	Pass	2:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:02pm
SRC	Pass	2:02pm
DET	Pass	2:02pm
BAR	Pass	2:02pm
BT	Pass	2:02pm

Blank Tests

Test	Status	Time
AIR	Pass	2:02pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:03pm

CRC Tests

Test	Status	Time
COMP	Pass	2:03pm
CAL	Pass	2:03pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County FRANKLIN Instrument Location FRANKLIN Co. JAIL

Instrument Serial No. 008933 285 TKEMP RD LOUISBURG, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Bud D Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: *008933* Test Record Number: *553*
Test Date: *06/20/2013* Test Time: *2:28pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:29pm
FLO	Pass	2:29pm
FC	Pass	2:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:29pm
SRC	Pass	2:29pm
DET	Pass	2:29pm
BAR	Pass	2:29pm
BT	Pass	2:29pm

Blank Tests

Test	Status	Time
AIR	Pass	2:30pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:30pm

CRC Tests

Test	Status	Time
COMP	Pass	2:30pm
CAL	Pass	2:30pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: *008933*
Test Date: *06/20/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SMITH, BRIAN D*
Permit Number: *08937E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG205401
Exp Date: 02/23/2014

Test	g/210L	Time
DIAG	Pass	2:19pm
AIR BLK	.00	2:20pm
ACCY CHK	.08	2:20pm
AIR BLK	.00	2:21pm
SUB TEST	**.00**	**2:22pm**
AIR BLK	.00	2:23pm
SUB TEST	**.00**	**2:24pm**
AIR BLK	.00	2:25pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County FRANKLIN Instrument Location FRANKLIN CO. JAIL

Instrument Serial No. 008942 285 T KEMP RD LOUISBURG, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: *008942*
Test Date: *06/20/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SMITH, BRIAN D*
Permit Number: *08937E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG205402
Exp Date: 02/23/2014

Test	g/210L	Time
DIAG	Pass	2:04pm
AIR BLK	.00	2:05pm
ACCY CHK	.08	2:06pm
AIR BLK	.00	2:07pm
SUB TEST	**.00**	**2:07pm**
AIR BLK	.00	2:08pm
SUB TEST	**.00**	**2:10pm**
AIR BLK	.00	2:10pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: *008942* Test Record Number: *598*
Test Date: *06/20/2013* Test Time: *2:14pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:14pm
FLO	Pass	2:14pm
FC	Pass	2:15pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:15pm
SRC	Pass	2:15pm
DET	Pass	2:15pm
BAR	Pass	2:15pm
BT	Pass	2:15pm

Blank Tests

Test	Status	Time
AIR	Pass	2:15pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:15pm

CRC Tests

Test	Status	Time
COMP	Pass	2:16pm
CAL	Pass	2:16pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County FRANKLIN Instrument Location FRANKLINTON PD

Instrument Serial No. 008815 #7 W. MASON ST. FRANKLINTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of JUNE, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

FRANKLIN COUNTY FRANKLINTON PD 340

Serial Number: *008815*
Test Date: *06/20/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SMITH, BRIAN D*
Permit Number: *08937E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206603
Exp Date: 03/06/2014

Test	g/210L	Time
DIAG	Pass	10:31am
AIR BLK	.00	10:32am
ACCY CHK	.08	10:32am
AIR BLK	.00	10:33am
SUB TEST	**.00**	**10:33am**
AIR BLK	.00	10:34am
SUB TEST	**.00**	**10:37am**
AIR BLK	.00	10:38am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FRANKLIN COUNTY FRANKLINTON PD 340

Serial Number: *008815* Test Record Number: *756*
Test Date: *06/20/2013* Test Time: *10:40am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:41am
FLO	Pass	10:41am
FC	Pass	10:41am

Temperature Tests

Test	Status	Time
FC1	Pass	10:41am
SRC	Pass	10:41am
DET	Pass	10:41am
BAR	Pass	10:41am
BT	Pass	10:41am

Blank Tests

Test	Status	Time
AIR	Pass	10:42am

Printer Tests

Test	Status	Time
PRNT	Pass	10:42am

CRC Tests

Test	Status	Time
COMP	Pass	10:42am
CAL	Pass	10:42am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Johnston Instrument Location Benson Police Dept.

Instrument Serial No. 008885 Benson, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY BENSON POLICE DEPT.
500

Serial Number: *008885*
Test Date: *06/27/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG125602
Exp Date: 09/13/2013

Test	g/210L	Time
DIAG	Pass	11:23am
AIR BLK	.00	11:24am
ACCY CHK	.07	11:24am
AIR BLK	.00	11:25am
SUB TEST	**.00**	**11:25am**
AIR BLK	.00	11:26am
SUB TEST	**.00**	**11:28am**
AIR BLK	.00	11:29am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY BENSON POLICE DEPT. 500

Serial Number: *008885* Test Record Number: *315*
Test Date: *06/27/2013* Test Time: *11:30am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:30am
FLO	Pass	11:30am
FC	Pass	11:30am

Temperature Tests

Test	Status	Time
FC1	Pass	11:30am
SRC	Pass	11:30am
DET	Pass	11:30am
BAR	Pass	11:30am
BT	Pass	11:30am

Blank Tests

Test	Status	Time
AIR	Pass	11:31am

Printer Tests

Test	Status	Time
PRNT	Pass	11:31am

CRC Tests

Test	Status	Time
COMP	Pass	11:31am
CAL	Pass	11:31am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BATMOBILE UNIT #7

Instrument Serial No. 008623 KNIGHTDALE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of June, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910
Serial Number: *008623* Test Record Number: *2722*
Test Date: *06/21/2013* Test Time: *10:27pm EDT*
System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:27pm
FLO	Pass	10:27pm
FC	Pass	10:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:27pm
SRC	Pass	10:27pm
DET	Pass	10:27pm
BAR	Pass	10:27pm
BT	Pass	10:27pm

Blank Tests

Test	Status	Time
AIR	Pass	10:28pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:28pm

CRC Tests

Test	Status	Time
COMP	Pass	10:28pm
CAL	Pass	10:28pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910
Serial Number: *008623*
Test Date: *06/21/2013*
Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	10:18pm
AIR BLK	.00	10:19pm
ACCY CHK	.07	10:19pm
AIR BLK	.00	10:20pm
SUB TEST	**.00**	**10:22pm**
AIR BLK	.00	10:23pm
SUB TEST	**.00**	**10:24pm**
AIR BLK	.00	10:25pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BATMOBILE UNIT #7

Instrument Serial No. 008760 KNIGHTDALE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official Certificate Number 636

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 920

Serial Number: *008760* Test Record Number: *452*
Test Date: *06/21/2013* Test Time: *10:36pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:36pm
FLO	Pass	10:36pm
FC	Pass	10:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:36pm
SRC	Pass	10:36pm
DET	Pass	10:36pm
BAR	Pass	10:36pm
BT	Pass	10:36pm

Blank Tests

Test	Status	Time
AIR	Pass	10:37pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:37pm

CRC Tests

Test	Status	Time
COMP	Pass	10:37pm
CAL	Pass	10:37pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 920

Serial Number: *008760*
Test Date: *06/21/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	10:26pm
AIR BLK	.00	10:27pm
ACCY CHK	.07	10:27pm
AIR BLK	.00	10:28pm
SUB TEST	**.00**	**10:29pm**
AIR BLK	.00	10:30pm
SUB TEST	**.00**	**10:32pm**
AIR BLK	.00	10:33pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Rockingham Instrument Location Rockingham Co. Jail

Instrument Serial No. 008796 Wentworth, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROCKINGHAM COUNTY ROCKINGHAM CO JAIL
780

Serial Number: *008796*
Test Date: *06/11/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	11:33am
AIR BLK	.00	11:34am
ACCY CHK	.07	11:34am
AIR BLK	.00	11:35am
SUB TEST	**.00**	**11:36am**
AIR BLK	.00	11:37am
SUB TEST	**.00**	**11:38am**
AIR BLK	.00	11:39am

Reported AC: .00 g/210L

L. Kevin Dean
Signature of Chemical Analyst

Court CVR

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROCKINGHAM COUNTY ROCKINGHAM CO JAIL 780

Serial Number: *008796* Test Record Number: *1339*
Test Date: *06/11/2013* Test Time: *11:40am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:40am
FLO	Pass	11:40am
FC	Pass	11:40am

Temperature Tests

Test	Status	Time
FC1	Pass	11:41am
SRC	Pass	11:41am
DET	Pass	11:41am
BAR	Pass	11:41am
BT	Pass	11:41am

Blank Tests

Test	Status	Time
AIR	Pass	11:41am

Printer Tests

Test	Status	Time
PRNT	Pass	11:41am

CRC Tests

Test	Status	Time
COMP	Pass	11:41am
CAL	Pass	11:41am

Preventive Maintenance
Status: Pass

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Rockingham Instrument Location Eden Police

Instrument Serial No. 008636 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROCKINGHAM COUNTY EDEN PD 780

Serial Number: *008636*
Test Date: *06/11/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	12:33pm
AIR BLK	.00	12:33pm
ACCY CHK	.07	12:34pm
AIR BLK	.00	12:35pm
SUB TEST	**.00**	**12:36pm**
AIR BLK	.00	12:36pm
SUB TEST	**.00**	**12:38pm**
AIR BLK	.00	12:39pm

Reported AC: .00 g/210L

L. Kevin Dean
Signature of Chemical Analyst

Court CVR

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROCKINGHAM COUNTY EDEN PD 780

Serial Number: *008636* Test Record Number: *1301*
Test Date: *06/11/2013* Test Time: *12:39pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:40pm
FLO	Pass	12:40pm
FC	Pass	12:40pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:40pm
SRC	Pass	12:40pm
DET	Pass	12:40pm
BAR	Pass	12:40pm
BT	Pass	12:40pm

Blank Tests

Test	Status	Time
AIR	Pass	12:41pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:41pm

CRC Tests

Test	Status	Time
COMP	Pass	12:41pm
CAL	Pass	12:41pm

Preventive Maintenance
Status: Pass

L. Kevin Dean

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Rockingham Instrument Location Reidsville Police Department

Instrument Serial No. 008784

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

THE GREAT SEAL OF THE STATE OF NORTH CAROLINA MAY 20, 1775 APRIL 12, 1776 ESSE QUAM VIDERI

L. Brian Dean — Signature of Certifying Official

642 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROCKINGHAM COUNTY REIDSVILLE PD 780

Serial Number: *008784*
Test Date: *06/11/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG300202
Exp Date: 01/02/2015

Test	g/210L	Time
DIAG	Pass	1:26pm
AIR BLK	.00	1:27pm
ACCY CHK	.07	1:27pm
AIR BLK	.00	1:28pm
SUB TEST	**.00**	**1:29pm**
AIR BLK	.00	1:29pm
SUB TEST	**.00**	**1:31pm**
AIR BLK	.00	1:32pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROCKINGHAM COUNTY REIDSVILLE PD 780

Serial Number: *008784* Test Record Number: *692*
Test Date: *06/11/2013* Test Time: *1:33pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:33pm
FLO	Pass	1:33pm
FC	Pass	1:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:33pm
SRC	Pass	1:33pm
DET	Pass	1:33pm
BAR	Pass	1:33pm
BT	Pass	1:33pm

Blank Tests

Test	Status	Time
AIR	Pass	1:34pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:34pm

CRC Tests

Test	Status	Time
COMP	Pass	1:34pm
CAL	Pass	1:34pm

Preventive Maintenance
Status: Pass

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location Greensboro PD

Instrument Serial No. 008725 100 Police Plaza, Greensboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

L. Kevin Leon
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY GREENSBORO PD 400

Serial Number: *008725*
Test Date: *06/13/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG205401
Exp Date: 02/23/2014

Test	g/210L	Time
DIAG	Pass	11:54am
AIR BLK	.00	11:55am
ACCY CHK	.07	11:56am
AIR BLK	.00	11:57am
SUB TEST	**.00**	**11:57am**
AIR BLK	.00	11:58am
SUB TEST	**.00**	**11:59am**
AIR BLK	.00	12:00pm

Reported AC: .00 g/210L

L. Kevin Dean
Signature of Chemical Analyst

Court CVR

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY GREENSBORO PD 400

Serial Number: *008725* Test Record Number: *2894*
Test Date: *06/13/2013* Test Time: *12:01pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:01pm
FLO	Pass	12:01pm
FC	Pass	12:01pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:01pm
SRC	Pass	12:01pm
DET	Pass	12:01pm
BAR	Pass	12:01pm
BT	Pass	12:01pm

Blank Tests

Test	Status	Time
AIR	Pass	12:02pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:02pm

CRC Tests

Test	Status	Time
COMP	Pass	12:02pm
CAL	Pass	12:02pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location UNC - Greensboro

Instrument Serial No. 008604 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY UNC-G POLICE DEPT 400

Serial Number: *008604*
Test Date: *06/13/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG125603
Exp Date: 09/13/2013

Test	g/210L	Time
DIAG	Pass	12:35pm
AIR BLK	.00	12:36pm
ACCY CHK	.08	12:36pm
AIR BLK	.00	12:38pm
SUB TEST	**.00**	**12:38pm**
AIR BLK	.00	12:39pm
SUB TEST	**.00**	**12:41pm**
AIR BLK	.00	12:41pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY UNC-G POLICE DEPT 400

Serial Number: *008604* Test Record Number: *1185*
Test Date: *06/13/2013* Test Time: *12:42pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:42pm
FLO	Pass	12:42pm
FC	Pass	12:43pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:43pm
SRC	Pass	12:43pm
DET	Pass	12:43pm
BAR	Pass	12:43pm
BT	Pass	12:43pm

Blank Tests

Test	Status	Time
AIR	Pass	12:43pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:43pm

CRC Tests

Test	Status	Time
COMP	Pass	12:43pm
CAL	Pass	12:43pm

Preventive Maintenance
Status: Pass

K. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Randolph Instrument Location Asheboro

Instrument Serial No. 008791 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

S. Kevin Deon
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY ARCHDALE PD 750

Serial Number: *008791*
Test Date: *06/13/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	2:38pm
AIR BLK	.00	2:39pm
ACCY CHK	.08	2:39pm
AIR BLK	.00	2:40pm
SUB TEST	**.00**	**2:41pm**
AIR BLK	.00	2:42pm
SUB TEST	**.00**	**2:43pm**
AIR BLK	.00	2:44pm

Reported AC: .00 g/210L

L. Kevin Dean
Signature of Chemical Analyst

Court CVR

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY ARCHDALE PD 750

Serial Number: *008791* Test Record Number: *804*
Test Date: *06/13/2013* Test Time: *2:45pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:45pm
FLO	Pass	2:45pm
FC	Pass	2:45pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:45pm
SRC	Pass	2:45pm
DET	Pass	2:45pm
BAR	Pass	2:45pm
BT	Pass	2:45pm

Blank Tests

Test	Status	Time
AIR	Pass	2:46pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:46pm

CRC Tests

Test	Status	Time
COMP	Pass	2:46pm
CAL	Pass	2:46pm

Preventive Maintenance
Status: Pass

L. Kevin Dean

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location High Point Jail

Instrument Serial No. 008655

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY HIGH POINT JAIL 401

Serial Number: *008655*
Test Date: *06/17/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG205401
Exp Date: 02/23/2014

Test	g/210L	Time
DIAG	Pass	2:07pm
AIR BLK	.00	2:08pm
ACCY CHK	.07	2:09pm
AIR BLK	.00	2:10pm
SUB TEST	**.00**	**2:11pm**
AIR BLK	.00	2:12pm
SUB TEST	**.00**	**2:13pm**
AIR BLK	.00	2:14pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY HIGH POINT JAIL 401

Serial Number: *008655* Test Record Number: *2002*
Test Date: *06/17/2013* Test Time: *2:14pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:15pm
FLO	Pass	2:15pm
FC	Pass	2:15pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:15pm
SRC	Pass	2:15pm
DET	Pass	2:15pm
BAR	Pass	2:15pm
BT	Pass	2:15pm

Blank Tests

Test	Status	Time
AIR	Pass	2:16pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:16pm

CRC Tests

Test	Status	Time
COMP	Pass	2:16pm
CAL	Pass	2:16pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location High Point Jail

Instrument Serial No. 008828

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY HIGH POINT JAIL 401

Serial Number: *008828*
Test Date: *06/17/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG205402
Exp Date: 02/23/2014

Test	g/210L	Time
DIAG	Pass	2:10pm
AIR BLK	.00	2:10pm
ACCY CHK	.08	2:11pm
AIR BLK	.00	2:12pm
SUB TEST	**.00**	**2:13pm**
AIR BLK	.00	2:14pm
SUB TEST	**.00**	**2:15pm**
AIR BLK	.00	2:16pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY HIGH POINT JAIL 401

Serial Number: *008828* Test Record Number: *1272*
Test Date: *06/17/2013* Test Time: *2:17pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:17pm
FLO	Pass	2:17pm
FC	Pass	2:17pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:17pm
SRC	Pass	2:17pm
DET	Pass	2:17pm
BAR	Pass	2:17pm
BT	Pass	2:17pm

Blank Tests

Test	Status	Time
AIR	Pass	2:18pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:18pm

CRC Tests

Test	Status	Time
COMP	Pass	2:18pm
CAL	Pass	2:18pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CRAVEN Instrument Location NEWBERN P.D.

Instrument Serial No. 008817

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CRAVEN COUNTY NEW BERN PD 240

Serial Number: *008817*
Test Date: *06/19/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203103
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	9:21am
AIR BLK	.00	9:22am
ACCY CHK	.07	9:22am
AIR BLK	.00	9:23am
SUB TEST	**.00**	**9:24am**
AIR BLK	.00	9:25am
SUB TEST	**.00**	**9:26am**
AIR BLK	.00	9:27am

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY NEW BERN PD 240

Serial Number: *008817* Test Record Number: *957*
Test Date: *06/19/2013* Test Time: *9:28am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:28am
FLO	Pass	9:28am
FC	Pass	9:28am

Temperature Tests

Test	Status	Time
FC1	Pass	9:28am
SRC	Pass	9:28am
DET	Pass	9:28am
BAR	Pass	9:28am
BT	Pass	9:28am

Blank Tests

Test	Status	Time
AIR	Pass	9:29am

Printer Tests

Test	Status	Time
PRNT	Pass	9:29am

CRC Tests

Test	Status	Time
COMP	Pass	9:29am
CAL	Pass	9:29am

Preventive Maintenance
Status: Pass

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Carteret Instrument Location Morehead City P.D.

Instrument Serial No. 008731

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Randy E Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CARTERET COUNTY MOREHEAD CITY PD 150

Serial Number: *008731*
Test Date: *06/18/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG124904
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	1:11pm
AIR BLK	.00	1:12pm
ACCY CHK	.08	1:12pm
AIR BLK	.00	1:13pm
SUB TEST	**.00**	**1:14pm**
AIR BLK	.00	1:15pm
SUB TEST	**.00**	**1:16pm**
AIR BLK	.00	1:17pm

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY MOREHEAD CITY PD 150

Serial Number: *008731* Test Record Number: *1366*
Test Date: *06/18/2013* Test Time: *1:18pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:18pm
FLO	Pass	1:18pm
FC	Pass	1:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:18pm
SRC	Pass	1:18pm
DET	Pass	1:18pm
BAR	Pass	1:18pm
BT	Pass	1:18pm

Blank Tests

Test	Status	Time
AIR	Pass	1:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:19pm

CRC Tests

Test	Status	Time
COMP	Pass	1:19pm
CAL	Pass	1:19pm

Preventive Maintenance
Status: Pass

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Carteret Instrument Location Emerald Isle P.D.

Instrument Serial No. 008620

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Ray E Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CARTERET COUNTY EMERALD ISLE PD 150

Serial Number: *008620*
Test Date: *06/18/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203103
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	11:46am
AIR BLK	.00	11:47am
ACCY CHK	.08	11:47am
AIR BLK	.00	11:49am
SUB TEST	**.00**	**11:49am**
AIR BLK	.00	11:50am
SUB TEST	**.00**	**11:52am**
AIR BLK	.00	11:53am

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY EMERALD ISLE PD 150

Serial Number: *008620* Test Record Number: *1506*
Test Date: *06/18/2013* Test Time: *11:53am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:54am
FLO	Pass	11:54am
FC	Pass	11:54am

Temperature Tests

Test	Status	Time
FC1	Pass	11:54am
SRC	Pass	11:54am
DET	Pass	11:54am
BAR	Pass	11:54am
BT	Pass	11:54am

Blank Tests

Test	Status	Time
AIR	Pass	11:54am

Printer Tests

Test	Status	Time
PRNT	Pass	11:54am

CRC Tests

Test	Status	Time
COMP	Pass	11:55am
CAL	Pass	11:55am

Preventive Maintenance
Status: Pass

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Onslow Instrument Location MCAS New River PMO

Instrument Serial No. 008922

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of June, 20 23 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ONSLOW COUNTY MCAS NEW RIVER 660

Serial Number: *008922*
Test Date: *06/11/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG124904
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	3:04pm
AIR BLK	.00	3:05pm
ACCY CHK	.08	3:05pm
AIR BLK	.00	3:06pm
SUB TEST	**.00**	**3:07pm**
AIR BLK	.00	3:07pm
SUB TEST	**.00**	**3:09pm**
AIR BLK	.00	3:10pm

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY MCAS NEW RIVER 660

Serial Number: *008922* Test Record Number: *239*
Test Date: *06/11/2013* Test Time: *3:10pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:11pm
FLO	Pass	3:11pm
FC	Pass	3:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:11pm
SRC	Pass	3:11pm
DET	Pass	3:11pm
BAR	Pass	3:11pm
BT	Pass	3:11pm

Blank Tests

Test	Status	Time
AIR	Pass	3:11pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:11pm

CRC Tests

Test	Status	Time
COMP	Pass	3:12pm
CAL	Pass	3:12pm

Preventive Maintenance
Status: Pass

Randy E Hall
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Onslow Instrument Location Onslow County

Instrument Serial No. 008932 Sheriff's Office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Randy E Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ONSLOW COUNTY ONSLOW COUNTY SD 660

Serial Number: *008932*
Test Date: *06/11/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG202602
Exp Date: 01/26/2014

Test	g/210L	Time
DIAG	Pass	2:03pm
AIR BLK	.00	2:04pm
ACCY CHK	.07	2:04pm
AIR BLK	.00	2:05pm
SUB TEST	**.00**	**2:06pm**
AIR BLK	.00	2:07pm
SUB TEST	**.00**	**2:09pm**
AIR BLK	.00	2:09pm

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY ONSLOW COUNTY SD 660

Serial Number: *008932* Test Record Number: *1888*
Test Date: *06/11/2013* Test Time: *2:17pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:18pm
FLO	Pass	2:18pm
FC	Pass	2:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:18pm
SRC	Pass	2:18pm
DET	Pass	2:18pm
BAR	Pass	2:18pm
BT	Pass	2:18pm

Blank Tests

Test	Status	Time
AIR	Pass	2:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:19pm

CRC Tests

Test	Status	Time
COMP	Pass	2:19pm
CAL	Pass	2:19pm

Preventive Maintenance
Status: Pass

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Onslow Instrument Location Onslow County

Instrument Serial No. 008931 Sheriff's Office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Randy E Hall — Signature of Certifying Official

354 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ONSLOW COUNTY ONSLOW COUNTY SD 660

Serial Number: *008931*
Test Date: *06/11/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206102
Exp Date: 03/01/2014

Test	g/210L	Time
DIAG	Pass	2:02pm
AIR BLK	.00	2:03pm
ACCY CHK	.08	2:03pm
AIR BLK	.00	2:04pm
SUB TEST	**.00**	**2:05pm**
AIR BLK	.00	2:06pm
SUB TEST	**.00**	**2:07pm**
AIR BLK	.00	2:08pm

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY ONSLOW COUNTY SD 660

Serial Number: *008931* Test Record Number: *1847*
Test Date: *06/11/2013* Test Time: *2:09pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:09pm
FLO	Pass	2:09pm
FC	Pass	2:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:09pm
SRC	Pass	2:09pm
DET	Pass	2:09pm
BAR	Pass	2:09pm
BT	Pass	2:09pm

Blank Tests

Test	Status	Time
AIR	Pass	2:10pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:10pm

CRC Tests

Test	Status	Time
COMP	Pass	2:10pm
CAL	Pass	2:10pm

Preventive Maintenance
Status: Pass

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ONSLOW Instrument Location JACKSONVILLE P.D.

Instrument Serial No. 008930

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ONSLOW COUNTY JACKSONVILLE PD 660

Serial Number: *008930*
Test Date: *06/11/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206102
Exp Date: 03/01/2014

Test	g/210L	Time
DIAG	Pass	1:33pm
AIR BLK	.00	1:34pm
ACCY CHK	.08	1:34pm
AIR BLK	.00	1:35pm
SUB TEST	**.00**	**1:36pm**
AIR BLK	.00	1:37pm
SUB TEST	**.00**	**1:38pm**
AIR BLK	.00	1:39pm

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY JACKSONVILLE PD 660

Serial Number: *008930* Test Record Number: *1968*
Test Date: *06/11/2013* Test Time: *1:40pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:40pm
FLO	Pass	1:40pm
FC	Pass	1:40pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:40pm
SRC	Pass	1:40pm
DET	Pass	1:40pm
BAR	Pass	1:40pm
BT	Pass	1:40pm

Blank Tests

Test	Status	Time
AIR	Pass	1:41pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:41pm

CRC Tests

Test	Status	Time
COMP	Pass	1:41pm
CAL	Pass	1:41pm

Preventive Maintenance
Status: Pass

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Onslow Instrument Location Camp Lejeune PMO

Instrument Serial No. 008920

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Randy E Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ONSLOW COUNTY CAMP LEJEUNE PMO 660

Serial Number: *008920*
Test Date: *06/11/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG202602
Exp Date: 01/26/2014

Test	g/210L	Time
DIAG	Pass	12:54pm
AIR BLK	.00	12:55pm
ACCY CHK	.08	12:55pm
AIR BLK	.00	12:56pm
SUB TEST	**.00**	**12:57pm**
AIR BLK	.00	12:57pm
SUB TEST	**.00**	**12:59pm**
AIR BLK	.00	1:00pm

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY CAMP LEJEUNE PMO 660

Serial Number: *008920* Test Record Number: *777*
Test Date: *06/11/2013* Test Time: *1:01pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:01pm
FLO	Pass	1:01pm
FC	Pass	1:01pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:01pm
SRC	Pass	1:01pm
DET	Pass	1:01pm
BAR	Pass	1:01pm
BT	Pass	1:01pm

Blank Tests

Test	Status	Time
AIR	Pass	1:02pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:02pm

CRC Tests

Test	Status	Time
COMP	Pass	1:02pm
CAL	Pass	1:02pm

Preventive Maintenance
Status: Pass

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Carteret Instrument Location Atlantic Beach P.D.

Instrument Serial No. 008785

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Randy E Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CARTERET COUNTY ATLANTIC BEACH PD 150

Serial Number: *008785*
Test Date: *06/18/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG202602
Exp Date: 01/26/2014

Test	g/210L	Time
DIAG	Pass	12:33pm
AIR BLK	.00	12:34pm
ACCY CHK	.07	12:35pm
AIR BLK	.00	12:36pm
SUB TEST	**.00**	**12:36pm**
AIR BLK	.00	12:37pm
SUB TEST	**.00**	**12:38pm**
AIR BLK	.00	12:39pm

Reported AC: .00 g/210L

RE Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY ATLANTIC BEACH PD 150

Serial Number: *008785* Test Record Number: *660*
Test Date: *06/18/2013* Test Time: *12:48pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:49pm
FLO	Pass	12:49pm
FC	Pass	12:49pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:49pm
SRC	Pass	12:49pm
DET	Pass	12:49pm
BAR	Pass	12:49pm
BT	Pass	12:49pm

Blank Tests

Test	Status	Time
AIR	Pass	12:49pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:49pm

CRC Tests

Test	Status	Time
COMP	Pass	12:50pm
CAL	Pass	12:50pm

Preventive Maintenance
Status: Pass

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CRAVEN Instrument Location MCAS CHERRY POINT PMO

Instrument Serial No. 010819

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CRAVEN COUNTY MCAS CHERRY POINT 240

Serial Number: *010819*
Test Date: *06/18/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG202602
Exp Date: 01/26/2014

Test	g/210L	Time
DIAG	Pass	2:04pm
AIR BLK	.00	2:04pm
ACCY CHK	.08	2:05pm
AIR BLK	.00	2:06pm
SUB TEST	**.00**	**2:06pm**
AIR BLK	.00	2:07pm
SUB TEST	**.00**	**2:09pm**
AIR BLK	.00	2:10pm

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY MCAS CHERRY POINT 240

Serial Number: *010819* Test Record Number: *313*
Test Date: *06/18/2013* Test Time: *2:11pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:12pm
FLO	Pass	2:12pm
FC	Pass	2:12pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:12pm
SRC	Pass	2:12pm
DET	Pass	2:12pm
BAR	Pass	2:12pm
BT	Pass	2:12pm

Blank Tests

Test	Status	Time
AIR	Pass	2:13pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:13pm

CRC Tests

Test	Status	Time
COMP	Pass	2:13pm
CAL	Pass	2:13pm

Preventive Maintenance
Status: Pass

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CRAVEN Instrument Location CRAVEN County

Instrument Serial No. 008732 SHERIFFS OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CRAVEN COUNTY CRAVEN COUNTY SD 240

Serial Number: *008732*
Test Date: *06/19/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG202602
Exp Date: 01/26/2014

Test	g/210L	Time
DIAG	Pass	9:55am
AIR BLK	.00	9:56am
ACCY CHK	.07	9:56am
AIR BLK	.00	9:57am
SUB TEST	**.00**	**9:58am**
AIR BLK	.00	9:58am
SUB TEST	**.00**	**10:00am**
AIR BLK	.00	10:01am

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY CRAVEN COUNTY SD 240

Serial Number: *008732* Test Record Number: *942*
Test Date: *06/19/2013* Test Time: *10:01am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:02am
FLO	Pass	10:02am
FC	Pass	10:02am

Temperature Tests

Test	Status	Time
FC1	Pass	10:02am
SRC	Pass	10:02am
DET	Pass	10:02am
BAR	Pass	10:02am
BT	Pass	10:02am

Blank Tests

Test	Status	Time
AIR	Pass	10:03am

Printer Tests

Test	Status	Time
PRNT	Pass	10:03am

CRC Tests

Test	Status	Time
COMP	Pass	10:03am
CAL	Pass	10:03am

Preventive Maintenance
Status: Pass

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Jones Instrument Location Jones County

Instrument Serial No. 008705 Sheriffs Office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

JONES COUNTY JONES COUNTY SD 510

Serial Number: *008705*
Test Date: *06/19/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203103
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	11:07am
AIR BLK	.00	11:08am
ACCY CHK	.08	11:08am
AIR BLK	.00	11:09am
SUB TEST	**.00**	**11:10am**
AIR BLK	.00	11:11am
SUB TEST	**.00**	**11:12am**
AIR BLK	.00	11:13am

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

JONES COUNTY JONES COUNTY SD 510

Serial Number: *008705* Test Record Number: *829*
Test Date: *06/19/2013* Test Time: *11:13am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:14am
FLO	Pass	11:14am
FC	Pass	11:14am

Temperature Tests

Test	Status	Time
FC1	Pass	11:14am
SRC	Pass	11:14am
DET	Pass	11:14am
BAR	Pass	11:14am
BT	Pass	11:14am

Blank Tests

Test	Status	Time
AIR	Pass	11:14am

Printer Tests

Test	Status	Time
PRNT	Pass	11:14am

CRC Tests

Test	Status	Time
COMP	Pass	11:15am
CAL	Pass	11:15am

Preventive Maintenance
Status: Pass

Randy E Hall
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pamlico Instrument Location Pamlico County

Instrument Serial No. 008640 Sheriff's Office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Ronny E Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PAMLICO COUNTY PAMLICO COUNTY SD 680

Serial Number: *008640*
Test Date: *06/19/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HALL, RANDY E*
Permit Number: *03462E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206102
Exp Date: 03/01/2014

Test	g/210L	Time
DIAG	Pass	12:31pm
AIR BLK	.00	12:31pm
ACCY CHK	.08	12:32pm
AIR BLK	.00	12:33pm
SUB TEST	**.00**	**12:33pm**
AIR BLK	.00	12:34pm
SUB TEST	**.00**	**12:36pm**
AIR BLK	.00	12:37pm

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PAMLICO COUNTY PAMLICO COUNTY SD 680

Serial Number: *008640* Test Record Number: *956*
Test Date: *06/19/2013* Test Time: *12:37pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:38pm
FLO	Pass	12:38pm
FC	Pass	12:38pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:38pm
SRC	Pass	12:38pm
DET	Pass	12:38pm
BAR	Pass	12:38pm
BT	Pass	12:38pm

Blank Tests

Test	Status	Time
AIR	Pass	12:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:38pm

CRC Tests

Test	Status	Time
COMP	Pass	12:39pm
CAL	Pass	12:39pm

Preventive Maintenance
Status: Pass

Randy E Hall
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pender Instrument Location Pender County

Instrument Serial No. 008948 Sheriffs Dept Annex

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PENDER COUNTY SHERIFF DEPT ANNEX 700

Serial Number: *008948*
Test Date: *06/05/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204602
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	3:30pm
AIR BLK	.00	3:31pm
ACCY CHK	.07	3:31pm
AIR BLK	.00	3:32pm
SUB TEST	**.00**	**3:33pm**
AIR BLK	.00	3:34pm
SUB TEST	**.00**	**3:35pm**
AIR BLK	.00	3:36pm

Reported AC: .00 g/210L

K.C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PENDER COUNTY SHERIFF DEPT ANNEX 700

Serial Number: *008948* Test Record Number: *605*
Test Date: *06/05/2013* Test Time: *3:36pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:37pm
FLO	Pass	3:37pm
FC	Pass	3:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:37pm
SRC	Pass	3:37pm
DET	Pass	3:37pm
BAR	Pass	3:37pm
BT	Pass	3:37pm

Blank Tests

Test	Status	Time
AIR	Pass	3:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:38pm

CRC Tests

Test	Status	Time
COMP	Pass	3:38pm
CAL	Pass	3:38pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pender Instrument Location Pender County Sheriffs office

Instrument Serial No. 008935

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible]
Signature of Certifying Official

609
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PENDER PENDER CO SD 700

Serial Number: *008935*
Test Date: *06/05/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204602
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	1:45pm
AIR BLK	.00	1:45pm
ACCY CHK	.07	1:46pm
AIR BLK	.00	1:47pm
SUB TEST	**.00**	**1:47pm**
AIR BLK	.00	1:48pm
SUB TEST	**.00**	**1:50pm**
AIR BLK	.00	1:51pm

Reported AC: .00 g/210L

K. C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PENDER PENDER CO SD 700

Serial Number: *008935* Test Record Number: *1391*
Test Date: *06/05/2013* Test Time: *1:51pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:52pm
FLO	Pass	1:52pm
FC	Pass	1:52pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:52pm
SRC	Pass	1:52pm
DET	Pass	1:52pm
BAR	Pass	1:52pm
BT	Pass	1:52pm

Blank Tests

Test	Status	Time
AIR	Pass	1:53pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:53pm

CRC Tests

Test	Status	Time
COMP	Pass	1:53pm
CAL	Pass	1:53pm

Preventive Maintenance
Status: Pass

K. C. Rhodes

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pender Instrument Location Pender County Sheriffs office

Instrument Serial No. 008946

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. Rhodes
Signature of Certifying Official

001
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PENDER PENDER COUNTY SD 700

Serial Number: *008946*
Test Date: *06/05/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204602
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	1:44pm
AIR BLK	.00	1:44pm
ACCY CHK	.08	1:45pm
AIR BLK	.00	1:46pm
SUB TEST	**.00**	**1:46pm**
AIR BLK	.00	1:47pm
SUB TEST	**.00**	**1:49pm**
AIR BLK	.00	1:50pm

Reported AC: .00 g/210L

K. C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PENDER PENDER COUNTY SD 700

Serial Number: *008946* Test Record Number: *675*
Test Date: *06/05/2013* Test Time: *1:51pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:51pm
FLO	Pass	1:51pm
FC	Pass	1:51pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:51pm
SRC	Pass	1:51pm
DET	Pass	1:51pm
BAR	Pass	1:51pm
BT	Pass	1:51pm

Blank Tests

Test	Status	Time
AIR	Pass	1:52pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:52pm

CRC Tests

Test	Status	Time
COMP	Pass	1:52pm
CAL	Pass	1:52pm

Preventive Maintenance
Status: Pass

K. C. Walker

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Duplin Instrument Location Wallace

Instrument Serial No. 008858 Police Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. Rhodes
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DUPLIN COUNTY WALLACE PD 300

Serial Number: *008858*
Test Date: *06/05/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	12:55pm
AIR BLK	.00	12:56pm
ACCY CHK	.07	12:56pm
AIR BLK	.00	12:57pm
SUB TEST	**.00**	**12:58pm**
AIR BLK	.00	12:59pm
SUB TEST	**.00**	**1:00pm**
AIR BLK	.00	1:01pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DUPLIN COUNTY WALLACE PD 300

Serial Number: *008858* Test Record Number: *577*
Test Date: *06/05/2013* Test Time: *1:02pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:02pm
FLO	Pass	1:02pm
FC	Pass	1:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:02pm
SRC	Pass	1:02pm
DET	Pass	1:02pm
BAR	Pass	1:02pm
BT	Pass	1:02pm

Blank Tests

Test	Status	Time
AIR	Pass	1:03pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:03pm

CRC Tests

Test	Status	Time
COMP	Pass	1:03pm
CAL	Pass	1:03pm

Preventive Maintenance
Status: Pass

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Duplin Instrument Location Warsaw

Instrument Serial No. 008874 Police Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] Signature of Certifying Official

601 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DUPLIN COUNTY WARSAW PD 300

Serial Number: *008874*
Test Date: *06/05/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	11:54am
AIR BLK	.00	11:55am
ACCY CHK	.07	11:55am
AIR BLK	.00	11:56am
SUB TEST	**.00**	**11:57am**
AIR BLK	.00	11:58am
SUB TEST	**.00**	**11:59am**
AIR BLK	.00	12:00pm

Reported AC: .00 g/210L

K. C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DUPLIN COUNTY WARSAW PD 300

Serial Number: *008874* Test Record Number: *269*
Test Date: *06/05/2013* Test Time: *12:00pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:01pm
FLO	Pass	12:01pm
FC	Pass	12:01pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:01pm
SRC	Pass	12:01pm
DET	Pass	12:01pm
BAR	Pass	12:01pm
BT	Pass	12:01pm

Blank Tests

Test	Status	Time
AIR	Pass	12:02pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:02pm

CRC Tests

Test	Status	Time
COMP	Pass	12:02pm
CAL	Pass	12:02pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Duplin Instrument Location Duplin County

Instrument Serial No. 008864 Sheriffs office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] — Signature of Certifying Official

601 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DUPLIN COUNTY DUPLIN CO SD 300

Serial Number: *008864*
Test Date: *06/05/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG303502
Exp Date: 02/04/2015

Test	g/210L	Time
DIAG	Pass	11:09am
AIR BLK	.00	11:10am
ACCY CHK	.08	11:10am
AIR BLK	.00	11:11am
SUB TEST	**.00**	**11:12am**
AIR BLK	.00	11:13am
SUB TEST	**.00**	**11:14am**
AIR BLK	.00	11:15am

Reported AC: .00 g/210L

K. C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DUPLIN COUNTY DUPLIN CO SD 300

Serial Number: *008864* Test Record Number: *1937*
Test Date: *06/05/2013* Test Time: *11:17am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:17am
FLO	Pass	11:17am
FC	Pass	11:18am

Temperature Tests

Test	Status	Time
FC1	Pass	11:18am
SRC	Pass	11:18am
DET	Pass	11:18am
BAR	Pass	11:18am
BT	Pass	11:18am

Blank Tests

Test	Status	Time
AIR	Pass	11:18am

Printer Tests

Test	Status	Time
PRNT	Pass	11:18am

CRC Tests

Test	Status	Time
COMP	Pass	11:18am
CAL	Pass	11:18am

Preventive Maintenance
Status: Pass

K. C. [illegible]
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Bladen Instrument Location Bladen County Sheriffs Office

Instrument Serial No. 008818

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. [illegible] Signature of Certifying Official

601 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: *008818*
Test Date: *06/04/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206102
Exp Date: 03/01/2014

Test	g/210L	Time
DIAG	Pass	1:02pm
AIR BLK	.00	1:03pm
ACCY CHK	.08	1:04pm
AIR BLK	.00	1:05pm
SUB TEST	**.00**	**1:05pm**
AIR BLK	.00	1:06pm
SUB TEST	**.00**	**1:07pm**
AIR BLK	.00	1:09pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: *008818* Test Record Number: *669*
Test Date: *06/04/2013* Test Time: *1:09pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:09pm
FLO	Pass	1:09pm
FC	Pass	1:10pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:10pm
SRC	Pass	1:10pm
DET	Pass	1:10pm
BAR	Pass	1:10pm
BT	Pass	1:10pm

Blank Tests

Test	Status	Time
AIR	Pass	1:10pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:10pm

CRC Tests

Test	Status	Time
COMP	Pass	1:10pm
CAL	Pass	1:10pm

Preventive Maintenance
Status: Pass

K.C. Bladen
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Bladen Instrument Location Bladen County

Instrument Serial No. 008894 Sheriff's office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible]
Signature of Certifying Official

1001
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: *008894*
Test Date: *06/04/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206602
Exp Date: 03/06/2014

Test	g/210L	Time
DIAG	Pass	12:55pm
AIR BLK	.00	12:56pm
ACCY CHK	.08	12:57pm
AIR BLK	.00	12:57pm
SUB TEST	**.00**	**12:58pm**
AIR BLK	.00	12:59pm
SUB TEST	**.00**	**1:00pm**
AIR BLK	.00	1:01pm

Reported AC: .00 g/210L

K. C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: *008894* Test Record Number: *573*
Test Date: *06/04/2013* Test Time: *1:02pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:02pm
FLO	Pass	1:02pm
FC	Pass	1:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:03pm
SRC	Pass	1:03pm
DET	Pass	1:03pm
BAR	Pass	1:03pm
BT	Pass	1:03pm

Blank Tests

Test	Status	Time
AIR	Pass	1:03pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:03pm

CRC Tests

Test	Status	Time
COMP	Pass	1:03pm
CAL	Pass	1:03pm

Preventive Maintenance
Status: Pass

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Sampson Instrument Location Sampson County

Instrument Serial No. 008877 Sheriffs office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

SAMPSON COUNTY SAMPSON COUNTY SD 810

Serial Number: *008877*
Test Date: *06/04/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG308702
Exp Date: 03/28/2015

Test	g/210L	Time
DIAG	Pass	11:31am
AIR BLK	.00	11:32am
ACCY CHK	.07	11:32am
AIR BLK	.00	11:33am
SUB TEST	**.00**	**11:34am**
AIR BLK	.00	11:35am
SUB TEST	**.00**	**11:36am**
AIR BLK	.00	11:37am

Reported AC: .00 g/210L

K. C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

SAMPSON COUNTY SAMPSON COUNTY SD 810

Serial Number: *008877* Test Record Number: *1384*
Test Date: *06/04/2013* Test Time: *11:37am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:38am
FLO	Pass	11:38am
FC	Pass	11:38am

Temperature Tests

Test	Status	Time
FC1	Pass	11:38am
SRC	Pass	11:38am
DET	Pass	11:38am
BAR	Pass	11:38am
BT	Pass	11:38am

Blank Tests

Test	Status	Time
AIR	Pass	11:39am

Printer Tests

Test	Status	Time
PRNT	Pass	11:39am

CRC Tests

Test	Status	Time
COMP	Pass	11:39am
CAL	Pass	11:39am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Sampson Instrument Location Sampson County

Instrument Serial No. 008825 Sheriffs office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] — Signature of Certifying Official

601 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

SAMPSON COUNTY SAMPSON CO SD 810

Serial Number: *008825*
Test Date: *06/04/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG202903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	11:32am
AIR BLK	.00	11:32am
ACCY CHK	.07	11:33am
AIR BLK	.00	11:34am
SUB TEST	**.00**	**11:34am**
AIR BLK	.00	11:35am
SUB TEST	**.00**	**11:37am**
AIR BLK	.00	11:38am

Reported AC: .00 g/210L

K. C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

SAMPSON COUNTY SAMPSON CO SD 810

Serial Number: *008825* Test Record Number: *1686*
Test Date: *06/04/2013* Test Time: *11:39am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:39am
FLO	Pass	11:39am
FC	Pass	11:39am

Temperature Tests

Test	Status	Time
FC1	Pass	11:39am
SRC	Pass	11:39am
DET	Pass	11:39am
BAR	Pass	11:39am
BT	Pass	11:39am

Blank Tests

Test	Status	Time
AIR	Pass	11:40am

Printer Tests

Test	Status	Time
PRNT	Pass	11:40am

CRC Tests

Test	Status	Time
COMP	Pass	11:40am
CAL	Pass	11:40am

Preventive Maintenance
Status: Pass

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location Wrightsville Beach

Instrument Serial No. 008667 Police Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. [signature] 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY WRIGHTSVILLE BCH PD
640

Serial Number: *008667*
Test Date: *06/03/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203903
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	1:52pm
AIR BLK	.00	1:52pm
ACCY CHK	.08	1:53pm
AIR BLK	.00	1:54pm
SUB TEST	**.00**	**1:54pm**
AIR BLK	.00	1:55pm
SUB TEST	**.00**	**1:57pm**
AIR BLK	.00	1:57pm

Reported AC: .00 g/210L

K. C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY WRIGHTSVILLE BCH PD 640

Serial Number: *008667* Test Record Number: *1204*
Test Date: *06/03/2013* Test Time: *2:00pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:00pm
FLO	Pass	2:00pm
FC	Pass	2:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:00pm
SRC	Pass	2:00pm
DET	Pass	2:00pm
BAR	Pass	2:00pm
BT	Pass	2:00pm

Blank Tests

Test	Status	Time
AIR	Pass	2:01pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:01pm

CRC Tests

Test	Status	Time
COMP	Pass	2:01pm
CAL	Pass	2:01pm

Preventive Maintenance
Status: Pass

K. C. Rhodes

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location Carolina Beach

Instrument Serial No. 008661 Police Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY CAROLINA BEACH PD
640

Serial Number: *008661*
Test Date: *06/03/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	12:52pm
AIR BLK	.00	12:53pm
ACCY CHK	.08	12:53pm
AIR BLK	.00	12:54pm
SUB TEST	**.00**	**12:55pm**
AIR BLK	.00	12:56pm
SUB TEST	**.00**	**12:57pm**
AIR BLK	.00	12:58pm

Reported AC: .00 g/210L

K. C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY CAROLINA BEACH PD 640

Serial Number: *008661* Test Record Number: *1584*
Test Date: *06/03/2013* Test Time: *12:59pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:59pm
FLO	Pass	12:59pm
FC	Pass	12:59pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:59pm
SRC	Pass	12:59pm
DET	Pass	12:59pm
BAR	Pass	12:59pm
BT	Pass	12:59pm

Blank Tests

Test	Status	Time
AIR	Pass	1:00pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:00pm

CRC Tests

Test	Status	Time
COMP	Pass	1:00pm
CAL	Pass	1:00pm

Preventive Maintenance
Status: Pass

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location Wilmington

Instrument Serial No. 008628 Police Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY WILMINGTON PD 640

Serial Number: *008628*
Test Date: *06/03/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206602
Exp Date: 03/06/2014

Test	g/210L	Time
DIAG	Pass	12:01pm
AIR BLK	.00	12:01pm
ACCY CHK	.07	12:02pm
AIR BLK	.00	12:03pm
SUB TEST	**.00**	**12:04pm**
AIR BLK	.00	12:04pm
SUB TEST	**.00**	**12:06pm**
AIR BLK	.00	12:07pm

Reported AC: .00 g/210L

K.C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY WILMINGTON PD 640

Serial Number: *008628* Test Record Number: *2780*
Test Date: *06/03/2013* Test Time: *12:07pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:08pm
FLO	Pass	12:08pm
FC	Pass	12:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:08pm
SRC	Pass	12:08pm
DET	Pass	12:08pm
BAR	Pass	12:08pm
BT	Pass	12:08pm

Blank Tests

Test	Status	Time
AIR	Pass	12:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:09pm

CRC Tests

Test	Status	Time
COMP	Pass	12:09pm
CAL	Pass	12:09pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location New Hanover

Instrument Serial No. 008626 Sheriffs office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. Moore
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY NEW HANOVER CO SD 640

Serial Number: *008626*
Test Date: *06/03/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG308702
Exp Date: 03/28/2015

Test	g/210L	Time
DIAG	Pass	11:15am
AIR BLK	.00	11:16am
ACCY CHK	.07	11:17am
AIR BLK	.00	11:18am
SUB TEST	**.00**	**11:18am**
AIR BLK	.00	11:19am
SUB TEST	**.00**	**11:21am**
AIR BLK	.00	11:22am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY NEW HANOVER CO SD 640

Serial Number: *008626* Test Record Number: *4431*
Test Date: *06/03/2013* Test Time: *11:23am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:23am
FLO	Pass	11:23am
FC	Pass	11:23am

Temperature Tests

Test	Status	Time
FC1	Pass	11:23am
SRC	Pass	11:23am
DET	Pass	11:23am
BAR	Pass	11:23am
BT	Pass	11:23am

Blank Tests

Test	Status	Time
AIR	Pass	11:24am

Printer Tests

Test	Status	Time
PRNT	Pass	11:24am

CRC Tests

Test	Status	Time
COMP	Pass	11:24am
CAL	Pass	11:24am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location New Hanover

Instrument Serial No. 008617 Sheriffs office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. Rhodes
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY NEW HANOVER CO SD 640

Serial Number: *008617*
Test Date: *06/03/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG308702
Exp Date: 03/28/2015

Test	g/210L	Time
DIAG	Pass	11:14am
AIR BLK	.00	11:14am
ACCY CHK	.07	11:15am
AIR BLK	.00	11:16am
SUB TEST	**.00**	**11:16am**
AIR BLK	.00	11:17am
SUB TEST	**.00**	**11:19am**
AIR BLK	.00	11:20am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY NEW HANOVER CO SD 640

Serial Number: *008617* Test Record Number: *2119*
Test Date: *06/03/2013* Test Time: *11:20am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:21am
FLO	Pass	11:21am
FC	Pass	11:21am

Temperature Tests

Test	Status	Time
FC1	Pass	11:21am
SRC	Pass	11:21am
DET	Pass	11:21am
BAR	Pass	11:21am
BT	Pass	11:21am

Blank Tests

Test	Status	Time
AIR	Pass	11:21am

Printer Tests

Test	Status	Time
PRNT	Pass	11:21am

CRC Tests

Test	Status	Time
COMP	Pass	11:22am
CAL	Pass	11:22am

Preventive Maintenance
Status: Pass

K. C. Wheeler
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pasquotank Instrument Location Elizabeth City PD

Instrument Serial No. 008950 302 E. Colonial Ave., Elizabeth City, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Kelly [illegible] — Signature of Certifying Official

643 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PASQUOTANK COUNTY ELIZABETH CITY PD
690

Serial Number: *008950*
Test Date: *06/20/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2011-08/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG120101
Exp Date: 07/20/2013

Test	g/210L	Time
DIAG	Pass	12:09pm
AIR BLK	.00	12:10pm
ACCY CHK	.07	12:10pm
AIR BLK	.00	12:11pm
SUB TEST	**.00**	**12:12pm**
AIR BLK	.00	12:13pm
SUB TEST	**.00**	**12:14pm**
AIR BLK	.00	12:15pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PASQUOTANK COUNTY ELIZABETH CITY PD 690

Serial Number: *008950* Test Record Number: *834*
Test Date: *06/20/2013* Test Time: *12:16pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:17pm
FLO	Pass	12:17pm
FC	Pass	12:17pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:17pm
SRC	Pass	12:17pm
DET	Pass	12:17pm
BAR	Pass	12:17pm
BT	Pass	12:17pm

Blank Tests

Test	Status	Time
AIR	Pass	12:18pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:18pm

CRC Tests

Test	Status	Time
COMP	Pass	12:18pm
CAL	Pass	12:18pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Lincoln Instrument Location Lincoln County Courthouse

Instrument Serial No. 008823 #1 Courthouse Square, Lincoln
704 - 732-9020

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4th day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

65Ø
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

LINCOLN COUNTY COURTHOUSE 540

Serial Number: *008823*
Test Date: *06/04/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206102
Exp Date: 03/14/2014

Test	g/210L	Time
DIAG	Pass	2:28pm
AIR BLK	.00	2:29pm
ACCY CHK	.08	2:30pm
AIR BLK	.00	2:30pm
SUB TEST	**.00**	**2:31pm**
AIR BLK	.00	2:32pm
SUB TEST	**.00**	**2:33pm**
AIR BLK	.00	2:34pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

LINCOLN COUNTY COURTHOUSE 540

Serial Number: *008823* Test Record Number: *1028*
Test Date: *06/04/2013* Test Time: *2:35pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:35pm
FLO	Pass	2:35pm
FC	Pass	2:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:36pm
SRC	Pass	2:36pm
DET	Pass	2:36pm
BAR	Pass	2:36pm
BT	Pass	2:36pm

Blank Tests

Test	Status	Time
AIR	Pass	2:36pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:36pm

CRC Tests

Test	Status	Time
COMP	Pass	2:36pm
CAL	Pass	2:36pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location CMPD-LEC

Instrument Serial No. 008594 601 E. Trade St., Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Mark D Hayes
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: *008594*
Test Date: *06/04/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203102
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	12:40pm
AIR BLK	.00	12:41pm
ACCY CHK	.07	12:41pm
AIR BLK	.00	12:42pm
SUB TEST	**.00**	**12:43pm**
AIR BLK	.00	12:44pm
SUB TEST	**.00**	**12:46pm**
AIR BLK	.00	12:47pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: *008594* Test Record Number: *1218*
Test Date: *06/04/2013* Test Time: *12:49pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:49pm
FLO	Pass	12:49pm
FC	Pass	12:49pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:49pm
SRC	Pass	12:49pm
DET	Pass	12:49pm
BAR	Pass	12:49pm
BT	Pass	12:49pm

Blank Tests

Test	Status	Time
AIR	Pass	12:50pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:50pm

CRC Tests

Test	Status	Time
COMP	Pass	12:50pm
CAL	Pass	12:50pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location CMPD- LEC

Instrument Serial No. 008691 601 E. Trade Street, Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Joseph E. Hat
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: *008691*
Test Date: *06/04/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204602
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	12:41pm
AIR BLK	.00	12:42pm
ACCY CHK	.08	12:43pm
AIR BLK	.00	12:44pm
SUB TEST	**.00**	**12:44pm**
AIR BLK	.00	12:45pm
SUB TEST	**.00**	**12:47pm**
AIR BLK	.00	12:48pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: *008691* Test Record Number: *3906*
Test Date: *06/04/2013* Test Time: *12:49pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:49pm
FLO	Pass	12:49pm
FC	Pass	12:49pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:49pm
SRC	Pass	12:49pm
DET	Pass	12:49pm
BAR	Pass	12:49pm
BT	Pass	12:49pm

Blank Tests

Test	Status	Time
AIR	Pass	12:50pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:50pm

CRC Tests

Test	Status	Time
COMP	Pass	12:50pm
CAL	Pass	12:50pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Mecklenburg County SD

Instrument Serial No. 008690 801 E. 4th Street, Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY SD 590

Serial Number: *008690*
Test Date: *06/04/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203103
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	11:45am
AIR BLK	.00	11:46am
ACCY CHK	.08	11:47am
AIR BLK	.00	11:48am
SUB TEST	**.00**	**11:49am**
AIR BLK	.00	11:49am
SUB TEST	**.00**	**11:51am**
AIR BLK	.00	11:52am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY SD 590

Serial Number: *008690* Test Record Number: *3742*
Test Date: *06/04/2013* Test Time: *11:53am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:53am
FLO	Pass	11:53am
FC	Pass	11:53am

Temperature Tests

Test	Status	Time
FC1	Pass	11:53am
SRC	Pass	11:53am
DET	Pass	11:53am
BAR	Pass	11:53am
BT	Pass	11:53am

Blank Tests

Test	Status	Time
AIR	Pass	11:54am

Printer Tests

Test	Status	Time
PRNT	Pass	11:54am

CRC Tests

Test	Status	Time
COMP	Pass	11:54am
CAL	Pass	11:54am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Iredell Instrument Location Mooresville PD

Instrument Serial No. 008685 750 W. Iredell Ave., Mooresville
704-664-3311

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

IREDELL COUNTY MOORESVILLE PD 480

Serial Number: *008685*
Test Date: *06/04/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203103
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	9:40am
AIR BLK	.00	9:41am
ACCY CHK	.08	9:41am
AIR BLK	.00	9:42am
SUB TEST	**.00**	**9:43am**
AIR BLK	.00	9:44am
SUB TEST	**.00**	**9:45am**
AIR BLK	.00	9:46am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY MOORESVILLE PD 480

Serial Number: *008685* Test Record Number: *1917*
Test Date: *06/04/2013* Test Time: *9:47am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:48am
FLO	Pass	9:48am
FC	Pass	9:48am

Temperature Tests

Test	Status	Time
FC1	Pass	9:48am
SRC	Pass	9:48am
DET	Pass	9:48am
BAR	Pass	9:48am
BT	Pass	9:48am

Blank Tests

Test	Status	Time
AIR	Pass	9:49am

Printer Tests

Test	Status	Time
PRNT	Pass	9:49am

CRC Tests

Test	Status	Time
COMP	Pass	9:49am
CAL	Pass	9:49am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Iredell Instrument Location Iredell County S.O.

Instrument Serial No. 008809 221 E. Water Street, Statesville

704-878-3131

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Joseph E. [illegible]
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

IREDELL COUNTY IREDELL COUNTY SD 480

Serial Number: *008809*
Test Date: *06/04/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204602
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	8:53am
AIR BLK	.00	8:54am
ACCY CHK	.07	8:55am
AIR BLK	.00	8:56am
SUB TEST	**.00**	**8:56am**
AIR BLK	.00	8:57am
SUB TEST	**.00**	**8:58am**
AIR BLK	.00	9:00am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY IREDELL COUNTY SD 480

Serial Number: *008809* Test Record Number: *2291*
Test Date: *06/04/2013* Test Time: *9:01am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:02am
FLO	Pass	9:02am
FC	Pass	9:02am

Temperature Tests

Test	Status	Time
FC1	Pass	9:02am
SRC	Pass	9:02am
DET	Pass	9:02am
BAR	Pass	9:02am
BT	Pass	9:02am

Blank Tests

Test	Status	Time
AIR	Pass	9:02am

Printer Tests

Test	Status	Time
PRNT	Pass	9:02am

CRC Tests

Test	Status	Time
COMP	Pass	9:03am
CAL	Pass	9:03am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Mecklenburg County S.D.

Instrument Serial No. 008694 801 E. 4th St., Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY SD 590

Serial Number: *008694*
Test Date: *06/13/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204602
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	10:17am
AIR BLK	.00	10:18am
ACCY CHK	.08	10:19am
AIR BLK	.00	10:20am
SUB TEST	**.00**	**10:20am**
AIR BLK	.00	10:21am
SUB TEST	**.00**	**10:23am**
AIR BLK	.00	10:24am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY SD 590

Serial Number: *008694* Test Record Number: *563*
Test Date: *06/13/2013* Test Time: *10:05am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:06am
FLO	Pass	10:06am
FC	Pass	10:06am

Temperature Tests

Test	Status	Time
FC1	Pass	10:06am
SRC	Pass	10:06am
DET	Pass	10:06am
BAR	Pass	10:06am
BT	Pass	10:06am

Blank Tests

Test	Status	Time
AIR	Pass	10:07am

Printer Tests

Test	Status	Time
PRNT	Pass	10:07am

CRC Tests

Test	Status	Time
COMP	Pass	10:07am
CAL	Pass	10:07am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Union Instrument Location Waxhaw PD

Instrument Serial No. 008598 703 W. South Main St., Waxhaw
704-843-0353

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

UNION COUNTY WAXHAW PD 890

Serial Number: *008598*
Test Date: *06/18/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG303502
Exp Date: 02/04/2015

Test	g/210L	Time
DIAG	Pass	11:15am
AIR BLK	.00	11:16am
ACCY CHK	.07	11:16am
AIR BLK	.00	11:17am
SUB TEST	**.00**	**11:18am**
AIR BLK	.00	11:19am
SUB TEST	**.00**	**11:21am**
AIR BLK	.00	11:21am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

UNION COUNTY WAXHAW PD 890

Serial Number: *008598* Test Record Number: *440*
Test Date: *06/18/2013* Test Time: *11:11am*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:12am
FLO	Pass	11:12am
FC	Pass	11:12am

Temperature Tests

Test	Status	Time
FC1	Pass	11:12am
SRC	Pass	11:12am
DET	Pass	11:12am
BAR	Pass	11:12am
BT	Pass	11:12am

Blank Tests

Test	Status	Time
AIR	Pass	11:12am

Printer Tests

Test	Status	Time
PRNT	Pass	11:12am

CRC Tests

Test	Status	Time
COMP	Pass	11:13am
CAL	Pass	11:13am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Pineville PD

Instrument Serial No. 008703 427 Main St., Pineville
704-889-2231

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY PINEVILLE PD 590

Serial Number: *008703*
Test Date: *06/18/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203103
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	10:06am
AIR BLK	.00	10:06am
ACCY CHK	.07	10:07am
AIR BLK	.00	10:08am
SUB TEST	**.00**	**10:08am**
AIR BLK	.00	10:09am
SUB TEST	**.00**	**10:11am**
AIR BLK	.00	10:12am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY PINEVILLE PD 590

Serial Number: *008703* Test Record Number: *5032*
Test Date: *06/18/2013* Test Time: *10:00am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:00am
FLO	Pass	10:01am
FC	Pass	10:01am

Temperature Tests

Test	Status	Time
FC1	Pass	10:01am
SRC	Pass	10:01am
DET	Pass	10:01am
BAR	Pass	10:01am
BT	Pass	10:01am

Blank Tests

Test	Status	Time
AIR	Pass	10:01am

Printer Tests

Test	Status	Time
PRNT	Pass	10:01am

CRC Tests

Test	Status	Time
COMP	Pass	10:02am
CAL	Pass	10:02am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cleveland Instrument Location Cleveland County SD

Instrument Serial No. 008893 100 Justice Place, Shelby
704-484-4888

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CLEVELAND COUNTY CLEVELAND COUNTY SD 220

Serial Number: *008893*
Test Date: *06/20/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123501
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	10:07am
AIR BLK	.00	10:08am
ACCY CHK	.08	10:08am
AIR BLK	.00	10:10am
SUB TEST	**.00**	**10:10am**
AIR BLK	.00	10:11am
SUB TEST	**.00**	**10:12am**
AIR BLK	.00	10:13am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CLEVELAND COUNTY CLEVELAND COUNTY SD 220

Serial Number: *008893* Test Record Number: *1146*
Test Date: *06/20/2013* Test Time: *10:15am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:15am
FLO	Pass	10:15am
FC	Pass	10:15am

Temperature Tests

Test	Status	Time
FC1	Pass	10:15am
SRC	Pass	10:15am
DET	Pass	10:15am
BAR	Pass	10:15am
BT	Pass	10:15am

Blank Tests

Test	Status	Time
AIR	Pass	10:16am

Printer Tests

Test	Status	Time
PRNT	Pass	10:16am

CRC Tests

Test	Status	Time
COMP	Pass	10:16am
CAL	Pass	10:16am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cleveland Instrument Location Cleveland County S.D. - Annex

Instrument Serial No. 008887 407 McBrayer St. Shelby
704-484-4888

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CLEVELAND COUNTY CLEVELAND SD-ANNEX 220

Serial Number: *008887*
Test Date: *06/20/2013*

Citation Number: *M0000000-0*
Subject's Name: *PREVENTIVE, MAINTENANCE*
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective: *02/01/2012-02/01/2014*

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206102
Exp Date: 03/01/2014

Test	g/210L	Time
DIAG	Pass	10:59am
AIR BLK	.00	10:59am
ACCY CHK	.07	11:00am
AIR BLK	.00	11:01am
SUB TEST	**.00**	**11:02am**
AIR BLK	.00	11:02am
SUB TEST	**.00**	**11:04am**
AIR BLK	.00	11:05am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CLEVELAND COUNTY CLEVELAND SD-ANNEX 220

Serial Number: *008887* Test Record Number: *1451*
Test Date: *06/20/2013* Test Time: *10:48am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:48am
FLO	Pass	10:48am
FC	Pass	10:48am

Temperature Tests

Test	Status	Time
FC1	Pass	10:49am
SRC	Pass	10:49am
DET	Pass	10:49am
BAR	Pass	10:49am
BT	Pass	10:49am

Blank Tests

Test	Status	Time
AIR	Pass	10:49am

Printer Tests

Test	Status	Time
PRNT	Pass	10:49am

CRC Tests

Test	Status	Time
COMP	Pass	10:49am
CAL	Pass	10:49am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Carteret Co. Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008734

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22TH day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

3682E
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CARTERET COUNTY BAT MOBILE UNIT 4 150

Serial Number: *008734*
Test Date: *06/22/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG920302
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	9:18pm
AIR BLK	.00	9:19pm
ACCY CHK	.07	9:20pm
AIR BLK	.00	9:21pm
SUB TEST	**.00**	**9:21pm**
AIR BLK	.00	9:22pm
SUB TEST	**.00**	**9:24pm**
AIR BLK	.00	9:25pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY BAT MOBILE UNIT 4 150

Serial Number: *008734* Test Record Number: *686*
Test Date: *06/22/2013* Test Time: *9:27pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:27pm
FLO	Pass	9:27pm
FC	Pass	9:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:27pm
SRC	Pass	9:27pm
DET	Pass	9:27pm
BAR	Pass	9:27pm
BT	Pass	9:27pm

Blank Tests

Test	Status	Time
AIR	Pass	9:28pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:28pm

CRC Tests

Test	Status	Time
COMP	Pass	9:28pm
CAL	Pass	9:28pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Carteret Co. Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008871

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

3682E
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CARTERET COUNTY BAT MOBILE UNIT 4 150

Serial Number: *008871*
Test Date: *06/22/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	9:23pm
AIR BLK	.00	9:24pm
ACCY CHK	.07	9:24pm
AIR BLK	.00	9:25pm
SUB TEST	**.00**	**9:26pm**
AIR BLK	.00	9:27pm
SUB TEST	**.00**	**9:29pm**
AIR BLK	.00	9:30pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY BAT MOBILE UNIT 4 150

Serial Number: *008871* Test Record Number: *670*
Test Date: *06/22/2013* Test Time: *9:32pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:32pm
FLO	Pass	9:32pm
FC	Pass	9:32pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:33pm
SRC	Pass	9:33pm
DET	Pass	9:33pm
BAR	Pass	9:33pm
BT	Pass	9:33pm

Blank Tests

Test	Status	Time
AIR	Pass	9:33pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:33pm

CRC Tests

Test	Status	Time
COMP	Pass	9:33pm
CAL	Pass	9:33pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Robeson Co. Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008734

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

Certificate Number 26825

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROBESON COUNTY BAT MOBILE UNIT 4 770

Serial Number: *008734*
Test Date: *06/07/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG920302
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	5:49pm
AIR BLK	.00	5:50pm
ACCY CHK	.07	5:51pm
AIR BLK	.00	5:51pm
SUB TEST	**.00**	**5:52pm**
AIR BLK	.00	5:53pm
SUB TEST	**.00**	**5:54pm**
AIR BLK	.00	5:55pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY BAT MOBILE UNIT 4 770

Serial Number: *008734* Test Record Number: *682*
Test Date: *06/07/2013* Test Time: *5:57pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:57pm
FLO	Pass	5:57pm
FC	Pass	5:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:57pm
SRC	Pass	5:57pm
DET	Pass	5:57pm
BAR	Pass	5:57pm
BT	Pass	5:57pm

Blank Tests

Test	Status	Time
AIR	Pass	5:58pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:58pm

CRC Tests

Test	Status	Time
COMP	Pass	5:58pm
CAL	Pass	5:58pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Robeson Co Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008871

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5th day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

4822E
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROBESON COUNTY BAT MOBILE UNIT 4 770

Serial Number: *008871*
Test Date: *06/07/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	5:47pm
AIR BLK	.00	5:48pm
ACCY CHK	.07	5:48pm
AIR BLK	.00	5:49pm
SUB TEST	**.00**	**5:50pm**
AIR BLK	.00	5:51pm
SUB TEST	**.00**	**5:52pm**
AIR BLK	.00	5:53pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY BAT MOBILE UNIT 4 770

Serial Number: *008871* Test Record Number: *662*
Test Date: *06/07/2013* Test Time: *5:54pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:54pm
FLO	Pass	5:54pm
FC	Pass	5:54pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:54pm
SRC	Pass	5:54pm
DET	Pass	5:54pm
BAR	Pass	5:54pm
BT	Pass	5:54pm

Blank Tests

Test	Status	Time
AIR	Pass	5:55pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:55pm

CRC Tests

Test	Status	Time
COMP	Pass	5:55pm
CAL	Pass	5:55pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County RANDOLPH Instrument Location LIBERTY POLICE DEPT.

Instrument Serial No. 008830 LIBERTY NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

RANDOLPH LIBERTY POLICE DEPT 750

Serial Number: *008830*
Test Date: *06/18/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309101
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	11:51am
AIR BLK	.00	11:52am
ACCY CHK	.08	11:52am
AIR BLK	.00	11:53am
SUB TEST	**.00**	**11:54am**
AIR BLK	.00	11:55am
SUB TEST	**.00**	**11:57am**
AIR BLK	.00	11:57am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

RANDOLPH LIBERTY POLICE DEPT 750

Serial Number: *008830* Test Record Number: *402*
Test Date: *06/18/2013* Test Time: *11:58am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:58am
FLO	Pass	11:58am
FC	Pass	11:58am

Temperature Tests

Test	Status	Time
FC1	Pass	11:58am
SRC	Pass	11:58am
DET	Pass	11:58am
BAR	Pass	11:58am
BT	Pass	11:58am

Blank Tests

Test	Status	Time
AIR	Pass	11:59am

Printer Tests

Test	Status	Time
PRNT	Pass	11:59am

CRC Tests

Test	Status	Time
COMP	Pass	11:59am
CAL	Pass	11:59am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CUMBERLAND Instrument Location FT. BRAGG

Instrument Serial No. 008908 P.M.O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of JUNE, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY FORT BRAGG, LEC. 250

Serial Number: *008908*
Test Date: *06/19/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RUSSELL, LARRY H*
Permit Number: *06108E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG205402
Exp Date: 02/23/2014

Test	g/210L	Time
DIAG	Pass	12:12pm
AIR BLK	.00	12:13pm
ACCY CHK	.07	12:14pm
AIR BLK	.00	12:15pm
SUB TEST	**.00**	**12:15pm**
AIR BLK	.00	12:16pm
SUB TEST	**.00**	**12:18pm**
AIR BLK	.00	12:18pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY FORT BRAGG, LEC. 250

Serial Number: *008908* Test Record Number: *1447*
Test Date: *06/19/2013* Test Time: *12:20pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:20pm
FLO	Pass	12:20pm
FC	Pass	12:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:21pm
SRC	Pass	12:21pm
DET	Pass	12:21pm
BAR	Pass	12:21pm
BT	Pass	12:21pm

Blank Tests

Test	Status	Time
AIR	Pass	12:21pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:21pm

CRC Tests

Test	Status	Time
COMP	Pass	12:21pm
CAL	Pass	12:21pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Harnett Instrument Location Dunn Police Dept.

Instrument Serial No. 008644 Dunn, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

HARNETT COUNTY DUNN POLICE DEPT. 420

Serial Number: *008644*
Test Date: *06/17/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG204603
Exp Date: 02/15/2014

Test	g/210L	Time
DIAG	Pass	10:28am
AIR BLK	.00	10:29am
ACCY CHK	.08	10:30am
AIR BLK	.00	10:31am
SUB TEST	**.00**	**10:32am**
AIR BLK	.00	10:33am
SUB TEST	**.00**	**10:34am**
AIR BLK	.00	10:35am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HARNETT COUNTY DUNN POLICE DEPT. 420

Serial Number: *008644* Test Record Number: *971*
Test Date: *06/17/2013* Test Time: *10:37am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:37am
FLO	Pass	10:37am
FC	Pass	10:37am

Temperature Tests

Test	Status	Time
FC1	Pass	10:37am
SRC	Pass	10:37am
DET	Pass	10:37am
BAR	Pass	10:37am
BT	Pass	10:37am

Blank Tests

Test	Status	Time
AIR	Pass	10:38am

Printer Tests

Test	Status	Time
PRNT	Pass	10:38am

CRC Tests

Test	Status	Time
COMP	Pass	10:38am
CAL	Pass	10:38am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Johnston Instrument Location Clayton Police Dept.

Instrument Serial No. 008658 Clayton, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of June, 20 13 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

Certificate Number 651

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY CLAYTON PD. 500

Serial Number: *008658*
Test Date: *06/18/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206103
Exp Date: 03/01/2014

Test	g/210L	Time
DIAG	Pass	9:31am
AIR BLK	.00	9:32am
ACCY CHK	.08	9:33am
AIR BLK	.00	9:34am
SUB TEST	**.00**	**9:34am**
AIR BLK	.00	9:35am
SUB TEST	**.00**	**9:37am**
AIR BLK	.00	9:37am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY CLAYTON PD. 500

Serial Number: *008658* Test Record Number: *904*
Test Date: *06/18/2013* Test Time: *9:39am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:39am
FLO	Pass	9:39am
FC	Pass	9:39am

Temperature Tests

Test	Status	Time
FC1	Pass	9:39am
SRC	Pass	9:39am
DET	Pass	9:39am
BAR	Pass	9:39am
BT	Pass	9:39am

Blank Tests

Test	Status	Time
AIR	Pass	9:40am

Printer Tests

Test	Status	Time
PRNT	Pass	9:40am

CRC Tests

Test	Status	Time
COMP	Pass	9:40am
CAL	Pass	9:40am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cumberland Instrument Location FT. Bragg P.M.O.

Instrument Serial No. 008903

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY FORT BRAGG LEC. 250

Serial Number: *008903*
Test Date: *06/19/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG205402
Exp Date: 02/23/2014

Test	g/210L	Time
DIAG	Pass	12:24pm
AIR BLK	.00	12:25pm
ACCY CHK	.08	12:25pm
AIR BLK	.00	12:27pm
SUB TEST	**.00**	**12:27pm**
AIR BLK	.00	12:28pm
SUB TEST	**.00**	**12:30pm**
AIR BLK	.00	12:30pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY FORT BRAGG LEC. 250

Serial Number: *008903* Test Record Number: *1142*
Test Date: *06/19/2013* Test Time: *12:31pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:32pm
FLO	Pass	12:32pm
FC	Pass	12:32pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:32pm
SRC	Pass	12:32pm
DET	Pass	12:32pm
BAR	Pass	12:32pm
BT	Pass	12:32pm

Blank Tests

Test	Status	Time
AIR	Pass	12:33pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:33pm

CRC Tests

Test	Status	Time
COMP	Pass	12:33pm
CAL	Pass	12:33pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Durham Instrument Location Bat Mobile Unit #5

Instrument Serial No. 008600 Durham

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY BAT MOBILE UNIT 5 310

Serial Number: *008600* Test Record Number: *1242*
Test Date: *06/15/2013* Test Time: *12:14am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:15am
FLO	Pass	12:15am
FC	Pass	12:15am

Temperature Tests

Test	Status	Time
FC1	Pass	12:15am
SRC	Pass	12:15am
DET	Pass	12:15am
BAR	Pass	12:15am
BT	Pass	12:15am

Blank Tests

Test	Status	Time
AIR	Pass	12:15am

Printer Tests

Test	Status	Time
PRNT	Pass	12:15am

CRC Tests

Test	Status	Time
COMP	Pass	12:15am
CAL	Pass	12:15am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

DURHAM COUNTY BAT MOBILE UNIT 5 310

Serial Number: *008600*
Test Date: *06/15/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203102
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	12:06am
AIR BLK	.00	12:07am
ACCY CHK	.07	12:07am
AIR BLK	.00	12:08am
SUB TEST	**.00**	**12:09am**
AIR BLK	.00	12:10am
SUB TEST	**.00**	**12:12am**
AIR BLK	.00	12:13am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County DURHAM Instrument Location Bat MD Bile Unit #5

Instrument Serial No. 008698 DURHAM

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of June, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

THE GREAT SEAL OF THE STATE OF NORTH CAROLINA
MAY 20, 1775
APRIL 12, 1776
ESSE QUAM VIDERI

Stephen O'Thomas 636

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY BAT MOBILE UNIT 5 310

Serial Number: *008698* Test Record Number: *971*
Test Date: *06/15/2013* Test Time: *12:17am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:17am
FLO	Pass	12:17am
FC	Pass	12:17am

Temperature Tests

Test	Status	Time
FC1	Pass	12:17am
SRC	Pass	12:17am
DET	Pass	12:17am
BAR	Pass	12:17am
BT	Pass	12:17am

Blank Tests

Test	Status	Time
AIR	Pass	12:18am

Printer Tests

Test	Status	Time
PRNT	Pass	12:18am

CRC Tests

Test	Status	Time
COMP	Pass	12:18am
CAL	Pass	12:18am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

DURHAM COUNTY BAT MOBILE UNIT 5 310

Serial Number: *008698*
Test Date: *06/15/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	12:09am
AIR BLK	.00	12:10am
ACCY CHK	.07	12:10am
AIR BLK	.00	12:11am
SUB TEST	**.00**	**12:12am**
AIR BLK	.00	12:13am
SUB TEST	**.00**	**12:14am**
AIR BLK	.00	12:15am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Durham Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 005788 Durham

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

1036
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY BAT MOBILE UNIT 5 310

Serial Number: *008788* Test Record Number: *825*
Test Date: *06/15/2013* Test Time: *12:16am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:16am
FLO	Pass	12:16am
FC	Pass	12:16am

Temperature Tests

Test	Status	Time
FC1	Pass	12:16am
SRC	Pass	12:16am
DET	Pass	12:16am
BAR	Pass	12:16am
BT	Pass	12:16am

Blank Tests

Test	Status	Time
AIR	Pass	12:17am

Printer Tests

Test	Status	Time
PRNT	Pass	12:17am

CRC Tests

Test	Status	Time
COMP	Pass	12:17am
CAL	Pass	12:17am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

DURHAM COUNTY BAT MOBILE UNIT 5 310

Serial Number: *008788*
Test Date: *06/15/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG123502
Exp Date: 08/23/2013

Test	g/210L	Time
DIAG	Pass	12:08am
AIR BLK	.00	12:09am
ACCY CHK	.07	12:09am
AIR BLK	.00	12:10am
SUB TEST	**.00**	**12:11am**
AIR BLK	.00	12:11am
SUB TEST	**.00**	**12:13am**
AIR BLK	.00	12:14am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cumberland Instrument Location Cumberland Co. Detention Facility

Instrument Serial No. 008632 Fayetteville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

THE GREAT SEAL OF THE STATE OF NORTH CAROLINA MAY 20, 1775 APRIL 12, 1776 ESSE QUAM VIDERI

Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008632*
Test Date: *06/12/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206103
Exp Date: 03/01/2014

Test	g/210L	Time
DIAG	Pass	12:00pm
AIR BLK	.00	12:01pm
ACCY CHK	.08	12:02pm
AIR BLK	.00	12:02pm
SUB TEST	**.00**	**12:03pm**
AIR BLK	.00	12:04pm
SUB TEST	**.00**	**12:05pm**
AIR BLK	.00	12:06pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008632* Test Record Number: *2479*
Test Date: *06/12/2013* Test Time: *12:07pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:08pm
FLO	Pass	12:08pm
FC	Pass	12:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:08pm
SRC	Pass	12:08pm
DET	Pass	12:08pm
BAR	Pass	12:08pm
BT	Pass	12:08pm

Blank Tests

Test	Status	Time
AIR	Pass	12:08pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:09pm

CRC Tests

Test	Status	Time
COMP	Pass	12:09pm
CAL	Pass	12:09pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cumberland Instrument Location Cumberland Co. Detention Facility

Instrument Serial No. 008614 Fayetteville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008614*
Test Date: *06/12/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG300202
Exp Date: 01/02/2015

Test	g/210L	Time
DIAG	Pass	11:57am
AIR BLK	.00	11:58am
ACCY CHK	.08	11:59am
AIR BLK	.00	12:00pm
SUB TEST	**.00**	**12:00pm**
AIR BLK	.00	12:01pm
SUB TEST	**.00**	**12:03pm**
AIR BLK	.00	12:04pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008614* Test Record Number: *2251*
Test Date: *06/12/2013* Test Time: *12:05pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:05pm
FLO	Pass	12:05pm
FC	Pass	12:05pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:05pm
SRC	Pass	12:05pm
DET	Pass	12:05pm
BAR	Pass	12:05pm
BT	Pass	12:05pm

Blank Tests

Test	Status	Time
AIR	Pass	12:06pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:06pm

CRC Tests

Test	Status	Time
COMP	Pass	12:06pm
CAL	Pass	12:06pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cumberland Instrument Location Cumberland Co. Detention Facility

Instrument Serial No. 008633 Fayetteville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008633*
Test Date: *06/12/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG303502
Exp Date: 02/04/2015

Test	g/210L	Time
DIAG	Pass	11:56am
AIR BLK	.00	11:57am
ACCY CHK	.07	11:57am
AIR BLK	.00	11:58am
SUB TEST	**.00**	**11:59am**
AIR BLK	.00	12:00pm
SUB TEST	**.00**	**12:02pm**
AIR BLK	.00	12:03pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008633* Test Record Number: *2503*
Test Date: *06/12/2013* Test Time: *12:04pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:04pm
FLO	Pass	12:04pm
FC	Pass	12:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:04pm
SRC	Pass	12:04pm
DET	Pass	12:04pm
BAR	Pass	12:04pm
BT	Pass	12:04pm

Blank Tests

Test	Status	Time
AIR	Pass	12:05pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:05pm

CRC Tests

Test	Status	Time
COMP	Pass	12:05pm
CAL	Pass	12:05pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cumberland Instrument Location Cumberland Co Detention Facility

Instrument Serial No. 008672

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008672*
Test Date: *06/12/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TRUDELL, SR., DANIEL T*
Permit Number: *21535E*
Effective:
09/01/2011-09/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG206103
Exp Date: 03/01/2014

Test	g/210L	Time
DIAG	Pass	12:03pm
AIR BLK	.00	12:04pm
ACCY CHK	.07	12:04pm
AIR BLK	.00	12:05pm
SUB TEST	**.00**	**12:06pm**
AIR BLK	.00	12:07pm
SUB TEST	**.00**	**12:08pm**
AIR BLK	.00	12:09pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: *008672* Test Record Number: *3548*
Test Date: *06/12/2013* Test Time: *12:10pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:11pm
FLO	Pass	12:11pm
FC	Pass	12:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:11pm
SRC	Pass	12:11pm
DET	Pass	12:11pm
BAR	Pass	12:11pm
BT	Pass	12:11pm

Blank Tests

Test	Status	Time
AIR	Pass	12:12pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:12pm

CRC Tests

Test	Status	Time
COMP	Pass	12:12pm
CAL	Pass	12:12pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT MOBILE UNIT #7

Instrument Serial No. 008623 Holly Springs

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7th day of June, 2012 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910
Serial Number: *008623* Test Record Number: *2712*
Test Date: *06/07/2013* Test Time: *11:13pm EDT*
System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:14pm
FLO	Pass	11:14pm
FC	Pass	11:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:14pm
SRC	Pass	11:14pm
DET	Pass	11:14pm
BAR	Pass	11:14pm
BT	Pass	11:14pm

Blank Tests

Test	Status	Time
AIR	Pass	11:15pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:15pm

CRC Tests

Test	Status	Time
COMP	Pass	11:15pm
CAL	Pass	11:15pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910
Serial Number: *008623*
Test Date: *06/07/2013*
Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	11:01pm
AIR BLK	.00	11:02pm
ACCY CHK	.07	11:03pm
AIR BLK	.00	11:04pm
SUB TEST	**.00**	**11:05pm**
AIR BLK	.00	11:06pm
SUB TEST	**.00**	**11:08pm**
AIR BLK	.00	11:09pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT MOBILE UNIT

Instrument Serial No. 008760 Holly Spaulding

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7TH day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008760* Test Record Number: *442*
Test Date: *06/07/2013* Test Time: *11:22pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:23pm
FLO	Pass	11:23pm
FC	Pass	11:23pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:23pm
SRC	Pass	11:23pm
DET	Pass	11:23pm
BAR	Pass	11:23pm
BT	Pass	11:23pm

Blank Tests

Test	Status	Time
AIR	Pass	11:23pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:24pm

CRC Tests

Test	Status	Time
COMP	Pass	11:24pm
CAL	Pass	11:24pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008760*
Test Date: *06/07/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	11:11pm
AIR BLK	.00	11:12pm
ACCY CHK	.07	11:13pm
AIR BLK	.00	11:13pm
SUB TEST	**.00**	**11:14pm**
AIR BLK	.00	11:15pm
SUB TEST	**.00**	**11:17pm**
AIR BLK	.00	11:18pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT MOBILE UNIT #7

Instrument Serial No. 008000 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008600* Test Record Number: *1238*
Test Date: *06/09/2013* Test Time: *12:42am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:43am
FLO	Pass	12:43am
FC	Pass	12:43am

Temperature Tests

Test	Status	Time
FC1	Pass	12:43am
SRC	Pass	12:43am
DET	Pass	12:43am
BAR	Pass	12:43am
BT	Pass	12:43am

Blank Tests

Test	Status	Time
AIR	Pass	12:44am

Printer Tests

Test	Status	Time
PRNT	Pass	12:44am

CRC Tests

Test	Status	Time
COMP	Pass	12:44am
CAL	Pass	12:44am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008600*
Test Date: *06/09/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203102
Exp Date: 01/31/2014

Test	g/210L	Time
DIAG	Pass	12:28am
AIR BLK	.00	12:29am
ACCY CHK	.07	12:29am
AIR BLK	.00	12:31am
SUB TEST	**.00**	**12:32am**
AIR BLK	.00	12:33am
SUB TEST	**.00**	**12:35am**
AIR BLK	.00	12:36am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT Mobile Unit #7

Instrument Serial No. 008577 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008577* Test Record Number: *871*
Test Date: *06/09/2013* Test Time: *1:43am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:44am
FLO	Pass	1:44am
FC	Pass	1:44am

Temperature Tests

Test	Status	Time
FC1	Pass	1:44am
SRC	Pass	1:44am
DET	Pass	1:44am
BAR	Pass	1:44am
BT	Pass	1:44am

Blank Tests

Test	Status	Time
AIR	Pass	1:45am

Printer Tests

Test	Status	Time
PRNT	Pass	1:45am

CRC Tests

Test	Status	Time
COMP	Pass	1:45am
CAL	Pass	1:45am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008577*
Test Date: *06/09/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	1:33am
AIR BLK	.00	1:34am
ACCY CHK	.07	1:35am
AIR BLK	.00	1:36am
SUB TEST	**.00**	**1:36am**
AIR BLK	.00	1:37am
SUB TEST	**.00**	**1:39am**
AIR BLK	.00	1:40am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT Mobile Unit #7

Instrument Serial No. 005623 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official 636 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910
Serial Number: *008623* Test Record Number: *2715*
Test Date: *06/09/2013* Test Time: *1:11am EDT*
System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:11am
FLO	Pass	1:11am
FC	Pass	1:11am

Temperature Tests

Test	Status	Time
FC1	Pass	1:11am
SRC	Pass	1:11am
DET	Pass	1:11am
BAR	Pass	1:11am
BT	Pass	1:11am

Blank Tests

Test	Status	Time
AIR	Pass	1:12am

Printer Tests

Test	Status	Time
PRNT	Pass	1:12am

CRC Tests

Test	Status	Time
COMP	Pass	1:12am
CAL	Pass	1:12am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910
Serial Number: *008623*
Test Date: *06/09/2013*
Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	12:59am
AIR BLK	.00	1:00am
ACCY CHK	.07	1:01am
AIR BLK	.00	1:02am
SUB TEST	**.00**	**1:03am**
AIR BLK	.00	1:04am
SUB TEST	**.00**	**1:06am**
AIR BLK	.00	1:06am

Reported AC: .00 ~~g/210L~~

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT Mobile Unit #7

Instrument Serial No. 8760 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox E-/I-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008760* Test Record Number: *446*
Test Date: *06/09/2013* Test Time: *1:23am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:23am
FLO	Pass	1:23am
FC	Pass	1:24am

Temperature Tests

Test	Status	Time
FC1	Pass	1:24am
SRC	Pass	1:24am
DET	Pass	1:24am
BAR	Pass	1:24am
BT	Pass	1:24am

Blank Tests

Test	Status	Time
AIR	Pass	1:24am

Printer Tests

Test	Status	Time
PRNT	Pass	1:24am

CRC Tests

Test	Status	Time
COMP	Pass	1:25am
CAL	Pass	1:25am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008760*
Test Date: *06/09/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
10/01/2011-10/01/2013

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	1:04am
AIR BLK	.00	1:05am
ACCY CHK	.07	1:05am
AIR BLK	.00	1:06am
SUB TEST	**.00**	**1:07am**
AIR BLK	.00	1:08am
SUB TEST	**.00**	**1:09am**
AIR BLK	.00	1:11am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Carteret Co. Instrument Location BAT Mobile Unit 4

Instrument Serial No. 00881

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CRAVEN COUNTY BAT MOBILE UNIT 4 240

Serial Number: *008871*
Test Date: *06/01/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG203902
Exp Date: 02/08/2014

Test	g/210L	Time
DIAG	Pass	10:03pm
AIR BLK	.00	10:04pm
ACCY CHK	.07	10:05pm
AIR BLK	.00	10:06pm
SUB TEST	**.00**	**10:06pm**
AIR BLK	.00	10:07pm
SUB TEST	**.00**	**10:09pm**
AIR BLK	.00	10:10pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY BAT MOBILE UNIT 4 240

Serial Number: *008871* Test Record Number: *659*
Test Date: *06/01/2013* Test Time: *10:10pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:11pm
FLO	Pass	10:11pm
FC	Pass	10:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:11pm
SRC	Pass	10:11pm
DET	Pass	10:11pm
BAR	Pass	10:11pm
BT	Pass	10:11pm

Blank Tests

Test	Status	Time
AIR	Pass	10:12pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:12pm

CRC Tests

Test	Status	Time
COMP	Pass	10:12pm
CAL	Pass	10:12pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Craven Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008434

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1st day of June, 2013 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

7682E
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CRAVEN COUNTY BAT MOBILE UNIT 4 240

Serial Number: *008734*
Test Date: *06/01/2013*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2012-02/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG920302
Exp Date: 09/06/2013

Test	g/210L	Time
DIAG	Pass	10:05pm
AIR BLK	.00	10:06pm
ACCY CHK	.07	10:07pm
AIR BLK	.00	10:08pm
SUB TEST	**.00**	**10:08pm**
AIR BLK	.00	10:09pm
SUB TEST	**.00**	**10:11pm**
AIR BLK	.00	10:12pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY BAT MOBILE UNIT 4 240

Serial Number: *008734* Test Record Number: *678*
Test Date: *06/01/2013* Test Time: *10:13pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:13pm
FLO	Pass	10:13pm
FC	Pass	10:13pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:13pm
SRC	Pass	10:13pm
DET	Pass	10:13pm
BAR	Pass	10:13pm
BT	Pass	10:13pm

Blank Tests

Test	Status	Time
AIR	Pass	10:14pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:14pm

CRC Tests

Test	Status	Time
COMP	Pass	10:14pm
CAL	Pass	10:14pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007