DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Gaston Instrument Location Gaston County SD

Instrument Serial No. 008643 425 N. Marietta St. Gastonia
704-869-6800

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

GASTON COUNTY GASTON COUNTY SD 350

Serial Number: *008643* Test Record Number: *1833*
Test Date: *05/19/2014* Test Time: *9:19am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:19am
FLO	Pass	9:19am
FC	Pass	9:20am

Temperature Tests

Test	Status	Time
FC1	Pass	9:20am
SRC	Pass	9:20am
DET	Pass	9:20am
BAR	Pass	9:20am
BT	Pass	9:20am

Blank Tests

Test	Status	Time
AIR	Pass	9:20am

Printer Tests

Test	Status	Time
PRNT	Pass	9:20am

CRC Tests

Test	Status	Time
COMP	Pass	9:20am
CAL	Pass	9:20am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

GASTON COUNTY GASTON COUNTY SD 350

Serial Number: *008643*
Test Date: *05/19/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
01/01/2014-01/01/2016

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG335201
Exp Date: 12/18/2015

Test	g/210L	Time
DIAG	Pass	10:19am
AIR BLK	.00	10:20am
ACCY CHK	.07	10:21am
AIR BLK	.00	10:22am
SUB TEST	**.00**	**10:22am**
AIR BLK	.00	10:23am
SUB TEST	**.00**	**10:25am**
AIR BLK	.00	10:26am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Stokes Instrument Location Stokes County Jail

Instrument Serial No. 008596 Danbury, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

STOKES COUNTY STOKES COUNTY JAIL 840

Serial Number: *008596*
Test Date: *05/23/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	2:06pm
AIR BLK	.00	2:07pm
ACCY CHK	.08	2:07pm
AIR BLK	.00	2:08pm
SUB TEST	**.00**	**2:09pm**
AIR BLK	.00	2:09pm
SUB TEST	**.00**	**2:11pm**
AIR BLK	.00	2:12pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

STOKES COUNTY STOKES COUNTY JAIL 840

Serial Number: *008596* Test Record Number: *665*
Test Date: *05/23/2014* Test Time: *2:14pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:14pm
FLO	Pass	2:14pm
FC	Pass	2:15pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:15pm
SRC	Pass	2:15pm
DET	Pass	2:15pm
BAR	Pass	2:15pm
BT	Pass	2:15pm

Blank Tests

Test	Status	Time
AIR	Pass	2:15pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:15pm

CRC Tests

Test	Status	Time
COMP	Pass	2:15pm
CAL	Pass	2:15pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Stokes Instrument Location King Police

Instrument Serial No. 008610 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

STOKES COUNTY KING PD 840

Serial Number: *008610*
Test Date: *05/22/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	3:54pm
AIR BLK	.00	3:55pm
ACCY CHK	.07	3:55pm
AIR BLK	.00	3:57pm
SUB TEST	**.00**	**3:57pm**
AIR BLK	.00	3:58pm
SUB TEST	**.00**	**4:00pm**
AIR BLK	.00	4:00pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

STOKES COUNTY KING PD 840

Serial Number: *008610* Test Record Number: *1446*
Test Date: *05/22/2014* Test Time: *4:01pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:02pm
FLO	Pass	4:02pm
FC	Pass	4:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:02pm
SRC	Pass	4:02pm
DET	Pass	4:02pm
BAR	Pass	4:02pm
BT	Pass	4:02pm

Blank Tests

Test	Status	Time
AIR	Pass	4:02pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:03pm

CRC Tests

Test	Status	Time
COMP	Pass	4:03pm
CAL	Pass	4:03pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Forsyth Instrument Location Forsyth County Detention

Instrument Serial No. 008659 Winston-Salem, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

FORSYTH COUNTY FORSYTH CO DETENTION
330

Serial Number: *008659*
Test Date: *05/22/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG404101
Exp Date: 02/10/2016

Test	g/210L	Time
DIAG	Pass	1:40pm
AIR BLK	.00	1:41pm
ACCY CHK	.07	1:42pm
AIR BLK	.00	1:42pm
SUB TEST	**.00**	**1:43pm**
AIR BLK	.00	1:44pm
SUB TEST	**.00**	**1:46pm**
AIR BLK	.00	1:46pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: *008659* Test Record Number: *2767*
Test Date: *05/22/2014* Test Time: *1:47pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:48pm
FLO	Pass	1:48pm
FC	Pass	1:48pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:48pm
SRC	Pass	1:48pm
DET	Pass	1:48pm
BAR	Pass	1:48pm
BT	Pass	1:48pm

Blank Tests

Test	Status	Time
AIR	Pass	1:48pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:49pm

CRC Tests

Test	Status	Time
COMP	Pass	1:49pm
CAL	Pass	1:49pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Forsyth Instrument Location Forsyth County Detention

Instrument Serial No. 005583 Winston-Salem, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

FORSYTH COUNTY FORSYTH CO DETENTION
330

Serial Number: *008583*
Test Date: *05/22/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG404101
Exp Date: 02/10/2016

Test	g/210L	Time
DIAG	Pass	1:28pm
AIR BLK	.00	1:28pm
ACCY CHK	.08	1:29pm
AIR BLK	.00	1:30pm
SUB TEST	**.00**	**1:30pm**
AIR BLK	.00	1:31pm
SUB TEST	**.00**	**1:33pm**
AIR BLK	.00	1:34pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: *008583* Test Record Number: *5000*
Test Date: *05/22/2014* Test Time: *1:35pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:36pm
FLO	Pass	1:36pm
FC	Pass	1:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:36pm
SRC	Pass	1:36pm
DET	Pass	1:36pm
BAR	Pass	1:36pm
BT	Pass	1:36pm

Blank Tests

Test	Status	Time
AIR	Pass	1:37pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:37pm

CRC Tests

Test	Status	Time
COMP	Pass	1:37pm
CAL	Pass	1:37pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Forsyth Instrument Location Forsyth County Detention

Instrument Serial No. 008925 Winston-Salem, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

FORSYTH COUNTY FORSYTH CO DETENTION
330

Serial Number: *008925*
Test Date: *05/22/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG404101
Exp Date: 02/10/2016

Test	g/210L	Time
DIAG	Pass	1:17pm
AIR BLK	.00	1:18pm
ACCY CHK	.08	1:19pm
AIR BLK	.00	1:19pm
SUB TEST	**.00**	**1:20pm**
AIR BLK	.00	1:21pm
SUB TEST	**.00**	**1:22pm**
AIR BLK	.00	1:23pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: *008925* Test Record Number: *566*
Test Date: *05/22/2014* Test Time: *1:26pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:26pm
FLO	Pass	1:26pm
FC	Pass	1:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:26pm
SRC	Pass	1:26pm
DET	Pass	1:26pm
BAR	Pass	1:26pm
BT	Pass	1:26pm

Blank Tests

Test	Status	Time
AIR	Pass	1:27pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:27pm

CRC Tests

Test	Status	Time
COMP	Pass	1:27pm
CAL	Pass	1:27pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Davie Instrument Location Davie County Jail

Instrument Serial No. 008905 Mocksville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DAVIE COUNTY DAVIE COUNTY JAIL 290

Serial Number: *008905*
Test Date: *05/21/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG323402
Exp Date: 08/22/2015

Test	g/210L	Time
DIAG	Pass	11:02am
AIR BLK	.00	11:03am
ACCY CHK	.08	11:04am
AIR BLK	.00	11:05am
SUB TEST	**.00**	**11:06am**
AIR BLK	.00	11:06am
SUB TEST	**.00**	**11:08am**
AIR BLK	.00	11:09am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIE COUNTY DAVIE COUNTY JAIL 290

Serial Number: *008905* Test Record Number: *1381*
Test Date: *05/21/2014* Test Time: *11:12am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:12am
FLO	Pass	11:12am
FC	Pass	11:12am

Temperature Tests

Test	Status	Time
FC1	Pass	11:12am
SRC	Pass	11:12am
DET	Pass	11:12am
BAR	Pass	11:12am
BT	Pass	11:12am

Blank Tests

Test	Status	Time
AIR	Pass	11:13am

Printer Tests

Test	Status	Time
PRNT	Pass	11:13am

CRC Tests

Test	Status	Time
COMP	Pass	11:13am
CAL	Pass	11:13am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Forsyth Instrument Location Kernersville Police Department

Instrument Serial No. 008650

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH COUNTY KERNERSVILLE PD 330

Serial Number: *008650*
Test Date: *05/19/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG319902
Exp Date: 07/18/2015

Test	g/210L	Time
DIAG	Pass	11:12am
AIR BLK	.00	11:12am
ACCY CHK	.08	11:13am
AIR BLK	.00	11:14am
SUB TEST	**.00**	**11:14am**
AIR BLK	.00	11:15am
SUB TEST	**.00**	**11:17am**
AIR BLK	.00	11:18am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY KERNERSVILLE PD 330

Serial Number: *008650* Test Record Number: *1048*
Test Date: *05/19/2014* Test Time: *11:21am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:21am
FLO	Pass	11:21am
FC	Pass	11:21am

Temperature Tests

Test	Status	Time
FC1	Pass	11:22am
SRC	Pass	11:22am
DET	Pass	11:22am
BAR	Pass	11:22am
BT	Pass	11:22am

Blank Tests

Test	Status	Time
AIR	Pass	11:22am

Printer Tests

Test	Status	Time
PRNT	Pass	11:22am

CRC Tests

Test	Status	Time
COMP	Pass	11:22am
CAL	Pass	11:22am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Surry Instrument Location Elkin Police

Instrument Serial No. 008926 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of MAY, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

SURRY COUNTY ELKIN PD 850

Serial Number: *008926*
Test Date: *05/14/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BENFIELD II, KENNETH R*
Permit Number: *22067E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG320602
Exp Date: 07/25/2015

Test	g/210L	Time
DIAG	Pass	2:54pm
AIR BLK	.00	2:54pm
ACCY CHK	.07	2:55pm
AIR BLK	.00	2:56pm
SUB TEST	**.00**	**2:57pm**
AIR BLK	.00	2:58pm
SUB TEST	**.00**	**2:59pm**
AIR BLK	.00	3:00pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

SURRY COUNTY ELKIN PD 850

Serial Number: *008926* Test Record Number: *603*
Test Date: *05/14/2014* Test Time: *3:01pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:02pm
FLO	Pass	3:02pm
FC	Pass	3:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:02pm
SRC	Pass	3:02pm
DET	Pass	3:02pm
BAR	Pass	3:02pm
BT	Pass	3:02pm

Blank Tests

Test	Status	Time
AIR	Pass	3:02pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:03pm

CRC Tests

Test	Status	Time
COMP	Pass	3:03pm
CAL	Pass	3:03pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location B.t Mobile Unit 5

Instrument Serial No. 0086W

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

659
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG BAT MOBILE UNIT 5 590

Serial Number: *008600*
Test Date: *05/08/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG322601
Exp Date: 08/14/2015

Test	g/210L	Time
DIAG	Pass	7:48pm
AIR BLK	.00	7:49pm
ACCY CHK	.07	7:50pm
AIR BLK	.00	7:51pm
SUB TEST	**.00**	**7:52pm**
AIR BLK	.00	7:53pm
SUB TEST	**.00**	**7:54pm**
AIR BLK	.00	7:55pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG BAT MOBILE UNIT 5 590

Serial Number: *008600* Test Record Number: *1421*
Test Date: *05/08/2014* Test Time: *8:00pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:00pm
FLO	Pass	8:00pm
FC	Pass	8:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:00pm
SRC	Pass	8:00pm
DET	Pass	8:00pm
BAR	Pass	8:00pm
BT	Pass	8:00pm

Blank Tests

Test	Status	Time
AIR	Pass	8:01pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:01pm

CRC Tests

Test	Status	Time
COMP	Pass	8:01pm
CAL	Pass	8:01pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Bat mobile Unit 5

Instrument Serial No. 008698

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG BAT MOBILE UNIT 5 590

Serial Number: *008698*
Test Date: *05/08/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG405702
Exp Date: 02/26/2016

Test	g/210L	Time
DIAG	Pass	7:44pm
AIR BLK	.00	7:45pm
ACCY CHK	.07	7:46pm
AIR BLK	.00	7:47pm
SUB TEST	**.00**	**7:48pm**
AIR BLK	.00	7:49pm
SUB TEST	**.00**	**7:51pm**
AIR BLK	.00	7:52pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG BAT MOBILE UNIT 5 590

Serial Number: *008698* Test Record Number: *1103*
Test Date: *05/08/2014* Test Time: *7:56pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:57pm
FLO	Pass	7:57pm
FC	Pass	7:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:57pm
SRC	Pass	7:57pm
DET	Pass	7:57pm
BAR	Pass	7:57pm
BT	Pass	7:57pm

Blank Tests

Test	Status	Time
AIR	Pass	7:58pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:58pm

CRC Tests

Test	Status	Time
COMP	Pass	7:58pm
CAL	Pass	7:58pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location But mobile Unit 5

Instrument Serial No. 008798

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG BAT MOBILE UNIT 5 590

Serial Number: *008788*
Test Date: *05/08/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309101
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	7:55pm
AIR BLK	.00	7:56pm
ACCY CHK	.07	7:57pm
AIR BLK	.00	7:58pm
SUB TEST	**.00**	**7:58pm**
AIR BLK	.00	7:59pm
SUB TEST	**.00**	**8:00pm**
AIR BLK	.00	8:01pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG BAT MOBILE UNIT 5 590

Serial Number: *008788* Test Record Number: *957*
Test Date: *05/08/2014* Test Time: *8:05pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:06pm
FLO	Pass	8:06pm
FC	Pass	8:06pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:06pm
SRC	Pass	8:06pm
DET	Pass	8:06pm
BAR	Pass	8:06pm
BT	Pass	8:06pm

Blank Tests

Test	Status	Time
AIR	Pass	8:06pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:07pm

CRC Tests

Test	Status	Time
COMP	Pass	8:07pm
CAL	Pass	8:07pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County IRedell Instrument Location Bat mobile Unit 5

Instrument Serial No. 008600

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

IREDELL BAT MOBILE UNIT 5 480

Serial Number: *008600*
Test Date: *05/25/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG322601
Exp Date: 08/14/2015

Test	g/210L	Time
DIAG	Pass	1:16pm
AIR BLK	.00	1:17pm
ACCY CHK	.07	1:18pm
AIR BLK	.00	1:19pm
SUB TEST	**.00**	**1:20pm**
AIR BLK	.00	1:21pm
SUB TEST	**.00**	**1:22pm**
AIR BLK	.00	1:23pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

IREDELL BAT MOBILE UNIT 5 480

Serial Number: *008600* Test Record Number: *1438*
Test Date: *05/25/2014* Test Time: *1:25pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:25pm
FLO	Pass	1:25pm
FC	Pass	1:25pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:26pm
SRC	Pass	1:26pm
DET	Pass	1:26pm
BAR	Pass	1:26pm
BT	Pass	1:26pm

Blank Tests

Test	Status	Time
AIR	Pass	1:26pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:26pm

CRC Tests

Test	Status	Time
COMP	Pass	1:26pm
CAL	Pass	1:26pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Iredell Instrument Location Butmobile Unit 5

Instrument Serial No. 068788

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

IREDELL BAT MOBILE UNIT 5 480

Serial Number: *008788*
Test Date: *05/25/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309101
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	1:19pm
AIR BLK	.00	1:20pm
ACCY CHK	.07	1:21pm
AIR BLK	.00	1:21pm
SUB TEST	**.00**	**1:22pm**
AIR BLK	.00	1:23pm
SUB TEST	**.00**	**1:24pm**
AIR BLK	.00	1:25pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

IREDELL BAT MOBILE UNIT 5 480

Serial Number: *008788* Test Record Number: *967*
Test Date: *05/25/2014* Test Time: *1:34pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:34pm
FLO	Pass	1:34pm
FC	Pass	1:35pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:35pm
SRC	Pass	1:35pm
DET	Pass	1:35pm
BAR	Pass	1:35pm
BT	Pass	1:35pm

Blank Tests

Test	Status	Time
AIR	Pass	1:35pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:35pm

CRC Tests

Test	Status	Time
COMP	Pass	1:35pm
CAL	Pass	1:35pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location Bat mobile 5

Instrument Serial No. 008698

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

GREENSBORO BAT MOBILE UNIT 5 400

Serial Number: *008698* Test Record Number: *1112*
Test Date: *05/23/2014* Test Time: *8:17pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:18pm
FLO	Pass	8:18pm
FC	Pass	8:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:18pm
SRC	Pass	8:18pm
DET	Pass	8:18pm
BAR	Pass	8:18pm
BT	Pass	8:18pm

Blank Tests

Test	Status	Time
AIR	Pass	8:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:19pm

CRC Tests

Test	Status	Time
COMP	Pass	8:19pm
CAL	Pass	8:19pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

GREENSBORO BAT MOBILE UNIT 5 400

Serial Number: *008698*
Test Date: *05/23/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG405702
Exp Date: 02/26/2016

Test	g/210L	Time
DIAG	Pass	8:08pm
AIR BLK	.00	8:09pm
ACCY CHK	.07	8:10pm
AIR BLK	.00	8:11pm
SUB TEST	**.00**	**8:12pm**
AIR BLK	.00	8:13pm
SUB TEST	**.00**	**8:15pm**
AIR BLK	.00	8:16pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location But Mobile Units

Instrument Serial No. 008788

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

659
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GREENSBORO BAT MOBILE UNIT 5 400

Serial Number: *008788*
Test Date: *05/23/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309101
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	8:06pm
AIR BLK	.00	8:07pm
ACCY CHK	.07	8:07pm
AIR BLK	.00	8:08pm
SUB TEST	**.00**	**8:08pm**
AIR BLK	.00	8:09pm
SUB TEST	**.00**	**8:11pm**
AIR BLK	.00	8:12pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GREENSBORO BAT MOBILE UNIT 5 400

Serial Number: *008788* Test Record Number: *962*
Test Date: *05/23/2014* Test Time: *8:13pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:13pm
FLO	Pass	8:13pm
FC	Pass	8:13pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:13pm
SRC	Pass	8:13pm
DET	Pass	8:13pm
BAR	Pass	8:13pm
BT	Pass	8:13pm

Blank Tests

Test	Status	Time
AIR	Pass	8:14pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:14pm

CRC Tests

Test	Status	Time
COMP	Pass	8:14pm
CAL	Pass	8:14pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location But mobile Unit 5

Instrument Serial No. 008400

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

GREENSBORO BAT MOBILE UNIT 5 400

Serial Number: *008600* Test Record Number: *1431*
Test Date: *05/23/2014* Test Time: *8:12pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:12pm
FLO	Pass	8:12pm
FC	Pass	8:12pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:13pm
SRC	Pass	8:13pm
DET	Pass	8:13pm
BAR	Pass	8:13pm
BT	Pass	8:13pm

Blank Tests

Test	Status	Time
AIR	Pass	8:13pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:13pm

CRC Tests

Test	Status	Time
COMP	Pass	8:13pm
CAL	Pass	8:13pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

GREENSBORO BAT MOBILE UNIT 5 400

Serial Number: *008600*
Test Date: *05/23/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG322601
Exp Date: 08/14/2015

Test	g/210L	Time
DIAG	Pass	8:04pm
AIR BLK	.00	8:05pm
ACCY CHK	.07	8:05pm
AIR BLK	.00	8:06pm
SUB TEST	**.00**	**8:07pm**
AIR BLK	.00	8:08pm
SUB TEST	**.00**	**8:10pm**
AIR BLK	.00	8:11pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Burke Instrument Location Bud Mobile Unit 5

Instrument Serial No. 008788

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BURKE BAT MOBILE UNIT 5 110

Serial Number: *008788*
Test Date: *05/24/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309101
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	7:37pm
AIR BLK	.00	7:38pm
ACCY CHK	.07	7:38pm
AIR BLK	.00	7:39pm
SUB TEST	**.00**	**7:40pm**
AIR BLK	.00	7:41pm
SUB TEST	**.00**	**7:42pm**
AIR BLK	.00	7:43pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BURKE BAT MOBILE UNIT 5 110

Serial Number: *008788* Test Record Number: *965*
Test Date: *05/24/2014* Test Time: *7:44pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:44pm
FLO	Pass	7:44pm
FC	Pass	7:44pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:45pm
SRC	Pass	7:45pm
DET	Pass	7:45pm
BAR	Pass	7:45pm
BT	Pass	7:45pm

Blank Tests

Test	Status	Time
AIR	Pass	7:45pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:45pm

CRC Tests

Test	Status	Time
COMP	Pass	7:45pm
CAL	Pass	7:45pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Burke Instrument Location But mobile Unit 5

Instrument Serial No. 008600

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of May, 20____ the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BURKE BAT MOBILE UNIT 5 110

Serial Number: *008600*
Test Date: *05/24/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG322601
Exp Date: 08/14/2015

Test	g/210L	Time
DIAG	Pass	7:35pm
AIR BLK	.00	7:36pm
ACCY CHK	.07	7:37pm
AIR BLK	.00	7:38pm
SUB TEST	**.00**	**7:39pm**
AIR BLK	.00	7:40pm
SUB TEST	**.00**	**7:41pm**
AIR BLK	.00	7:42pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BURKE BAT MOBILE UNIT 5 110

Serial Number: *008600* Test Record Number: *1435*
Test Date: *05/24/2014* Test Time: *7:47pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:47pm
FLO	Pass	7:47pm
FC	Pass	7:48pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:48pm
SRC	Pass	7:48pm
DET	Pass	7:48pm
BAR	Pass	7:48pm
BT	Pass	7:48pm

Blank Tests

Test	Status	Time
AIR	Pass	7:48pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:48pm

CRC Tests

Test	Status	Time
COMP	Pass	7:49pm
CAL	Pass	7:49pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Bat mobile Unit 5

Instrument Serial No. 008600

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG BAT MOBILE UNIT 5 590

Serial Number: *008600*
Test Date: *05/30/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG322601
Exp Date: 08/14/2015

Test	g/210L	Time
DIAG	Pass	7:09pm
AIR BLK	.00	7:10pm
ACCY CHK	.07	7:11pm
AIR BLK	.00	7:12pm
SUB TEST	**.00**	**7:13pm**
AIR BLK	.00	7:14pm
SUB TEST	**.00**	**7:15pm**
AIR BLK	.00	7:16pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG BAT MOBILE UNIT 5 590

Serial Number: *008600* Test Record Number: *1441*
Test Date: *05/30/2014* Test Time: *7:21pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:21pm
FLO	Pass	7:21pm
FC	Pass	7:22pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:22pm
SRC	Pass	7:22pm
DET	Pass	7:22pm
BAR	Pass	7:22pm
BT	Pass	7:22pm

Blank Tests

Test	Status	Time
AIR	Pass	7:22pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:22pm

CRC Tests

Test	Status	Time
COMP	Pass	7:23pm
CAL	Pass	7:23pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location But Mobile Unit 5

Instrument Serial No. DD 8788

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG BAT MOBILE UNIT 5 590

Serial Number: *008788*
Test Date: *05/30/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309101
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	7:12pm
AIR BLK	.00	7:13pm
ACCY CHK	.07	7:14pm
AIR BLK	.00	7:15pm
SUB TEST	**.00**	**7:15pm**
AIR BLK	.00	7:16pm
SUB TEST	**.00**	**7:18pm**
AIR BLK	.00	7:19pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG BAT MOBILE UNIT 5 590

Serial Number: *008788* Test Record Number: *970*
Test Date: *05/30/2014* Test Time: *7:20pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:20pm
FLO	Pass	7:20pm
FC	Pass	7:20pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:20pm
SRC	Pass	7:20pm
DET	Pass	7:20pm
BAR	Pass	7:20pm
BT	Pass	7:20pm

Blank Tests

Test	Status	Time
AIR	Pass	7:21pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:21pm

CRC Tests

Test	Status	Time
COMP	Pass	7:21pm
CAL	Pass	7:21pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Bat mobile Unit 5

Instrument Serial No. 008690

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of May, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG BAT MOBILE UNIT 5 590

Serial Number: *008698*
Test Date: *05/30/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG405702
Exp Date: 02/26/2016

Test	g/210L	Time
DIAG	Pass	7:17pm
AIR BLK	.00	7:18pm
ACCY CHK	.07	7:19pm
AIR BLK	.00	7:19pm
SUB TEST	**.00**	**7:20pm**
AIR BLK	.00	7:21pm
SUB TEST	**.00**	**7:23pm**
AIR BLK	.00	7:23pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG BAT MOBILE UNIT 5 590

Serial Number: *008698* Test Record Number: *1116*
Test Date: *05/30/2014* Test Time: *7:29pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:29pm
FLO	Pass	7:29pm
FC	Pass	7:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:30pm
SRC	Pass	7:30pm
DET	Pass	7:30pm
BAR	Pass	7:30pm
BT	Pass	7:30pm

Blank Tests

Test	Status	Time
AIR	Pass	7:30pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:30pm

CRC Tests

Test	Status	Time
COMP	Pass	7:30pm
CAL	Pass	7:30pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Onslow Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008871

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ONSLOW COUNTY BAT MOBILE UNIT 4 660

Serial Number: *008871*
Test Date: *05/31/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2014-02/01/2016

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG400603
Exp Date: 01/06/2016

Test	g/210L	Time
DIAG	Pass	11:05am
AIR BLK	.00	11:05am
ACCY CHK	.07	11:06am
AIR BLK	.00	11:07am
SUB TEST	**.00**	**11:08am**
AIR BLK	.00	11:09am
SUB TEST	**.00**	**11:10am**
AIR BLK	.00	11:11am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY BAT MOBILE UNIT 4 660

Serial Number: *008871* Test Record Number: *784*
Test Date: *05/31/2014* Test Time: *11:12am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:12am
FLO	Pass	11:12am
FC	Pass	11:12am

Temperature Tests

Test	Status	Time
FC1	Pass	11:13am
SRC	Pass	11:13am
DET	Pass	11:13am
BAR	Pass	11:13am
BT	Pass	11:13am

Blank Tests

Test	Status	Time
AIR	Pass	11:13am

Printer Tests

Test	Status	Time
PRNT	Pass	11:13am

CRC Tests

Test	Status	Time
COMP	Pass	11:13am
CAL	Pass	11:13am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Onslow Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008734

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

Certificate Number 654

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ONSLOW COUNTY BAT MOBILE UNIT 4 660

Serial Number: *008734*
Test Date: *05/31/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, GRAYHAM C*
Permit Number: *7682E*
Effective:
02/01/2014-02/01/2016

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG322601
Exp Date: 08/14/2015

Test	g/210L	Time
DIAG	Pass	10:41am
AIR BLK	.00	10:42am
ACCY CHK	.08	10:42am
AIR BLK	.00	10:43am
SUB TEST	**.00**	**10:44am**
AIR BLK	.00	10:45am
SUB TEST	**.00**	**10:46am**
AIR BLK	.00	10:47am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY BAT MOBILE UNIT 4 660

Serial Number: *008734* Test Record Number: *829*
Test Date: *05/31/2014* Test Time: *10:51am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:51am
FLO	Pass	10:51am
FC	Pass	10:51am

Temperature Tests

Test	Status	Time
FC1	Pass	10:51am
SRC	Pass	10:51am
DET	Pass	10:51am
BAR	Pass	10:51am
BT	Pass	10:51am

Blank Tests

Test	Status	Time
AIR	Pass	10:52am

Printer Tests

Test	Status	Time
PRNT	Pass	10:52am

CRC Tests

Test	Status	Time
COMP	Pass	10:52am
CAL	Pass	10:52am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Alamance Instrument Location Alamance Co. Jail

Instrument Serial No. 008913 109 S. Maple St.
Graham, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

655
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: *008913*
Test Date: *05/30/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *FARLEY, CYNTHIA D*
Permit Number: *24123E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG320602
Exp Date: 07/25/2015

Test	g/210L	Time
DIAG	Pass	11:00am
AIR BLK	.00	11:01am
ACCY CHK	.08	11:02am
AIR BLK	.00	11:03am
SUB TEST	**.00**	**11:04am**
AIR BLK	.00	11:04am
SUB TEST	**.00**	**11:06am**
AIR BLK	.00	11:07am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: *008913* Test Record Number: *1884*
Test Date: *05/30/2014* Test Time: *11:09am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:09am
FLO	Pass	11:09am
FC	Pass	11:09am

Temperature Tests

Test	Status	Time
FC1	Pass	11:09am
SRC	Pass	11:09am
DET	Pass	11:09am
BAR	Pass	11:09am
BT	Pass	11:09am

Blank Tests

Test	Status	Time
AIR	Pass	11:10am

Printer Tests

Test	Status	Time
PRNT	Pass	11:10am

CRC Tests

Test	Status	Time
COMP	Pass	11:10am
CAL	Pass	11:10am

Preventive Maintenance
Status: Pass

Dale Farley
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Alamance Instrument Location Alamance Co. Jail

Instrument Serial No. 008853 109 S. Maple St.
Graham, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

655
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II Subject Test

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: *008853*
Test Date: *05/30/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *FARLEY, CYNTHIA D*
Permit Number: *24123E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG322601
Exp Date: 08/14/2015

Test	g/210L	Time
DIAG	Pass	10:36am
AIR BLK	.00	10:36am
ACCY CHK	.08	10:37am
AIR BLK	.00	10:38am
SUB TEST	**.00**	**10:39am**
AIR BLK	.00	10:40am
SUB TEST	**.00**	**10:42am**
AIR BLK	.00	10:43am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: *008853* Test Record Number: *1413*
Test Date: *05/30/2014* Test Time: *10:44am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:45am
FLO	Pass	10:45am
FC	Pass	10:45am

Temperature Tests

Test	Status	Time
FC1	Pass	10:45am
SRC	Pass	10:45am
DET	Pass	10:45am
BAR	Pass	10:45am
BT	Pass	10:45am

Blank Tests

Test	Status	Time
AIR	Pass	10:45am

Printer Tests

Test	Status	Time
PRNT	Pass	10:45am

CRC Tests

Test	Status	Time
COMP	Pass	10:46am
CAL	Pass	10:46am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Nash Instrument Location Nash Co. Jail

Instrument Serial No. 008630 222 W. Washington St.
Nashville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Dale Farley — Signature of Certifying Official

655 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NASH COUNTY NASH COUNTY JAIL 630

Serial Number: *008630*
Test Date: *05/29/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *FARLEY, CYNTHIA D*
Permit Number: *24123E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG322601
Exp Date: 08/14/2015

Test	g/210L	Time
DIAG	Pass	2:39pm
AIR BLK	.00	2:40pm
ACCY CHK	.08	2:40pm
AIR BLK	.00	2:41pm
SUB TEST	**.00**	**2:42pm**
AIR BLK	.00	2:43pm
SUB TEST	**.00**	**2:45pm**
AIR BLK	.00	2:46pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NASH COUNTY NASH COUNTY JAIL 630

Serial Number: *008630* Test Record Number: *3026*
Test Date: *05/29/2014* Test Time: *2:47pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:47pm
FLO	Pass	2:47pm
FC	Pass	2:47pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:47pm
SRC	Pass	2:47pm
DET	Pass	2:47pm
BAR	Pass	2:47pm
BT	Pass	2:47pm

Blank Tests

Test	Status	Time
AIR	Pass	2:48pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:48pm

CRC Tests

Test	Status	Time
COMP	Pass	2:48pm
CAL	Pass	2:48pm

Preventive Maintenance
Status: Pass

Dale Farley
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Nash Instrument Location Rocky Mount P.D.

Instrument Serial No. 008740 #1 Government Plaza
Rocky Mount NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

655
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: *008740*
Test Date: *05/29/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *FARLEY, CYNTHIA D*
Permit Number: *24123E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG409709
Exp Date: 04/07/2016

Test	g/210L	Time
DIAG	Pass	11:10am
AIR BLK	.00	11:11am
ACCY CHK	.08	11:11am
AIR BLK	.00	11:12am
SUB TEST	**.00**	**11:13am**
AIR BLK	.00	11:14am
SUB TEST	**.00**	**11:16am**
AIR BLK	.00	11:17am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: *008740* Test Record Number: *501*
Test Date: *05/29/2014* Test Time: *11:18am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:19am
FLO	Pass	11:19am
FC	Pass	11:19am

Temperature Tests

Test	Status	Time
FC1	Pass	11:19am
SRC	Pass	11:19am
DET	Pass	11:19am
BAR	Pass	11:19am
BT	Pass	11:19am

Blank Tests

Test	Status	Time
AIR	Pass	11:19am

Printer Tests

Test	Status	Time
PRNT	Pass	11:20am

CRC Tests

Test	Status	Time
COMP	Pass	11:20am
CAL	Pass	11:20am

Preventive Maintenance
Status: Pass

C Dale Farley

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Wake Co. Detention Cntr.

Instrument Serial No. 008826 3301 Hammond Rd. Raleigh, N C

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

655
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WAKE COUNTY DETENTION CENTER 910
Serial Number: *008826*
Test Date: *05/28/2014*
Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *FARLEY, CYNTHIA D*
Permit Number: *24123E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG409709
Exp Date: 04/07/2016

Test	g/210L	Time
DIAG	Pass	1:46pm
AIR BLK	.00	1:47pm
ACCY CHK	.08	1:47pm
AIR BLK	.00	1:49pm
SUB TEST	**.00**	**1:50pm**
AIR BLK	.00	1:50pm
SUB TEST	**.00**	**1:52pm**
AIR BLK	.00	1:53pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY DETENTION CENTER 910
Serial Number: *008826* Test Record Number: *6960*
Test Date: *05/28/2014* Test Time: *1:54pm EDT*
System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:54pm
FLO	Pass	1:54pm
FC	Pass	1:54pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:54pm
SRC	Pass	1:54pm
DET	Pass	1:54pm
BAR	Pass	1:54pm
BT	Pass	1:54pm

Blank Tests

Test	Status	Time
AIR	Pass	1:55pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:55pm

CRC Tests

Test	Status	Time
COMP	Pass	1:55pm
CAL	Pass	1:55pm

Preventive Maintenance
Status: Pass

Dale Farley
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Apex P.S.S. #4

Instrument Serial No. 008621 1615 E Williams St.
Apex, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

655
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WAKE COUNTY APEX PD
Serial Number: *008621*
Test Date: *05/28/2014*
Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *FARLEY, CYNTHIA D*
Permit Number: *24123E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG323402
Exp Date: 08/22/2015

Test	g/210L	Time
DIAG	Pass	10:05am
AIR BLK	.00	10:06am
ACCY CHK	.08	10:06am
AIR BLK	.00	10:07am
SUB TEST	**.00**	**10:08am**
AIR BLK	.00	10:09am
SUB TEST	**.00**	**10:11am**
AIR BLK	.00	10:12am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY APEX PD
Serial Number: *008621* Test Record Number: *1582*
Test Date: *05/28/2014* Test Time: *10:15am EDT*
System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:15am
FLO	Pass	10:15am
FC	Pass	10:16am

Temperature Tests

Test	Status	Time
FC1	Pass	10:16am
SRC	Pass	10:16am
DET	Pass	10:16am
BAR	Pass	10:16am
BT	Pass	10:16am

Blank Tests

Test	Status	Time
AIR	Pass	10:16am

Printer Tests

Test	Status	Time
PRNT	Pass	10:16am

CRC Tests

Test	Status	Time
COMP	Pass	10:16am
CAL	Pass	10:16am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Nash Instrument Location Rocky Mount P.D.

Instrument Serial No. 008741 #1 Government Plaza
Rocky Mount, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

655
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: *008741*
Test Date: *05/29/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *FARLEY, CYNTHIA D*
Permit Number: *24123E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG335201
Exp Date: 12/18/2015

Test	g/210L	Time
DIAG	Pass	10:50am
AIR BLK	.00	10:51am
ACCY CHK	.07	10:52am
AIR BLK	.00	10:53am
SUB TEST	**.00**	**10:54am**
AIR BLK	.00	10:55am
SUB TEST	**.00**	**10:57am**
AIR BLK	.00	10:58am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: *008741* Test Record Number: *1473*
Test Date: *05/29/2014* Test Time: *10:59am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:00am
FLO	Pass	11:00am
FC	Pass	11:00am

Temperature Tests

Test	Status	Time
FC1	Pass	11:00am
SRC	Pass	11:00am
DET	Pass	11:00am
BAR	Pass	11:00am
BT	Pass	11:00am

Blank Tests

Test	Status	Time
AIR	Pass	11:00am

Printer Tests

Test	Status	Time
PRNT	Pass	11:01am

CRC Tests

Test	Status	Time
COMP	Pass	11:01am
CAL	Pass	11:01am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Washington Instrument Location Washington Co. S.D.

Instrument Serial No. 008829 Adams St., Plymouth, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of MAY, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WASHINGTON COUNTY SHERIFF'S OFFICE 930

Serial Number: *008829*
Test Date: *05/15/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, LINDA A*
Permit Number: *11646E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	12:05pm
AIR BLK	.00	12:06pm
ACCY CHK	.08	12:07pm
AIR BLK	.00	12:08pm
SUB TEST	**.00**	**12:09pm**
AIR BLK	.00	12:10pm
SUB TEST	**.00**	**12:11pm**
AIR BLK	.00	12:12pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Linda A Keesler
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WASHINGTON COUNTY SHERIFF'S OFFICE 930

Serial Number: *008829* Test Record Number: *556*
Test Date: *05/15/2014* Test Time: *12:13pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:14pm
FLO	Pass	12:14pm
FC	Pass	12:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:14pm
SRC	Pass	12:14pm
DET	Pass	12:14pm
BAR	Pass	12:14pm
BT	Pass	12:14pm

Blank Tests

Test	Status	Time
AIR	Pass	12:14pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:14pm

CRC Tests

Test	Status	Time
COMP	Pass	12:15pm
CAL	Pass	12:15pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Lenoir Instrument Location Lenoir Co. S.O.

Instrument Serial No. 008639 130 Queen St., Kinston, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16th day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

Certificate Number 647

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

LENOIR COUNTY LENOIR CO SO 530

Serial Number: *008639*
Test Date: *05/16/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, LINDA A*
Permit Number: *11646E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	12:37pm
AIR BLK	.00	12:38pm
ACCY CHK	.07	12:39pm
AIR BLK	.00	12:40pm
SUB TEST	**.00**	**12:40pm**
AIR BLK	.00	12:41pm
SUB TEST	**.00**	**12:43pm**
AIR BLK	.00	12:44pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Linda A. Keesler
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

LENOIR COUNTY LENOIR CO SO 530

Serial Number: *008639* Test Record Number: *2271*
Test Date: *05/16/2014* Test Time: *12:46pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:46pm
FLO	Pass	12:46pm
FC	Pass	12:46pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:47pm
SRC	Pass	12:47pm
DET	Pass	12:47pm
BAR	Pass	12:47pm
BT	Pass	12:47pm

Blank Tests

Test	Status	Time
AIR	Pass	12:47pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:47pm

CRC Tests

Test	Status	Time
COMP	Pass	12:47pm
CAL	Pass	12:47pm

Preventive Maintenance
Status: Pass

Linda A. Kaese

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Lenoir Instrument Location Kinston P.D.

Instrument Serial No. DD 8624 205 E. King ST., Kinston, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

LENOIR COUNTY KINSTON PD 530

Serial Number: *008624*
Test Date: *05/16/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, LINDA A*
Permit Number: *11646E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG400603
Exp Date: 01/06/2016

Test	g/210L	Time
DIAG	Pass	1:22pm
AIR BLK	.00	1:23pm
ACCY CHK	.07	1:23pm
AIR BLK	.00	1:24pm
SUB TEST	**.00**	**1:25pm**
AIR BLK	.00	1:26pm
SUB TEST	**.00**	**1:27pm**
AIR BLK	.00	1:28pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

LENOIR COUNTY KINSTON PD 530

Serial Number: *008624* Test Record Number: *1364*
Test Date: *05/16/2014* Test Time: *1:29pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:30pm
FLO	Pass	1:30pm
FC	Pass	1:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:30pm
SRC	Pass	1:30pm
DET	Pass	1:30pm
BAR	Pass	1:30pm
BT	Pass	1:30pm

Blank Tests

Test	Status	Time
AIR	Pass	1:30pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:31pm

CRC Tests

Test	Status	Time
COMP	Pass	1:31pm
CAL	Pass	1:31pm

Preventive Maintenance
Status: Pass

Linda Reese

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Camden Instrument Location Camden Co. S.D.

Instrument Serial No. 008940 113 Hwy 343, Camden, N.C

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CAMDEN COUNTY CAMDEN CO SO 140

Serial Number: *008940*
Test Date: *05/22/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, LINDA A*
Permit Number: *11646E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG335201
Exp Date: 12/18/2015

Test	g/210L	Time
DIAG	Pass	11:56am
AIR BLK	.00	11:57am
ACCY CHK	.07	11:58am
AIR BLK	.00	11:59am
SUB TEST	**.00**	**12:00pm**
AIR BLK	.00	12:00pm
SUB TEST	**.00**	**12:02pm**
AIR BLK	.00	12:03pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Linda A. Keesler
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CAMDEN COUNTY CAMDEN CO SO 140

Serial Number: *008940* Test Record Number: *678*
Test Date: *05/22/2014* Test Time: *12:04pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:04pm
FLO	Pass	12:04pm
FC	Pass	12:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:04pm
SRC	Pass	12:04pm
DET	Pass	12:04pm
BAR	Pass	12:04pm
BT	Pass	12:04pm

Blank Tests

Test	Status	Time
AIR	Pass	12:05pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:05pm

CRC Tests

Test	Status	Time
COMP	Pass	12:05pm
CAL	Pass	12:05pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Hertford Instrument Location Ahoskie P.D.

Instrument Serial No. 008848 705 W. Main St., Ahoskie, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

HERTFORD COUNTY AHOSKIE PD 450

Serial Number: *008848*
Test Date: *05/23/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, LINDA A*
Permit Number: *11646E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG335201
Exp Date: 12/18/2015

Test	g/210L	Time
DIAG	Pass	10:53am
AIR BLK	.00	10:54am
ACCY CHK	.08	10:55am
AIR BLK	.00	10:55am
SUB TEST	**.00**	**10:56am**
AIR BLK	.00	10:57am
SUB TEST	**.00**	**10:58am**
AIR BLK	.00	10:59am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HERTFORD COUNTY AHOSKIE PD 450

Serial Number: *008848* Test Record Number: *971*
Test Date: *05/23/2014* Test Time: *11:00am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:01am
FLO	Pass	11:01am
FC	Pass	11:01am

Temperature Tests

Test	Status	Time
FC1	Pass	11:01am
SRC	Pass	11:01am
DET	Pass	11:01am
BAR	Pass	11:01am
BT	Pass	11:01am

Blank Tests

Test	Status	Time
AIR	Pass	11:02am

Printer Tests

Test	Status	Time
PRNT	Pass	11:02am

CRC Tests

Test	Status	Time
COMP	Pass	11:02am
CAL	Pass	11:02am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Hertford Instrument Location Murfreesboro P.D.

Instrument Serial No. DD8906 115 E. Broad St., Murfreesboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of MAY, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

HERTFORD COUNTY MURFREESBORO PD 450

Serial Number: *008906*
Test Date: *05/23/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, LINDA A*
Permit Number: *11646E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG315701
Exp Date: 06/06/2015

Test	g/210L	Time
DIAG	Pass	12:00pm
AIR BLK	.00	12:01pm
ACCY CHK	.08	12:02pm
AIR BLK	.00	12:03pm
SUB TEST	**.00**	**12:04pm**
AIR BLK	.00	12:05pm
SUB TEST	**.00**	**12:06pm**
AIR BLK	.00	12:07pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HERTFORD COUNTY MURFREESBORO PD 450

Serial Number: *008906* Test Record Number: *459*
Test Date: *05/23/2014* Test Time: *12:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:08pm
FLO	Pass	12:08pm
FC	Pass	12:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:09pm
SRC	Pass	12:09pm
DET	Pass	12:09pm
BAR	Pass	12:09pm
BT	Pass	12:09pm

Blank Tests

Test	Status	Time
AIR	Pass	12:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:09pm

CRC Tests

Test	Status	Time
COMP	Pass	12:09pm
CAL	Pass	12:09pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pitt Instrument Location Pitt Co. Detention Center

Instrument Serial No. 008646 124 Detention Dr., Greenville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PITT COUNTY PITT CO DETENTION 730

Serial Number: *008646*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, LINDA A*
Permit Number: *11646E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG312802
Exp Date: 05/08/2015

Test	g/210L	Time
DIAG	Pass	12:00pm
AIR BLK	.00	12:01pm
ACCY CHK	.07	12:02pm
AIR BLK	.00	12:03pm
SUB TEST	**.00**	**12:03pm**
AIR BLK	.00	12:04pm
SUB TEST	**.00**	**12:06pm**
AIR BLK	.00	12:07pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Linda A. Keesler
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY PITT CO DETENTION 730

Serial Number: *008646* Test Record Number: *2582*
Test Date: *05/28/2014* Test Time: *12:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:08pm
FLO	Pass	12:08pm
FC	Pass	12:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:08pm
SRC	Pass	12:08pm
DET	Pass	12:08pm
BAR	Pass	12:08pm
BT	Pass	12:08pm

Blank Tests

Test	Status	Time
AIR	Pass	12:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:09pm

CRC Tests

Test	Status	Time
COMP	Pass	12:09pm
CAL	Pass	12:09pm

Preventive Maintenance
Status: Pass

Linda A. Keese
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pitt Instrument Location Pitt Co. Detention Center

Instrument Serial No. 008608 124 Detention Dr., Greenville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of MAY, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

047
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PITT COUNTY PITT CO DETENTION 730

Serial Number: *008668*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, LINDA A*
Permit Number: *11646E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG312802
Exp Date: 05/08/2015

Test	g/210L	Time
DIAG	Pass	12:09pm
AIR BLK	.00	12:10pm
ACCY CHK	.08	12:10pm
AIR BLK	.00	12:12pm
SUB TEST	**.00**	**12:13pm**
AIR BLK	.00	12:14pm
SUB TEST	**.00**	**12:17pm**
AIR BLK	.00	12:18pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY PITT CO DETENTION 730

Serial Number: *008668* Test Record Number: *2344*
Test Date: *05/28/2014* Test Time: *12:19pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:19pm
FLO	Pass	12:19pm
FC	Pass	12:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:19pm
SRC	Pass	12:19pm
DET	Pass	12:19pm
BAR	Pass	12:19pm
BT	Pass	12:19pm

Blank Tests

Test	Status	Time
AIR	Pass	12:20pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:20pm

CRC Tests

Test	Status	Time
COMP	Pass	12:20pm
CAL	Pass	12:20pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pitt Instrument Location Pitt Co. Detention Center

Instrument Serial No. 008662 124 Detention Dr., Greenville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PITT COUNTY PITT CO DETENTION 730

Serial Number: *008662*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, LINDA A*
Permit Number: *11646E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG400603
Exp Date: 01/06/2016

Test	g/210L	Time
DIAG	Pass	12:17pm
AIR BLK	.00	12:18pm
ACCY CHK	.07	12:18pm
AIR BLK	.00	12:19pm
SUB TEST	**.00**	**12:21pm**
AIR BLK	.00	12:21pm
SUB TEST	**.00**	**12:25pm**
AIR BLK	.00	12:26pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY PITT CO DETENTION 730

Serial Number: *008662* Test Record Number: *817*
Test Date: *05/28/2014* Test Time: *12:27pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:27pm
FLO	Pass	12:27pm
FC	Pass	12:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:28pm
SRC	Pass	12:28pm
DET	Pass	12:28pm
BAR	Pass	12:28pm
BT	Pass	12:28pm

Blank Tests

Test	Status	Time
AIR	Pass	12:28pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:28pm

CRC Tests

Test	Status	Time
COMP	Pass	12:28pm
CAL	Pass	12:28pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Moore Instrument Location Bat mobile Unit 2

Instrument Serial No. DD8929 Moore Co SO

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Sonya B Skinner
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MOORE COUNTY BAT MOBILE UNIT 2 620

Serial Number: *008929*
Test Date: *05/31/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Female*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SKINNER, TONYA B*
Permit Number: *13651E*
Effective:
10/01/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG402703
Exp Date: 01/27/2016

Test	g/210L	Time
DIAG	Pass	9:56pm
AIR BLK	.00	9:57pm
ACCY CHK	.07	9:58pm
AIR BLK	.00	9:58pm
SUB TEST	**.00**	**10:00pm**
AIR BLK	.00	10:01pm
SUB TEST	**.00**	**10:03pm**
AIR BLK	.00	10:04pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Tonya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MOORE COUNTY BAT MOBILE UNIT 2 620

Serial Number: *008929* Test Record Number: *723*
Test Date: *05/31/2014* Test Time: *10:05pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:05pm
FLO	Pass	10:05pm
FC	Pass	10:05pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:05pm
SRC	Pass	10:05pm
DET	Pass	10:05pm
BAR	Pass	10:05pm
BT	Pass	10:05pm

Blank Tests

Test	Status	Time
AIR	Pass	10:06pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:06pm

CRC Tests

Test	Status	Time
COMP	Pass	10:06pm
CAL	Pass	10:06pm

Preventive Maintenance
Status: Pass

Donya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Moore Instrument Location Bat Mobile Unit 2

Instrument Serial No 008601 Moore Co SO

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Donya B Skinner
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MOORE COUNTY BATMOBILE UNIT 2 620

Serial Number: *008601*
Test Date: *05/31/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Female*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SKINNER, TONYA B*
Permit Number: *13651E*
Effective:
10/01/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG402703
Exp Date: 01/27/2016

Test	g/210L	Time
DIAG	Pass	10:13pm
AIR BLK	.00	10:14pm
ACCY CHK	.07	10:15pm
AIR BLK	.00	10:16pm
SUB TEST	**.00**	**10:16pm**
AIR BLK	.00	10:17pm
SUB TEST	**.00**	**10:19pm**
AIR BLK	.00	10:19pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Tonya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MOORE COUNTY BATMOBILE UNIT 2 620

Serial Number: *008601* Test Record Number: *905*
Test Date: *05/31/2014* Test Time: *10:22pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:22pm
FLO	Pass	10:22pm
FC	Pass	10:22pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:22pm
SRC	Pass	10:22pm
DET	Pass	10:22pm
BAR	Pass	10:22pm
BT	Pass	10:22pm

Blank Tests

Test	Status	Time
AIR	Pass	10:23pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:23pm

CRC Tests

Test	Status	Time
COMP	Pass	10:23pm
CAL	Pass	10:23pm

Preventive Maintenance
Status: Pass

Donya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Harnett Instrument Location Bat Mobile Unit 2

Instrument Serial No 008929 Coats PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Donna B Skinner
Signature of Certifying Official

6441
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

HARNETT COUNTY BAT MOBILE UNIT 2 420

Serial Number: *008929*
Test Date: *05/30/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Female*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SKINNER, TONYA B*
Permit Number: *13651E*
Effective:
10/01/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG402703
Exp Date: 01/27/2016

Test	g/210L	Time
DIAG	Pass	10:12pm
AIR BLK	.00	10:13pm
ACCY CHK	.07	10:14pm
AIR BLK	.00	10:15pm
SUB TEST	**.00**	**10:15pm**
AIR BLK	.00	10:16pm
SUB TEST	**.00**	**10:18pm**
AIR BLK	.00	10:19pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Tonya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HARNETT COUNTY BAT MOBILE UNIT 2 420

Serial Number: *008929* Test Record Number: *720*
Test Date: *05/30/2014* Test Time: *10:24pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:25pm
FLO	Pass	10:25pm
FC	Pass	10:25pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:25pm
SRC	Pass	10:25pm
DET	Pass	10:25pm
BAR	Pass	10:25pm
BT	Pass	10:25pm

Blank Tests

Test	Status	Time
AIR	Pass	10:26pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:26pm

CRC Tests

Test	Status	Time
COMP	Pass	10:26pm
CAL	Pass	10:26pm

Preventive Maintenance
Status: Pass

Donya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mitchell Instrument Location Spruce Pine PD

Instrument Serial No. 008226 Spruce Pine

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official 649 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: *008726*
Test Date: *05/09/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BURNETTE, ANTHONY J*
Permit Number: *11304E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309105
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	11:17am
AIR BLK	.00	11:18am
ACCY CHK	.08	11:19am
AIR BLK	.00	11:20am
SUB TEST	**.00**	**11:20am**
AIR BLK	.00	11:21am
SUB TEST	**.00**	**11:23am**
AIR BLK	.00	11:24am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: *008726* Test Record Number: *672*
Test Date: *05/09/2014* Test Time: *11:25am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:25am
FLO	Pass	11:25am
FC	Pass	11:25am

Temperature Tests

Test	Status	Time
FC1	Pass	11:25am
SRC	Pass	11:25am
DET	Pass	11:25am
BAR	Pass	11:25am
BT	Pass	11:25am

Blank Tests

Test	Status	Time
AIR	Pass	11:26am

Printer Tests

Test	Status	Time
PRNT	Pass	11:26am

CRC Tests

Test	Status	Time
COMP	Pass	11:26am
CAL	Pass	11:26am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Bertie Instrument Location Bertie Co. S.O.

Instrument Serial No. 008897 104 Dundee St., Windsor, N.C

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Kelly [illegible]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BERTIE COUNTY BERTIE CO SO 070

Serial Number: *008897*
Test Date: *05/22/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG335201
Exp Date: 12/18/2015

Test	g/210L	Time
DIAG	Pass	12:03pm
AIR BLK	.00	12:03pm
ACCY CHK	.08	12:04pm
AIR BLK	.00	12:05pm
SUB TEST	**.00**	**12:06pm**
AIR BLK	.00	12:07pm
SUB TEST	**.00**	**12:08pm**
AIR BLK	.00	12:09pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BERTIE COUNTY BERTIE CO SO 070

Serial Number: *008897* Test Record Number: *854*
Test Date: *05/22/2014* Test Time: *12:11pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:11pm
FLO	Pass	12:11pm
FC	Pass	12:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:11pm
SRC	Pass	12:11pm
DET	Pass	12:11pm
BAR	Pass	12:11pm
BT	Pass	12:11pm

Blank Tests

Test	Status	Time
AIR	Pass	12:12pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:12pm

CRC Tests

Test	Status	Time
COMP	Pass	12:12pm
CAL	Pass	12:12pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wayne Instrument Location Wayne Co. Detention Center

Instrument Serial No. 008649 207 E. Chestnut St, Goldsboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Kelly M
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: *008649*
Test Date: *05/27/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG315701
Exp Date: 06/06/2015

Test	g/210L	Time
DIAG	Pass	10:55am
AIR BLK	.00	10:56am
ACCY CHK	.08	10:57am
AIR BLK	.00	10:58am
SUB TEST	**.00**	**10:59am**
AIR BLK	.00	11:00am
SUB TEST	**.00**	**11:01am**
AIR BLK	.00	11:02am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: *008649* Test Record Number: *2535*
Test Date: *05/27/2014* Test Time: *11:03am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:04am
FLO	Pass	11:04am
FC	Pass	11:04am

Temperature Tests

Test	Status	Time
FC1	Pass	11:04am
SRC	Pass	11:04am
DET	Pass	11:04am
BAR	Pass	11:04am
BT	Pass	11:04am

Blank Tests

Test	Status	Time
AIR	Pass	11:05am

Printer Tests

Test	Status	Time
PRNT	Pass	11:05am

CRC Tests

Test	Status	Time
COMP	Pass	11:05am
CAL	Pass	11:05am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Beaufort Instrument Location Belhaven Police Dept.

Instrument Serial No. 008928 Belhaven, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

6413
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BEAUFORT COUNTY BELHAVEN PD 060

Serial Number: *008928*
Test Date: *05/29/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG400603
Exp Date: 01/06/2016

Test	g/210L	Time
DIAG	Pass	10:39am
AIR BLK	.00	10:40am
ACCY CHK	.08	10:40am
AIR BLK	.00	10:41am
SUB TEST	**.00**	**10:42am**
AIR BLK	.00	10:43am
SUB TEST	**.00**	**10:44am**
AIR BLK	.00	10:45am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BEAUFORT COUNTY BELHAVEN PD 060

Serial Number: *008928* Test Record Number: *227*
Test Date: *05/29/2014* Test Time: *10:46am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:46am
FLO	Pass	10:46am
FC	Pass	10:46am

Temperature Tests

Test	Status	Time
FC1	Pass	10:46am
SRC	Pass	10:46am
DET	Pass	10:46am
BAR	Pass	10:46am
BT	Pass	10:46am

Blank Tests

Test	Status	Time
AIR	Pass	10:47am

Printer Tests

Test	Status	Time
PRNT	Pass	10:47am

CRC Tests

Test	Status	Time
COMP	Pass	10:47am
CAL	Pass	10:47am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Martin Instrument Location Martin Co. S.O.

Instrument Serial No. D08912 305 E. Main St, Williamston, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Kelly [illegible]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MARTIN COUNTY SHERIFF'S OFFICE 570

Serial Number: *008912*
Test Date: *05/22/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309105
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	1:21pm
AIR BLK	.00	1:21pm
ACCY CHK	.08	1:22pm
AIR BLK	.00	1:23pm
SUB TEST	**.00**	**1:24pm**
AIR BLK	.00	1:25pm
SUB TEST	**.00**	**1:26pm**
AIR BLK	.00	1:27pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MARTIN COUNTY SHERIFF'S OFFICE 570

Serial Number: *008912* Test Record Number: *835*
Test Date: *05/22/2014* Test Time: *1:29pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:30pm
FLO	Pass	1:30pm
FC	Pass	1:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:30pm
SRC	Pass	1:30pm
DET	Pass	1:30pm
BAR	Pass	1:30pm
BT	Pass	1:30pm

Blank Tests

Test	Status	Time
AIR	Pass	1:30pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:30pm

CRC Tests

Test	Status	Time
COMP	Pass	1:31pm
CAL	Pass	1:31pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT MOBILE UNIT 7

Instrument Serial No. 008704 Holly Springs

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008704* Test Record Number: *265*
Test Date: *05/30/2014* Test Time: *10:20pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:20pm
FLO	Pass	10:20pm
FC	Pass	10:20pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:20pm
SRC	Pass	10:20pm
DET	Pass	10:20pm
BAR	Pass	10:20pm
BT	Pass	10:20pm

Blank Tests

Test	Status	Time
AIR	Pass	10:21pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:21pm

CRC Tests

Test	Status	Time
COMP	Pass	10:21pm
CAL	Pass	10:21pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008704*
Test Date: *05/30/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG308002
Exp Date: 03/21/2015

Test	g/210L	Time
DIAG	Pass	10:08pm
AIR BLK	.00	10:09pm
ACCY CHK	.08	10:09pm
AIR BLK	.00	10:10pm
SUB TEST	**.00**	**10:11pm**
AIR BLK	.00	10:12pm
SUB TEST	**.00**	**10:14pm**
AIR BLK	.00	10:15pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Bat Mobile Unit #7

Instrument Serial No. 008778 Holly Springs

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008778* Test Record Number: *1152*
Test Date: *05/30/2014* Test Time: *10:13pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:13pm
FLO	Pass	10:13pm
FC	Pass	10:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:14pm
SRC	Pass	10:14pm
DET	Pass	10:14pm
BAR	Pass	10:14pm
BT	Pass	10:14pm

Blank Tests

Test	Status	Time
AIR	Pass	10:14pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:14pm

CRC Tests

Test	Status	Time
COMP	Pass	10:15pm
CAL	Pass	10:15pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008778*
Test Date: *05/30/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG404101
Exp Date: 02/10/2016

Test	g/210L	Time
DIAG	Pass	10:05pm
AIR BLK	.00	10:06pm
ACCY CHK	.07	10:06pm
AIR BLK	.00	10:07pm
SUB TEST	**.00**	**10:08pm**
AIR BLK	.00	10:09pm
SUB TEST	**.00**	**10:11pm**
AIR BLK	.00	10:12pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT Mobile Unit #7

Instrument Serial No. 008760 Holly Springs

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official 626 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008760* Test Record Number: *613*
Test Date: *05/30/2014* Test Time: *10:15pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:15pm
FLO	Pass	10:15pm
FC	Pass	10:15pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:15pm
SRC	Pass	10:15pm
DET	Pass	10:15pm
BAR	Pass	10:15pm
BT	Pass	10:15pm

Blank Tests

Test	Status	Time
AIR	Pass	10:16pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:16pm

CRC Tests

Test	Status	Time
COMP	Pass	10:16pm
CAL	Pass	10:16pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008760*
Test Date: *05/30/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	10:06pm
AIR BLK	.00	10:07pm
ACCY CHK	.08	10:08pm
AIR BLK	.00	10:09pm
SUB TEST	**.00**	**10:09pm**
AIR BLK	.00	10:10pm
SUB TEST	**.00**	**10:12pm**
AIR BLK	.00	10:13pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location BAT MOBILE UNIT #7

Instrument Serial No. 008612 Holly Springs

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

634
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008612* Test Record Number: *1496*
Test Date: *05/30/2014* Test Time: *11:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:08pm
FLO	Pass	11:08pm
FC	Pass	11:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:08pm
SRC	Pass	11:08pm
DET	Pass	11:08pm
BAR	Pass	11:08pm
BT	Pass	11:08pm

Blank Tests

Test	Status	Time
AIR	Pass	11:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:09pm

CRC Tests

Test	Status	Time
COMP	Pass	11:09pm
CAL	Pass	11:09pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008612*
Test Date: *05/30/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG303502
Exp Date: 02/04/2015

Test	g/210L	Time
DIAG	Pass	11:00pm
AIR BLK	.00	11:01pm
ACCY CHK	.07	11:01pm
AIR BLK	.00	11:02pm
SUB TEST	**.00**	**11:03pm**
AIR BLK	.00	11:04pm
SUB TEST	**.00**	**11:05pm**
AIR BLK	.00	11:06pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County RANDOLPH Instrument Location LIBERTY Police DEPT.

Instrument Serial No. 008830 LIBERTY N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

RANDOLPH LIBERTY POLICE DEPT 750

Serial Number: *008830*
Test Date: *05/27/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309101
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	9:53am
AIR BLK	.00	9:54am
ACCY CHK	.08	9:55am
AIR BLK	.00	9:56am
SUB TEST	**.00**	**9:56am**
AIR BLK	.00	9:57am
SUB TEST	**.00**	**9:59am**
AIR BLK	.00	9:59am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

RANDOLPH LIBERTY POLICE DEPT 750

Serial Number: *008830* Test Record Number: *454*
Test Date: *05/27/2014* Test Time: *10:00am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:01am
FLO	Pass	10:01am
FC	Pass	10:01am

Temperature Tests

Test	Status	Time
FC1	Pass	10:01am
SRC	Pass	10:01am
DET	Pass	10:01am
BAR	Pass	10:01am
BT	Pass	10:01am

Blank Tests

Test	Status	Time
AIR	Pass	10:01am

Printer Tests

Test	Status	Time
PRNT	Pass	10:01am

CRC Tests

Test	Status	Time
COMP	Pass	10:02am
CAL	Pass	10:02am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County DAVIDSON Instrument Location DAVIDSON Co. JAIL

Instrument Serial No. 008845 LEXINGTON, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of MAY, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Leon
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY DAVIDSON CO JAIL 280

Serial Number: *008845*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG320602
Exp Date: 07/25/2015

Test	g/210L	Time
DIAG	Pass	4:23pm
AIR BLK	.00	4:24pm
ACCY CHK	.07	4:25pm
AIR BLK	.00	4:26pm
SUB TEST	**.00**	**4:27pm**
AIR BLK	.00	4:28pm
SUB TEST	**.00**	**4:29pm**
AIR BLK	.00	4:30pm

Reported AC: .00 g/210L

L. Kevin Dean
Signature of Chemical Analyst

Court CVR

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY DAVIDSON CO JAIL 280

Serial Number: *008845* Test Record Number: *1686*
Test Date: *05/28/2014* Test Time: *4:32pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:32pm
FLO	Pass	4:32pm
FC	Pass	4:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:33pm
SRC	Pass	4:33pm
DET	Pass	4:33pm
BAR	Pass	4:33pm
BT	Pass	4:33pm

Blank Tests

Test	Status	Time
AIR	Pass	4:33pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:33pm

CRC Tests

Test	Status	Time
COMP	Pass	4:33pm
CAL	Pass	4:33pm

Preventive Maintenance
Status: Pass

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Davidson Instrument Location Lexington Police Department

Instrument Serial No. 008883

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

THE GREAT SEAL OF THE STATE OF NORTH CAROLINA MAY 20, 1775 APRIL 12, 1776 ESSE QUAM VIDERI

L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY LEXINGTON PD 280

Serial Number: *008883*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG320602
Exp Date: 07/25/2015

Test	g/210L	Time
DIAG	Pass	3:45pm
AIR BLK	.00	3:46pm
ACCY CHK	.08	3:46pm
AIR BLK	.00	3:48pm
SUB TEST	**.00**	**3:49pm**
AIR BLK	.00	3:50pm
SUB TEST	**.00**	**3:51pm**
AIR BLK	.00	3:52pm

Reported AC: .00 g/210L

L. Kevin Dean
Signature of Chemical Analyst

Court CVR

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY LEXINGTON PD 280

Serial Number: *008883* Test Record Number: *1274*
Test Date: *05/28/2014* Test Time: *3:53pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:53pm
FLO	Pass	3:53pm
FC	Pass	3:53pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:54pm
SRC	Pass	3:54pm
DET	Pass	3:54pm
BAR	Pass	3:54pm
BT	Pass	3:54pm

Blank Tests

Test	Status	Time
AIR	Pass	3:54pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:54pm

CRC Tests

Test	Status	Time
COMP	Pass	3:54pm
CAL	Pass	3:54pm

Preventive Maintenance
Status: Pass

L. Kevin Leon
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Davidson Instrument Location Thomasville Police Department

Instrument Serial No. 008872

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY THOMASVILLE PD 280

Serial Number: *008872*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG320602
Exp Date: 07/25/2015

Test	g/210L	Time
DIAG	Pass	2:19pm
AIR BLK	.00	2:19pm
ACCY CHK	.07	2:20pm
AIR BLK	.00	2:21pm
SUB TEST	**.00**	**2:22pm**
AIR BLK	.00	2:22pm
SUB TEST	**.00**	**2:24pm**
AIR BLK	.00	2:25pm

Reported AC: .00 g/210L

L. Kevin Dean
Signature of Chemical Analyst

Court CVR

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY THOMASVILLE PD 280

Serial Number: *008872* Test Record Number: *1102*
Test Date: *05/28/2014* Test Time: *2:25pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:26pm
FLO	Pass	2:26pm
FC	Pass	2:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:26pm
SRC	Pass	2:26pm
DET	Pass	2:26pm
BAR	Pass	2:26pm
BT	Pass	2:26pm

Blank Tests

Test	Status	Time
AIR	Pass	2:26pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:26pm

CRC Tests

Test	Status	Time
COMP	Pass	2:27pm
CAL	Pass	2:27pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location Greensboro Jail

Instrument Serial No. 008790

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: *008790*
Test Date: *05/27/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG315701
Exp Date: 06/06/2015

Test	g/210L	Time
DIAG	Pass	3:06pm
AIR BLK	.00	3:06pm
ACCY CHK	.07	3:07pm
AIR BLK	.00	3:08pm
SUB TEST	**.00**	**3:09pm**
AIR BLK	.00	3:10pm
SUB TEST	**.00**	**3:11pm**
AIR BLK	.00	3:12pm

Reported AC: .00 g/210L

L. Kevin Dean
Signature of Chemical Analyst

Court CVR

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: *008790* Test Record Number: *4283*
Test Date: *05/27/2014* Test Time: *3:13pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:13pm
FLO	Pass	3:13pm
FC	Pass	3:13pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:13pm
SRC	Pass	3:13pm
DET	Pass	3:13pm
BAR	Pass	3:13pm
BT	Pass	3:13pm

Blank Tests

Test	Status	Time
AIR	Pass	3:14pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:14pm

CRC Tests

Test	Status	Time
COMP	Pass	3:14pm
CAL	Pass	3:14pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Bat Mobile Unit 2

Instrument Serial No. 008929 Garner PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

1044
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 2 910

Serial Number: *008929*
Test Date: *05/16/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Female*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SKINNER, TONYA B*
Permit Number: *13651E*
Effective:
10/01/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG402703
Exp Date: 01/27/2016

Test	g/210L	Time
DIAG	Pass	9:51pm
AIR BLK	.00	9:52pm
ACCY CHK	.07	9:53pm
AIR BLK	.00	9:54pm
SUB TEST	**.00**	**9:54pm**
AIR BLK	.00	9:55pm
SUB TEST	**.00**	**9:57pm**
AIR BLK	.00	9:58pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Tonya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 2 910

Serial Number: *008929* Test Record Number: *706*
Test Date: *05/16/2014* Test Time: *9:59pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:00pm
FLO	Pass	10:00pm
FC	Pass	10:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:00pm
SRC	Pass	10:00pm
DET	Pass	10:00pm
BAR	Pass	10:00pm
BT	Pass	10:00pm

Blank Tests

Test	Status	Time
AIR	Pass	10:01pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:01pm

CRC Tests

Test	Status	Time
COMP	Pass	10:01pm
CAL	Pass	10:01pm

Preventive Maintenance
Status: Pass

Donya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Bat Mobile Unit 2

Instrument Serial No. 008601 Garner PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

441
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WAKE COUNTY BATMOBILE UNIT 2 910

Serial Number: *008601*
Test Date: *05/16/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Female*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SKINNER, TONYA B*
Permit Number: *13651E*
Effective:
10/01/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG402703
Exp Date: 01/27/2016

Test	g/210L	Time
DIAG	Pass	9:54pm
AIR BLK	.00	9:55pm
ACCY CHK	.07	9:55pm
AIR BLK	.00	9:56pm
SUB TEST	**.00**	**9:57pm**
AIR BLK	.00	9:58pm
SUB TEST	**.00**	**9:59pm**
AIR BLK	.00	10:00pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Tonya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BATMOBILE UNIT 2 910

Serial Number: *008601* Test Record Number: *897*
Test Date: *05/16/2014* Test Time: *10:01pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:02pm
FLO	Pass	10:02pm
FC	Pass	10:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:02pm
SRC	Pass	10:02pm
DET	Pass	10:02pm
BAR	Pass	10:02pm
BT	Pass	10:02pm

Blank Tests

Test	Status	Time
AIR	Pass	10:02pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:02pm

CRC Tests

Test	Status	Time
COMP	Pass	10:03pm
CAL	Pass	10:03pm

Preventive Maintenance
Status: Pass

Tonya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Bat Mobile Unit 2

Instrument Serial No 008736 Garner PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

1011
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 2 910

Serial Number: *008736*
Test Date: *05/16/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Female*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SKINNER, TONYA B*
Permit Number: *13651E*
Effective:
10/01/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG409709
Exp Date: 04/07/2016

Test	g/210L	Time
DIAG	Pass	10:08pm
AIR BLK	.00	10:10pm
ACCY CHK	.08	10:10pm
AIR BLK	.00	10:11pm
SUB TEST	**.00**	**10:12pm**
AIR BLK	.00	10:13pm
SUB TEST	**.00**	**10:14pm**
AIR BLK	.00	10:15pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 2 910

Serial Number: *008736* Test Record Number: *693*
Test Date: *05/16/2014* Test Time: *10:24pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:25pm
FLO	Pass	10:25pm
FC	Pass	10:25pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:25pm
SRC	Pass	10:25pm
DET	Pass	10:25pm
BAR	Pass	10:25pm
BT	Pass	10:25pm

Blank Tests

Test	Status	Time
AIR	Pass	10:26pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:26pm

CRC Tests

Test	Status	Time
COMP	Pass	10:26pm
CAL	Pass	10:26pm

Preventive Maintenance
Status: Pass

Donya B Skinner
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Lincoln Instrument Location Lincoln County Courthouse

Instrument Serial No. 008823 #1 Courthouse Square, Lincolnton

704-732-9020

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Joseph E Hutch
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

LINCOLN COUNTY COURTHOUSE 540

Serial Number: *008823*
Test Date: *05/23/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective:
10/01/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG404101
Exp Date: 02/10/2016

Test	g/210L	Time
DIAG	Pass	10:06am
AIR BLK	.00	10:07am
ACCY CHK	.07	10:08am
AIR BLK	.00	10:09am
SUB TEST	**.00**	**10:09am**
AIR BLK	.00	10:11am
SUB TEST	**.00**	**10:12am**
AIR BLK	.00	10:13am

Reported AC: .00 g/210L

J. E. Hutch

Signature of Chemical Analyst

Court CVR

J. E. Hutch

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

LINCOLN COUNTY COURTHOUSE 540

Serial Number: *008823* Test Record Number: *1100*
Test Date: *05/23/2014* Test Time: *10:14am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:15am
FLO	Pass	10:15am
FC	Pass	10:15am

Temperature Tests

Test	Status	Time
FC1	Pass	10:15am
SRC	Pass	10:15am
DET	Pass	10:15am
BAR	Pass	10:15am
BT	Pass	10:15am

Blank Tests

Test	Status	Time
AIR	Pass	10:15am

Printer Tests

Test	Status	Time
PRNT	Pass	10:15am

CRC Tests

Test	Status	Time
COMP	Pass	10:16am
CAL	Pass	10:16am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Pineville PD

Instrument Serial No. 008703 427 Main Street, Pineville
704-889-2231

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY PINEVILLE PD 590

Serial Number: *008703*
Test Date: *05/21/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective:
10/01/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG335201
Exp Date: 12/18/2015

Test	g/210L	Time
DIAG	Pass	11:04am
AIR BLK	.00	11:05am
ACCY CHK	.08	11:06am
AIR BLK	.00	11:07am
SUB TEST	**.00**	**11:07am**
AIR BLK	.00	11:08am
SUB TEST	**.00**	**11:10am**
AIR BLK	.00	11:11am

Reported AC: .00 g/210L

J. E. Hutch

Signature of Chemical Analyst

Court CVR

J. E. Hutch

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive M ntenance

MECKLENBURG COUNTY PINEVIL PD 590

Serial Number: *008703* Test Rec d Number: *5238*
Test Date: *05/21/2014* Test Ti : *11:12am EDT*

System Check: *Passe*

Baseline Tests

Test	Status	Ti
IR	Pass	11 2am
FLO	Pass	11 2am
FC	Pass	11 2am

Temperature Tests

Test	Status	Ti
FC1	Pass	11 2am
SRC	Pass	11 2am
DET	Pass	11 2am
BAR	Pass	11 2am
BT	Pass	11 2am

Blank Tests

Test	Status	Ti
AIR	Pass	11 3am

Printer Tests

Test	Status	Ti
PRNT	Pass	11 3am

CRC Tests

Test	Status	Ti
COMP	Pass	11 3am
CAL	Pass	11 3am

Preventive Maintenan
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Matthews PD

Instrument Serial No. 008699 1201 Crews Road, Matthews
704-847-4069

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Joseph E [illegible] 650

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY MATTHEWS PD 590

Serial Number: *008699*
Test Date: *05/21/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective:
10/01/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	10:19am
AIR BLK	.00	10:20am
ACCY CHK	.08	10:21am
AIR BLK	.00	10:22am
SUB TEST	**.00**	**10:22am**
AIR BLK	.00	10:23am
SUB TEST	**.00**	**10:25am**
AIR BLK	.00	10:26am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY MATTHEWS PD 590

Serial Number: *008699* Test Record Number: *2142*
Test Date: *05/21/2014* Test Time: *10:27am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:27am
FLO	Pass	10:27am
FC	Pass	10:27am

Temperature Tests

Test	Status	Time
FC1	Pass	10:27am
SRC	Pass	10:27am
DET	Pass	10:27am
BAR	Pass	10:27am
BT	Pass	10:27am

Blank Tests

Test	Status	Time
AIR	Pass	10:28am

Printer Tests

Test	Status	Time
PRNT	Pass	10:28am

CRC Tests

Test	Status	Time
COMP	Pass	10:28am
CAL	Pass	10:28am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cleveland Instrument Location Cleveland County SD - Annex

Instrument Serial No. 008887 407 McBrayer Street, Shelby
704-484-4888

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20th day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Joseph E. Hunter
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CLEVELAND COUNTY CLEVELAND SD-ANNEX 220

Serial Number: *008887*
Test Date: *05/20/2014*

Citation Number: *M0000000-0*
Subject's Name: *PREVENTIVE, MAINTENANCE*
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective: *10/01/2013-10/01/2015*

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	11:24am
AIR BLK	.00	11:24am
ACCY CHK	.08	11:25am
AIR BLK	.00	11:26am
SUB TEST	**.00**	**11:27am**
AIR BLK	.00	11:28am
SUB TEST	**.00**	**11:29am**
AIR BLK	.00	11:30am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CLEVELAND COUNTY CLEVELAND SD-ANNEX 220

Serial Number: *008887* Test Record Number: *1718*
Test Date: *05/20/2014* Test Time: *11:31am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:32am
FLO	Pass	11:32am
FC	Pass	11:32am

Temperature Tests

Test	Status	Time
FC1	Pass	11:32am
SRC	Pass	11:32am
DET	Pass	11:32am
BAR	Pass	11:32am
BT	Pass	11:32am

Blank Tests

Test	Status	Time
AIR	Pass	11:32am

Printer Tests

Test	Status	Time
PRNT	Pass	11:33am

CRC Tests

Test	Status	Time
COMP	Pass	11:33am
CAL	Pass	11:33am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cleveland Instrument Location Cleveland County SD

Instrument Serial No. 008893 100 Justice Place, Shelby
704-484-4888

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Joseph E. Hutchins
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CLEVELAND COUNTY CLEVELAND COUNTY SD 220

Serial Number: *008893*
Test Date: *05/20/2014*

Citation Number: *M0000000-0*
Subject's Name: *PREVENTIVE, MAINTENANCE*
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective: *10/01/2013-10/01/2015*

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG320602
Exp Date: 07/25/2015

Test	g/210L	Time
DIAG	Pass	11:00am
AIR BLK	.00	11:01am
ACCY CHK	.08	11:01am
AIR BLK	.00	11:03am
SUB TEST	**.00**	**11:03am**
AIR BLK	.00	11:04am
SUB TEST	**.00**	**11:05am**
AIR BLK	.00	11:06am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CLEVELAND COUNTY CLEVELAND COUNTY SD 220

Serial Number: *008893* Test Record Number: *1230*
Test Date: *05/20/2014* Test Time: *11:07am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:08am
FLO	Pass	11:08am
FC	Pass	11:08am

Temperature Tests

Test	Status	Time
FC1	Pass	11:08am
SRC	Pass	11:08am
DET	Pass	11:08am
BAR	Pass	11:08am
BT	Pass	11:08am

Blank Tests

Test	Status	Time
AIR	Pass	11:09am

Printer Tests

Test	Status	Time
PRNT	Pass	11:09am

CRC Tests

Test	Status	Time
COMP	Pass	11:09am
CAL	Pass	11:09am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Union Instrument Location Union County SD

Instrument Serial No. 008876 3344 Presson Road, Monroe
704-283-3770

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Joseph E. Hutchins
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

UNION COUNTY UNION COUNTY SD 890

Serial Number: *008876*
Test Date: *05/11/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective:
10/01/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG322601
Exp Date: 08/14/2015

Test	g/210L	Time
DIAG	Pass	4:05pm
AIR BLK	.00	4:06pm
ACCY CHK	.08	4:07pm
AIR BLK	.00	4:08pm
SUB TEST	**.00**	**4:09pm**
AIR BLK	.00	4:10pm
SUB TEST	**.00**	**4:11pm**
AIR BLK	.00	4:12pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

UNION COUNTY UNION COUNTY SD 890

Serial Number: *008876* Test Record Number: *3332*
Test Date: *05/11/2014* Test Time: *4:13pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:14pm
FLO	Pass	4:14pm
FC	Pass	4:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:14pm
SRC	Pass	4:14pm
DET	Pass	4:14pm
BAR	Pass	4:14pm
BT	Pass	4:14pm

Blank Tests

Test	Status	Time
AIR	Pass	4:15pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:15pm

CRC Tests

Test	Status	Time
COMP	Pass	4:15pm
CAL	Pass	4:15pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Cornelius PD

Instrument Serial No. 008692 21440 Catawba Avenue, Cornelius
704-892-1363

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11th day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Joseph E. Hutchins
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY CORNELIUS PD 590

Serial Number: *008692*
Test Date: *05/11/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HUTCHINSON, JOSEPH E*
Permit Number: *19951E*
Effective:
10/01/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG315701
Exp Date: 06/06/2015

Test	g/210L	Time
DIAG	Pass	2:32pm
AIR BLK	.00	2:33pm
ACCY CHK	.08	2:34pm
AIR BLK	.00	2:35pm
SUB TEST	**.00**	**2:36pm**
AIR BLK	.00	2:36pm
SUB TEST	**.00**	**2:38pm**
AIR BLK	.00	2:39pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY CORNELIUS PD 590

Serial Number: *008692* Test Record Number: *2212*
Test Date: *05/11/2014* Test Time: *2:40pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:40pm
FLO	Pass	2:40pm
FC	Pass	2:41pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:41pm
SRC	Pass	2:41pm
DET	Pass	2:41pm
BAR	Pass	2:41pm
BT	Pass	2:41pm

Blank Tests

Test	Status	Time
AIR	Pass	2:41pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:41pm

CRC Tests

Test	Status	Time
COMP	Pass	2:41pm
CAL	Pass	2:41pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Greene Instrument Location Greene Co. S.O.

Instrument Serial No. DU 8670 301 N. Greene St., Snow Hill, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GREENE COUNTY GREENE CO SO 390

Serial Number: *008670*
Test Date: *05/19/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG315701
Exp Date: 06/06/2015

Test	g/210L	Time
DIAG	Pass	10:43am
AIR BLK	.00	10:43am
ACCY CHK	.07	10:44am
AIR BLK	.00	10:45am
SUB TEST	**.00**	**10:46am**
AIR BLK	.00	10:47am
SUB TEST	**.00**	**10:48am**
AIR BLK	.00	10:49am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GREENE COUNTY GREENE CO SO 390

Serial Number: *008670* Test Record Number: *1423*
Test Date: *05/19/2014* Test Time: *10:50am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:50am
FLO	Pass	10:50am
FC	Pass	10:51am

Temperature Tests

Test	Status	Time
FC1	Pass	10:51am
SRC	Pass	10:51am
DET	Pass	10:51am
BAR	Pass	10:51am
BT	Pass	10:51am

Blank Tests

Test	Status	Time
AIR	Pass	10:51am

Printer Tests

Test	Status	Time
PRNT	Pass	10:51am

CRC Tests

Test	Status	Time
COMP	Pass	10:52am
CAL	Pass	10:52am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pitt Instrument Location Ayden P.D.

Instrument Serial No. 008666 4144 West Ave., Ayden, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Kelly L

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PITT AYDEN PD 730

Serial Number: *008666*
Test Date: *05/19/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	9:28am
AIR BLK	.00	9:29am
ACCY CHK	.08	9:29am
AIR BLK	.00	9:30am
SUB TEST	**.00**	**9:31am**
AIR BLK	.00	9:32am
SUB TEST	**.00**	**9:33am**
AIR BLK	.00	9:34am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PITT AYDEN PD 730

Serial Number: *008666* Test Record Number: *737*
Test Date: *05/19/2014* Test Time: *9:35am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:35am
FLO	Pass	9:35am
FC	Pass	9:36am

Temperature Tests

Test	Status	Time
FC1	Pass	9:36am
SRC	Pass	9:36am
DET	Pass	9:36am
BAR	Pass	9:36am
BT	Pass	9:36am

Blank Tests

Test	Status	Time
AIR	Pass	9:36am

Printer Tests

Test	Status	Time
PRNT	Pass	9:36am

CRC Tests

Test	Status	Time
COMP	Pass	9:36am
CAL	Pass	9:36am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wilson Instrument Location Wilson Co Detention Center

Instrument Serial No. 008652 100 E. Green St, Wilson, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WILSON COUNTY WILSON CO DETENTION 970

Serial Number: *008652*
Test Date: *05/16/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG317801
Exp Date: 06/27/2015

Test	g/210L	Time
DIAG	Pass	10:12am
AIR BLK	.00	10:13am
ACCY CHK	.07	10:14am
AIR BLK	.00	10:15am
SUB TEST	**.00**	**10:16am**
AIR BLK	.00	10:17am
SUB TEST	**.00**	**10:19am**
AIR BLK	.00	10:19am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WILSON COUNTY WILSON CO DETENTION 970

Serial Number: *008652* Test Record Number: *2507*
Test Date: *05/16/2014* Test Time: *10:21am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:22am
FLO	Pass	10:22am
FC	Pass	10:22am

Temperature Tests

Test	Status	Time
FC1	Pass	10:22am
SRC	Pass	10:22am
DET	Pass	10:22am
BAR	Pass	10:22am
BT	Pass	10:22am

Blank Tests

Test	Status	Time
AIR	Pass	10:23am

Printer Tests

Test	Status	Time
PRNT	Pass	10:23am

CRC Tests

Test	Status	Time
COMP	Pass	10:23am
CAL	Pass	10:23am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Beaufort Instrument Location Beaufort Co. Courthouse

Instrument Serial No. DD8588 102 E. 2nd St., Washington, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: *008588*
Test Date: *05/20/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	12:28pm
AIR BLK	.00	12:29pm
ACCY CHK	.08	12:30pm
AIR BLK	.00	12:30pm
SUB TEST	**.00**	**12:31pm**
AIR BLK	.00	12:32pm
SUB TEST	**.00**	**12:34pm**
AIR BLK	.00	12:35pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: *008588* Test Record Number: *771*
Test Date: *05/20/2014* Test Time: *12:36pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:36pm
FLO	Pass	12:36pm
FC	Pass	12:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:36pm
SRC	Pass	12:36pm
DET	Pass	12:36pm
BAR	Pass	12:36pm
BT	Pass	12:36pm

Blank Tests

Test	Status	Time
AIR	Pass	12:37pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:37pm

CRC Tests

Test	Status	Time
COMP	Pass	12:37pm
CAL	Pass	12:37pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Beaufort Instrument Location Beaufort Co. Courthouse

Instrument Serial No. 008909 102 E. 2nd St., Washington, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: *008909*
Test Date: *05/15/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	10:51am
AIR BLK	.00	10:52am
ACCY CHK	.07	10:53am
AIR BLK	.00	10:54am
SUB TEST	**.00**	**10:55am**
AIR BLK	.00	10:56am
SUB TEST	**.00**	**10:57am**
AIR BLK	.00	10:58am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: *008909* Test Record Number: *1907*
Test Date: *05/15/2014* Test Time: *10:59am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:59am
FLO	Pass	10:59am
FC	Pass	10:59am

Temperature Tests

Test	Status	Time
FC1	Pass	11:00am
SRC	Pass	11:00am
DET	Pass	11:00am
BAR	Pass	11:00am
BT	Pass	11:00am

Blank Tests

Test	Status	Time
AIR	Pass	11:00am

Printer Tests

Test	Status	Time
PRNT	Pass	11:00am

CRC Tests

Test	Status	Time
COMP	Pass	11:00am
CAL	Pass	11:00am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wilson Instrument Location Wilson Co Detention Center

Instrument Serial No. 008627 100 E. Green St, Wilson, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WILSON COUNTY WILSON CO DETENTION 970

Serial Number: *008627*
Test Date: *05/16/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG312802
Exp Date: 05/08/2015

Test	g/210L	Time
DIAG	Pass	9:57am
AIR BLK	.00	9:58am
ACCY CHK	.07	9:58am
AIR BLK	.00	9:59am
SUB TEST	**.00**	**10:00am**
AIR BLK	.00	10:01am
SUB TEST	**.00**	**10:03am**
AIR BLK	.00	10:03am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WILSON COUNTY WILSON CO DETENTION 970

Serial Number: *008627* Test Record Number: *1639*
Test Date: *05/16/2014* Test Time: *10:05am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:05am
FLO	Pass	10:05am
FC	Pass	10:05am

Temperature Tests

Test	Status	Time
FC1	Pass	10:06am
SRC	Pass	10:06am
DET	Pass	10:06am
BAR	Pass	10:06am
BT	Pass	10:06am

Blank Tests

Test	Status	Time
AIR	Pass	10:06am

Printer Tests

Test	Status	Time
PRNT	Pass	10:06am

CRC Tests

Test	Status	Time
COMP	Pass	10:06am
CAL	Pass	10:06am

Preventive Maintenance
Status: Pass

Kelly [illegible]
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Beaufort Instrument Location Beaufort Co. Courthouse

Instrument Serial No. 008851 102 E. 2nd St, Washington, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: *008851*
Test Date: *05/15/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	11:05am
AIR BLK	.00	11:06am
ACCY CHK	.08	11:07am
AIR BLK	.00	11:08am
SUB TEST	**.00**	**11:08am**
AIR BLK	.00	11:09am
SUB TEST	**.00**	**11:11am**
AIR BLK	.00	11:12am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: *008851* Test Record Number: *487*
Test Date: *05/15/2014* Test Time: *11:14am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:14am
FLO	Pass	11:14am
FC	Pass	11:14am

Temperature Tests

Test	Status	Time
FC1	Pass	11:14am
SRC	Pass	11:14am
DET	Pass	11:14am
BAR	Pass	11:14am
BT	Pass	11:14am

Blank Tests

Test	Status	Time
AIR	Pass	11:15am

Printer Tests

Test	Status	Time
PRNT	Pass	11:15am

CRC Tests

Test	Status	Time
COMP	Pass	11:15am
CAL	Pass	11:15am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County JOHNSTON Instrument Location SELMA Police Dept

Instrument Serial No. 008595 SELMA N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY SELMA PD 500

Serial Number: *008595*
Test Date: *05/12/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG317801
Exp Date: 06/27/2015

Test	g/210L	Time
DIAG	Pass	9:37am
AIR BLK	.00	9:38am
ACCY CHK	.07	9:39am
AIR BLK	.00	9:40am
SUB TEST	**.00**	**9:41am**
AIR BLK	.00	9:41am
SUB TEST	**.00**	**9:43am**
AIR BLK	.00	9:44am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY SELMA PD 500

Serial Number: *008595* Test Record Number: *741*
Test Date: *05/12/2014* Test Time: *9:44am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:45am
FLO	Pass	9:45am
FC	Pass	9:45am

Temperature Tests

Test	Status	Time
FC1	Pass	9:45am
SRC	Pass	9:45am
DET	Pass	9:45am
BAR	Pass	9:45am
BT	Pass	9:45am

Blank Tests

Test	Status	Time
AIR	Pass	9:45am

Printer Tests

Test	Status	Time
PRNT	Pass	9:45am

CRC Tests

Test	Status	Time
COMP	Pass	9:46am
CAL	Pass	9:46am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County JOHNSTON Instrument Location BENSON Police DEPT

Instrument Serial No. 008885 BENSON N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY BENSON POLICE DEPT.
500

Serial Number: *008885*
Test Date: *05/12/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG317801
Exp Date: 06/27/2015

Test	g/210L	Time
DIAG	Pass	10:29am
AIR BLK	.00	10:30am
ACCY CHK	.07	10:31am
AIR BLK	.00	10:32am
SUB TEST	**.00**	**10:32am**
AIR BLK	.00	10:33am
SUB TEST	**.00**	**10:35am**
AIR BLK	.00	10:36am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY BENSON POLICE DEPT. 500

Serial Number: *008885* Test Record Number: *358*
Test Date: *05/12/2014* Test Time: *10:36am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:37am
FLO	Pass	10:37am
FC	Pass	10:37am

Temperature Tests

Test	Status	Time
FC1	Pass	10:37am
SRC	Pass	10:37am
DET	Pass	10:37am
BAR	Pass	10:37am
BT	Pass	10:37am

Blank Tests

Test	Status	Time
AIR	Pass	10:38am

Printer Tests

Test	Status	Time
PRNT	Pass	10:38am

CRC Tests

Test	Status	Time
COMP	Pass	10:38am
CAL	Pass	10:38am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County JOHNSTON Instrument Location CLAYTON POLICE DEPT.

Instrument Serial No. 008658 CLAYTON N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY CLAYTON PD. 500

Serial Number: *008658*
Test Date: *05/12/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG404101
Exp Date: 02/10/2016

Test	g/210L	Time
DIAG	Pass	11:25am
AIR BLK	.00	11:26am
ACCY CHK	.08	11:27am
AIR BLK	.00	11:28am
SUB TEST	**.00**	**11:28am**
AIR BLK	.00	11:29am
SUB TEST	**.00**	**11:31am**
AIR BLK	.00	11:31am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY CLAYTON PD. 500

Serial Number: *008658* Test Record Number: *1008*
Test Date: *05/12/2014* Test Time: *11:33am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:34am
FLO	Pass	11:34am
FC	Pass	11:34am

Temperature Tests

Test	Status	Time
FC1	Pass	11:34am
SRC	Pass	11:34am
DET	Pass	11:34am
BAR	Pass	11:34am
BT	Pass	11:34am

Blank Tests

Test	Status	Time
AIR	Pass	11:34am

Printer Tests

Test	Status	Time
PRNT	Pass	11:34am

CRC Tests

Test	Status	Time
COMP	Pass	11:35am
CAL	Pass	11:35am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County LEE Instrument Location LEE CO. JAIL

Instrument Serial No. 008645 SANFORD N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

LEE COUNTY LEE CO. LEC. 520

Serial Number: *008645*
Test Date: *05/13/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG300202
Exp Date: 01/02/2015

Test	g/210L	Time
DIAG	Pass	11:43am
AIR BLK	.00	11:44am
ACCY CHK	.07	11:44am
AIR BLK	.00	11:45am
SUB TEST	**.00**	**11:46am**
AIR BLK	.00	11:46am
SUB TEST	**.00**	**11:48am**
AIR BLK	.00	11:49am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

LEE COUNTY LEE CO. LEC. 520

Serial Number: *008645* Test Record Number: *1359*
Test Date: *05/13/2014* Test Time: *11:50am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:50am
FLO	Pass	11:50am
FC	Pass	11:50am

Temperature Tests

Test	Status	Time
FC1	Pass	11:50am
SRC	Pass	11:50am
DET	Pass	11:50am
BAR	Pass	11:50am
BT	Pass	11:50am

Blank Tests

Test	Status	Time
AIR	Pass	11:51am

Printer Tests

Test	Status	Time
PRNT	Pass	11:51am

CRC Tests

Test	Status	Time
COMP	Pass	11:51am
CAL	Pass	11:51am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County HARNETT Instrument Location DUNN Police DEPT.

Instrument Serial No. 008644 DUNN N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

HARNETT COUNTY DUNN POLICE DEPT. 420

Serial Number: *008644*
Test Date: *05/14/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG312802
Exp Date: 05/08/2015

Test	g/210L	Time
DIAG	Pass	8:49am
AIR BLK	.00	8:50am
ACCY CHK	.08	8:50am
AIR BLK	.00	8:51am
SUB TEST	**.00**	**8:52am**
AIR BLK	.00	8:53am
SUB TEST	**.00**	**8:55am**
AIR BLK	.00	8:56am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HARNETT COUNTY DUNN POLICE DEPT. 420

Serial Number: *008644* Test Record Number: *1054*
Test Date: *05/14/2014* Test Time: *8:58am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:58am
FLO	Pass	8:58am
FC	Pass	8:58am

Temperature Tests

Test	Status	Time
FC1	Pass	8:59am
SRC	Pass	8:59am
DET	Pass	8:59am
BAR	Pass	8:59am
BT	Pass	8:59am

Blank Tests

Test	Status	Time
AIR	Pass	8:59am

Printer Tests

Test	Status	Time
PRNT	Pass	8:59am

CRC Tests

Test	Status	Time
COMP	Pass	8:59am
CAL	Pass	8:59am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ROBENSON Instrument Location PEMBROKE Police DEPT.

Instrument Serial No. 008837 PEMBROKE, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROBESON COUNTY PEMBROKE POLICE DEPT
770

Serial Number: *008837*
Test Date: *05/15/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	10:21am
AIR BLK	.00	10:22am
ACCY CHK	.07	10:22am
AIR BLK	.00	10:23am
SUB TEST	**.00**	**10:24am**
AIR BLK	.00	10:24am
SUB TEST	**.00**	**10:27am**
AIR BLK	.00	10:27am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY PEMBROKE POLICE DEPT 770

Serial Number: *008837* Test Record Number: *561*
Test Date: *05/15/2014* Test Time: *10:29am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:29am
FLO	Pass	10:29am
FC	Pass	10:29am

Temperature Tests

Test	Status	Time
FC1	Pass	10:29am
SRC	Pass	10:29am
DET	Pass	10:29am
BAR	Pass	10:29am
BT	Pass	10:29am

Blank Tests

Test	Status	Time
AIR	Pass	10:30am

Printer Tests

Test	Status	Time
PRNT	Pass	10:30am

CRC Tests

Test	Status	Time
COMP	Pass	10:30am
CAL	Pass	10:30am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ROBESON Instrument Location RED SPRINGS P.D.

Instrument Serial No. 008857 RED SPRINGS P.D.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MAY, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROBESON COUNTY RED SPRINGS PD 770

Serial Number: *008857*
Test Date: *05/15/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG300202
Exp Date: 01/02/2015

Test	g/210L	Time
DIAG	Pass	11:12am
AIR BLK	.00	11:13am
ACCY CHK	.07	11:13am
AIR BLK	.00	11:14am
SUB TEST	**.00**	**11:15am**
AIR BLK	.00	11:16am
SUB TEST	**.00**	**11:17am**
AIR BLK	.00	11:18am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY RED SPRINGS PD 770

Serial Number: *008857* Test Record Number: *397*
Test Date: *05/15/2014* Test Time: *11:20am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:20am
FLO	Pass	11:20am
FC	Pass	11:20am

Temperature Tests

Test	Status	Time
FC1	Pass	11:20am
SRC	Pass	11:20am
DET	Pass	11:20am
BAR	Pass	11:20am
BT	Pass	11:20am

Blank Tests

Test	Status	Time
AIR	Pass	11:21am

Printer Tests

Test	Status	Time
PRNT	Pass	11:21am

CRC Tests

Test	Status	Time
COMP	Pass	11:21am
CAL	Pass	11:21am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ROBESON Instrument Location ST. PAULS Police DEPT.

Instrument Serial No. 008814 ST. PAULS N.C

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ROBESON COUNTY ST. PAULS PD. 770

Serial Number: *008814*
Test Date: *05/15/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name:
QUARANTELLO, NICHOLAS J
Permit Number: *21536E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309105
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	12:04pm
AIR BLK	.00	12:04pm
ACCY CHK	.08	12:05pm
AIR BLK	.00	12:06pm
SUB TEST	**.00**	**12:07pm**
AIR BLK	.00	12:07pm
SUB TEST	**.00**	**12:09pm**
AIR BLK	.00	12:10pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY ST. PAULS PD. 770

Serial Number: *008814* Test Record Number: *480*
Test Date: *05/15/2014* Test Time: *12:10pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:11pm
FLO	Pass	12:11pm
FC	Pass	12:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:11pm
SRC	Pass	12:11pm
DET	Pass	12:11pm
BAR	Pass	12:11pm
BT	Pass	12:11pm

Blank Tests

Test	Status	Time
AIR	Pass	12:12pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:12pm

CRC Tests

Test	Status	Time
COMP	Pass	12:12pm
CAL	Pass	12:12pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Mobile Unit #7

Instrument Serial No. 008704 Wake Forest

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008704* Test Record Number: *260*
Test Date: *05/02/2014* Test Time: *11:55pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:56pm
FLO	Pass	11:56pm
FC	Pass	11:56pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:56pm
SRC	Pass	11:56pm
DET	Pass	11:56pm
BAR	Pass	11:56pm
BT	Pass	11:56pm

Blank Tests

Test	Status	Time
AIR	Pass	11:57pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:57pm

CRC Tests

Test	Status	Time
COMP	Pass	11:57pm
CAL	Pass	11:57pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910

Serial Number: *008704*
Test Date: *05/02/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG308002
Exp Date: 03/21/2015

Test	g/210L	Time
DIAG	Pass	11:38pm
AIR BLK	.00	11:39pm
ACCY CHK	.08	11:40pm
AIR BLK	.00	11:41pm
SUB TEST	**.00**	**11:41pm**
AIR BLK	.00	11:42pm
SUB TEST	**.00**	**11:44pm**
AIR BLK	.00	11:45pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT MOBILE UNIT #7

Instrument Serial No. 008623 WAKE FOREST

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 7 910
Serial Number: *008623* Test Record Number: *2873*
Test Date: *05/02/2014* Test Time: *11:48pm EDT*
System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:49pm
FLO	Pass	11:49pm
FC	Pass	11:49pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:49pm
SRC	Pass	11:49pm
DET	Pass	11:49pm
BAR	Pass	11:49pm
BT	Pass	11:49pm

Blank Tests

Test	Status	Time
AIR	Pass	11:50pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:50pm

CRC Tests

Test	Status	Time
COMP	Pass	11:50pm
CAL	Pass	11:50pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 7 910
Serial Number: *008623*
Test Date: *05/02/2014*
Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *MORGART, STEPHEN G*
Permit Number: *9372E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG305202
Exp Date: 02/21/2015

Test	g/210L	Time
DIAG	Pass	11:34pm
AIR BLK	.00	11:35pm
ACCY CHK	.07	11:35pm
AIR BLK	.00	11:36pm
SUB TEST	**.00**	**11:37pm**
AIR BLK	.00	11:38pm
SUB TEST	**.00**	**11:40pm**
AIR BLK	.00	11:41pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Iredell Instrument Location Iredell County S.D.

Instrument Serial No. 008809 221 E. Water St. Statesville
704-878-3131

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13th day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY IREDELL COUNTY SD 480

Serial Number: *008809* Test Record Number: *2624*
Test Date: *05/13/2014* Test Time: *11:42am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:43am
FLO	Pass	11:43am
FC	Pass	11:43am

Temperature Tests

Test	Status	Time
FC1	Pass	11:43am
SRC	Pass	11:43am
DET	Pass	11:43am
BAR	Pass	11:43am
BT	Pass	11:43am

Blank Tests

Test	Status	Time
AIR	Pass	11:44am

Printer Tests

Test	Status	Time
PRNT	Pass	11:44am

CRC Tests

Test	Status	Time
COMP	Pass	11:44am
CAL	Pass	11:44am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

IREDELL COUNTY IREDELL COUNTY SD 480

Serial Number: *008809*
Test Date: *05/13/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
01/01/2014-01/01/2016

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG335201
Exp Date: 12/18/2015

Test	g/210L	Time
DIAG	Pass	11:48am
AIR BLK	.00	11:49am
ACCY CHK	.07	11:49am
AIR BLK	.00	11:50am
SUB TEST	**.00**	**11:51am**
AIR BLK	.00	11:52am
SUB TEST	**.00**	**11:53am**
AIR BLK	.00	11:55am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Iredell Instrument Location Statesville PD

Instrument Serial No. 008619 330 S. Todd St., Statesville
704-878-3406

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY STATESVILLE PD 480

Serial Number: *008619* Test Record Number: *1009*
Test Date: *05/13/2014* Test Time: *12:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:09pm
FLO	Pass	12:09pm
FC	Pass	12:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:09pm
SRC	Pass	12:09pm
DET	Pass	12:09pm
BAR	Pass	12:09pm
BT	Pass	12:09pm

Blank Tests

Test	Status	Time
AIR	Pass	12:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:09pm

CRC Tests

Test	Status	Time
COMP	Pass	12:10pm
CAL	Pass	12:10pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

IREDELL COUNTY STATESVILLE PD 480

Serial Number: *008619*
Test Date: *05/13/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
01/01/2014-01/01/2016

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG308702
Exp Date: 03/28/2015

Test	g/210L	Time
DIAG	Pass	12:12pm
AIR BLK	.00	12:13pm
ACCY CHK	.08	12:13pm
AIR BLK	.00	12:14pm
SUB TEST	**.00**	**12:15pm**
AIR BLK	.00	12:15pm
SUB TEST	**.00**	**12:17pm**
AIR BLK	.00	12:18pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Iredell Instrument Location Mooresville PD

Instrument Serial No. 008685 750 W. Iredell Ave., Mooresville
704-664-311

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY MOORESVILLE PD 480

Serial Number: *008685* Test Record Number: *2159*
Test Date: *05/13/2014* Test Time: *1:36pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:37pm
FLO	Pass	1:37pm
FC	Pass	1:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:37pm
SRC	Pass	1:37pm
DET	Pass	1:37pm
BAR	Pass	1:37pm
BT	Pass	1:37pm

Blank Tests

Test	Status	Time
AIR	Pass	1:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:38pm

CRC Tests

Test	Status	Time
COMP	Pass	1:38pm
CAL	Pass	1:38pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

IREDELL COUNTY MOORESVILLE PD 480

Serial Number: *008685*
Test Date: *05/13/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
01/01/2014-01/01/2016

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG322601
Exp Date: 08/14/2015

Test	g/210L	Time
DIAG	Pass	1:40pm
AIR BLK	.00	1:41pm
ACCY CHK	.08	1:42pm
AIR BLK	.00	1:43pm
SUB TEST	**.00**	**1:43pm**
AIR BLK	.00	1:44pm
SUB TEST	**.00**	**1:45pm**
AIR BLK	.00	1:46pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Mecklenburg County SD

Instrument Serial No. 008690 801 E 4th St, Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY SHERIFFS DEPARTMENT 590

Serial Number: *008690* Test Record Number: *4348*
Test Date: *05/15/2014* Test Time: *1:38pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:38pm
FLO	Pass	1:38pm
FC	Pass	1:38pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:39pm
SRC	Pass	1:39pm
DET	Pass	1:39pm
BAR	Pass	1:39pm
BT	Pass	1:39pm

Blank Tests

Test	Status	Time
AIR	Pass	1:39pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:39pm

CRC Tests

Test	Status	Time
COMP	Pass	1:39pm
CAL	Pass	1:39pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY SHERIFFS DEPARTMENT
590

Serial Number: *008690*
Test Date: *05/15/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
01/01/2014-01/01/2016

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG335201
Exp Date: 12/18/2015

Test	g/210L	Time
DIAG	Pass	1:47pm
AIR BLK	.00	1:48pm
ACCY CHK	.08	1:49pm
AIR BLK	.00	1:50pm
SUB TEST	**.00**	**1:50pm**
AIR BLK	.00	1:51pm
SUB TEST	**.00**	**1:52pm**
AIR BLK	.00	1:53pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location CMPD-LEC

Instrument Serial No. 008594 601 E. Trade St., Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: *008594* Test Record Number: *1973*
Test Date: *05/15/2014* Test Time: *3:02pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:02pm
FLO	Pass	3:02pm
FC	Pass	3:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:02pm
SRC	Pass	3:02pm
DET	Pass	3:02pm
BAR	Pass	3:02pm
BT	Pass	3:02pm

Blank Tests

Test	Status	Time
AIR	Pass	3:03pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:03pm

CRC Tests

Test	Status	Time
COMP	Pass	3:03pm
CAL	Pass	3:03pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: *008594*
Test Date: *05/15/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
01/01/2014-01/01/2016

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG405702
Exp Date: 02/26/2016

Test	g/210L	Time
DIAG	Pass	3:07pm
AIR BLK	.00	3:08pm
ACCY CHK	.07	3:08pm
AIR BLK	.00	3:09pm
SUB TEST	**.00**	**3:10pm**
AIR BLK	.00	3:11pm
SUB TEST	**.00**	**3:12pm**
AIR BLK	.00	3:13pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location CMPD LEC

Instrument Serial No. 008691 601 E. Trade St., Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: *008691* Test Record Number: *4626*
Test Date: *05/15/2014* Test Time: *2:37pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:37pm
FLO	Pass	2:37pm
FC	Pass	2:38pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:38pm
SRC	Pass	2:38pm
DET	Pass	2:38pm
BAR	Pass	2:38pm
BT	Pass	2:38pm

Blank Tests

Test	Status	Time
AIR	Pass	2:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:38pm

CRC Tests

Test	Status	Time
COMP	Pass	2:38pm
CAL	Pass	2:38pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: *008691*
Test Date: *05/15/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *HAYS, MARK D*
Permit Number: *15924E*
Effective:
01/01/2014-01/01/2016

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG404101
Exp Date: 02/10/2016

Test	g/210L	Time
DIAG	Pass	2:50pm
AIR BLK	.00	2:51pm
ACCY CHK	.07	2:51pm
AIR BLK	.00	2:52pm
SUB TEST	**.00**	**2:53pm**
AIR BLK	.00	2:54pm
SUB TEST	**.00**	**2:55pm**
AIR BLK	.00	2:56pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Martin Instrument Location Martin Co. S.O.

Instrument Serial No. 008918 305 E. Main St., Williamston, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7th day of MAY, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MARTIN COUNTY SHERIFF'S OFFICE 570

Serial Number: *008918*
Test Date: *05/07/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *KEESLER, LINDA A*
Permit Number: *11646E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309105
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	10:55am
AIR BLK	.00	10:56am
ACCY CHK	.08	10:56am
AIR BLK	.00	10:57am
SUB TEST	**.00**	**10:58am**
AIR BLK	.00	10:59am
SUB TEST	**.00**	**11:00am**
AIR BLK	.00	11:01am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MARTIN COUNTY SHERIFF'S OFFICE 570

Serial Number: *008918* Test Record Number: *333*
Test Date: *05/07/2014* Test Time: *11:03am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:04am
FLO	Pass	11:04am
FC	Pass	11:04am

Temperature Tests

Test	Status	Time
FC1	Pass	11:04am
SRC	Pass	11:04am
DET	Pass	11:04am
BAR	Pass	11:04am
BT	Pass	11:04am

Blank Tests

Test	Status	Time
AIR	Pass	11:05am

Printer Tests

Test	Status	Time
PRNT	Pass	11:05am

CRC Tests

Test	Status	Time
COMP	Pass	11:05am
CAL	Pass	11:05am

Preventive Maintenance
Status: Pass

Fred A. Kerse
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Perquimans Instrument Location Perquimans Co. S.O.

Instrument Serial No. 008921 110 N. Church St., Hertford, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PERQUIMANS COUNTY PERQUIMANS CO SO 710

Serial Number: *008921*
Test Date: *05/13/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *GUARD, KELLY G*
Permit Number: *12955E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	11:06am
AIR BLK	.00	11:07am
ACCY CHK	.08	11:07am
AIR BLK	.00	11:08am
SUB TEST	**.00**	**11:09am**
AIR BLK	.00	11:10am
SUB TEST	**.00**	**11:11am**
AIR BLK	.00	11:12am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PERQUIMANS COUNTY PERQUIMANS CO SO 710

Serial Number: *008921* Test Record Number: *490*
Test Date: *05/13/2014* Test Time: *11:14am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:14am
FLO	Pass	11:14am
FC	Pass	11:14am

Temperature Tests

Test	Status	Time
FC1	Pass	11:14am
SRC	Pass	11:14am
DET	Pass	11:14am
BAR	Pass	11:14am
BT	Pass	11:14am

Blank Tests

Test	Status	Time
AIR	Pass	11:15am

Printer Tests

Test	Status	Time
PRNT	Pass	11:15am

CRC Tests

Test	Status	Time
COMP	Pass	11:15am
CAL	Pass	11:15am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Wake Co. Detention Center

Instrument Serial No. 008924 3301 Hammond Rd.
Raleigh, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Dale Farley
Signature of Certifying Official

655
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WAKE COUNTY DETENTION CENTER 910

Serial Number: *008924*
Test Date: *05/06/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *FARLEY, CYNTHIA D*
Permit Number: *24123E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG405702
Exp Date: 02/26/2016

Test	g/210L	Time
DIAG	Pass	9:39am
AIR BLK	.00	9:40am
ACCY CHK	.07	9:41am
AIR BLK	.00	9:42am
SUB TEST	**.00**	**9:43am**
AIR BLK	.00	9:44am
SUB TEST	**.00**	**9:46am**
AIR BLK	.00	9:47am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY DETENTION CENTER 910

Serial Number: *008924* Test Record Number: *648*
Test Date: *05/06/2014* Test Time: *9:48am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:48am
FLO	Pass	9:48am
FC	Pass	9:48am

Temperature Tests

Test	Status	Time
FC1	Pass	9:48am
SRC	Pass	9:48am
DET	Pass	9:48am
BAR	Pass	9:48am
BT	Pass	9:48am

Blank Tests

Test	Status	Time
AIR	Pass	9:49am

Printer Tests

Test	Status	Time
PRNT	Pass	9:49am

CRC Tests

Test	Status	Time
COMP	Pass	9:49am
CAL	Pass	9:49am

Preventive Maintenance
Status: Pass

C. Dale Farley

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wilkes Instrument Location Bat mobile Unit 5

Instrument Serial No. 008698

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

WILKES BAT MOBILE UNIT 5 960

Serial Number: *008698* Test Record Number: *1098*
Test Date: *05/03/2014* Test Time: *8:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:09pm
FLO	Pass	8:09pm
FC	Pass	8:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:09pm
SRC	Pass	8:09pm
DET	Pass	8:09pm
BAR	Pass	8:09pm
BT	Pass	8:09pm

Blank Tests

Test	Status	Time
AIR	Pass	8:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:09pm

CRC Tests

Test	Status	Time
COMP	Pass	8:10pm
CAL	Pass	8:10pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

WILKES BAT MOBILE UNIT 5 960

Serial Number: *008698*
Test Date: *05/03/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG405702
Exp Date: 02/26/2016

Test	g/210L	Time
DIAG	Pass	8:00pm
AIR BLK	.00	8:01pm
ACCY CHK	.07	8:02pm
AIR BLK	.00	8:03pm
SUB TEST	**.00**	**8:04pm**
AIR BLK	.00	8:05pm
SUB TEST	**.00**	**8:06pm**
AIR BLK	.00	8:07pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cleveland Instrument Location Bat mobile Unit 5

Instrument Serial No. 008600

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

CLEVELAND BAT MOBILE UNIT 5 220

Serial Number: *008600* Test Record Number: *1426*
Test Date: *05/10/2014* Test Time: *8:24pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:25pm
FLO	Pass	8:25pm
FC	Pass	8:25pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:25pm
SRC	Pass	8:25pm
DET	Pass	8:25pm
BAR	Pass	8:25pm
BT	Pass	8:25pm

Blank Tests

Test	Status	Time
AIR	Pass	8:26pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:26pm

CRC Tests

Test	Status	Time
COMP	Pass	8:26pm
CAL	Pass	8:26pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

CLEVELAND BAT MOBILE UNIT 5 220

Serial Number: *008600*
Test Date: *05/10/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG322601
Exp Date: 08/14/2015

Test	g/210L	Time
DIAG	Pass	8:16pm
AIR BLK	.00	8:18pm
ACCY CHK	.07	8:18pm
AIR BLK	.00	8:19pm
SUB TEST	**.00**	**8:20pm**
AIR BLK	.00	8:21pm
SUB TEST	**.00**	**8:22pm**
AIR BLK	.00	8:23pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Cleveland Instrument Location Bat Mobile Units

Instrument Serial No. 008698

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Preventive Maintenance

CLEVELAND BAT MOBILE UNIT 5 220

Serial Number: *008698* Test Record Number: *1109*
Test Date: *05/10/2014* Test Time: *8:23pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:23pm
FLO	Pass	8:23pm
FC	Pass	8:23pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:23pm
SRC	Pass	8:23pm
DET	Pass	8:23pm
BAR	Pass	8:23pm
BT	Pass	8:23pm

Blank Tests

Test	Status	Time
AIR	Pass	8:24pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:24pm

CRC Tests

Test	Status	Time
COMP	Pass	8:24pm
CAL	Pass	8:24pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

CLEVELAND BAT MOBILE UNIT 5 220

Serial Number: *008698*
Test Date: *05/10/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *TOWERY, CHAD V*
Permit Number: *26632E*
Effective:
10/18/2013-10/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG405702
Exp Date: 02/26/2016

Test	g/210L	Time
DIAG	Pass	8:14pm
AIR BLK	.00	8:15pm
ACCY CHK	.07	8:16pm
AIR BLK	.00	8:17pm
SUB TEST	**.00**	**8:18pm**
AIR BLK	.00	8:19pm
SUB TEST	**.00**	**8:20pm**
AIR BLK	.00	8:21pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Madison Instrument Location Mars Hill PD

Instrument Serial No. 008582 Mars Hill, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MADISON COUNTY MARS HILL PD 560

Serial Number: *008582*
Test Date: *05/08/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BURNETTE, ANTHONY J*
Permit Number: *11304E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG303502
Exp Date: 02/04/2015

Test	g/210L	Time
DIAG	Pass	4:28pm
AIR BLK	.00	4:29pm
ACCY CHK	.07	4:29pm
AIR BLK	.00	4:30pm
SUB TEST	**.00**	**4:31pm**
AIR BLK	.00	4:32pm
SUB TEST	**.00**	**4:33pm**
AIR BLK	.00	4:34pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MADISON COUNTY MARS HILL PD 560

Serial Number: *008582* Test Record Number: *991*
Test Date: *05/08/2014* Test Time: *4:35pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:35pm
FLO	Pass	4:35pm
FC	Pass	4:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:36pm
SRC	Pass	4:36pm
DET	Pass	4:36pm
BAR	Pass	4:36pm
BT	Pass	4:36pm

Blank Tests

Test	Status	Time
AIR	Pass	4:36pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:36pm

CRC Tests

Test	Status	Time
COMP	Pass	4:36pm
CAL	Pass	4:36pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Watuga Instrument Location Watauga Co. Jail

Instrument Serial No. 008715 Boone, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WATAUGA COUNTY WATAUGA JAIL 940

Serial Number: *008715*
Test Date: *05/07/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BURNETTE, ANTHONY J*
Permit Number: *11304E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG300202
Exp Date: 01/02/2015

Test	g/210L	Time
DIAG	Pass	6:04pm
AIR BLK	.00	6:05pm
ACCY CHK	.07	6:06pm
AIR BLK	.00	6:07pm
SUB TEST	**.00**	**6:07pm**
AIR BLK	.00	6:08pm
SUB TEST	**.00**	**6:10pm**
AIR BLK	.00	6:11pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WATAUGA COUNTY WATAUGA JAIL 940

Serial Number: *008715* Test Record Number: *1426*
Test Date: *05/07/2014* Test Time: *6:12pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:12pm
FLO	Pass	6:12pm
FC	Pass	6:13pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:13pm
SRC	Pass	6:13pm
DET	Pass	6:13pm
BAR	Pass	6:13pm
BT	Pass	6:13pm

Blank Tests

Test	Status	Time
AIR	Pass	6:13pm

Printer Tests

Test	Status	Time
PRNT	Pass	6:13pm

CRC Tests

Test	Status	Time
COMP	Pass	6:13pm
CAL	Pass	6:13pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CASWELL Instrument Location CASWELL CO. DETENTION CTR

Instrument Serial No. 008593 231 COUNTY PARK RD YANCEYVILLE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 02 day of MAY, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CASWELL COUNTY DETENTION CENTER 160

Serial Number: *008593*
Test Date: *05/02/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SMITH, BRIAN D*
Permit Number: *8937E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG335201
Exp Date: 12/18/2015

Test	g/210L	Time
DIAG	Pass	12:38pm
AIR BLK	.00	12:39pm
ACCY CHK	.08	12:40pm
AIR BLK	.00	12:41pm
SUB TEST	**.00**	**12:41pm**
AIR BLK	.00	12:43pm
SUB TEST	**.00**	**12:44pm**
AIR BLK	.00	12:45pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CASWELL COUNTY DETENTION CENTER 160

Serial Number: *008593* Test Record Number: *1089*
Test Date: *05/02/2014* Test Time: *12:56pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:56pm
FLO	Pass	12:56pm
FC	Pass	12:56pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:57pm
SRC	Pass	12:57pm
DET	Pass	12:57pm
BAR	Pass	12:57pm
BT	Pass	12:57pm

Blank Tests

Test	Status	Time
AIR	Pass	12:57pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:57pm

CRC Tests

Test	Status	Time
COMP	Pass	12:57pm
CAL	Pass	12:57pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location CARY PD

Instrument Serial No. 008587 120 WILKINSON AVE. CARY, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 01 day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

WAKE COUNTY CARY PD 910

Serial Number: *008587*
Test Date: *05/01/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *SMITH, BRIAN D*
Permit Number: *8937E*
Effective:
08/01/2013-08/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG409709
Exp Date: 04/07/2016

Test	g/210L	Time
DIAG	Pass	3:11pm
AIR BLK	.00	3:11pm
ACCY CHK	.08	3:12pm
AIR BLK	.00	3:13pm
SUB TEST	**.00**	**3:14pm**
AIR BLK	.00	3:14pm
SUB TEST	**.00**	**3:16pm**
AIR BLK	.00	3:17pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY CARY PD 910

Serial Number: *008587* Test Record Number: *2639*
Test Date: *05/01/2014* Test Time: *3:19pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:19pm
FLO	Pass	3:19pm
FC	Pass	3:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:20pm
SRC	Pass	3:20pm
DET	Pass	3:20pm
BAR	Pass	3:20pm
BT	Pass	3:20pm

Blank Tests

Test	Status	Time
AIR	Pass	3:20pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:20pm

CRC Tests

Test	Status	Time
COMP	Pass	3:20pm
CAL	Pass	3:20pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Durham Instrument Location Durham Co. Jail

Instrument Serial No. 008891 217 S. Mangum St.
Durham, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1st day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Dale Farley
Signature of Certifying Official

655
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: *008891*
Test Date: *05/01/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *FARLEY, CYNTHIA D*
Permit Number: *24123E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG405702
Exp Date: 02/26/2016

Test	g/210L	Time
DIAG	Pass	3:36pm
AIR BLK	.00	3:37pm
ACCY CHK	.08	3:37pm
AIR BLK	.00	3:39pm
SUB TEST	**.00**	**3:40pm**
AIR BLK	.00	3:40pm
SUB TEST	**.00**	**3:42pm**
AIR BLK	.00	3:43pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: *008891* Test Record Number: *2581*
Test Date: *05/01/2014* Test Time: *3:46pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:46pm
FLO	Pass	3:46pm
FC	Pass	3:46pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:46pm
SRC	Pass	3:46pm
DET	Pass	3:46pm
BAR	Pass	3:46pm
BT	Pass	3:46pm

Blank Tests

Test	Status	Time
AIR	Pass	3:47pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:47pm

CRC Tests

Test	Status	Time
COMP	Pass	3:47pm
CAL	Pass	3:47pm

Preventive Maintenance
Status: Pass

Dale Farley
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Durham Instrument Location Durham Co. Jail

Instrument Serial No. 008878 217 S. Mangum St.
Durham, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1st day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

655
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: *008878*
Test Date: *05/01/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *FARLEY, CYNTHIA D*
Permit Number: *24123E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	3:02pm
AIR BLK	.00	3:03pm
ACCY CHK	.07	3:04pm
AIR BLK	.00	3:05pm
SUB TEST	**.00**	**3:06pm**
AIR BLK	.00	3:07pm
SUB TEST	**.00**	**3:11pm**
AIR BLK	.00	3:12pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: *008878* Test Record Number: *2858*
Test Date: *05/01/2014* Test Time: *3:13pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:13pm
FLO	Pass	3:13pm
FC	Pass	3:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:14pm
SRC	Pass	3:14pm
DET	Pass	3:14pm
BAR	Pass	3:14pm
BT	Pass	3:14pm

Blank Tests

Test	Status	Time
AIR	Pass	3:14pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:14pm

CRC Tests

Test	Status	Time
COMP	Pass	3:14pm
CAL	Pass	3:14pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location HIGH Point JAIL

Instrument Serial No. 008865

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MAY, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

GUILFORD COUNTY HIGH POINT JAIL 401

Serial Number: *008865*
Test Date: *05/15/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *DEAN, L K*
Permit Number: *11598E*
Effective:
06/01/2013-06/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG323402
Exp Date: 08/22/2015

Test	g/210L	Time
DIAG	Pass	10:30am
AIR BLK	.00	10:30am
ACCY CHK	.07	10:31am
AIR BLK	.00	10:32am
SUB TEST	**.00**	**10:32am**
AIR BLK	.00	10:33am
SUB TEST	**.00**	**10:35am**
AIR BLK	.00	10:36am

Reported AC: .00 g/210L

L. Kevin Dean
Signature of Chemical Analyst

Court CVR

L. Kevin Dean
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY HIGH POINT JAIL 401

Serial Number: *008865* Test Record Number: *279*
Test Date: *05/15/2014* Test Time: *10:26am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:27am
FLO	Pass	10:27am
FC	Pass	10:27am

Temperature Tests

Test	Status	Time
FC1	Pass	10:27am
SRC	Pass	10:27am
DET	Pass	10:27am
BAR	Pass	10:27am
BT	Pass	10:27am

Blank Tests

Test	Status	Time
AIR	Pass	10:27am

Printer Tests

Test	Status	Time
PRNT	Pass	10:27am

CRC Tests

Test	Status	Time
COMP	Pass	10:28am
CAL	Pass	10:28am

Preventive Maintenance
Status: Pass

L. Kevin Deer
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ONSLOW Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 RICHLANDS, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of MAY, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ONSLOW COUNTY BAT MOBILE UNIT 3 660

Serial Number: *008707*
Test Date: *05/02/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309101
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	11:13pm
AIR BLK	.00	11:14pm
ACCY CHK	.08	11:14pm
AIR BLK	.00	11:15pm
SUB TEST	**.00**	**11:16pm**
AIR BLK	.00	11:17pm
SUB TEST	**.00**	**11:19pm**
AIR BLK	.00	11:20pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY BAT MOBILE UNIT 3 660

Serial Number: *008707* Test Record Number: *1995*
Test Date: *05/02/2014* Test Time: *11:21pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:22pm
FLO	Pass	11:22pm
FC	Pass	11:22pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:22pm
SRC	Pass	11:22pm
DET	Pass	11:22pm
BAR	Pass	11:22pm
BT	Pass	11:22pm

Blank Tests

Test	Status	Time
AIR	Pass	11:23pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:23pm

CRC Tests

Test	Status	Time
COMP	Pass	11:23pm
CAL	Pass	11:23pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CRAVEN Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 RIVERBEND, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CRAVEN COUNTY BAT MOBILE UNIT 3 240

Serial Number: *008647*
Test Date: *05/03/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG320602
Exp Date: 07/25/2015

Test	g/210L	Time
DIAG	Pass	11:01pm
AIR BLK	.00	11:02pm
ACCY CHK	.08	11:02pm
AIR BLK	.00	11:03pm
SUB TEST	**.00**	**11:04pm**
AIR BLK	.00	11:05pm
SUB TEST	**.00**	**11:06pm**
AIR BLK	.00	11:07pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY BAT MOBILE UNIT 3 240

Serial Number: *008647* Test Record Number: *1936*
Test Date: *05/03/2014* Test Time: *11:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:08pm
FLO	Pass	11:08pm
FC	Pass	11:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:08pm
SRC	Pass	11:08pm
DET	Pass	11:08pm
BAR	Pass	11:08pm
BT	Pass	11:08pm

Blank Tests

Test	Status	Time
AIR	Pass	11:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:09pm

CRC Tests

Test	Status	Time
COMP	Pass	11:09pm
CAL	Pass	11:09pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CARTERET Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008616 NEWPORT, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CARTERET COUNTY BAT MOBILE UNIT 3 150

Serial Number: *008616*
Test Date: *05/17/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG320602
Exp Date: 07/25/2015

Test	g/210L	Time
DIAG	Pass	9:02pm
AIR BLK	.00	9:03pm
ACCY CHK	.08	9:04pm
AIR BLK	.00	9:05pm
SUB TEST	**.00**	**9:05pm**
AIR BLK	.00	9:06pm
SUB TEST	**.00**	**9:07pm**
AIR BLK	.00	9:08pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY BAT MOBILE UNIT 3 150

Serial Number: *008616* Test Record Number: *1907*
Test Date: *05/17/2014* Test Time: *9:10pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:10pm
FLO	Pass	9:10pm
FC	Pass	9:10pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:10pm
SRC	Pass	9:10pm
DET	Pass	9:10pm
BAR	Pass	9:10pm
BT	Pass	9:10pm

Blank Tests

Test	Status	Time
AIR	Pass	9:11pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:11pm

CRC Tests

Test	Status	Time
COMP	Pass	9:11pm
CAL	Pass	9:11pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CARTERET Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 NEWPORT, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MAY, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CARTERET COUNTY BAT MOBILE UNIT 3 150

Serial Number: *008647*
Test Date: *05/17/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG320602
Exp Date: 07/25/2015

Test	g/210L	Time
DIAG	Pass	8:59pm
AIR BLK	.00	9:00pm
ACCY CHK	.08	9:01pm
AIR BLK	.00	9:01pm
SUB TEST	**.00**	**9:02pm**
AIR BLK	.00	9:03pm
SUB TEST	**.00**	**9:05pm**
AIR BLK	.00	9:05pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY BAT MOBILE UNIT 3 150

Serial Number: *008647* Test Record Number: *1940*
Test Date: *05/17/2014* Test Time: *9:07pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:07pm
FLO	Pass	9:07pm
FC	Pass	9:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:08pm
SRC	Pass	9:08pm
DET	Pass	9:08pm
BAR	Pass	9:08pm
BT	Pass	9:08pm

Blank Tests

Test	Status	Time
AIR	Pass	9:08pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:08pm

CRC Tests

Test	Status	Time
COMP	Pass	9:09pm
CAL	Pass	9:09pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County CARTERET Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 NEWPORT, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

CARTERET COUNTY BAT MOBILE UNIT 3 150

Serial Number: *008707*
Test Date: *05/17/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309101
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	8:57pm
AIR BLK	.00	8:58pm
ACCY CHK	.08	8:59pm
AIR BLK	.00	9:00pm
SUB TEST	**.00**	**9:00pm**
AIR BLK	.00	9:01pm
SUB TEST	**.00**	**9:03pm**
AIR BLK	.00	9:03pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY BAT MOBILE UNIT 3 150

Serial Number: *008707* Test Record Number: *2000*
Test Date: *05/17/2014* Test Time: *9:04pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:05pm
FLO	Pass	9:05pm
FC	Pass	9:05pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:05pm
SRC	Pass	9:05pm
DET	Pass	9:05pm
BAR	Pass	9:05pm
BT	Pass	9:05pm

Blank Tests

Test	Status	Time
AIR	Pass	9:06pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:06pm

CRC Tests

Test	Status	Time
COMP	Pass	9:06pm
CAL	Pass	9:06pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County BRUNSWICK Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 LELAND, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of May, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 3 090

Serial Number: *008707*
Test Date: *05/24/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *BARNES, ALVIN R*
Permit Number: *15671E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309101
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	12:04am
AIR BLK	.00	12:05am
ACCY CHK	.08	12:05am
AIR BLK	.00	12:06am
SUB TEST	**.00**	**12:07am**
AIR BLK	.00	12:08am
SUB TEST	**.00**	**12:09am**
AIR BLK	.00	12:10am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 3 090

Serial Number: *008707* Test Record Number: *2003*
Test Date: *05/24/2014* Test Time: *12:11am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:11am
FLO	Pass	12:11am
FC	Pass	12:11am

Temperature Tests

Test	Status	Time
FC1	Pass	12:12am
SRC	Pass	12:12am
DET	Pass	12:12am
BAR	Pass	12:12am
BT	Pass	12:12am

Blank Tests

Test	Status	Time
AIR	Pass	12:12am

Printer Tests

Test	Status	Time
PRNT	Pass	12:12am

CRC Tests

Test	Status	Time
COMP	Pass	12:12am
CAL	Pass	12:12am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Macon Instrument Location Macon Co. Magistrate

Instrument Serial No. 008795 Highlands, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

MACON COUNTY MACON CO MAGISTRATE 550

Serial Number: *008795*
Test Date: *05/08/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *CUTLER, DANIEL R*
Permit Number: *8457E*
Effective:
10/01/2013-10/01/2015

Officer's Name: *NONE,*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG335201
Exp Date: 12/18/2015

Test	g/210L	Time
DIAG	Pass	10:51am
AIR BLK	.00	10:52am
ACCY CHK	.08	10:53am
AIR BLK	.00	10:54am
SUB TEST	**.00**	**10:54am**
AIR BLK	.00	10:55am
SUB TEST	**.00**	**10:57am**
AIR BLK	.00	10:57am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MACON COUNTY MACON CO MAGISTRATE 550

Serial Number: *008795* Test Record Number: *338*
Test Date: *05/08/2014* Test Time: *10:58am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:59am
FLO	Pass	10:59am
FC	Pass	10:59am

Temperature Tests

Test	Status	Time
FC1	Pass	10:59am
SRC	Pass	10:59am
DET	Pass	10:59am
BAR	Pass	10:59am
BT	Pass	10:59am

Blank Tests

Test	Status	Time
AIR	Pass	11:00am

Printer Tests

Test	Status	Time
PRNT	Pass	11:00am

CRC Tests

Test	Status	Time
COMP	Pass	11:00am
CAL	Pass	11:00am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pender Instrument Location Pender County

Instrument Serial No. 008946 Sheriff Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. [illegible] 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PENDER PENDER COUNTY SD 700

Serial Number: *008946*
Test Date: *05/29/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG335201
Exp Date: 12/18/2015

Test	g/210L	Time
DIAG	Pass	8:41am
AIR BLK	.00	8:41am
ACCY CHK	.08	8:42am
AIR BLK	.00	8:43am
SUB TEST	**.00**	**8:43am**
AIR BLK	.00	8:44am
SUB TEST	**.00**	**8:46am**
AIR BLK	.00	8:47am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PENDER PENDER COUNTY SD 700

Serial Number: *008946* Test Record Number: *705*
Test Date: *05/29/2014* Test Time: *8:47am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:48am
FLO	Pass	8:48am
FC	Pass	8:48am

Temperature Tests

Test	Status	Time
FC1	Pass	8:48am
SRC	Pass	8:48am
DET	Pass	8:48am
BAR	Pass	8:48am
BT	Pass	8:48am

Blank Tests

Test	Status	Time
AIR	Pass	8:48am

Printer Tests

Test	Status	Time
PRNT	Pass	8:48am

CRC Tests

Test	Status	Time
COMP	Pass	8:49am
CAL	Pass	8:49am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pender Instrument Location Pender County

Instrument Serial No. 008901 Sheriff Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PENDER COUNTY PENDER CO SD 700

Serial Number: *008901*
Test Date: *05/29/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG335201
Exp Date: 12/18/2015

Test	g/210L	Time
DIAG	Pass	8:39am
AIR BLK	.00	8:40am
ACCY CHK	.07	8:41am
AIR BLK	.00	8:42am
SUB TEST	**.00**	**8:42am**
AIR BLK	.00	8:43am
SUB TEST	**.00**	**8:45am**
AIR BLK	.00	8:46am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PENDER COUNTY PENDER CO SD 700

Serial Number: *008901* Test Record Number: *625*
Test Date: *05/29/2014* Test Time: *8:46am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:47am
FLO	Pass	8:47am
FC	Pass	8:47am

Temperature Tests

Test	Status	Time
FC1	Pass	8:47am
SRC	Pass	8:47am
DET	Pass	8:47am
BAR	Pass	8:47am
BT	Pass	8:47am

Blank Tests

Test	Status	Time
AIR	Pass	8:48am

Printer Tests

Test	Status	Time
PRNT	Pass	8:48am

CRC Tests

Test	Status	Time
COMP	Pass	8:48am
CAL	Pass	8:48am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location Wilmington

Instrument Serial No. 008628 Police Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. Rhodes
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY WILMINGTON PD 640

Serial Number: *008628*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG405702
Exp Date: 02/26/2016

Test	g/210L	Time
DIAG	Pass	4:07pm
AIR BLK	.00	4:07pm
ACCY CHK	.07	4:08pm
AIR BLK	.00	4:09pm
SUB TEST	**.00**	**4:09pm**
AIR BLK	.00	4:10pm
SUB TEST	**.00**	**4:11pm**
AIR BLK	.00	4:12pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY WILMINGTON PD 640

Serial Number: *008628* Test Record Number: *3112*
Test Date: *05/28/2014* Test Time: *4:13pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:14pm
FLO	Pass	4:14pm
FC	Pass	4:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:14pm
SRC	Pass	4:14pm
DET	Pass	4:14pm
BAR	Pass	4:14pm
BT	Pass	4:14pm

Blank Tests

Test	Status	Time
AIR	Pass	4:15pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:15pm

CRC Tests

Test	Status	Time
COMP	Pass	4:15pm
CAL	Pass	4:15pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location New Hanover County Sheriff Dept

Instrument Serial No. 008617

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY NEW HANOVER CO SD 640

Serial Number: *008617*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name: *PREVENTIVE, MAINTENANCE*
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective: *09/01/2013-09/01/2015*

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG409709
Exp Date: 04/07/2016

Test	g/210L	Time
DIAG	Pass	3:01pm
AIR BLK	.00	3:02pm
ACCY CHK	.07	3:02pm
AIR BLK	.00	3:03pm
SUB TEST	**.00**	**3:04pm**
AIR BLK	.00	3:05pm
SUB TEST	**.00**	**3:06pm**
AIR BLK	.00	3:07pm

Reported AC: .00 g/210L

K. C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY NEW HANOVER CO SD 640

Serial Number: *008617* Test Record Number: *2310*
Test Date: *05/28/2014* Test Time: *3:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:09pm
FLO	Pass	3:09pm
FC	Pass	3:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:09pm
SRC	Pass	3:09pm
DET	Pass	3:09pm
BAR	Pass	3:09pm
BT	Pass	3:09pm

Blank Tests

Test	Status	Time
AIR	Pass	3:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:10pm

CRC Tests

Test	Status	Time
COMP	Pass	3:10pm
CAL	Pass	3:10pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location New Hanover Co. Sheriff Dept.

Instrument Serial No. D08626

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K.C. [illegible]
Signature of Certifying Official

60
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY NEW HANOVER CO SD 640

Serial Number: *008626*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG404101
Exp Date: 02/10/2016

Test	g/210L	Time
DIAG	Pass	3:00pm
AIR BLK	.00	3:00pm
ACCY CHK	.07	3:01pm
AIR BLK	.00	3:02pm
SUB TEST	**.00**	**3:03pm**
AIR BLK	.00	3:04pm
SUB TEST	**.00**	**3:05pm**
AIR BLK	.00	3:06pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY NEW HANOVER CO SD-640

Serial Number: *008626* Test Record Number: *6140*
Test Date: *05/28/2014* Test Time: *3:08pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:08pm
FLO	Pass	3:08pm
FC	Pass	3:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:09pm
SRC	Pass	3:09pm
DET	Pass	3:09pm
BAR	Pass	3:09pm
BT	Pass	3:09pm

Blank Tests

Test	Status	Time
AIR	Pass	3:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:09pm

CRC Tests

Test	Status	Time
COMP	Pass	3:09pm
CAL	Pass	3:09pm

Preventive Maintenance
Status: Pass

K.C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pender Instrument Location Pender County

Instrument Serial No. 008948 Sheriff Dept Annex

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Mock
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

PENDER COUNTY SHERIFF DEPT ANNEX 700

Serial Number: *008948*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG321904
Exp Date: 08/07/2015

Test	g/210L	Time
DIAG	Pass	2:01pm
AIR BLK	.00	2:01pm
ACCY CHK	.08	2:02pm
AIR BLK	.00	2:03pm
SUB TEST	**.00**	**2:03pm**
AIR BLK	.00	2:04pm
SUB TEST	**.00**	**2:05pm**
AIR BLK	.00	2:06pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PENDER COUNTY SHERIFF DEPT ANNEX 700

Serial Number: *008948* Test Record Number: *668*
Test Date: *05/28/2014* Test Time: *2:07pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:07pm
FLO	Pass	2:07pm
FC	Pass	2:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:08pm
SRC	Pass	2:08pm
DET	Pass	2:08pm
BAR	Pass	2:08pm
BT	Pass	2:08pm

Blank Tests

Test	Status	Time
AIR	Pass	2:08pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:08pm

CRC Tests

Test	Status	Time
COMP	Pass	2:08pm
CAL	Pass	2:08pm

Preventive Maintenance
Status: Pass

K. C. Rhodes

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location Wrightsville Beach

Instrument Serial No. 008667 Police Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. Wheeler — Signature of Certifying Official

601 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY WRIGHTSVILLE BCH PD
640

Serial Number: *008667*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG321904
Exp Date: 08/07/2015

Test	g/210L	Time
DIAG	Pass	1:07pm
AIR BLK	.00	1:07pm
ACCY CHK	.08	1:08pm
AIR BLK	.00	1:09pm
SUB TEST	**.00**	**1:09pm**
AIR BLK	.00	1:10pm
SUB TEST	**.00**	**1:12pm**
AIR BLK	.00	1:12pm

Reported AC: .00 g/210L

K.C. Rhodes
Signature of Chemical Analyst

Court CVR

K.C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY WRIGHTSVILLE BCH PD 640

Serial Number: *008667* Test Record Number: *1320*
Test Date: *05/28/2014* Test Time: *1:13pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:13pm
FLO	Pass	1:13pm
FC	Pass	1:13pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:14pm
SRC	Pass	1:14pm
DET	Pass	1:14pm
BAR	Pass	1:14pm
BT	Pass	1:14pm

Blank Tests

Test	Status	Time
AIR	Pass	1:14pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:14pm

CRC Tests

Test	Status	Time
COMP	Pass	1:14pm
CAL	Pass	1:14pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County New Hanover Instrument Location Carolina Beach

Instrument Serial No. 008661 Police Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



S. C. [illegible] 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY CAROLINA BEACH PD
640

Serial Number: *008661*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG308702
Exp Date: 03/28/2015

Test	g/210L	Time
DIAG	Pass	11:59am
AIR BLK	.00	11:59am
ACCY CHK	.07	12:00pm
AIR BLK	.00	12:01pm
SUB TEST	**.00**	**12:01pm**
AIR BLK	.00	12:02pm
SUB TEST	**.00**	**12:04pm**
AIR BLK	.00	12:04pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY CAROLINA BEACH PD 640

Serial Number: *008661* Test Record Number: *1785*
Test Date: *05/28/2014* Test Time: *12:05pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:05pm
FLO	Pass	12:05pm
FC	Pass	12:05pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:05pm
SRC	Pass	12:05pm
DET	Pass	12:05pm
BAR	Pass	12:05pm
BT	Pass	12:05pm

Blank Tests

Test	Status	Time
AIR	Pass	12:06pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:06pm

CRC Tests

Test	Status	Time
COMP	Pass	12:06pm
CAL	Pass	12:06pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Instrument Location Brunswick Co.

Instrument Serial No. 008602 Sheriff Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] — Signature of Certifying Official

601 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BRUNSWICK CO SD 090

Serial Number: *008602*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG309101
Exp Date: 04/01/2015

Test	g/210L	Time
DIAG	Pass	10:16am
AIR BLK	.00	10:17am
ACCY CHK	.08	10:18am
AIR BLK	.00	10:19am
SUB TEST	**.00**	**10:20am**
AIR BLK	.00	10:20am
SUB TEST	**.00**	**10:22am**
AIR BLK	.00	10:23am

Reported AC: .00 g/210L

K. C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BRUNSWICK CO SD 090

Serial Number: *008602* Test Record Number: *2823*
Test Date: *05/28/2014* Test Time: *10:24am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:25am
FLO	Pass	10:25am
FC	Pass	10:25am

Temperature Tests

Test	Status	Time
FC1	Pass	10:25am
SRC	Pass	10:25am
DET	Pass	10:25am
BAR	Pass	10:25am
BT	Pass	10:25am

Blank Tests

Test	Status	Time
AIR	Pass	10:26am

Printer Tests

Test	Status	Time
PRNT	Pass	10:26am

CRC Tests

Test	Status	Time
COMP	Pass	10:26am
CAL	Pass	10:26am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Instrument Location Brunswick Co.

Instrument Serial No. 008585 Sheriff Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BRUNSWICK CO SD 090

Serial Number: *008585*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG409709
Exp Date: 04/17/2016

Test	g/210L	Time
DIAG	Pass	10:16am
AIR BLK	.00	10:16am
ACCY CHK	.07	10:17am
AIR BLK	.00	10:18am
SUB TEST	**.00**	**10:19am**
AIR BLK	.00	10:20am
SUB TEST	**.00**	**10:21am**
AIR BLK	.00	10:22am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BRUNSWICK CO SD 090

Serial Number: *008585* Test Record Number: *2729*
Test Date: *05/28/2014* Test Time: *10:22am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:23am
FLO	Pass	10:23am
FC	Pass	10:23am

Temperature Tests

Test	Status	Time
FC1	Pass	10:23am
SRC	Pass	10:23am
DET	Pass	10:23am
BAR	Pass	10:23am
BT	Pass	10:23am

Blank Tests

Test	Status	Time
AIR	Pass	10:24am

Printer Tests

Test	Status	Time
PRNT	Pass	10:24am

CRC Tests

Test	Status	Time
COMP	Pass	10:24am
CAL	Pass	10:24am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Instrument Location Oak Island

Instrument Serial No. 008648 Police Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY OAK ISLAND PD 090

Serial Number: *008648*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG404101
Exp Date: 02/10/2016

Test	g/210L	Time
DIAG	Pass	9:05am
AIR BLK	.00	9:06am
ACCY CHK	.07	9:06am
AIR BLK	.00	9:07am
SUB TEST	**.00**	**9:08am**
AIR BLK	.00	9:08am
SUB TEST	**.00**	**9:10am**
AIR BLK	.00	9:11am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY OAK ISLAND PD 090

Serial Number: *008648* Test Record Number: *1199*
Test Date: *05/28/2014* Test Time: *9:13am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:13am
FLO	Pass	9:13am
FC	Pass	9:13am

Temperature Tests

Test	Status	Time
FC1	Pass	9:13am
SRC	Pass	9:13am
DET	Pass	9:13am
BAR	Pass	9:13am
BT	Pass	9:13am

Blank Tests

Test	Status	Time
AIR	Pass	9:14am

Printer Tests

Test	Status	Time
PRNT	Pass	9:14am

CRC Tests

Test	Status	Time
COMP	Pass	9:14am
CAL	Pass	9:14am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Instrument Location Sunset Beach

Instrument Serial No. 008874 Police Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY SUNSET BEACH PD 090

Serial Number: *008874*
Test Date: *05/28/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG321904
Exp Date: 08/07/2015

Test	g/210L	Time
DIAG	Pass	7:59am
AIR BLK	.00	8:00am
ACCY CHK	.08	8:01am
AIR BLK	.00	8:02am
SUB TEST	**.00**	**8:02am**
AIR BLK	.00	8:03am
SUB TEST	**.00**	**8:05am**
AIR BLK	.00	8:05am

Reported AC: .00 g/210L

K. C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY SUNSET BEACH PD 090

Serial Number: *008874* Test Record Number: *353*
Test Date: *05/28/2014* Test Time: *8:06am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:06am
FLO	Pass	8:06am
FC	Pass	8:06am

Temperature Tests

Test	Status	Time
FC1	Pass	8:06am
SRC	Pass	8:06am
DET	Pass	8:06am
BAR	Pass	8:06am
BT	Pass	8:06am

Blank Tests

Test	Status	Time
AIR	Pass	8:07am

Printer Tests

Test	Status	Time
PRNT	Pass	8:07am

CRC Tests

Test	Status	Time
COMP	Pass	8:07am
CAL	Pass	8:07am

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Columbus Instrument Location Columbus Co.

Instrument Serial No. 008875 Sheriff Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. Moore
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

COLUMBUS COUNTY COLUMBUS COUNTY SD 230

Serial Number: *008875*
Test Date: *05/27/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG315701
Exp Date: 06/06/2015

Test	g/210L	Time
DIAG	Pass	3:31pm
AIR BLK	.00	3:31pm
ACCY CHK	.08	3:32pm
AIR BLK	.00	3:33pm
SUB TEST	**.00**	**3:33pm**
AIR BLK	.00	3:34pm
SUB TEST	**.00**	**3:36pm**
AIR BLK	.00	3:36pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

COLUMBUS COUNTY COLUMBUS COUNTY SD 230

Serial Number: *008875* Test Record Number: *1247*
Test Date: *05/27/2014* Test Time: *3:38pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:39pm
FLO	Pass	3:39pm
FC	Pass	3:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:39pm
SRC	Pass	3:39pm
DET	Pass	3:39pm
BAR	Pass	3:39pm
BT	Pass	3:39pm

Blank Tests

Test	Status	Time
AIR	Pass	3:40pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:40pm

CRC Tests

Test	Status	Time
COMP	Pass	3:40pm
CAL	Pass	3:40pm

Preventive Maintenance
Status: Pass

K. C. [illegible]
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Columbus Instrument Location Columbus Co. Sheriff Dept.

Instrument Serial No. 008886

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. Thorne
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

COLUMBUS COUNTY COLUMBUS COUNTY SD 230

Serial Number: *008886*
Test Date: *05/27/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG405702
Exp Date: 02/26/2016

Test	g/210L	Time
DIAG	Pass	3:32pm
AIR BLK	.00	3:32pm
ACCY CHK	.08	3:33pm
AIR BLK	.00	3:34pm
SUB TEST	**.00**	**3:34pm**
AIR BLK	.00	3:36pm
SUB TEST	**.00**	**3:37pm**
AIR BLK	.00	3:38pm

Reported AC: .00 g/210L

K.C. Rhodes
Signature of Chemical Analyst

Court CVR

K.C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

COLUMBUS COUNTY COLUMBUS COUNTY SD 230

Serial Number: *008886* Test Record Number: *967*
Test Date: *05/27/2014* Test Time: *3:38pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:39pm
FLO	Pass	3:39pm
FC	Pass	3:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:39pm
SRC	Pass	3:39pm
DET	Pass	3:39pm
BAR	Pass	3:39pm
BT	Pass	3:39pm

Blank Tests

Test	Status	Time
AIR	Pass	3:39pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:40pm

CRC Tests

Test	Status	Time
COMP	Pass	3:40pm
CAL	Pass	3:40pm

Preventive Maintenance
Status: Pass

K.C. [signature]
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Bladen Instrument Location Bladen Co.

Instrument Serial No. 008894 Sheriff Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: *008894*
Test Date: *05/27/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG404101
Exp Date: 02/10/2016

Test	g/210L	Time
DIAG	Pass	2:05pm
AIR BLK	.00	2:05pm
ACCY CHK	.08	2:06pm
AIR BLK	.00	2:07pm
SUB TEST	**.00**	**2:07pm**
AIR BLK	.00	2:08pm
SUB TEST	**.00**	**2:10pm**
AIR BLK	.00	2:10pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: *008894* Test Record Number: *614*
Test Date: *05/27/2014* Test Time: *2:20pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:20pm
FLO	Pass	2:20pm
FC	Pass	2:20pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:21pm
SRC	Pass	2:21pm
DET	Pass	2:21pm
BAR	Pass	2:21pm
BT	Pass	2:21pm

Blank Tests

Test	Status	Time
AIR	Pass	2:21pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:21pm

CRC Tests

Test	Status	Time
COMP	Pass	2:21pm
CAL	Pass	2:21pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Bladen Instrument Location Bladen Co

Instrument Serial No. 008818 Sheriff Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] 601

Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: *008818*
Test Date: *05/27/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTNENACE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG404101
Exp Date: 02/10/2016

Test	g/210L	Time
DIAG	Pass	2:06pm
AIR BLK	.00	2:07pm
ACCY CHK	.07	2:07pm
AIR BLK	.00	2:08pm
SUB TEST	**.00**	**2:09pm**
AIR BLK	.00	2:10pm
SUB TEST	**.00**	**2:11pm**
AIR BLK	.00	2:12pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: *008818* Test Record Number: *805*
Test Date: *05/27/2014* Test Time: *2:19pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:19pm
FLO	Pass	2:19pm
FC	Pass	2:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:20pm
SRC	Pass	2:20pm
DET	Pass	2:20pm
BAR	Pass	2:20pm
BT	Pass	2:20pm

Blank Tests

Test	Status	Time
AIR	Pass	2:20pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:20pm

CRC Tests

Test	Status	Time
COMP	Pass	2:20pm
CAL	Pass	2:20pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Sampson Instrument Location Sampson County Sheriff Dept.

Instrument Serial No. 008825

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

Certificate Number 601

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

SAMPSON COUNTY SAMPSON CO SD 810

Serial Number: *008825*
Test Date: *05/27/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG315701
Exp Date: 06/06/2015

Test	g/210L	Time
DIAG	Pass	12:39pm
AIR BLK	.00	12:40pm
ACCY CHK	.07	12:40pm
AIR BLK	.00	12:41pm
SUB TEST	**.00**	**12:42pm**
AIR BLK	.00	12:42pm
SUB TEST	**.00**	**12:44pm**
AIR BLK	.00	12:45pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

SAMPSON COUNTY SAMPSON CO SD 810

Serial Number: *008825* Test Record Number: *1788*
Test Date: *05/27/2014* Test Time: *12:46pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:46pm
FLO	Pass	12:46pm
FC	Pass	12:46pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:46pm
SRC	Pass	12:46pm
DET	Pass	12:46pm
BAR	Pass	12:46pm
BT	Pass	12:46pm

Blank Tests

Test	Status	Time
AIR	Pass	12:47pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:47pm

CRC Tests

Test	Status	Time
COMP	Pass	12:47pm
CAL	Pass	12:47pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Sampson Instrument Location Sampson County Sheriff Dept.

Instrument Serial No. 008877

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible] — Signature of Certifying Official

601 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

SAMPSON COUNTY SAMPSON COUNTY SD 810

Serial Number: *008877*
Test Date: *05/27/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG405702
Exp Date: 02/26/2016

Test	g/210L	Time
DIAG	Pass	12:29pm
AIR BLK	.00	12:30pm
ACCY CHK	.08	12:30pm
AIR BLK	.00	12:31pm
SUB TEST	**.00**	**12:32pm**
AIR BLK	.00	12:33pm
SUB TEST	**.00**	**12:34pm**
AIR BLK	.00	12:35pm

Reported AC: .00 g/210L

K. C. Rhodes
Signature of Chemical Analyst

Court CVR

K. C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

SAMPSON COUNTY SAMPSON COUNTY SD 810

Serial Number: *008877* Test Record Number: *1658*
Test Date: *05/27/2014* Test Time: *12:36pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:37pm
FLO	Pass	12:37pm
FC	Pass	12:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:37pm
SRC	Pass	12:37pm
DET	Pass	12:37pm
BAR	Pass	12:37pm
BT	Pass	12:37pm

Blank Tests

Test	Status	Time
AIR	Pass	12:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:38pm

CRC Tests

Test	Status	Time
COMP	Pass	12:38pm
CAL	Pass	12:38pm

Preventive Maintenance
Status: Pass

K.C. [illegible]
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Duplin Instrument Location Wallace PD

Instrument Serial No. 008858

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

K. C. [illegible]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

DUPLIN COUNTY WALLACE PD 300

Serial Number: *008858*
Test Date: *05/29/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *RHODES, KENNETH C*
Permit Number: *5329E*
Effective:
09/01/2013-09/01/2015

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG409709
Exp Date: 04/07/2016

Test	g/210L	Time
DIAG	Pass	9:34am
AIR BLK	.00	9:35am
ACCY CHK	.07	9:35am
AIR BLK	.00	9:36am
SUB TEST	**.00**	**9:37am**
AIR BLK	.00	9:38am
SUB TEST	**.00**	**9:39am**
AIR BLK	.00	9:40am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DUPLIN COUNTY WALLACE PD 300

Serial Number: *008858* Test Record Number: *644*
Test Date: *05/29/2014* Test Time: *9:48am EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:48am
FLO	Pass	9:48am
FC	Pass	9:48am

Temperature Tests

Test	Status	Time
FC1	Pass	9:48am
SRC	Pass	9:48am
DET	Pass	9:48am
BAR	Pass	9:48am
BT	Pass	9:48am

Blank Tests

Test	Status	Time
AIR	Pass	9:49am

Printer Tests

Test	Status	Time
PRNT	Pass	9:49am

CRC Tests

Test	Status	Time
COMP	Pass	9:49am
CAL	Pass	9:49am

Preventive Maintenance
Status: Pass

K.C. Rhodes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Alamance Instrument Location Burlington P.D.

Instrument Serial No. 008812 267 W. Front St.
Burlington, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Signature of Certifying Official

655
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: *008812*
Test Date: *05/30/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *FARLEY, CYNTHIA D*
Permit Number: *24123E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG326006
Exp Date: 09/17/2015

Test	g/210L	Time
DIAG	Pass	12:17pm
AIR BLK	.00	12:18pm
ACCY CHK	.07	12:19pm
AIR BLK	.00	12:20pm
SUB TEST	**.00**	**12:21pm**
AIR BLK	.00	12:22pm
SUB TEST	**.00**	**12:23pm**
AIR BLK	.00	12:24pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: *008812* Test Record Number: *1984*
Test Date: *05/30/2014* Test Time: *12:27pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:27pm
FLO	Pass	12:27pm
FC	Pass	12:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:27pm
SRC	Pass	12:27pm
DET	Pass	12:27pm
BAR	Pass	12:27pm
BT	Pass	12:27pm

Blank Tests

Test	Status	Time
AIR	Pass	12:28pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:28pm

CRC Tests

Test	Status	Time
COMP	Pass	12:28pm
CAL	Pass	12:28pm

Preventive Maintenance
Status: Pass

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Alamance Instrument Location Burlington P.D.

Instrument Serial No. 008907 267 W. Front St.
Burlington, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of May, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.

Dale Farley — Signature of Certifying Official

655 — Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

DHHS 4080 (11/07)

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: *008907*
Test Date: *05/30/2014*

Citation Number: *M0000000-0*
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: *11/11/1911*
Subject's Sex: *Male*
Driver's License State: *XX*
Driver's License Number: *NONE*

Analyst's Name: *FARLEY, CYNTHIA D*
Permit Number: *24123E*
Effective:
12/01/2012-12/01/2014

Officer's Name: *NONE, NONE*
Type of Agency: *FTA*
Agency: *DHHS*
Test Type: *Breath Test*

Lot Number: AG320602
Exp Date: 07/25/2015

Test	g/210L	Time
DIAG	Pass	12:01pm
AIR BLK	.00	12:02pm
ACCY CHK	.08	12:02pm
AIR BLK	.00	12:04pm
SUB TEST	**.00**	**12:04pm**
AIR BLK	.00	12:05pm
SUB TEST	**.00**	**12:07pm**
AIR BLK	.00	12:08pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: *008907* Test Record Number: *621*
Test Date: *05/30/2014* Test Time: *12:10pm EDT*

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:10pm
FLO	Pass	12:10pm
FC	Pass	12:10pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:10pm
SRC	Pass	12:10pm
DET	Pass	12:10pm
BAR	Pass	12:10pm
BT	Pass	12:10pm

Blank Tests

Test	Status	Time
AIR	Pass	12:11pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:11pm

CRC Tests

Test	Status	Time
COMP	Pass	12:11pm
CAL	Pass	12:11pm

Preventive Maintenance
Status: Pass

Dale Farley
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007